

Customer Repair Report

Model	Serial Number	Manufacture Da	te	RMA Number	IRIDEX Service Contact	Completion Date	
OTHER	YH13733TC	C-D		03086	Mark W		
Customer Name		Customer Number	Repa	ir Type			
Chatham Eye Associates		25610		Warranty Marranty Upgrade Other			

Customer Description of Problem

E05111 and skip pulses

Technician Product Evaluate:

UNABLE TO TESTED CONTROL BOX NON-STOP BLINKING BAD COMPUTER BOARD.

Service Performed on Unit: Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price	
88356		1	COMPUTER HDD	\$	1,950.00
88351		2	COMPUTER MOUNT	\$	138.00
88352		1	CABLE ADAPTER, TxCell CPU	\$	46.16
88261		1	Cable Assy. USB Two Ports.	\$	83.00
TECHNICIAN HQ		5	Technician Labor	\$	1,425.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$	196.70

TOTAL CHARGES \$ 3,838.86

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:	
Signature:	Date