



## Customer Repair Report

Model OTHER	Serial Number IQG81303C-D	Manufacture Date	RMA Number 02110	IRIDEX Service Contact Mark W	Completion Date
Customer Name UCI Medical Center		Customer Number 15040	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

### Customer Description of Problem

Low Power

### Technician Product Evaluate:

Need to Replaced Assy optics combiner , fine adjust screws, realigned the laser head, and recalibrated the unit.

### Service Performed on Unit:

Waiting for customer approval

### ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
50299		4	CREW 1/4-80 FINE ADJUST BALL END HEX SET THORLABS: F25SS05	\$ 75.04
65994		1	MOUNT COMBINER	\$ 91.00
65001		1	OPTIC COMBINER R532/T635 & 1064	\$ 140.00
TECHNICIAN HQ		10	Technician Labor	\$ 2,850.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$ 84.54

---

**TOTAL CHARGES** **\$ 3,240.58**

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges.  
Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: