



## Customer Repair Report

Model OTHER	Serial Number 1404	Manufacture Date	RMA Number 03037	IRIDEX Service Contact Mark W	Completion Date
Customer Name Memorial Hospital		Customer Number 16252	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

### Customer Description of Problem

didn't pass at the annual PM

### Technician Product Evaluate:

PM FAIL FOUND LOW POWER AND DIPLAY HAS BROKEN LIO AND AIMING POT. NEED POWER DIODE AIMING DIODE SET SCREW AND TWO POT

### Service Performed on Unit:

Waiting for customer approval

### ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
11224		1	Laser Diode 808nm 3W	\$ 2,500.00
12958		1	AIMING DIODE 650nm ASSY PLATE INSULATOR & DIODE	\$ 192.16
50299		4	CREW 1/4-80 FINE ADJUST BALL END HEX SET THORLABS: F25SS05	\$ 75.04
10096		1	0K 1TRN 1/2W PNL MNTPREC ELEC: SPRU5031S28 DIGIKEY: SPRU50	\$ 34.32
10097		1	5K 1TRN 1/2W PNL MNT PREC ELEC: SPRU2521S28 DIGIKEY: SPRU25	\$ 38.60
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$ 206.50

---

**TOTAL CHARGES** **\$ 3,046.62**

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: