

Customer Repair Report

24360 Salient Medical Warranty non-warranty Upgrade Other	Model		Serial Number	Manufacture Date		RMA Number	IRIDEX Service Contact	Completion Date	
24360 Salient iviedical Warranty non-warranty Upgrade Other	OTHER		YEZ15668TC			01994	Mark W		
24360 Warranty Inon-warranty Upgrade Uther	Customer Name		Customer Number Repair Type						
Solutions	2	24360	Sai	Salutions Warranty Marranty Upgrade Other					

Customer Description of Problem

PER PHYSICIAN LOW POWER

Technician Product Evaluate:

NEED TO REPLACE FIBER. WILL TESTED, ALIGNED, CALIBRATED, CLEAN, AND INSPECTED.

Service Performed on Unit: Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description		Total Price	
70208		1	ASSY JUMPER RFID SLA 9FT EXTENDED JACKET	\$	984.00	
12757		1	THUMBSCREW M3.5 ZEISS SLAMM	\$	50.00	
TECHNICIAN HQ		4	Technician Labor	\$	1,140.00	

TOTAL CHARGES

2,174.00

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:

Signature: Date: