



Customer Repair Report

Model OTHER	Serial Number EZZ13063TC	Manufacture Date	RMA Number 02003	IRIDEX Service Contact Mark W	Completion Date
Customer Name Lawton Indian Hospital		Customer Number 24407	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

Customer Description of Problem

No communicatin between the IQ& TxCell E05111

Technician Product Evaluate:

UNABLE TOTESTED CONTROL BOX NON-STOP BLINKING DO TO COMPUTER BOARD NOT WORKING GALVO'S ARE DOWN REV AND PARTS NEED TO BE UPGRADED.

Service Performed on Unit:

Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
70431		1	Bracket Cable Guide	\$ 32.72
70171		1	GALVO DRIVER SET (6SIRIDEX2)	\$ 4,030.20
70406		1	EXTENSION COLLAR STRAIN RELIEF TxCELL	\$ 133.00
70407		1	SEPARATOR CABLE	\$ 120.56
70318		1	STRAIN RELIEF CABLE CLAMP	\$ 46.80
70076		1	ASSY SCANNING INPUT LENS	\$ 531.84
12757		1	THUMBSCREW M3.5 ZEISS SLAMM	\$ 62.50
88356		1	COMPUTER HDD	\$ 1,950.00
88351		2	COMPUTER MOUNT	\$ 138.00
88352		1	CABLE ADAPTER, TxCell CPU	\$ 43.16
88261		1	Cable Assy. USB Two Ports.	\$ 91.04
70117		2	AIMING BEAM FLEX PLATE	\$ 216.64
TECHNICIAN HQ		8	Technician Labor	\$ 2,280.00
BILLABLE FREIGHT				1 \$ 188.28

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges.
Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: