

Customer Repair Report

Model		Serial Number	Manufacture Dat	e	RMA Number	IRIDEX Service Contact	Completion Date	
OTHER		1170219			02097	Mark W		
Customer Name		(Customer Number	Repai	ir Type			
Trimedx			300381		Warranty non-warranty Upgrade Other			

Customer Description of Problem

Preventative Maintenance

Technician Product Evaluate:

Cannot test due to missing assembly fiber optic, extension cable, halogen bulb, broken flexure mirror support, bent control stem, will need adjustmet and relignment. Control stem been modifies will not ease to remove. Broken heine headband omega H500.

Service Performed on Unit: Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description		Total Price	
88080		1	Jumper Assy Mulitlam LIO+	\$	984.00	
87603		1	LIO EXTENSION CABLE(CS)	\$	235.00	
12685		2	CONTROL STEM KNOB L & R VIEWING P/N: 11.65 38	\$	220.00	
12687		1	BEAM CONTROL STEMS HEINE SUPPORT KNOBS P/N: 11.65 15	\$	125.00	
10784		1	LIO Halogen Bulb-10W/6V Heine: X-004.88.068	\$	105.00	
11481		1	DUST CAP W/LANYARD DART: IRIS DWG:	\$	3.00	
TECHNICIAN HQ		12	Technician Labor	\$	3,420.00	
13354		1	Heine LIO Headset Assy IRIDEX: 13354	\$	600.00	
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$	196.70	

TOTAL CHARGES	\$ 5,888.70

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:	
Signature:	Date: