



## Customer Repair Report

Model OTHER	Serial Number GH14081TC	Manufacture Date	RMA Number 02700	IRIDEX Service Contact Mark W	Completion Date
Customer Name Tennessee Eye Care Center		Customer Number 16058	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

### Customer Description of Problem

LED BLINKING

### Technician Product Evaluate:

UNABLE TO TESTED CONTROL BOX NON-STOP BLINKING BAD COMPUTER BOARD. BROKEN BEZEL.

### Service Performed on Unit:

Waiting for customer approval

### ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
70358		1	INTERFACE BOX BEZEL LEDPLASTIC	\$ 373.00
88356		1	COMPUTER HDD	\$ 1,950.00
88351		2	COMPUTER MOUNT	\$ 138.00
88352		1	CABLE ADAPTER, TxCell CPU	\$ 46.16
88261		1	Cable Assy. USB Two Ports.	\$ 83.00
TECHNICIAN HQ		6	Technician Labor	\$ 1,710.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$ 206.50

**TOTAL CHARGES** **\$ 4,506.66**

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: