

MAINTENANCE WORK ORDER

Work Order No. : WO-3591
Date Issued : 22 September 2018
Priority : ROUTINE
Status : COMPLETED
Department : Instrumentation & Electrical
Location : Bay 4, Zone C
Equipment Tag : GB-14 (Draeger Polytron 2 XP Combustible Gas Detector)
Assigned To : Anil Mehta, Technician (I&E)
Supervised By : Rajiv Sharma, Maintenance Supervisor

1. WORK DESCRIPTION

Scheduled 6-monthly calibration of fixed combustible gas detector GB-14, Bay 4, Zone C.
Previous calibration: 14 March 2018 (WO-3441).

2. WORK PERFORMED

Date Completed : 22 September 2018, 09:45 IST

Step 1 ? Bump test: Response to 50% LEL methane standard: 49% LEL. PASS.
Step 2 ? Zero calibration: Pre-cal zero: 0.0% LEL. No adjustment required.
Step 3 ? Span calibration using 50% LEL methane (Cert. No. CAL-2018-MET-4419).
Post-calibration reading: 50.0% LEL. PASS.
Step 4 ? Visual inspection: Housing intact. No issues found.
Step 5 ? Sensor age: 1 year 6 months (installed March 2017). Within 24-month limit.

3. NOTES

Calibration completed satisfactorily. No abnormalities noted.
Next calibration due: March 2019.
Sensor element replacement due: March 2019 (24-month limit from March 2017 install).

4. SIGNATURES

Technician : Anil Mehta Date: 22/09/2018
Supervisor : Rajiv Sharma Date: 22/09/2018

NOTE: Next calibration (March 2019) was MISSED ? not performed until after the near-miss incident of 14 March 2019 (NMR-2019-047). This gap was identified as Immediate Cause CA-(a) in NMR-2019-047.

CROSS-REFERENCE: WO-3441 (previous cal), NMR-2019-047 (next significant event)