

Quality Assurance of a Diagnostic X-ray Machine



Date of experiment: 08 January 2026

Medical Physics Laboratory II

Name: Souvik Das

Roll No: 241126007



**National Institute of Science Education and Research
Jatani, Odisha, India**

Date of submission: January 21, 2026

Contents

1	Objective	1
2	Apparatus	1
2.1	Cobia Smart R/F	1
2.2	RTI Scatter Probe	1
2.3	The Pro-RF BasicTool phantom	1
3	Theory	2
3.1	X-ray tube exposure parameter affect the output	2
3.2	Congruence of Optical and Radiation Fields	4
3.3	Central beam alignment	4
3.4	Focal spot size	4
3.5	Exposure time	4
3.6	Applied tube potential (kVp)	4
3.7	Total filtration	5
3.8	Linearity of the mA loading station	5
3.9	Consistency (reproducibility) of radiation output	5
3.10	Radiation leakage through tube housing	5
3.11	Measurement of Low & High Contrast sensitivity	5
3.12	Radiation protection survey	6
4	Observation	6
4.1	DETAILS OF THE DIAGNOSTIC X-RAY EQUIPMENT	6
4.2	CONGRUENCE OF RADIATION & OPTICAL FIELD	6
4.3	CENTRAL BEAM ALIGNMENT	7
4.4	EFFECTIVE FOCAL SPOT MEASUREMENT	7
4.5	ACCURACY OF OPERATING POTENTIAL	8
4.6	LINEARITY OF (mA/mAs) LOADING STATIONS	8
4.7	OUTPUT CONSISTENCY	8
4.8	CONSTRAST SENSITIVITY	9
4.9	TUBE HOUSING LEAKAGE	9
4.10	DETAILS OF RADIATION PROTECTION SURVEY OF THE INSTALLATION	9
5	Conclusion	10
5.1	Equipment Performance and Accuracy	10
5.2	Output Consistency and Linearity	10
5.3	Image Resolution and Safety	10
5.4	Final Summary	10

1 Objective

To perform the quality assurance tests of a Diagnostic X-Ray Machine.

2 Apparatus

- Diagnostic X-Ray Machine
- X-Ray QA Kit. (Equipment Details given below)

2.1 Cobia Smart R/F

The Cobia Smart R/F is a user-friendly, all-in-one X-ray QA (Quality Assurance) meter from RTI Group designed for straightforward testing of conventional radiography and fluoroscopy units, as well as intraoral dental and CBCT systems. It provides instant and accurate readings of key X-ray parameters without requiring user adjustments.

2.2 RTI Scatter Probe

The RTI Scatter Probe is a rugged, solid-state detector designed for high-sensitivity measurements of scatter and leakage radiation in diagnostic X-ray environments. Unlike traditional pressurized ion chambers, it requires no warm-up time and is independent of temperature or pressure variations.


Ocean Next Software is associated with the above QA tools for evaluation and analysis of the test results.



Cobia Smart R/F



RTI Scatter Probe



Pro-RF Basic tool

Figure 1: X-Ray QA Kit Components

2.3 The Pro-RF BasicTool phantom

The Pro-RF Basic tool is a quality assurance phantom designed for diagnostic X-ray machines. It is used to evaluate various parameters of X-ray imaging systems, including spatial resolution, contrast resolution, and image uniformity. The phantom typically consists of different test patterns and materials that simulate human tissue, allowing for comprehensive assessment of the imaging system's performance.

The Pro-RF Basic tool includes features such as:

- outside dimensions: 345×290 mm
- high contrast array of 5 and 10 mm uniformly pitched mesh
- Highlight and Lowlight details
- 0.6 to 5.0 LP/mm resolution pattern
- 7×11.0 mm diameter low contrast details
- 11×0.5 mm diameter high contrast details
- extra stand for the spatial resolution pattern for focal spot size evaluation

- cone for perpendicular X-Ray beam control in the range of $0^\circ \div 1.5^\circ$
- an area of uniform attenuation suitable for noise measurements
- 1 mm Cu filter for HVL measurements

3 Theory




Figure 2: X-ray tube schematic diagram. A: Anode target, B: Bearing, C: Cathode filament, D: Envelope, E: Housing, F: Target, G: Stator, H: Rotor, I: Oil, J: Window [1]

3.1 X-ray tube exposure parameter affect the output

1. Tube voltage (kV): Changing tube voltage changes the total number of photons emitted (quantity) and the photon energies (quality). This is because although the number of bombarding electrons is the same, the probability of X-ray production increases with increased tube voltage. Increasing the kV may result in the appearance of more characteristic X-ray photons and the shifting of the mean photon energy to higher energy (keV) (see Figure). The relationship between the X-ray outputs with tube voltage is often approximated as follows:

$$\text{Output} \propto \text{kV}^2$$

2. Tube current (mA): Increasing the tube current results in an increase of the amount of electrons used to produce X-rays (Figure). Thus, there is an increase in the quantity of the electrons as well as the X-ray photons produced. The X-ray quality remains essentially the same, i.e. the maximum photon energy and the effective photon energy remains unchanged.

$$\text{Output} \propto \text{mA}$$

3. Time of exposure (s): Time of exposure determines the number of electrons striking the anode. The number of X-ray photon increases with longer exposure time.

$$\text{Output} \propto \text{exposure time}$$

4. Filtration: Filtration changes the quantity and quality of the X-ray tube output. Filtration can be due to inherent filters such as oil and the window of the tube housing. Filtration can also be due to additional filters placed in the beam path. With increased filtration, the number of X-ray photons

is reduced and the mean energy of the X-ray energy spectrum shifts to a higher energy (Figure). This is because a proportion of the low energy photons is attenuated or filtered. K-edge filters are special filters that behave as a bandpass filter to allow the transmission of X-ray photons of a certain energy window.

5. Target material: The probability of X-ray production, or X-ray yield, increases with the use of X-ray targets made of high atomic number (Z) materials (Figure). X-ray output generally increases in the total number of photons produced. Characteristic X-ray photons with higher energies are also produced using a target material with higher Z.




Figure 3: Tube Voltage vs Output




Figure 4: Tube Current vs Output




Figure 5: Filtration vs Output




Figure 6: Target Material vs Output

The Quality assurance (QA) of diagnostic X-ray machines is a critical aspect of medical physics that ensures the safety, reliability, and optimal performance of these imaging devices. Regular QA procedures help in maintaining image quality, minimizing radiation exposure to patients and staff, and ensuring compliance with regulatory standards. Diagnostic X-ray machines are widely used in medical imaging for various applications, including radiography, fluoroscopy, and computed tomography (CT). The QA process involves a series of tests and evaluations designed to assess the performance of the X-ray equipment and its components. These tests typically include assessments of image quality parameters such as spatial resolution, contrast resolution, noise levels, and uniformity. Additionally, QA procedures evaluate the accuracy of radiation dose delivery, beam alignment, and mechanical integrity of the equipment. Regular QA checks are essential to identify any deviations from established performance standards, which may arise due to equipment wear and tear, calibration drift, or other factors. By promptly addressing these issues, healthcare facilities can ensure that diagnostic X-ray machines operate within acceptable limits, thereby enhancing patient safety and diagnostic accuracy. The responsibility of quality control rests upon the Medical Physicist and the Technical staff of the diagnostic department. The test procedure involves the following steps: Carry out the QA tests as per the prescribed procedures, Record the results, Analyze the results, Seek corrective and preventive measures if the results are not satisfactory, Repeat the tests until satisfactory results are obtained.

In QA tests, the accuracy and consistency of various parameters, which influence the quality of diagnostic images and patient dose, are checked. The procedure for each experiment is given below.

- Congruence of optical and radiation fields
- Central beam alignment
- Focal spot size
- Exposure time
- Applied tube potential (kVp)
- Total filtration
- Linearity of mA loading stations
- Consistency of radiation output
- Radiation leakage through tube housing
- Measurement of Low & High Contrast sensitivity
- Radiation protection survey

3.2 Congruence of Optical and Radiation Fields

The congruence of the optical and radiation fields is a crucial quality assurance test for diagnostic X-ray machines. This test ensures that the area indicated by the optical light field accurately represents the area exposed to X-rays during imaging. Proper alignment between these fields is essential for accurate patient positioning and optimal image quality, as any misalignment can lead to unintended radiation exposure to non-targeted areas and compromise diagnostic outcomes.

Place the phantom on the table, between the image detector and the X-ray tube in such a way that the phantom so its center and main axis align with the markers of the apparatus. Limit the light field to the chosen rectangular marking on the phantom surface. Make the exposure. Measure the distance between the light field borders and actual X-ray beam borders in perpendicular and parallel directions. Use a clearly marked scale on the phantom. The sum of differences between the light and the X-ray field in perpendicular and parallel directions should not be greater than 2% of the FFD (Focal spot to Film distance).

3.3 Central beam alignment

The image may be distorted if the X-ray beam is not perpendicular to the image receptor (Detector). For this test, screw the optional cone on the phantom. The optional cone has a stainless- steel ball on the top and a circular steel patch at the bottom. Make the exposure. If the image of the ball lies entirely within the image of the circular patch, then the deviation of the beam from the tube axis is not more significant than 1.5°.

3.4 Focal spot size

The ability to resolve the smallest image size (i.e., Spatial Resolution) in a radiograph depends on the focal spot size. Since the focal spot size may be altered due to the bombardment of electrons with the target, it must be checked periodically to ensure that it is within acceptable limits. Focal spot size is evaluated on the principle of minimum resolution and a Bar/hole test pattern is employed for evaluating focal spot size. At minimum resolution, the edge gradient (penumbra) of one pattern of the pair merges with the image of the other, and the images of both patterns of the pair cannot be resolved separately. When this happens, focal spot size (f), line width, and Magnification (M) are related as follows:

$$f = \frac{M}{M - 1} \times \text{Line width}$$

Where, M = Source-to-film distance / Source-to-pattern distance

To perform this test, a resolution bar pattern is mounted on the phantom such that the source-to-pattern distance is three times the pattern-to-film/detector distance. The resolution bar pattern consists of several groups of lines (slits) of sizes gradually decreasing in dimension. Expose the bar pattern. Perform two measurements in two bar pattern positions: parallel and perpendicular to the bottom edge of the phantom. A given pattern must be resolved in both directions to count. From the resolved Line Width, the focal spot size can be calculated using the above formula.

3.5 Exposure time

The accuracy of the exposure time is essential for obtaining consistent image quality and controlling patient dose. If the exposure time of the X-ray diagnostic unit is not in order, the radiograph can be underexposed or overexposed. There may be a change in the adequate tissue contrast resolution in the image. The Cobia Smart R/F detector can measure the exposure time.

3.6 Applied tube potential (kVp)

The applied kilovoltage affects the quality and quantity of X-rays reaching the image receptor (Detector) and hence the contrast and density of the radiograph. Since any variation from the set kVp can affect the quality of the radiograph, the kVp settings must be checked periodically. kVp can be measured by the Cobia Smart R/F.

3.7 Total filtration

A specific minimum filtration must be present in the X-ray tube to cut off low-energy components from the X-ray beam. These low-energy components do not contribute to image formation but result in unnecessary patient exposure. If the filtration is too high, the attenuation will be more and image contrast will be reduced. Therefore, the total filtration for the X-ray tube should be optimum for patient safety and image quality. For this purpose, regulatory bodies (AERB) recommend total filtration for X-Ray machines for different maximum-rated tube potentials. Total filtration includes inherent filtration and added filtration. Hence total filtration evaluation is necessary to verify whether the added filtration is adequate. If not adequate, additional filtration must be provided for the x-ray tube. The Cobia Smart R/F detector can measure the total filtration.

3.8 Linearity of the mA loading station

Keeping the kVp and time constant, the radiation output is measured at different mA stations. Measurement for each mA station is to be repeated for several times to eliminate statistical errors. Each mA station reading is averaged, and the coefficient of linearity (COL) is evaluated from average mR/mAs or mGy/ mAs (X) as follows. The tolerance is 0.1.

$$\text{Coefficient of Linearity} = \frac{x_{max} - x_{min}}{x_{max} + x_{min}}$$

3.9 Consistency (reproducibility) of radiation output

Keeping fixed mA and time, radiation output is measured at various available kVp stations. The average of mR/mAs (mGy/mAs) is calculated (X). Consistency at each kVp station is checked by evaluating the Coefficient of Variation (COV). The COV should not exceed 0.05.

$$COV = \frac{1}{X} \sqrt{\frac{(X_i - X)^2}{n - 1}}$$

Where, X_i = individual reading, n = number of readings

3.10 Radiation leakage through tube housing

As per AERB Safety Code on Diagnostic X-ray Equipment and installations, every housing for medical diagnostic X-ray equipment shall be so constructed that the leakage radiation through the protective tube housing in any direction shall not exceed an air kerma of 1.0mGy (about 114mR) in one hour at 1.0m from the x-ray target. The measurement conditions are given below.

- Averaged over an area not larger than 100 cm²
- No linear dimension greater than 20cm
- Operating at maximum rated kVp and for the maximum rated current at that kVp.

The radiation leakage measurement is carried out with an ionization radiation survey meter. For checking the leakage radiation, the collimator of the tube housing is fully closed. The operating time should be greater than the time constant of the survey meter. The exposure rate at one meter from the target is measured at different locations (anode side, cathode side, front back, and top) from the tube housing and collimator. For the maximum leakage rate (mR/h) for both tube housing and collimator, leakage in one hour is computed based on the workload of the machine. 180mA-min in one hour is taken as the maximum workload of a diagnostic machine. Hence leakage in one hour will be:

$$\text{Leakage in one hour (mR)} = \text{Leakage rate (mR/h)} \times \frac{180}{\text{Operating mA-min}}$$

3.11 Measurement of Low & High Contrast sensitivity

Low contrast sensitivity refers to the ability of a system to visualize low-contrast objects or structures. In clinical practice, patient diagnoses are frequently made on the perception of the fluoroscopic image of small low-contrast anatomic structures. The number of visible low-contrast objects should not change over time.

The high contrast sensitivity allows for a simple evaluation of system resolution to ensure that optimal anatomic detail is visible to the practitioner. High contrast resolution is essential for all fluoroscopic procedures. Using the zoom tool, assess the Line Pairs per millimeter resolution. It is the highest number next to the group of lines where three separate lines can be distinguished.

3.12 Radiation protection survey

A radiation protection survey is a series of measurements of radiation levels at various locations around the diagnostic X-Ray machine installation. This is done to check whether the radiation levels around the building are within the permissible limits as mandated by the Competent Authority (AERB). A pressurized ion chamber-based survey meter is used to measure the radiation levels.

4 Observation

4.1 DETAILS OF THE DIAGNOSTIC X-RAY EQUIPMENT

Name of the Institution and City	NISER Health Center, Jatani
Type of Equipment	Medical X-Ray Collimator
Model Name	SR-102
Name of the Manufacturer	Hangzhou Sailray Imp & Exp Co. Ltd
Name(s) of Person(s) testing the equipment	Souvik Das and Deepak Saini
Dates and Duration of the Tests	08/01/2026, in 2 days

4.2 CONGRUENCE OF RADIATION & OPTICAL FIELD

Operating parameters:

FFD(cm) = 100 kV = 60 mAs = 5

Shift in the edges of the radiation field.

Dimensions (cm)	Observed shift	% of FFD	Tolerance	Remark
X + X'	1.05	1.05 %	2% of FFD	Within Tol.
Y + Y'	0.5	0.5 %	2% of FFD	Within Tol.




Figure 7: X-ray image of Pro-RF basic phantom.

4.3 CENTRAL BEAM ALIGNMENT

Operating parameters:

FFD(cm) = 100 kV = 60 mAs = 5

Observe the images of the two steel balls on the radiograph and evaluate the tilt in the central beam.

Observed tilt =	< 1.5°	Remark:	Within tol.
-----------------	--------	---------	-------------

Tolerance: Central Beam Alignment < 1.5°




Figure 8: Verification of central beam alignment using cone test.

4.4 EFFECTIVE FOCAL SPOT MEASUREMENT

FFD(cm) = 60 Magnification = M = 80/60

	The stated value of Focal Spot Size (mm x mm)	Direction	Resolution line pair (lp/mm)	The measured value of Focal Spot size (mm x mm)	Tolerance:	Tolerance	Remarks
Large Focus	2	Parallel	1.6	2.50	1. + 0.5 f for f < 0.08 mm 2. + 0.4 f for 0.8 ≤ f ≤ 1.5 mm 3. + 0.3 f for f > 1.5 mm	0.6	Within tol.
		Perpendicular	2	2.00		0.6	Within tol.
Small Focus	1	Parallel	2.2	1.82		0.4	Not in tol.
		Perpendicular	2.2	1.82		0.4	Not in tol.




Figure 9: Parallel to the anode-cathode axis




Figure 10: Perpendicular to the anode-cathode axis



4.5 ACCURACY OF OPERATING POTENTIAL

Applied kVp	Set time(ms)	Measured values									
		mA stations									
		Minimum mA station(10 mA)		Routinely used mA station(125 mA)		Maximum mA station(200 mA)					
		kVp	Time(ms)	kVp	Time(ms)	kVp	Time(ms)	Average kVp	% Error in kVp	Average Time(ms)	%Error in Timer
60	50	60.85	50.57	60.30	49.37	60.20	49.46	60.45	0.75	49.80	0.40
80	50	80.20	51.19	80.10	49.62	80.20	49.60	80.17	0.21	50.14	0.27
100	50	101.20	51.89	100.30	49.81	100.30	49.73	100.60	0.60	50.48	0.95
120	50	120.90	52.58	119.30	49.98	119.50	50.38	119.90	0.08	50.98	1.96
Tolerance for kVp: ± 5 kV Tolerance for Timer: % Error: $\pm 10\%$ Total Filtration (measurement at maximum kVp): 3.2 mm of AL Tolerance for Total Filtration: 1.5 mm Al for for kV ≤ 70 , 2.0mm Al for kV ≤ 100 , 2.5 mm Al for kV > 100											


So both the kVp and mAs are within the acceptable limits.

4.6 LINEARITY OF (mA/mAs) LOADING STATIONS

Operating parameters:

FFD(cm) = 100 kV = 120 Time(s) = 0.05

mA applied	Radiation output (mGy)			Average Output	mGy/mAs (x)	Coefficient of Linearity (CoL)
	Reading 1	Reading 2	Reading 3			
250	1.426	1.429	1.434	1.429667	0.114373	0.010652
200	1.145	1.151	1.155	1.150333	0.115033	
160	0.8906	0.8929	0.9036	0.895700	0.111963	
Tolerance: COL < 0.1						



4.7 OUTPUT CONSISTENCY

Applied kV	mAs	Radiation Output (mGy)					Average (X)	Coefficient of Variation (COV)
		1	2	3	4	5		
80	10	0.5488	0.5476	0.5479	0.5475	0.5484	0.54804	0.001004
120	5	0.6085	0.6086	0.6085	0.6091	0.6112	0.60918	0.00190
Tolerance: CoV < 0.05								



4.8 CONTRAST SENSITIVITY

LOW CONTRAST SENSITIVITY

The diameter of the smallest size hole resolved on the monitor	1.0 mm hole pattern is resolved
Recommended performance standard	3.0 mm hole pattern must be resolved

HIGH CONTRAST SENSITIVITY

Bar strips resolved on the monitor (lp/mm)	2.2 lp/mm
Recommended performance standard	1.5 lp/mm pattern must be resolved

4.9 TUBE HOUSING LEAKAGE

FDD(cm) = 100 kVp(max) = 120 mA = 250 Time(s) = 0.004

Location(at 1 m from the focus)	Exposure level (μGy/hr)						Remarks
	Left	Right	Front	Back	Top	Back	
Tube(100 cm ²)	1097	1641	142.6	37.08	10970	20.52	
Collimator(10 cm ²)	16.56	16.56	11.16	11.52	21.96	-	

Tolerance: Maximum leakage radiation level at 1 meter from the focus should be ≤ 1 mGy (114mR) in one hr

Work load = 180 mA-min in one hour.

$$\text{Leakage radiation in one hour} = \frac{180 \text{ mAmin in 1hr} \times \text{Max Exposure level (mG/hr)}}{60 \times \text{mA used for measurement}}$$

Maximum radiation leakage from tube housing = **131.64 μGy/hr**. And maximum radiation leakage from tube collimator = **0.19872 μGy/hr**.

4.10 DETAILS OF RADIATION PROTECTION SURVEY OF THE INSTALLATION

kVp = 120, mA = 100 and time = 1 s.

Location	Max. Radiation level (mR/hr)	Weekly Radiation Level(mR/week)
Control console (Operator Position)	1.720	288.960
Behind windows (if applicable)	1.250	210.000
Outside the patient entrance door	0.026	4.368
Patient Waiting Area	0.022	3.696
AERB Limit	For the location of Radiation Workers: 20 mSv in a year (40 mR/week)	
	For the location of Member of Public: 1 mSv in a year (2mR/week)	

5 Conclusion

The Quality Assurance (QA) program for the diagnostic X-ray machine was successfully completed to ensure high-quality diagnostic imaging and patient safety. Based on the evaluation of the experimental data, the following conclusions were reached:

5.1 Equipment Performance and Accuracy

- **Field Congruence:** The congruence of the optical and radiation fields was found to be within 2% of the Focal spot to Film Distance (FFD), ensuring the clinical area of interest is accurately captured.
- **Beam Alignment:** Central beam alignment was verified using the cone test, with an observed tilt of less than 1.5° , meeting the required tolerance to prevent image distortion.
- **Filtration and kVp:** The total filtration was measured at 3.2 mm Al, satisfying the AERB recommendation for machines operating at maximum rated tube potentials. The kVp and timer accuracy remained within the ± 5 kV and $\pm 10\%$ error margins, respectively.

5.2 Output Consistency and Linearity

- **mA Linearity:** The Coefficient of Linearity (COL) was calculated as 0.010652, which is well within the acceptable limit of 0.1, indicating consistent radiation output across different mA stations.
- **Reproducibility:** The Coefficient of Variation (COV) for radiation output consistency was measured at 0.001004 at 80 kV and 0.00190 at 120 kV, both significantly lower than the 0.05 threshold.

5.3 Image Resolution and Safety

- **Spatial Resolution:** The system demonstrated a high-contrast sensitivity of 2.2 lp/mm , exceeding the minimum performance standard of 1.5 lp/mm .
- **Radiation Leakage:** Maximum leakage radiation from the tube housing and collimator was computed based on a 180 mA-min workload, resulting in **131.64 μGy** and **0.19872 μGy** in one hour, respectively. These values are well below the limit of 114 mR (1.0 mGy) per hour.
- **Protection Survey:** Radiation levels at the control console and public areas were confirmed to be within the permissible weekly limits for both radiation workers and the general public.

5.4 Final Summary

The diagnostic X-ray machine complies with all regulatory standards and performance criteria. It is concluded that the machine is operating optimally, providing reliable diagnostic information with minimum radiation risk to patients and staff.

References

- [1] K. H. Ng, R. Hill, A. Perkins, J. H. D. Wong, G. Clarke, C. H. Yeong, and N. M. Ung, *Problems and Solutions in Medical Physics: Three Volume Set*. Series in Medical Physics and Biomedical Engineering, CRC Press, 1st ed., 2022.