

Quality Assurance of a Diagnostic CT Machine




Date of experiment: 08 January 2026

Medical Physics Laboratory II

Name: Souvik Das

Roll No: 241126007



**National Institute of Science Education and Research
Jatani, Odisha, India**

Date of submission: April 26, 2026

Contents

1	Objective	1
2	Apparatus	1
3	Theory	1
3.1	Introduction to Computed Tomography Quality Assurance	1
3.2	Basic Principles of Computed Tomography (CT)	1
3.3	Description of the CT Machine Components	1
3.3.1	The Scanning Gantry	2
3.3.2	The Examination Table (Couch)	2
3.3.3	The Computer and Reconstruction System	2
3.3.4	The Control Console	2
3.4	Facility Design and CT Room Layout (AERB Guidelines)	3
3.5	Quality Assurance (QA) Parameters	4
3.5.1	Mechanical and Geometric Tests	4
3.5.2	Tests for High-Frequency Generators	4
3.5.3	Image Quality Evaluation	4
3.5.4	CT Dosimetry and Radiation Safety	5
4	Observation	5
5	Conclusion	5

1 Objective

To comprehensively evaluate the mechanical and dosimetric integrity of a diagnostic Computed Tomography (CT) scanner in accordance with established Quality Assurance (QA) protocols.

2 Apparatus

- Diagnostic CT Scanner
- Standardized CT QA Phantoms (Water, Slab, Catphan or equivalent)
- Positioning lasers and metric measuring tools

3 Theory

3.1 Introduction to Computed Tomography Quality Assurance

A Quality Assurance (QA) program in diagnostic radiology is a comprehensive set of protocols and procedures designed to ensure that diagnostic images are of sufficiently high quality to provide adequate diagnostic information, achieved at the lowest possible cost and with the least possible radiation exposure to the patient [1].

For a Computed Tomography (CT) unit, QA protocols must strictly address image quality, geometric alignment, and dosimetric output. Furthermore, when CT images are utilized for radiotherapy treatment planning (CT-Simulation), the spatial integrity of the scanner and the consistency of electron density values across the CT, Treatment Planning System (TPS), and Digitally Reconstructed Radiographs (DRRs) must be rigorously verified [2].

Quality assurance testing guarantees two fundamental tenets of radiation protection:

1. The radiation dose received by the patient for diagnostic purposes is optimally minimized (ALARA principle).
2. High image quality facilitates accurate diagnosis, thereby scientifically justifying the radiation exposure.

3.2 Basic Principles of Computed Tomography (CT)

Computed Tomography (CT) is a highly sophisticated medical imaging modality that utilizes X-ray radiation to produce cross-sectional images (slices) of the human body. Unlike conventional projection radiography, where a three-dimensional anatomical volume is superimposed onto a two-dimensional plane, CT generates true cross-sectional maps of X-ray attenuation coefficients. It accomplishes this via computationally reconstructing multiple angular projections acquired as an X-ray tube and detector array rotate around the patient.

The fundamental principle revolves around the measurement of radiation attenuation. An X-ray beam is restricted by collimators into a narrow fan or cone shape. As this beam transverses the patient, its intensity is attenuated through photoelectric absorption and Compton scattering. The transmitted beam intensity is measured by an array of detectors situated directly opposite the X-ray tube. The data set representing the attenuation profile from a single angle is termed a "projection." The gantry rotates seamlessly, establishing millions of discrete measurements over 360°. A computer system applies complex mathematical algorithms—most notably, Filtered Back Projection (FBP) or modern Iterative Reconstruction (IR) techniques—to map these attenuation measurements into a matrix of pixels. Each pixel is assigned a CT number (Hounsfield Unit, HU) indicative of the linear attenuation coefficient of the corresponding tissue voxel. This rigorous mathematical foundation not only eliminates structural superimposition but dramatically empowers the detection of subtle low-contrast differences among soft tissues compared to standard radiography.

3.3 Description of the CT Machine Components

A diagnostic state-of-the-art CT machine is an amalgamation of intricate mechanical, electrical, hardware, and software subsystems functioning in extraordinary synchrony. The core components include the Gantry, Examination Table, the Computer System, and the Control Console.

3.3.1 The Scanning Gantry

The gantry is the dominant mechanical structure in the CT suite and houses the principal X-ray generation and detection assemblies. It typically features a central aperture (commonly 70 to 80 cm in diameter) through which the patient couch freely translates. To accommodate different anatomical orientations, the entire gantry structure can often be tilted (usually up to $\pm 30^\circ$) from the vertical axis.

- **X-ray Tube:** Modern CT tubes are engineered to sustain immense thermal loads resulting from continuous, high-mA scanning required in multislice or helical CTs. They utilize rotating tungsten-rhenium anodes with large anode heat capacities and high cooling rates. The focal spot size directly bounds the system's geometric resolution; hence, many tubes incorporate "flying focal spot" technology, alternating the electron beam position to double the sampling frequency and reduce aliasing artifacts.
- **Detector Array:** Operating directly opposite the X-ray tube, the detector array captures exiting attenuation photon flux and transduces it into proportional electronic signals. Contemporary multi-detector CT (MDCT) units utilize solid-state scintillation detectors (such as Gadolinium Oxy-sulfide, $\text{Gd}_2\text{O}_2\text{S}$, or specialized ceramics). These rare-earth scintillators absorb X-rays efficiently to produce visible light flashes, which are instantaneously registered by integrated photodiodes. Essential characteristics of CT detectors include high geometrical efficiency, superior capture efficiency, negligible afterglow, and high dynamic range.
- **Data Acquisition System (DAS):** Positioned intimately within the gantry alongside the detectors, the DAS is responsible for rapidly measuring the minute analog photoelectric signals emanating from the detectors. These weak currents are amplified and converted mathematically into discrete digital data streams via analog-to-digital converters (ADCs). The colossal data packages are subsequently relayed to the computer system.
- **Slip Ring Technology:** This revolution enabled continuous helical (spiral) scanning. Unlike early generations where cables tethered the rotating components—necessitating a scan-and-stop operation to unwind—slip rings utilize conductive concentric tracks and metallic brushes. This mechanism transmits electrical power to the rotating tube and smoothly transfers continuous streams of digital data back to the stationary computer framework, facilitating unceasing sub-second rotations.
- **Collimators:** Beam collimation dictates the dose profile and z-axis slice width. *Pre-patient collimators* shape the primary X-ray fan/cone, controlling patient dose by eliminating penumbral, divergent edges, and establishing the dose profile width ($N \cdot w$). *Post-patient collimators* are situated precisely over the detector faces, significantly minimizing the detection of scattered photon trajectories which would drastically degrade image contrast.

3.3.2 The Examination Table (Couch)

The patient table provides immense structural rigidity combined with remarkable positional precision. During helical acquisitions, the table uniformly feeds the patient into the aperture simultaneously with continuous X-ray tube rotation. The bed must be fabricated from low Z-material (often carbon fiber) to avert significant X-ray attenuation and beam-hardening artifacts where the beam crosses through the structure. It must maintain mechanical indexing accuracy up to fraction-of-a-millimeter precision beneath weight loads exceeding 200 kg.

3.3.3 The Computer and Reconstruction System

The high-performance computer infrastructure bridges raw output and displayable image. It resolves millions of simultaneous linear equations generated by the DAS in essentially real time. It governs the entire hardware, stores retrospective raw data, interfaces via DICOM to the PACS (Picture Archiving and Communication System), and enables 3-dimensional multi-planar rendering, fluoroscopic CT, and complex dosimetric records.

3.3.4 The Control Console

The control console acts as the solitary user interface situated safely behind the lead-glass partition. It consists of multiple monitors, dedicated keypads for parameter manipulation (e.g., kV, mA, slice

thickness, pitch), and communication systems coupling the operator visually and audibly to the patient via intercoms.

3.4 Facility Design and CT Room Layout (AERB Guidelines)

Since computed tomography utilizes high-energy X-rays and large cumulative doses, structural radiation shielding and logical facility architecture are imperative for protecting staff, public, and adjacent zones against secondary scattered rays. In India, the Atomic Energy Regulatory Board (AERB) stringently regulates these layouts.

The basic guidelines explicitly prescribe:

- **Minimum Room Size:** The room housing the CT gantry must be sufficiently large to maneuver patients completely (minimum 25 sq. m usually depending on unit dimensions; minimizing dimensions around 4m). This safely dilutes in-room scatter by the inverse-square law.
- **Shielding Integrity:** The primary mechanism of exposure to individuals outside the CT room is scattered radiation from the patient's body and the gantry itself. Walls typically must consist of robust masonry (e.g., 9-inch solid brick) or be internally lined with specified lead equivalence (frequently 2 mm Pb). The access doors strictly require equivalent wood-and-lead sandwich protection (2 mm lead equivalence) to ensure zero weak points.
- **Control Console Placement and Viewing Glass:** To prevent occupational exposure, the technologist must operate the scanner entirely from a distinct control room or a heavily shielded alcove. Direct visual observation is uncompromisingly established using a specialized lead-glass window. As per the AERB Model Layout in the appended document, the viewing window is specified as 100 cm × 80 cm maintaining a strict **2.0 mm lead equivalence**.
- **Operational Layout:** The layout must minimize the scatter angle. In the prescribed structural blueprint (refer to Figure 1), the Examination Table (1) moves perpendicularly relative to the structural geometry, while the Control Unit (3) is protected dynamically from immediate sideways scattering. Warning fixtures, prominently a bright *Red Light* and radiation warning placards strictly adorn the external facade to deter entry during beam-on periods.




Figure 1: Model Room Layout for CT Scan Installation as prescribed by AERB. Components include: 1. CT gantry, 2. Examination Table, 3. Control Unit, 4. Electronics, 5. Viewing Glass (100 cm × 80 cm of 2.0 mm lead equivalence).

3.5 Quality Assurance (QA) Parameters

As per the guidelines mandated by regulatory bodies such as the Atomic Energy Regulatory Board (AERB), comprehensive acceptance testing upon new equipment installation and periodic QA are strict requirements. The QA tests for CT equipment are broadly classified into four primary categories: Mechanical Tests, Tests for High-Frequency Generators, Image Quality Parameters, and Radiation Safety [3].

3.5.1 Mechanical and Geometric Tests

The geometric fidelity of the scanner is critical to avoid image artifacts and ensure precise localization.

- **Alignment of the Table to Gantry:** This test ensures that the longitudinal axis of the patient table is perfectly horizontal and aligned with the vertical line passing through the scanner's rotational axis. A mismatch introduces severe image artifacts, particularly for bulky patients. The acceptable tolerance is ± 5 mm of the table midline.
- **Gantry Tilt Accuracy:** Most modern CT scanners allow non-orthogonal scanning by tilting the gantry. The accuracy of the displayed tilt angle is assessed by supporting an envelope-wrapped X-ray film vertically at the gantry end of the couch. The tolerance is typically $\pm 2^\circ$ [3].
- **Table Indexing Accuracy and Increment:** Couch translation must be precisely synchronized with the X-ray tube rotation. Using a ruler or tape measure placed alongside the table, the indicated degree of couch movement is compared with the actual physical distance moved. Tolerances specify differences extending over ± 1 mm.
- **Laser Positioning Accuracy:** Both internal gantry lasers and external wall-mounted alignment lasers guarantee iso-centering the patient. Max acceptable limit is typically ± 2 mm.

3.5.2 Tests for High-Frequency Generators

The X-ray generator parameters directly dictate the photon fluence and beam quality, directly impacting patient dose and image contrast.

- **Accuracy of Operating Potential (kVp):** The applied kilovoltage affects both the quality and quantity of the X-rays. Variations from the prescribed kVp will negatively alter image contrast. Performance bounds fall within $\pm 5\%$.
- **Reproducibility of Radiation Output:** Keeping mA and time fixed, the output is measured repeatedly at available kVp stations. Consistency is evaluated via the Coefficient of Variation (COV), maintaining $\text{COV} \leq 0.05$.
- **Linearity of Radiation Output (mA/mAs Linearity):** Keeping the kVp and exposure time constant, radiation output is measured across various mA stations. The Coefficient of Linearity (COL) is calculated spanning contiguous values and must fall under ≤ 0.1 .
- **Half-Value Layer (HVL) / Total Filtration:** Added filtration removes low-energy X-ray photons that do not possess sufficient energy to penetrate the patient and reach the detectors.

3.5.3 Image Quality Evaluation

Image quality parameters are assessed using standardized test phantoms of human shape or specific uniform test objects provided by the manufacturer.

- **CT Number Accuracy and Uniformity:** CT numbers (Hounsfield Units) depend on tube voltage, beam filtration, and object attenuation. By definition, the CT number of water is 0.
 - *Accuracy:* The measured CT number in the central Region of Interest (ROI) of a water phantom must be 0 ± 4 HU.
 - *Uniformity:* The CT number of a homogeneous object should remain constant across the entire field of view. The deviation between any peripheral ROI and the central ROI must not exceed ± 4 HU.

- **High Contrast Resolution (Spatial Resolution):** Defines the minimum size of a detail visualized with a contrast $> 10\%$. Expressed as spatial frequency (line pairs/cm), aiming normally for > 15 lp/cm.
- **Low Contrast Resolution:** This metric represents the smallest visible object at a specific low contrast level. It determines the system's capacity to differentiate tissues with minimal density differences relative to the surrounding background.
- **Slice Thickness (Z-axis Resolution):** Assesses whether the effective physical slice thickness matches the nominally selected thickness on the console. Tolerances define an error bracket ≤ 1 mm for sub-millimeter slices and $\pm 10\%$ for > 2 mm.

3.5.4 CT Dosimetry and Radiation Safety

The primary objective of these metrics aims exactly mapping standard absorbed doses.

- **Computed Tomography Dose Index ($CTDI_{100}$):** Incorporates practical parameters using a standard 100 mm pencil ionization chamber internally fixed inside standard body phantoms registering core dose lengths mathematically via:

$$CTDI_{100} = \frac{1}{N \cdot w} \int_{-50}^{50} D(z) dz \quad (1)$$

- **Weighted CTDI ($CTDI_w$):** Captures differences over standard volumetric bodies:

$$CTDI_w = \frac{1}{3}CTDI_{100, \text{Center}} + \frac{2}{3}CTDI_{100, \text{Periphery}} \quad (2)$$

- **Volume CTDI ($CTDI_{vol}$):** Scales linearly utilizing standard overlap metrics or Pitch factor (p):

$$CTDI_{vol} = \frac{CTDI_w}{\text{Pitch}} \quad (3)$$

- **Dose Length Product (DLP):** Standard metric summing absolutely up along structural axial axes L :

$$DLP = CTDI_{vol} \times L \quad (\text{mGy} \cdot \text{cm}) \quad (4)$$

4 Observation

5 Conclusion

A comprehensive understanding of the governing physical principles and stringent Quality Assurance protocols ensures optimal clinical outcomes and fundamental radiation safety.

References

- [1] J. T. Bushberg, J. A. Seibert, E. M. Leidholdt Jr, and J. M. Boone, *The Essential Physics of Medical Imaging*. Lippincott Williams & Wilkins, 2011.
- [2] S. Mutic, J. R. Palta, E. K. Butker, I. J. Das, M. S. Huq, L.-N. D. Loo, B. J. Salter, C. H. McCollough, and J. Van Dyk, "Quality assurance for computed-tomography simulators and the computed-tomography-simulation process: Report of the aapm radiation therapy committee task group no. 66," Tech. Rep. 10, 2003.
- [3] A. T. G. 39, "Specification and acceptance testing of computed tomography scanners," tech. rep., AAPM Report No. 39, 1993.