

Cs-137 A = 3.7 MBq Sept 2022 Distance = 50 cm

True dose rate (uR/h)	Average background counts(B) = 16.04						Corrected dose rate	Uncertainty
	1	2	3	4	5	Avg		
118.95	112	124	117	147	136	119.5	103.46	13.02227827
	126	109	115	122	129			
	124	106	121	115	114			
	119	112	130	117	116			
	113	111	120	119	125			
	111	115	123	121	117			
	117.5	112.8	121.0	123.5	122.8			



True dose rate (uR/h)	Average background counts(B) = 16.04						Corrected dose rate	Uncertainty
	1	2	3	4	5	Avg		
118.95	122	130	118	125	113	129.3	113.26	4.783522488
	131	138	131	128	123			
	136	143	125	130	122			
	140	135	123	122	135			
	139	124	117	135	137			
	137	129	123	138	130			
	134.2	133.2	122.8	129.7	126.7			



True dose rate (uR/h)	Average background counts(B) = 16.04						Corrected dose rate	Uncertainty
	1	2	3	4	5	Avg		
118.95	121	117	139	120	119	119.9	103.86	12.68600252
	118	125	132	106	120			
	120	128	124	105	118			
	113	122	128	109	119			
	112	120	130	110	122			
	118	119	126	117	119			
	117.0	121.8	129.8	111.2	119.5			

