

Dose Calibrator



Date of experiment: 08 January 2026

Medical Physics Laboratory II

Name: Souvik Das

Roll No: 241126007



National Institute of Science Education and Research
Jatani, Odisha, India

Date of submission: March 31, 2026

Contents

1	Objective	1
2	Apparatus	1
3	Theory	1
3.1	Ionization Chamber Mechanics	1
3.2	Routine Quality Assurance	1
4	Observation	1
5	Conclusion	1



1 Objective

To perform comprehensive Quality Assurance (QA) on a radionuclide dose calibrator, including assessments of precision, accuracy, system linearity, and geometric dependence.

2 Apparatus

- Radionuclide Dose Calibrator (Well-type Ionization Chamber)
- NIST-traceable standard check sources (e.g., ^{137}Cs , ^{57}Co)
- Radiopharmaceutical samples of varying activities (e.g., ^{99m}Tc)
- Calibration Canister and Vials/Syringes

3 Theory

The dose calibrator is the ultimate gatekeeper in nuclear medicine, acting as the final metrological check before radiopharmaceuticals are administered to patients. Assuring the highest degree of accuracy in these devices directly governs patient safety and the diagnostic or therapeutic efficacy of the procedure [1].

3.1 Ionization Chamber Mechanics

Dose calibrators are fundamentally deep, well-type ionization chambers, characteristically pressurized with Argon gas to 10–30 atmospheres [2]. The “well” geometry provides a near- 4π solid angle coverage, resulting in remarkably high geometric efficiency. When ionizing radiation traverses the pressurized gas, it generates electron-ion pairs. A stable, high potential difference across the concentric cylindrical electrodes sweeps these charge carriers away before recombination occurs, operating securely in the ionization saturation region (Townsend plateau).

The resulting ionization current—often on the order of pico-to-microamperes—is proportional to the absolute activity of the sample, scaled by the specific gamma-ray constant of the radionuclide. A highly sensitive electrometer digitizes this current, mapping it to user-selected isotope settings [3].

3.2 Routine Quality Assurance

The American Association of Physicists in Medicine (AAPM) outlines four mandatory QA procedures for dose calibrators [2]:

1. **Accuracy (Annually):** Validates the system against NIST-traceable reference sources (e.g., ^{57}Co , ^{137}Cs , ^{133}Ba). Measurements must lie within $\pm 5\%$ of the certified decay-corrected activity.
2. **Constancy (Daily):** Guarantees day-to-day reproducibility. A long-lived source (^{137}Cs) is measured across all frequently used isotope channels.
3. **Linearity (Quarterly):** Verifies that the electrometer responds linearly across the entire spectrum of clinical activities (from several Curies down to microCuries). This is commonly performed via the intrinsic decay method of ^{99m}Tc or utilizing calibrated attenuation sleeves (Calicheck).
4. **Geometry (Installation/Repair):** Assesses variations in response due to sample volume, container shape (syringe vs. vial), and position within the well.

4 Observation

5 Conclusion

A comprehensive understanding of the governing physical principles and stringent Quality Assurance protocols ensures optimal clinical outcomes and fundamental radiation safety.

References

- [1] S. R. Cherry, J. A. Sorenson, and M. E. Phelps, *Physics in Nuclear Medicine*. Elsevier Health Sciences, 2012.
- [2] B. E. Zimmerman *et al.*, “The selection, use, calibration, and quality assurance of radionuclide calibrators used in nuclear medicine,” tech. rep., AAPM Report No. 181, 2012.
- [3] G. F. Knoll, *Radiation Detection and Measurement*. John Wiley & Sons, 2010.