

Thyroid uptake



Date of experiment: 08 January 2026

Medical Physics Laboratory II

Name: Souvik Das

Roll No: 241126007



National Institute of Science Education and Research
Jatani, Odisha, India

Date of submission: March 31, 2026

Contents

1	Objective	1
2	Apparatus	1
3	Theory	1
3.1	Detection Physics: NaI(Tl) Scintillators	1
3.2	Quality Assurance of the Uptake Probe	1
3.3	Uptake Calculation	1
4	Observation	1
5	Conclusion	1

1 Objective

To calibrate a NaI(Tl) scintillation thyroid uptake probe, execute standard performance QA metrics, and quantitatively assess radioiodine fractional uptake from patient data.

2 Apparatus

- Thyroid Uptake Probe Unit (e.g., CAPTUS 4000e)
- Multi-Channel Analyzer (MCA) software interface
- Certified Check Sources: ^{137}Cs , ^{152}Eu , or ^{133}Ba
- Standard Lucite Neck Phantom

3 Theory

The clinical evaluation of thyroid physiology heavily relies on the gland's inherent mechanism of trapping and organifying iodine. By administering a known activity of a radioactive iodine isotope—such as ^{131}I or ^{123}I —the fractional uptake can be quantitatively determined. This provides profound insight into hyperthyroid or hypothyroid conditions [1].

3.1 Detection Physics: NaI(Tl) Scintillators

Thyroid uptake probes universally employ Thallium-activated Sodium Iodide [NaI(Tl)] scintillation crystals. Incident gamma rays (e.g., the 364 keV emission of ^{131}I) interact via the photoelectric effect and Compton scattering within the crystal lattice. The energy deposited raises electrons to the conduction band, which migrate to Thallium activation centers and subsequently emit visible photons (approximately 415 nm) upon de-excitation [2].

A photomultiplier tube (PMT) coupled to the crystal converts these scintillations into an amplified electrical pulse. The pulse amplitude is strictly proportional to the energy deposited by the incident gamma ray, allowing for energy discrimination via a Multi-Channel Analyzer (MCA).

3.2 Quality Assurance of the Uptake Probe

To ensure clinical integrity, the system must undergo stringent QA:

- **Energy Calibration and Resolution:** Performed using a standard ^{137}Cs source (662 keV). The resolution, $R = (\text{FWHM}/E_0) \times 100$, typically must not exceed 9.5% for flat-faced probes [1].
- **Constancy:** A daily verification to ensure that the measured activity of a long-lived standard source aligns with its theoretical decay-corrected value.
- **Chi-Square (χ^2) Test:** Evaluates the statistical integrity of the counting system, assessing whether observed counts follow a true Poisson distribution. It is formulated as:

$$\chi^2 = \sum_{i=1}^N \frac{(O_i - \bar{O})^2}{\bar{O}} \quad (1)$$

3.3 Uptake Calculation

Thyroid uptake is expressed as the ratio of background-corrected net neck counts to the standard dose counts:

$$\% \text{ Uptake} = \left(\frac{\text{Neck Counts} - \text{Thigh (Background) Counts}}{\text{Standard Capsule Counts} - \text{Room Background}} \right) \times 100 \quad (2)$$

4 Observation

5 Conclusion

A comprehensive understanding of the governing physical principles and stringent Quality Assurance protocols ensures optimal clinical outcomes and fundamental radiation safety.



References

- [1] S. R. Cherry, J. A. Sorenson, and M. E. Phelps, *Physics in Nuclear Medicine*. Elsevier Health Sciences, 2012.
- [2] G. F. Knoll, *Radiation Detection and Measurement*. John Wiley & Sons, 2010.