

IFU QC Automation Report

Source file: faulty_IFU_document.pdf
Generated: 30 Jul 2026, 22:04

Result	Failures	Warnings
FAIL	6	3

1. Page Number Verification

Severity	Page	Issue
FAIL	2	Declared total pages (5) does not match actual document length (4)
FAIL	4	No page number found matching expected pattern 'Page\s+(\d+)\s+of\s+(\d+)'
FAIL	—	Duplicate page number '2' found on physical pages [2, 3]
FAIL	—	Missing page number(s): [3, 4]

2. Manufacturer Information

Severity	Page	Issue
FAIL	—	Manufacturer address does not match exactly
WARN	—	Multiple distinct email addresses found, verify consistency: ['sales@acmemedical.example', 'support@acmemedical.example']

3. Regulatory Symbols

Severity	Page	Issue
WARN	—	Label/caption for symbol 'Temperature limit' not found in extracted text — verify the symbol graphic is present manually or via image-based check
WARN	—	Label/caption for symbol 'Keep dry' not found in extracted text — verify the symbol graphic is present manually or via image-based check

4. Date Verification

Severity	Page	Issue
FAIL	—	Required date field 'Effective date' not found in document