

# Remote Photoplethysmography (rPPG) using RGB Camera using Peak Detection Algorithm for Measurement of Heart Rate Variability (HRV)

Prof. Prasad AM<sup>1</sup>, VinayAchari N<sup>2</sup>, Tanmaya Thejana K Nazare<sup>3</sup>

<sup>1</sup>UG Student, Department of Computer Science and Engineering, Dayananda Sagar College of Engineering, Bengaluru, Karnataka, India

<sup>2</sup>Associate Professor, Department of Computer Science and Engineering, Dayananda Sagar College of Engineering, Bengaluru, Karnataka, India

\*\*\*

**Abstract** -In this Literature Survey, a method for digital digicam based entirely coronary heart rate detection, often called Photoplethysmography Imaging (PPG), is demonstrated. The pulse signals are obtained from RGB videos of a human faces taken using an available digital digicams in low-light situations. The set of principles for estimating coronary HR is primarily according to the beat to beat examination of the obtained PPG sign, which allows for the estimate of psychophysiological factors such as coronary heart rate variability. We create an optical rPPG sign version from the optical properties like human pores on skin, allowing origins of the RIPPG sign and movement artefacts to be precisely defined. The skin's place of interest (ROI) is marked as lambertian radiator. The impact of ROI monitoring is assessed using radiometry. We recommend an adaptive colour distinction operation among the inexperienced and purple channels to reduce movement artefacts by thinking of a virtual colour digital digicam as a simple- spectrometer. Then we provide adaptive bandpass clearout to remove residual rPPG movement artefacts based on spectral characteristics of photoplethysmography indicators. To improve the RIPPG sign quality, we combine ROI selection at the person's cheeks with accelerated strong functions factor monitoring. In comparison to contemporary face video- based fully rPPG procedures, experimental data reveal that the proposed rPPG can achieve significantly increased overall performance in obtaining coronary HR in shifting people. The findings show that PPGI could be a viable option for conveying crucial signal data to drivers in automobiles.

**Key Words:** Remote Photoplethysmography, Heart-Rate (HR), Heart-Rate Variability (HRV), Respiration-Rate (RR), region-of-interest (ROI), Contactless HR detecting; Beat to beat analysis; Online Heart-rate monitoring; Driver-state monitoring system.

## 1. INTRODUCTION

Recently, a number of courses have focused on photoplethysmography imaging (PPG) as a method for achieving contactless crucial signal monitoring, launching a vast field of tracking and biofeedback applications. The colour changes caused by blood-perfusions can be observed by analysing and processing videos of a skin's location. Data about a subject's respiration rate is included in these colour

modifications. HRV (heart-rate variability) is a very important and primary biometric indication to indicate existence of multiple disorders. Obtaining real- time HRV information for patients and other individuals can be difficult, since many people cannot own expensive clinical devices or appropriate tracking device. Traditional tiny heart rate detection mechanisms, such as oximeters and smartwatches, need regular skin-contact and can be an added cost for many people. Verkruysse described an early method for detecting pulse rate and respiration rate in video recordings of the human face. The recovered signal from the green digital digicam channel is bandpass filtered, allowing the heartbeat and respiration rate to be located within the frequency spectrum using a Fast Fourier Transform. We will use statistical BSS(Blind Source Separation) noise detection algorithms to segregate and separate preferred blood-pulse indications and background noise, using those character inputs as neutral sources. In order to improve the accuracy of HRV results, we can divide our face ROI into a number of tiny and neutral zones. Faraway photo-plethysmography is another name for non- invasive vital signal tracking (r-PPG). Because of the tracking of remotely captured video using virtual digital digicam, the term r-PPG was developed.

## 2. LITERATURE SURVEY

TimonBlöcher, "An online PPG approach for camera- based HR monitoring using beat to beat detection", (2017- IEEE )

A method for digital digicam based entirely coronary heart rate detection, often called Photoplethysmography Imaging (PPG), is demonstrated. The pulse signals are obtained from RGB videos of a human faces taken using an available digital digicams in low-light situations. The set of principles for estimating coronary HR is primarily according to the beat to beat examination of the obtained PPG sign, which allows for the estimate of psychophysiological factors such as coronary heart rate variability. We create an optical rPPG sign version from the optical properties like human pores on skin, allowing origins of the RIPPG sign and movement artefacts to be precisely defined. The skin's place of interest (ROI) is marked as lambertian radiator. The impact of ROI monitoring is assessed using radiometry. We recommend an adaptive colour distinction operation among the inexperienced and purple channels to reduce movement

artefacts by thinking of a virtual colour digital digicam as a simple- spectrometer. Then we provide adaptive bandpass clearout to remove residual rPPG movement artefacts based on spectral characteristics of photoplethysmography indicators. To improve the RPPG sign quality, we combine ROI selection at the person's cheeks with accelerated strong functions factor monitoring. In comparison to contemporary face video-based fully rPPG procedures, experimental data reveal that the proposed rPPG can achieve significantly increased overall performance in obtaining coronary HR in shifting people. The findings show that PPGI could be a viable option for conveying crucial signal data to drivers in automobiles.

*Kazi Shafiu Alam "Remote Heart Rate and Heart Rate Variability Detection and Monitoring from Face Video with Minimum Resources", (2020 IEEE )*

This document includes cloud derived heart-rate (HR) and heart-rate-variability (HRV). This is a description of the monitor. Monitor someone's HR and HRV using very little effort on the user's side. This method of HR and HRV monitoring does not need touch sensing devices or pricey equipment. It also doesn't need a high res camera to record the subject's face on video. To video record faces, this HR surveillance method just uses a web-cam on your PC or a camera on a portable mobile devices such as phones and tablets. Before uploading the video segment to server, system uses user's device's resources to record and encode it. On the server, all operations for determining HR and HRV take place. The number of faces recognised will be tiny relative to the number of frames if recorded in the dark or if the video isn't well focused on subject's face. This will not be sufficient to calculate HR variability information. Future-Research should concentrate on face and colour detection in order to ensure there is a constant number of subject's faces in all undesirable conditions. The prototype gadget was successfully tested and shown good performance. It can deliver high-quality photoplethysmogram data while uniformly illuminating human skin. The device's blood volume signal has a high temporal-resolution . The capacity to discern higher-order pulse pulses. The prototype gadget had a very good signal-to-noise ratio and could be used for tracking change in human blood volume in the palm of hand.

*Edgars Kviesis-Kipge "Remote-photoplethysmography (rPPG) device for monitoring of blood volume changes with high temporal resolution", (2016-IEEE)*

An electrical board with twelve precisely regulated brilliant near-infrared LED lights placed in a circle, a high- speed video cam, and a battery charging circuit design to make up the prototype gadget. The device was evaluated in the lab and found to be suitable for noncontact monitoring of changes in human blood volume in the palm. The prototype gadget was successful and showed excellent performance. It provides consistent lighting for human skin, resulting in

high-quality photoplethysmogram data. The device's blood-volume signals have a high-temporal resolution and the capacity to observe higher-order pulse pulses. Overall, the prototype gadget demonstrated an excellent signal-to-noise ratio and may be utilised to track changes in human blood volume in the hand.

*Alexandra Dunaeva " Video-Analysis Methods for Remote Measurement of Respiration Characteristics(HR) and Heart Rate Variability(HRV)", (2020-IEEE)*

The focus of the paper is on video sequence processing methods for analysing a patient's physiological features in the harsh environment of an ICU. We provide a methodology of forming HRV using a (rPPG) approach, and we compare the received signals to original ECG signals. The rPPG signal exhibits high accuracy in the extremely low frequency band, according to experimental data. Furthermore, we demonstrate that breathing assessment techniques allow for the assessment of the person's breathing as well as forced-breathing during mechanical-ventilation for identifying the respiratory-rate and time when the person's own breathing is restored. For that purpose, a video sequence processing technique that was created for non-contact evaluation of respiratory parameters in patients undergoing surgical resuscitation in the critical care unit is proposed. The suggested technique is suitable for producing sufficiently informative signals that may extract information on the patient's breathing frequency and qualitative features, according to experimental validation.

*Sungjun Kwon "Region of Interest analysis for remote photoplethysmography(rPPG) on facial video", (2015- IEEE )*

Due to the development of camera-equipped gadgets in ordinary living contexts, camera-based Remote Photo Pletismography (PPG) has become popular. There is a fantastic potential to employ physiological monitoring on a daily basis. The region of interest (ROI) in camera-based remote PPG (rPPG) monitoring is related to signal quality and the computing burden of the signal extraction procedure. For computationally effective rPPG extraction, defining the best ROI for the body while limiting size is critical. We investigated the face region in depth in this work and discovered a computationally efficient ROI for facial rPPG extraction. The signal to noise ratio was compared to high correlation area ratio, and the mean and the standard deviation (SD) of SNR, as well as the correlation coefficient, were used to assess signal quality in each region. The findings suggest that the forehead and both cheeks are particularly attractive candidates for a computationally efficient ROI. The signal quality from the lips and chin, on the other hand, was poor. To obtain an efficient ROI, nose and nose have limits. The signal quality from the lips and chin, on the other hand, was poor. To obtain an efficient ROI, nose and nose have limits. The face was separated into seven halves (forehead, left and right cheeks, nose, mouth, nose,

chin, as shown in Figure 3). The area ratio of the mouth and chin (lower region) was low, while the area ratio of the forehead, nose, nose, and both cheek sections (high area) was very large.

Po-Wei Huang "A Heart Rate Monitoring Framework for Real-World Drivers Using Remote Photoplethysmography", (IEEE-2020)

A driver is Remote PhotoPlethysmography (rPPG). However, difficulties such as driver behaviour, motion artefacts, and changes in illumination levels all make heart rate monitoring difficult when driving. Heart rate periodicity is an often utilised assumption when dealing with interference (or spectral sparsity). Several strategies increase stability at the price of heart rate variability sensitivity monitoring. Adaptive spectrum filter banks (ADs) give explanatory spectral filter bank tuning options to establish a compromise between robustness and sensitivity, and statistical-signal processing (SSP) and Monte-Carlo simulations are used to determine outlier probability. It is offered as a new method of providing. Driving circumstances may be monitored from afar. In addition, we will develop an operational database with over 23 hours of data to test the suggested method. The effect of driving behaviours (both amateur and professional), vehicle type (compact cars and buses), and route on rPPG will also be investigated. The mean absolute errors for scenarios "passenger," "small car," and "bus" are 3.42, 7.86 and 5.03 bpm respectively, when compared to the most recent rPPG in driving situations. Furthermore, AD placed first and third in the inaugural Remote Physiological Signal Acquisition (RePSS) competition, which has a low computing complexity. The suggested technique, which is based on SSP and Monte-Carlo simulation, generates a probabilistic model that allows for a better balance of robustness and sensitivity. The HR MAE will be decreased to 3.42 bpm, 7.86 bpm, and 5.03 bpm for passenger cars, compact cars, and buses, respectively. The suggested approach outperforms benchmarking techniques and leads to improved driving situations, according to comparisons across and within datasets.

R. M. Fouad," Optimizing RPPG using Adaptive Skin-Segmentation, for Real-Time HR Monitoring", (2019-IEEE)

The total heart rate (HR) algorithm is evaluated for accuracy and dependability. Did. HR data is not collected in areas where there is no skin. However, there are few studies that address the problem of non-skin pixels in ROIs. To begin, this study explores how to improve quality of remote Photoplethysmography signals by removing non skin pixels from the Region of Interest. Use of skin- segmentation for ROI determination is shown to be feasible. Then, we compare this strategy to our earlier real time rPPG model. Furthermore, we investigate the impact of signal fusion in extraction of Heart Rate from 3 ROIs. Then , we provide a detailed description of all the models in the algorithm for

identification of faces, face-tracking, skin-detection, and blind-signal separation that have been investigated. Finally, compare the rPPG results to the ground truth data from a commercially available pulse oximeter. The suggested technique considerably increases the quality of rPPG technology, according to simulation findings.

Sebastian Zaunseder "Heart Beat Detection and Analysis from Videos", (2014-IEEE)

Photoplethysmography using a camera allows for remote monitoring of cardiovascular parameters. Heart rate detection using a remote photoplethysmogram (rPPG) utilising a wavelet-based detection technique and time delay estimation is described in this white paper. We show that, depending on the region of interest chosen, we can obtain identification accuracy of 95 percent to 99 percent based on experimental data of 18 healthy volunteers at rest. Time delay estimations have been found to increase acquisition time accuracy. Non-quiet detection and automated identification of accessible signal segments will be the subject of future research. I showed you how to use rPPG to identify individual beats and align them. Despite certain limits, our research has demonstrated that extremely precise single heart rate detection in rPPG is possible, and that TDE aids temporal alignment. To deal with moderate movement circumstances in the future, the suggested processing method should be combined with appropriate pretreatment techniques, such as chrominance based algorithms.

Wenjin Wang "Algorithmic Principles of Remote- PPG", (2016 IEEE)

In order to obtain a deeper grasp of the computational concepts underpinning multi-wavelength remotes, this study merges the important optical and physiological aspects of skin responses into mathematics. An algorithm is shown. Photoplethysmography should be improved (rPPG). This model is used to explain the different momentum extraction decisions made by existing rPPG algorithms. The model's findings may be utilised to create rPPG solutions that are both robust and application- specific. This is demonstrated by creating an alternate rPPG approach for momentum extraction that employs a projection plane perpendicular to the skin colour. The proposed model may be utilised to understand the relative benefits of the various rPPG approaches presented based on the major benchmarks. This model is used to determine the similarities and differences in existing rPPG techniques for extracting the pulse. Our findings show that combining the model with different assumptions allows for the creation of a variety of pulse extraction algorithms. We also propose an alternative method called POS, which is similar to CHROM but changes the order in which the main expected distortions are reduced using different priors. To corroborate our knowledge, a large benchmark with a variety of obstacles is run on current and newly suggested rPPG approaches.



Philipp V. Rouast, "HR measurement using low-cost RGB face video"(All content following this page was uploaded by Philipp V. Rouast on 22 October 2017. Research Gate)

Remote photoplethysmography (rPPG) is a technique for measuring heart rate from a distance utilizing less expensive camera. In this paper, one can examine the evolution of rPPG from its inception in early 2008. This paper characterizes current rPPG techniques thus developing a framework that shows how modular phases are organised. Developers could utilize this categorization and create rPPG programs and algorithms suiting their requirements based on this framework. Researchers can enhance certain elements of a rPPG algorithm by starting with the evaluated and categorized methods. Clearly, remote HRM can be done with low-cost video equipment, and prior research suggest that rPPG is becoming more sophisticated. rPPG for HR and HRV have a wide range of application. The increase in researches over the course of this study implies that trustworthy rPPG algorithms are becoming more popular.

Litong Feng "Dynamic ROI based on K-Means for Remote PPG", (2018 IEEE)

Contactless human vital sign monitoring is possible using remote imaging photoplethysmography (RIPPG). Though the remote operating mode is convenient in rPPG, the quality of the PPG signal is limited by the remote-nature of the equipment. In clinical uses of PPG, increasing the accuracy and PPG signal quality is a must requirement. Because the RIPPG's region of interest (ROI) varies from just a point to a larger area, there is a new method to improve signal quality by refining and enhancing the region of interest. Here we present a dynamic Region of Interest for PPG in this study, which automatically picks the person's skin areas that correlate to high-quality PPG signals. To begin, a fixed ROI is partitioned into non-overlapping parts. Two characteristics are then supplied to perform no-reference quality evaluation for RIPPG signals from various blocks. In a two-dimensional feature space, K-means clustering is then used. A dynamic ROI for a video segment may be computed based on the clustering result, which is updated every two seconds. Nineteen healthy individuals were recruited to test the proposed ROI selection method on the face and palmar regions. Experimental findings of heart rate measurement reveal that the suggested dynamic ROI method for RIPPG may effectively increase RIPPG signal quality when compared to state-of-the-art RIPPG ROI methods.

### 3. CONCLUSIONS

In this job, we present a PPGI system that uses a commercially accessible camera and ambient light to extract human pulse rate online. The ROI is the region on the forehead. Several image and signal processing algorithms are included in the proposed algorithm. Independent component analysis and spatial filtering To estimate the needed values, a peak detection technique was created utilising the signal's

Hilbert transform. This method enables for real time estimations of heart rate from beat to beat. Signal quality is continually checked using three separate artefact indicators to mitigate the impacts of motion and light artefacts. When compared to the ECG system under laboratory circumstances, the results of eight participants' tests revealed improved accuracy in pulse peak recognition and heart rate estimate. Future studies will include looking into ways to decrease such artefacts, such as picture stabilisation technologies. Another objective is to combine with an IR-based system for vital sign-based nocturnal tiredness detection and porting to embedded devices.

### REFERENCES

- [1] Timon Blöcher, "An online PPGI approach for camera based heart rate monitoring using beat-to-beat detection", (2017 IEEE ).
- [2] Kazi Shafiul Alam "Remote Heart Rate and Heart Rate Variability Detection and Monitoring from Face Video with Minimum Resources", (2020 IEEE ).
- [3] Edgars Kviesis-Kipge "Remote photoplethysmography device for monitoring of blood volume changes with high temporal resolution", (2016 IEEE)
- [4] Alexandra Dunaeva "Video Analysis Methods for Remote Measurement of Respiration Characteristics and Heart Rate Variability", (2020 IEEE)
- [5] Sungjun Kwon "ROI analysis for remote photoplethysmography on facial video", (2015 IEEE)
- [6] Po-Wei Huang "A Heart Rate Monitoring Framework for Real-World Drivers Using Remote Photoplethysmography", (IEEE JOURNAL OF BIOMEDICAL AND HEALTH INFORMATICS, VOL. , NO. , 2020 )
- [7] R.M.Fouad, "Optimizing rPhotoplethysmography Using Adaptive Skin Segmentation for Real-Time Heart Rate Monitoring", (2019 IEEE )
- [8] Sebastian Zaunseder "Heart Beat Detection and Analysis from Videos", (2014 IEEE).
- [9] Wenjin Wang "Algorithmic Principles of Remote-PPG", (2016 IEEE )
- [10] Philipp V. Rouast, Remote heart rate measurement using low-cost RGB face video (All content following this page was uploaded by Philipp V. Rouast on 22 October 2017. Research Gate)
- [11] Wenjin Wang "Novel Algorithm for Remote Photoplethysmography: Spatial Subspace Rotation ", (2015 IEEE)

[12] Litong Feng "DYNAMIC ROI BASED ON K-MEANSFOR REMOTE PHOTOPLETHYSMOGRAPHY", ( 2018 IEEE ).

[13] D.J. McDuff, J.R. Estepp, A.M. Piasecki, and E.B. Blackford; "A survey of remote optical photoplethysmographic imaging methods," in Engineering in Medicine and Biology Society (EMBC), 37<sup>th</sup> Annual International Conference of the IEEE, pp. 6398–6404, August 2015.

[14] J.T.B. Moyle; "Pulse Oximetry", 2nd Edn. Published by BMJ Books, London, 2002.

[15] Y. Sun and N. Thakor; "Photoplethysmography revisited: from contact to noncontact, from point to imaging," in IEEE Transactions on Biomedical Engineering, vol. 63, pp. 463-477, 2016