URITECT Unresolved Evidence Gaps

Read-only audit of the current repository implementation
Repository snapshot: 11 October 2026 | Frozen app: 1.5.11+28

Highest-priority gaps

Rank	Issue	Affected component	Consequence	Required action	Expertise
1	No prospective culture-confirmed validation	Bayesian/clinical outputs	No diagnostic accuracy or calibration claim	Independent protocol and outcome study	Clinical/statistical
2	Physical strip/IFU not identified	All ten analytes and timing	Mappings and 60-second assumptions unverified	Record manufacturer, lot, expiry, IFU, chart, timings	Two MedTechs
3	Exact-build device checklist all NOT_RUN	Offline, camera, history, safety, latency	No field acceptance evidence	Run D01–D15 on two phones	Mobile/QA
4	Male residual LRs and female LR multiplication unresolved	Bayesian models	Potential evidence mismatch/calibration error	Statistical review with retain/revise/withhold decision	Statistician
5	Confirmation data not untouched; rare abnormal support	Classifier	Inflated/uncertain generalization	Independent leakage-controlled evaluation	ML/statistical
6	Expert and physician reviews pending	Content, renal/safety, ISO quality	No professional endorsement	Complete signed current-build review	MedTech/physician/IT
7	Manuscript evidence sources/version drift	Thesis documents	Unsupported or outdated claims	Run consistency checklist	Thesis authors

Required phrase for absent evidence

NOT VERIFIED — supporting evidence not located.