

Gene Pulser® Electroprotocols

* We recommend adapting this protocol to use the Gene Pulser electroporation buffer (catalog #165-2676, 165-2677), which increases cell viability and transfection efficiency in mammalian cell lines.

Cell Type Mammalian, adherent

Species Used Rat, L-6, myoblast

Molecules Electroporated DNA: supercoiled DNA used for transient transfections.

Before the Pulse

Cell Growth Medium DMEM, 10% Fetal Calf Serum (FCS) (GIBCO/BRL, Sigma)

Growth Phase at Harvest 50 to 70% confluency

Pre-pulse Incubation 4° C, 10 min. (optional: add 50 µl FCS if using HeBS as electroporation media; 50 µl salmon sperm DNA for transient transfections).

Wash Solution Wash two times in electroporation buffer.

The Pulse

Instruments Used Gene Pulser® apparatus & Capacitance

Electroporation Temperature Room temperature

Electroporation Medium* HEPES Buffered Saline, 6mM glucose, (optional: add 50 µl FCS, 50 µl salmon sperm DNA).

Cell Density 5 x 10 (6) cells / pulse

Volume of Cells 0.5 ml

DNA Concentration 10 µg / pulse

DNA Resuspension Buffer Not given; final volume: 0.8 ml

Volume of DNA Not given; final volume: 0.8 ml

Cuvette Gap 0.4 cm

Voltage 0.350 kV

Field Strength 0.875 kV/cm

Capacitor 960 µF

Resistor (Pulse Controller) Ω none

Time Constant 25.0 msec

After the Pulse

Outgrowth Medium F12, 10% Fetal Calf Serum (FCS)

Relevant Publications and/or Comments

Note: exponential values designated in parentheses.

HBS: 10mM HEPES, pH 7.2, 150 mM NaCl, 5 mM CaCl₂

Outgrowth Temperature 37 °C

Length of Incubation 48 to 72 hrs.

Selection Method or Assay Used Transient assays

Electroporation Efficiency Not given

Per Cent Survival about 50%

Name of Submitter Jackie Beall

Institution Address University of Adelaide
Department of Biochemistry
P.O. Box 498
Adelaide 5001
Australia

Survey Number

149