

Cell Type	Bacterial, gram positive	Molecules Electroporated	DNA: plasmid, pAL 5000
Species Used	<i>Mycobacterium smegmatis</i> , ATCC 607		

Before the Pulse

Cell Growth Medium	Mycobacteria, 7H11 (Difco)	Growth Phase at Harvest	O.D. (600) = 0.5 to 1.0
		Pre-pulse Incubation	None

Wash Solution 272 mM sucrose, 7 mM potassium phosphate, 1 mM MgCl₂, pH 7.0

The Pulse

Instruments Used Gene Pulser® apparatus

Electroporation Temperature	Room temperature		
Electroporation Medium	272 mM sucrose, 7 mM potassium phosphate, 1 mM MgCl ₂ , pH 7.0	Cuvette Gap	0.2 cm
Cell Density	1 X 10 ⁹ (9) cells / ml	Voltage	2.5 kV
Volume of Cells	400 µl	Field Strength	12.5 kV/cm
DNA Concentration	10 µg / ml		
DNA Resuspension Buffer	TE buffer (10 mM Tris, 1 mM EDTA, pH 8.0)	Capacitor	25 µF
Volume of DNA	10 to 20 µl	Resistor	(Pulse Controller) 200 Ω
		Time Constant	4.5 to 5.0 msec

After the Pulse

Outgrowth Medium Mycobacteria, 7H11 (Difco)

Relevant Publications and/or Comments

Note: exponential values designated in parentheses.

Outgrowth Temperature	37 °C
Length of Incubation	15 to 21 days
Selection Method or Assay Used	Not given
Electroporation Efficiency	Not given
Per Cent Survival	Not given

Name of Submitter Dr. Gustavo Ortega

Institution Address
 Instituto De Biomedicina
 Ingeniería Genética
 Postal 4043
 Caracas, Venezuela 1010A

Survey Number
 080