

Cell Type	Bacterial, gram positive	Molecules Electroporated	DNA: plasmids, various sizes
Species Used	<i>Mycobacterium smegmatis</i>		

Before the Pulse

Cell Growth Medium	Middlebrook 7H9 Broth (Difco) [MB]	Growth Phase at Harvest	O.D. (600) = mid- log
		Pre-pulse Incubation	None

Wash Solution 10% glycerol

The Pulse

Instruments Used Gene Pulser® apparatus

Electroporation Temperature Room temperature

Electroporation Medium 10% glycerol

Cuvette Gap 0.2 cm

Cell Density Not given

Voltage 1.25 kV

Volume of Cells 100 µl

Field Strength 6.25 kV/cm

DNA Concentration varies

Capacitor 25 µF

DNA Resuspension Buffer 400 to 600 µl MB

Resistor (Pulse Controller) 800 Ω

Volume of DNA Not given

Time Constant 9 msec

After the Pulse

Outgrowth Medium MB

Relevant Publications and/or Comments

Note: exponential values designated in parentheses.
Ref: Lee, M.H., Pascopella, L., Jacobs, Jr., W.R. and Hatfull, G.F., *PNAS* **88**:3111-5, 1991.
 See also: Snapper, *et.al.* *PNAS* **85**:6987-6991, 1988.

Outgrowth Temperature 37 °C

Length of Incubation 2 hours

Selection Method or Assay Used MB / usually kanamycin sensitive

Electroporation Efficiency Usually very good

Per Cent Survival Not given

Name of Submitter Lisa Peterson / Research Technician

Institution Address University of Pittsburgh,
 Biology Department, 376 Crawford,
 Pittsburgh, PA 15260

Survey Number
 079