

US Patent & Trademark Office

Patent Public Search | Text View

United States Patent Application Publication

20250249288

Kind Code

A1

Publication Date

August 07, 2025

Inventor(s)

Meir; Ivan et al.

METHOD OF CALIBRATION OF A STEREOSCOPIC CAMERA SYSTEM FOR USE WITH A RADIO THERAPY TREATMENT APPARATUS

Abstract

The disclosed calibration method includes a calibration phantom positioned on an adjustable table on the surface of a mechanical couch, with the phantom's centre at an estimated location for the iso-centre of a radio therapy treatment apparatus. The calibration phantom is then irradiated using the apparatus, and the relative location of the center of the calibration phantom and the iso-centre of the apparatus is determined by analyzing images of the irradiation of the calibration phantom. The calibration phantom is then repositioned by the mechanical couch applying an offset corresponding to the determined relative location of the centre of the calibration phantom and the iso-centre of the apparatus to the calibration phantom. Images of the relocated calibration phantom are obtained, to which the offset has been applied, and the obtained images are processed to set the co-ordinate system of a stereoscopic camera system relative to the iso-centre of the apparatus.

Inventors: Meir; Ivan (London, GB), Allen; Martin (London, GB), Hale; Gideon (London, GB), Smith; Norman (London, GB), Howe; Robert (Barcelona, ES)

Applicant: Vision RT Limited (London, GB)

Family ID: 49081425

Appl. No.: 19/051239

Filed: February 12, 2025

Foreign Application Priority Data

GB	1312808.7	Jul. 17, 2013
WO	PCT/GB2014/052125	Jul. 11, 2014

Related U.S. Application Data

parent US continuation 18487148 20231016 parent-grant-document US 12251580 child US 19051239
parent US continuation 18124273 20230321 parent-grant-document US 12220601 child US 18487148
parent US continuation 17159845 20210127 parent-grant-document US 11633629 child US 18124273
parent US continuation 16578905 20190923 parent-grant-document US 10933258 child US 17159845
parent US continuation 15946362 20180405 parent-grant-document US 10456601 child US 16578905
parent US continuation 14899112 20151216 parent-grant-document US 9962561 child US 15946362

Publication Classification

Int. Cl.: **A61N5/10** (20060101); **A61B6/58** (20240101); **G06T7/00** (20170101); **G06T7/80** (20170101)

U.S. Cl.:

CPC **A61N5/1075** (20130101); **A61B6/584** (20130101); **A61N5/1049** (20130101); **G06T7/0012** (20130101); **G06T7/85** (20170101); **A61N2005/1059** (20130101); **A61N2005/1076** (20130101); **G06T2207/30004** (20130101)

Background/Summary

CROSS REFERENCE TO RELATED APPLICATIONS

[0001] Any and all application for which a foreign or domestic priority claim is identified in the Application Data Sheet as filed with the present application are hereby incorporated by reference under 37 CFR 1.57.

BACKGROUND

[0002] The present invention concerns a method of calibration of a stereoscopic camera system. In particular, embodiments of the present invention concern a method of calibration of a stereoscopic camera system for use with a radio therapy treatment apparatus.

[0003] Radiotherapy consists of projecting, onto a predetermined region of a patient's body, a radiation beam so as to destroy or eliminate tumours existing therein. Such treatment is usually carried out periodically and repeatedly. At each medical intervention, the radiation source must be positioned with respect to the patient in order to irradiate the selected region with the highest possible accuracy to avoid radiating adjacent tissue on which radiation beams would be harmful.

[0004] A fundamental problem with radiotherapy is the need to position the patient in the same position, when obtaining diagnostic images and each of the subsequent times when radiation is applied to the patient's body. It is to that end that Vision RT have developed an image processing system for monitoring and positioning patients described in U.S. Pat. Nos. 7,348,974, 7,889,906, 8,135,201, and pending U.S. patent application Ser. No. 12/379,108, published as U.S. 2009/0187112.

[0005] In use, in the Vision RT system, images of a patient on a mechanical couch are obtained by a set of stereoscopic cameras which are then processed to generate a 3D wire mesh model of the surface of a patient being monitored. This 3D wire mesh model is compared with a reference

surface created during treatment planning. The relative positioning of the model and the reference surface is compared and used to generate instructions for the mechanical couch to position the couch, vertically, laterally and rotationally so as to match the surfaces and hence locate the patient reliably in the same location relative to the iso-centre of a treatment apparatus. Subsequently during treatment, the position of a patient is continually monitored and if for any reason the patient moves or repositions themselves, this can be detected and appropriate action can be taken if necessary.

[0006] There are several sources of uncertainty in radio therapy treatment systems such as errors in patient positioning, target localization, and dose delivery. It is practically impossible to achieve perfect alignment mainly due to the presence of several geometric errors in the system. One of the critical geometric errors in radio therapy treatments is uncertainty in localizing the radiation field centre, which directly affects the dosimetric accuracy and results in incorrect tumour targeting that may lead to the delivery of inadequate dose to the lesion and/or serious damage to the healthy adjacent tissues. Therefore, it is necessary to develop methods to reduce the probability of such errors by extensive and efficient quality assurance programs to ensure high-level geometric accuracy of the treatment.

[0007] Originally, the primary method for iso-centre verification in radiotherapy was to measure the distance between the tip of a mechanical pointer mounted on the gantry head of a treatment apparatus and a fixed point mounted on the treatment table. Such a method was manual, laborious and time-consuming. The accuracy of the method depended upon the human observer and was also limited by size of the tip of pointer used.

[0008] An improved technique was introduced by Lutz, Winston and Maleki at Harvard Medical School in 1988 which is described in Lutz W, Winston K R, Maleki N. A system for stereotactic radiosurgery with a linear accelerator. *Int J Radiat Oncol Biol Phys.* 1988; 14(2):373-81. In the Winston-Lutz system a calibration phantom comprising a small metallic ball made of steel, titanium or tungsten is fixed on the treatment table by a locking mechanism. The phantom position is adjustable in three directions by means of a micrometer tool. The collimator used for radiotherapy is attached to the gantry head and the ball is placed as closely as possible to the iso-centre by aligning the marks on the phantom with the treatment room lasers. The collimated beam is used to expose a radiographic test film mounted perpendicular to the beam direction on a stand behind the ball. Differences between the centre of the sphere shadow and the field centre identifies the differences between the true iso-centre and the iso-centre as indicated by the treatment room lasers. The offset is read on each film using transparent template guidance scales or scanning the film and software analysis.

[0009] A mathematical method for analysing Winston Lutz images was developed and is described in Low D A, Li Z, Drzymala R E. Minimization of target positioning error in accelerator-based radiosurgery. *Med Phys.* 1995; 22(4):443-48 which used the film-measured iso-centre positional errors for eight gantry angle and couch settings to find the suitable offset for the phantom stand to minimize the distance between the treatment apparatus iso-centre and the target. A similar aim was followed by Grimm et al., who developed an algorithm to reconstruct the Winston-Lutz phantom ball locus in three dimensions from two-dimensional film images taken at certain couch and gantry angles and combined them with the images of lasers taken by digital cameras. This approach is described in Grimm J, Grimm S L, Das I J, et al. A quality assurance method with sub-millimetre accuracy for stereotactic linear accelerators. *J Appl Clin Med Phys.* 2011; 12(1):182-98.

[0010] A further example of automated processing of phantom images is described in E Schrieblmann, E Elder and T Fox, Automated Quality Assurance for Image-Guided Radiation Therapy, *J Appl Clin Med Phys.* 2009;10(1):71-79 which discusses the automation of Quality Assurance methods to ensure that a megavoltage (MV) treatment beam coincides with an integrated kilovoltage (kV) or volumetric cone beam CT. In the paper, a calibration cube is described as being located at the estimated location of treatment room iso-centre using laser markings. Images of the irradiation of the cube are then obtained and processed to determine the

extent the cube as positioned is offset from the iso-centre as identified by the MV, kV and volumetric cone beams.

[0011] Calibration of stereoscopic camera systems for use in treatment rooms has developed alongside methods for identifying the iso-centre of a treatment apparatus. Calibration techniques used to calibrate the Vision RT patient monitoring system are described in U.S. Pat. Nos. 7,348,974 and 7,889,906.

[0012] As is described in U.S. Pat. Nos. 7,348,974 and 7,889,906 a calibration sheet comprising a 40*40 cm sheet of flat rigid material such as aluminium or steel on which a pattern revealing a 20*20 matrix of circles at known positions on the surface of the sheet is provided. Additionally, towards the centre of the calibration sheet are four smaller markers adjacent to four circles the centres of which together identify the four corners of a square of known size.

[0013] Images of the calibration sheet are obtained and processed to identify within the image the positions of the four markers in the images and their associated circles. From the relative positions of circles identified by the markers in the images, a projective transformation is determined which accounts for the estimated centres of the identified circles defining the corners of a parallelogram in the image which arises due to the relative orientation of the calibration sheet and the camera obtaining the image. The calculated transform is then applied to each of the identified circles in turn to transform the oval shapes of the circles. More accurate estimates of the positions of the centres of the four circles are then determined by identifying the centres of the transformed circles and utilising an inverse transform to determine the corresponding position of the estimated circle centre in the original image.

[0014] When the co-ordinates for all the centres of each of the representations of the circles on the calibration sheet have been calculated for an image, the relative orientation of the different cameras can then be calculated from the relative positions of these points in the images and the known relative locations of these circles on the surface of the calibration sheet as is described in detail in "A Versatile Camera Calibration Technique for High-Accuracy 3D Machine Vision Metrology Using Off the Shelf TV Cameras and Lenses", Roger Tsai, IEEE Journal of Robotics and Automation, Vol. Ra-3, No. 4, August 1987. Further from the relative positions of the points in the individual images internal camera parameters such as the focal length and radial distortion within the camera images can also be determined.

[0015] Having determined the locations of the stereoscopic cameras, and any radial distortions present in the camera images, the positioning of the cameras relative to the iso-centre of the treatment apparatus is then determined. This is achieved by imaging a calibration cube of known size which is positioned on a treatment apparatus at a position with its centre at the iso-centre of the treatment apparatus as indicated by the co-incidence of marks on the exterior of the cube with the projection of the laser cross hairs which intersect at the iso-centre.

[0016] The images of the calibration cube are processed utilising the previously obtained measurements of the relative locations of the cameras and any data about the existence of any distortion present in the images to generate a 3D computer model of the surface of the cube. Since the cube has known dimensions and is at a known location and in a known orientation relative to the iso-centre of the treatment apparatus as indicated by the laser cross-hairs, a comparison between the generated 3D model and the known parameters for the size and position of the calibration cube enables measurements made in the co-ordinate system of the modelling software to be converted into real world measurements in the treatment room relative to the treatment iso-centre.

[0017] Although the conventional approach to calibrating a stereoscopic camera system for use with a radio therapy treatment apparatus is highly accurate, further improvements in accuracy are desirable.

SUMMARY

[0018] In accordance with one aspect of the present invention there is provided a method of

calibration of a stereoscopic camera system for use with a radio therapy treatment apparatus.

[0019] In the method, initially a calibration phantom is positioned with the phantom's centre at an estimated location for the iso-centre of a radio therapy treatment apparatus. The calibration phantom is then irradiated using the radio therapy treatment apparatus. The relative location of the centre of the calibration phantom and the iso-centre of the radio therapy treatment is then determined by analysing images of the irradiation of the calibration phantom. The calibration phantom is then repositioned, applying an offset corresponding to the determined relative location of the centre of the calibration phantom and the iso-centre of the radio therapy treatment apparatus to the calibration phantom. The stereoscopic camera system then proceeds to obtain images of the relocated calibration phantom to which the offset has been applied, processing the images to set the co-ordinate system of the stereoscopic camera system relative to the iso-centre of the radio therapy treatment apparatus.

[0020] Thus in this way the co-ordinate system for the stereoscopic camera system is established based directly on the positioning of a calibration phantom where the calibration phantom's location is directly linked to a determination of the iso-centre of a treatment apparatus.

[0021] In some embodiments, the stereoscopic camera system may additionally be utilised to monitor the repositioning of the calibration phantom. This may involve processing images of a relocated calibration phantom and comparing a generated 3D computer model of the surface of the repositioned calibration phantom with a 3D computer model of the expected surface of the calibration phantom positioned at the iso-centre of the treatment apparatus.

[0022] In such embodiments the 3D computer model of the expected surface of the calibration phantom positioned at the iso-centre of the treatment apparatus may comprise a stored 3D computer model of the surface of the calibration phantom.

[0023] Alternatively, the stereoscopic camera system may be utilised to generate a 3D computer model of the expected surface of the calibration phantom positioned at the iso-centre of the treatment apparatus. In such a system initially a model of the surface of the phantom may be generated by processing images of the calibration phantom positioned at the estimated location for the iso-centre of the radio therapy treatment apparatus. A model of the expected surface of the calibration phantom may then be generated by applying an offset corresponding to the determined relative location of the centre of the calibration phantom and the iso-centre of the radio therapy treatment apparatus to the generated model.

[0024] In some embodiments, a calibration phantom may be positioned at the estimated location of the iso-centre of a treatment apparatus by being placed on a mechanical couch. In such embodiment, relocating the calibration phantom may be achieved by instructing the mechanical couch to offset the current location of the calibration phantom by an offset corresponding to the determined relative location of the centre of the calibration phantom and the iso-centre of the radio therapy treatment apparatus.

[0025] In some embodiments an adjustable table may be used to level a calibration phantom. The adjustable table may be placed on the surface of a mechanical couch. In such embodiments the calibration phantom may be fixed on the adjustable table with the adjustable table being levelled thereby levelling the calibration phantom. In such embodiments the adjustable table may include a spirit level to facilitate levelling a calibration phantom by adjusting the adjustable table.

[0026] In embodiments of the claimed invention a calibration phantom may comprise a calibration phantom containing one or more calibration targets. In embodiments where the calibration phantom contains multiple calibration targets, preferably the calibration targets are arranged asymmetrically within the calibration phantom so assist with the analysis of images of an irradiated phantom.

[0027] Analysing images of the irradiation of the calibration phantom may comprise determining the relative location of the centre of the calibration phantom and the iso-centre of the radio therapy treatment apparatus by comparing the locations of the one or more targets in images of the irradiated calibration phantom irradiated with the treatment apparatus being in a number of

different positions.

[0028] In some embodiments the relocation of a calibration phantom may be undertaken iteratively. That is to say that having relocated a calibration phantom, the relocated calibration phantom may be irradiated and imaged whilst in its new location. The relative location of the centre of the relocated calibration phantom and the iso-centre of the radio therapy treatment apparatus can then be determined by analysing images of the irradiation of the relocated calibration phantom. The calibration phantom may then be further relocated by applying an offset corresponding to the determined relative location of the centre of the relocated calibration phantom and the iso-centre of the radio therapy treatment apparatus to the relocated calibration phantom. The co-ordinate system of the stereoscopic camera system may then be set by processing obtained images of the calibration phantom as relocated.

[0029] In some embodiments multiple calibration phantoms may be used. More specifically, an initial estimate of the location of the iso-centre of a treatment apparatus may be obtained by irradiating a first calibration phantom and positioning the calibration phantom on the basis of analysis of images of the irradiated calibration phantom. The first calibration phantom could then be replaced and further adjustments of position may be determined utilising that second calibration phantom.

[0030] In such systems initially a first phantom containing multiple calibration targets could be used. Where such targets are arranged asymmetrically within the body of the calibration phantom, processing images of the irradiated phantom can identify translations and rotations for positioning the centre of a phantom to coincide with the iso-centre of a treatment apparatus. Having done so, fine adjustments may then be based on the use of a second calibration phantom using containing a single small target positioned in the centre of a calibration phantom thereby improving accuracy.

[0031] In accordance with another aspect of the present invention there is provided a method of calibrating a stereoscopic camera system for use with a radio therapy treatment apparatus, the method comprising: positioning a calibration phantom with the phantom's centre at an estimated location for the iso-centre of a radio therapy treatment apparatus; utilizing a stereoscopic camera system to obtain images of the phantom and processing the images to generate a 3D computer model of the surface of the phantom; irradiating the calibration phantom using the radio therapy treatment apparatus and determining the relative location of the centre of the calibration phantom and the iso-centre of the radio therapy treatment apparatus by analysing images of the irradiation of the calibration phantom; and setting the co-ordinate system of the stereoscopic camera system on the basis of the determined relative location of the centre of the calibration phantom and the iso-centre of the radio therapy treatment apparatus and a determined location of the centre of the modelled phantom.

[0032] In such an embodiment setting the co-ordinate system of the stereoscopic camera system may comprise: determining the position of the centre of the modelled phantom relative to a current origin for the co-ordinate system for the 3D computer model and adjusting the co-ordinate system for the 3D computer model by the sum of the transformation required to make the current origin for the co-ordinate system for the 3D computer model coincide with the determined position of the centre of the modelled phantom and the transformation required to make location of the centre of the calibration phantom coincident with the iso-centre of the radio therapy treatment apparatus.

[0033] The determination of the transformation required to make the current origin for the co-ordinate system for the 3D computer model coincide with the determined position of the centre of the modelled phantom may in some embodiments be determined by storing a 3D computer model of a surface of a calibration phantom with the centre of the calibration model coincident with the centre of the calibration phantom; and determining the transformation required to make the current origin for the co-ordinate system for the 3D computer model coincide with the determined position of the centre of the modelled phantom on the basis of a comparison of the modelled surface of the phantom generated on the basis of images obtained by the stereoscopic camera with the stored

model.

[0034] The above described systems of calibration have a number of advantages over conventional systems for calibrating stereoscopic camera systems for use with radio therapy apparatus.

[0035] Conventionally, the iso-centre of a treatment apparatus is established using an approach such as suggested by Winston Lutz. That location was then normally marked using a laser system so that intersecting planes of laser light coincided at the identified iso-centre and laser light system was then utilised to position a calibration cube for calibrating a stereoscopic camera system. However, the use of the laser light system has the potential to introduce additional errors into the calibration and therefore introduce errors into the accuracy of the patient positioning system. Further the use of a laser light system also requires that the accuracy with which the laser lights identify the iso-centre of a treatment apparatus must be repeatedly checked and confirmed in case this varies over time. Instead in accordance with the claimed invention, the calibration of a stereoscopic camera system is achieved directly through imaging and irradiating a calibration phantom without having to rely upon the positioning of a phantom using lasers. This both increases the accuracy of the calibration and also avoids the need to calibrate the positioning lasers so that they correctly identify the position of the iso-centre over time.

[0036] Further details of the present invention will become apparent with reference to the following description and drawings in which:

Description

BRIEF DESCRIPTION OF THE FIGURES

[0037] FIG. 1 is a schematic perspective view of a stereoscopic camera system and a treatment apparatus; and

[0038] FIG. 2 is a flow diagram of a method of calibration of a stereoscopic camera system in accordance with the present invention.

DETAILED DESCRIPTION OF EXEMPLARY EMBODIMENTS

[0039] FIG. 1 is a schematic perspective view of a treatment system. The treatment system includes a set of stereoscopic cameras **10** that are connected by wiring (not shown) to a computer **12**. The computer **12** is also connected to treatment apparatus **14** such as a linear accelerator for applying radiotherapy. A mechanical couch **16** is provided as part of the treatment apparatus. The treatment apparatus **14** and the mechanical couch **16** are arranged such that, under the control of the computer **12**, the relative positions of the mechanical couch **16** and the treatment apparatus **14** may be varied, laterally, vertically, longitudinally and rotationally as is indicated in the figure by the arrows adjacent the couch. Additionally, the treatment apparatus **14** is also able to rotate about an axis as indicated by the arrows in the centre of the apparatus in the figure.

[0040] The computer **12** is arranged to process images obtained by the stereoscopic cameras **10** to generate 3D wire mesh computer models of the surfaces of objects and people lying on the mechanical couch **16**. When appropriately calibrated, a comparison of surface of a patient lying on the couch **16** with a reference surface of the same patient generated during the planning stage of treatment can enable the computer **12** to determine a transformation to match the two surfaces and generate a set of instructions to reposition the mechanical couch **16** so as to minimise differences between the surfaces. When repositioned, the patient can be treated using the treatment apparatus **14** in the knowledge that the patient is in the same position as the position used to determine treatment planning.

[0041] In order to minimise errors in the application of radiotherapy and patient positioning it is essential that the stereoscopic camera system **10** and the computer **12** are calibrated to the highest accuracy. In particular, it is necessary that the co-ordinate system used to generate 3D wire mesh models of the surface of a patient is matched both with the iso-centre of the treatment apparatus **14**

and the axes of motion of the mechanical couch **16**. In most embodiments, the iso-centre of the treatment apparatus **14** will be the location of the zero co-ordinates for the modelling system and the axes of the co-ordinate system for the modelling system will be selected to correspond to the axes of motion (i.e. the lateral, vertical and horizontal axes of motion) of the mechanical couch **16**. This then simplifies the conversion of transformations to match model surfaces in model space with instructions for relocating the mechanical couch **16** to match the position of a patient so that they can be treated using the treatment apparatus **14**.

[0042] FIG. **2** is a flow diagram of a method of calibrating the treatment system of FIG. **1**.

[0043] To calibrate the system, initially (s1) the relative locations of the stereoscopic cameras in the stereoscopic camera system **10** are determined by imaging a calibration sheet such as is described in U.S. Pat. Nos. 7,348,974 and 7,889,906. That is to say that a flat rigid calibration sheet such as a 70*70 cm sheet of material such as aluminum or steel on which a pattern revealing a 3D matrix of circles at known positions on the surface of the sheet is placed on the mechanical couch and imaged by the stereoscopic cameras **10**. On the sheet in addition to the pattern of circles are four smaller markers adjacent to four circles the centres of which together identify the four corners of a square of known size.

[0044] The images of the sheet obtained by the cameras are then processed by the computer **12** to identify within the image the positions of the four markers in the images. From the relative positions of circles identified by the markers in the images, for each image a first projective transformation is determined which accounts for the estimated centres of the identified circles defining the corners of a projected distorted square in the image which arises due to the relative orientation of the calibration sheet and the camera obtaining the image.

[0045] The calculated transform is then utilised to determine estimated three dimensional co-ordinate centres of each of the circles identified by markers. These calculated co-ordinates then identify an estimated location and orientation for the plane corresponding to the surface of the calibration sheet relative to the position from which an image has been obtained.

[0046] Each pixel in the image obtained by the cameras **10** is then processed in turn to determine where within the plane containing the estimated positions of the circle centres, each pixel corresponds. The estimated circle centres are then processed in turn and the pixels in an image corresponding to points lying within a pre-determined distance from each circle centre in the calculated plane are then identified. The grey scale values for each of the pixels in each set are then utilised to determine an improved estimate of the co-ordinates for the circle centres. x and y co-ordinates for the positions of the points in the estimated plane including the surface of the calibration sheet each pixel represents within the set is determined. These calculated x and y co-ordinates are then utilised to estimate an improved estimate of the x,y co-ordinates of the circle centre using the following equations:

$$[00001] x = \frac{\text{Math. } \Sigma gx}{\text{Math. } \Sigma g} \quad y = \frac{\text{Math. } \Sigma gy}{\text{Math. } \Sigma g}$$

[0047] where Σg is the sum of all pixel values in the set identified for a particular circle centre, Σgx is the sum of the grey scale pixel values for a pixel multiplied by the x co-ordinates for those pixels and Σgy is the sum of the grey scale pixel values for a pixel multiplied by the y co-ordinates for those pixels, and where the colour of the circle is associated with a high grey scale value and the colour of the background corresponding to the calibration sheet is associated with a low grey scale value.

[0048] Co-ordinates for the point within the image corresponding to the new estimated circle centres are then determined from these x,y co-ordinates and these updated estimates of the centres of the marked circles are then utilised to determine a more accurate estimated transformation to account for the location and orientation of the calibration sheet. The above process can then be repeated until an accurate estimate of the actual circle centre positions is made and the true transform required to account for the relative orientation of the calibration sheet is determined.

[0049] Using the final determined transform, the expected positions of all of the circles on the sheet appearing in the image are then calculated, the portions of the images in the vicinity of each of the estimated circle centres are then processed individually in the same way as described above. For each of the circles a set of pixels is identified corresponding to points within a pre-set distance to the circle centre and then an improved circle centre co-ordinate is calculated using the grey scale values and co-ordinate values as described above.

[0050] When the co-ordinates for all the centres of each of the representations of the circles on the calibration sheet have been calculated for an image, the relative orientation of the different cameras within the stereoscopic camera system **10** can then be calculated from the relative positions of these points in the images and the known relative locations of these circles on the surface of the calibration sheet as is described in detail in "A Versatile Camera Calibration Technique for High-Accuracy 3D Machine Vision Metrology Using Off the Shelf TV Cameras and Lenses", Roger Tsai, IEEE Journal of Robotics and Automation, Vol. Ra-3, No. 4, August 1987. Further from the relative positions of the points in the individual images internal camera parameters such as the focal length and radial distortion within the camera images can also be determined.

[0051] Thus by performing this initial calibration step, the computer **12** is provided with data which enables the computer to convert stereoscopic images obtained by the stereoscopic camera system **10** into 3D computer wire mesh models of the surfaces of imaged objects. Further, since the initial calibration is performed relative to a calibration sheet of known dimensions, the computer is also provided with data identifying the relationship between distances in the computer models and real world measurements. However, at this stage there is no means by which the computer **12** can relate such measurements to the position of the iso-centre of the treatment room or the axes of motion of the mechanical couch.

[0052] Having determined the relative locations and relevant internal parameters of the cameras of the stereoscopic camera system **10**, a calibration phantom **18** is then placed (s2) at the estimated iso-centre for the treatment room. The phantom **18** in this embodiment comprises a machined cube of known dimensions of a low density material with radiographic attenuation comparable to water and a density comparable to that of human tissue. In this embodiment, this is achieved by the phantom **18** being a machined polystyrene cube with the surface finish of the cube painted to provide optimal imagery for imaging by the stereoscopic camera system **10**. In other embodiments other materials could be used. However, in any embodiment the materials used should be selected to facilitate imaging using imaging techniques such as kV, stereo x-ray, MV or portal imaging, Cone Beam CT, and conventional axial and helical CT.

[0053] A known asymmetrical arrangement of targets is provided in a fixed orientation within the body of the phantom **18**. These targets may comprise small metallic or ceramic balls. The balls may be made of steel, titanium or tungsten or may be made of a ceramic material or any other material which provides a good contrast to the material forming the body of the phantom **18** when the phantom is imaged. Most suitable materials will be materials which have good radiation stability which are able to withstand considerable radiation dose before undergoing substantial change where the relative density and radiographic attenuation of the targets provided within the body of the phantom **18** is selected to assist with the subsequent imaging of the phantom **18**. It is preferable that any phantom **18** contains an asymmetrical arrangement of targets as the asymmetry of the targets enables the orientation of the phantom **18** to be determined from images of the phantom **18**, enabling the orientation of the phantom **18** with respect to the treatment co-ordinate system to be determined or confirmed if already known.

[0054] Suitable phantoms **18** with an asymmetric arrangement of imaging targets would include the Penta-Guide phantom (Modus Medical Devices, London Ontario, Canada) which consists of an acrylic cube of 16 cm size with five internal air pockets orientated in a unique pattern.

[0055] The phantom **18** is located at an estimated location for the iso-centre by being placed on an adjustable table **20** mounted on the mechanical couch **16**. The adjustable table **20** in addition to

fixing the location of the phantom **18** on the mechanical couch **16** is arranged to have adjustable feet so that minor adjustments to the orientation of the phantom can be made. This is desirable because the axes of motion of a typical mechanical couch **16** are limited to rotation and lateral, vertical and horizontal motion and hence motion of a mechanical couch **16** can only approximate changes in for example pitch or roll. In some embodiments the adjustable table **20** may include a spirit level which can indicate to a user when the table **20** and hence the phantom **18** is level.

[0056] So as not to impact the quality of the radiographic images, from any potential x-ray/photon beam scatter, the adjustable table **20** is constructed from materials with low density/radiographic attenuation. In addition, in this embodiment the table **20** has a clear aperture in the path of the vertical treatment beam, so as not to adversely affect the quality of the radiographic images obtained when the phantom **18** is irradiated.

[0057] In some embodiments, the adjustable table **20** may be arranged so that it can only be attached to the mechanical couch **16** at a number of fixed locations and in fixed orientations relative to the surface of the mechanical couch **16**. Similarly, the adjustable table **20** may also be arranged so that a phantom **18** may only be placed on the table **20** in one or more fixed orientations. Thus for example the adjustable table **20** may include a recess for receiving the phantom **18** in a particular orientation. The attachment of the adjustable table **20** to the mechanical couch **16** and the fixing of the position and orientation of the phantom **18** relative to the adjustable table **20** assists with fixing the positioning and orientation of the phantom relative to the iso-centre and co-ordinate system for the treatment room. In any event the adjustment table **20** should be constructed so as not to slip or move on the surface of the mechanical couch **16**.

[0058] In some embodiments, the adjustable table may incorporate a micrometer positioning tool for precisely moving the phantom laterally, longitudinally and vertically. The advantage of such a positioning tool would be that it would be easier to make fine amendments to the position of the phantom using the tool rather than the mechanical couch **16** since the relocation would only involve movement of the phantom and not the entire couch **16**.

[0059] In some embodiments a phantom which incorporated such a tool might be utilized in place of a separate phantom **18** and adjustable table **20**. An example of such a phantom would be the Varian Exact T Look-Bar (MedTec, Orange City, IA, USA) which consists of five markers embedded in a tissue equivalent material which is mountable on a couch via a cam lock mechanism which incorporates a micrometer tool.

[0060] Having positioned the phantom **18** with its centre point at the estimated treatment room iso-centre and if necessary levelled the phantom **18** by adjusting the feet of the adjustable table **20**, images of the phantom **18** are obtained (s3) using the stereoscopic cameras **10**. These images are then processed by the computer **12** using the previously obtained information about the relative locations and internal parameters of the cameras of the stereoscopic camera system **10** to create a 3D computer model of the surface of the phantom **18** in the current 3D model space.

[0061] The phantom **18** is then (s4) irradiated using the treatment apparatus **14** from a variety of angles and images of the projected radiation are obtained in a conventional manner in a similar way to the conventional Winston-Lutz test using a digital camera **22** arranged at a fixed point opposite the gantry head of the treatment apparatus **14** irradiating the phantom **18**. The images obtained by the digital camera **22** when the gantry head of the treatment apparatus **14** is orientated in different locations are then processed by the computer **12** in a conventional way to determine from differences between the appearance of the phantom **18** in the images and in particular shadows of the targets contained within the phantom **18** and compared with the expected locations of shadows of the targets within the phantom on the assumption that the phantom **18** is correctly placed at the iso-centre. From this processing, the computer **12** determines the extent to which the centre of the phantom is offset from the true iso-centre of the treatment apparatus **14**.

[0062] The computer **12** then (s5) proceeds to apply the determined offset to the modelled surface of the phantom **18** to generate a reference surface for the correct positioning of the phantom **18**.

Having done so, the computer instructs the mechanical couch **16** to reposition itself laterally, vertically, longitudinally and rotationally so as to relocate the phantom **18** to the position of the true iso-centre. If the phantom **18** is correctly levelled prior to imaging an exact realignment of the phantom **18** should be achievable purely through lateral, vertical, longitudinal and rotational transformations.

[0063] As the phantom **18** is repositioned, the location of the phantom **18** is monitored (s6) using the stereoscopic camera system **10** with the computer **12** processing images of the phantom **18** to generate a wire mesh models of the surface of the phantom **18** as it is repositioned. Any differences between the modelled surface and the original surface to which the required offset for relocating the phantom **18** to the true iso-centre can be displayed on the screen of the computer **12** to provide visual feedback of errors which remain.

[0064] Monitoring of the motion of the phantom **18** provides an opportunity for confirming that the axes of the co-ordinate system being used to generate wire mesh models corresponds to the axes of the co-ordinate system for the mechanical couch **16**. More specifically, the analysis of the images obtained by the digital camera **22** will provide an offset which is the offset for re-positioning the phantom **18** at the true iso-centre for the treatment room. Instructing the mechanical couch **16** to reposition the phantom **18** in a manner corresponding to the offset should result in a motion with corresponds which causes the phantom **18** to be correctly positioned. In addition, the motion of the phantom **18** due to the repositioning of the couch **16** should also correspond to an equivalent relocation of the surface of the phantom **18** as monitored by the stereoscopic camera system **10**. Additionally, this monitoring of the surface of the phantom **18** can highlight the existence of any further minor adjustments to the position or orientation of the phantom **18** such as might be achieved by adjusting the adjustable table **20** after the phantom **18** has been repositioned by the mechanical couch **16**.

[0065] When the phantom **18** located with its centre point at the estimated position of the iso-centre, the stereoscopic camera **10** system can then use the images of the phantom **18** of known dimensions at a known location together with the known relative locations of the cameras of the stereoscopic camera system **10** to calibrate the co-ordinate system for the 3D surface models generated by the computer **12** relative to the actual location of the iso-centre. This is typically achieved by redefining the co-ordinate system for the computer model so that the origin for the computer model co-ordinate system coincides with the centre of the phantom **18** as determined from the modelled surface of the phantom **18** in its new position.

[0066] In embodiments of the present invention, steps S3-S6 can be repeated if desired to obtain further confirmation to the relative position of the phantom **18** to the iso-centre. That is to say, after having relocated the mechanical couch **16** based on the determined offset based on images obtained by the digital camera **22** and monitoring that the phantom **18** has been correctly relocated using the stereoscopic camera system **10**, the repositioned phantom **18** could then be re-irradiated and further images of the irradiated phantom **18** could then be obtained to confirm that the phantom **18** was indeed located at the iso-centre and if this was not the case further adjustments in the position of the phantom **18** could be made.

[0067] For additional accuracy in a second iteration, a phantom **18** could be utilised comprising a single target where the phantom **18** comprises a cube of known dimensions with a single spherical target embedded in the cube where the sphere is located at the centre of the cube. Such a phantom **18** could then be imaged and adjustments made on the basis of analysis of images obtained by the digital camera **22** when the phantom is irradiated. The benefit of such a phantom **18** is that the target sphere could made to be very small and thereby assist with locating the iso-centre with greater accuracy. The disadvantage of such a phantom **18** would be that if only a single target sphere is present, images of the obtained phantom **18** will lack information about the orientation of the phantom **18** which can be derived where the phantom **18** contains an asymmetric arrangement of targets.

[0068] A suitable test phantom which consists of one central 2 mm radio-opaque marker embedded in tissue equivalent material would be the Varian cube (Varian Medical Systems, Palo Alto, CA, 15 USA).

[0069] In embodiments where multiple phantoms **18** are utilised, it is beneficial that an adjustable table **20** is used which permits attachment of a phantom **18** in a single orientation. In such a system, initially a phantom **18** containing multiple targets could be used to identify the location of the iso-centre of the treatment apparatus. Using the multi-target phantom would have the advantage that because of the asymmetric arrangement of target spheres is apparent, it is possible to determine the orientation of the phantom **18** as it appears in images obtained by the digital camera **22**. Having done so and having repositioned the phantom **18** based on analysis of the images, the multi-target phantom **18** could then be replaced with a single target phantom **18** and final minor adjustments could be made to the calibration using the single target phantom. When changing the phantom **18** the stereoscopic camera system **10** could be utilised to ensure that no change in position occurred when the phantoms **18** were exchanged. In such embodiments, it would be beneficial if the multiple phantoms **18** could be marked in a manner (e.g. colour coded) which enabled the individual phantoms **18** to be easily distinguished from one another.

[0070] In the above embodiments the calibration of a camera system has been described based upon the imaging of a phantom **18** which has been located to the iso-centre for a treatment room. It will be appreciated that in some embodiments, calibration could be achieved without physically relocating a phantom **18**. More specifically, as described, a phantom **18** could be positioned at the estimated location for an iso-centre and imaged using the stereoscopic camera system **18** and then irradiated with images being obtained using the digital camera **22**. The images obtained by the digital camera **22** could then be processed to determine the offset of the phantom relative to the actual iso-centre for the treatment room. Similarly, images of the phantom **18** obtained by the stereoscopic camera system **10** could be processed to generate a 3D computer model of the phantom at its current location. An adjustment for the co-ordinate system for the stereoscopic camera **10** could then be calculated by adding the transform calculated by processing the 3D model to relocate the centre of the model of phantom **18** with the origin of the axes for the computer model and the offset as determined by processing the images obtained by the digital camera **22**.

Claims

1. (canceled)
2. A camera system for use with a radiotherapy treatment system having a coordinate system, said camera system comprising: a camera configured to image a physical calibration phantom positioned in a camera coordinate system; and at least one processor configured to: receive image data of the physical calibration phantom from the camera assembly; receive coordinate data indicative of the physical calibration phantom's position and orientation with respect to the radiotherapy treatment system; determine, based on the image data and the coordinate data, an offset between the camera coordinate system and the radiotherapy treatment system's coordinate system; and compensate, based on the offset, the camera coordinate system for provision of an updated camera coordinate system.
3. The camera system of claim 2, wherein the image data is optical image data.
4. The camera system of claim 2, wherein the at least one processor is configured to receive the coordinate data from the radiotherapy treatment system.
5. The camera system of claim 2, wherein the at least one processor is configured to generate a 3D model based on the image data.
6. The camera system of claim 2, wherein the at least one processor is configured to output updated data indicative of the updated camera coordinate system.
7. The camera system according to claim 2, wherein to determine the offset comprises to compare

the received image data to reference data.

8. The camera system according to claim 2, wherein the at least one processor is configured to calculate, based on the offset, a transformation between the camera coordinate system and the second coordinate system.
 9. The camera system according to claim 8, wherein to compensate the camera coordinate system comprises to apply the transformation to the camera coordinate system.
 10. The camera system according to claim 8, wherein the transformation aligns the camera coordinate system with the second coordinate system.
 11. The camera system according to claim 2, wherein the at least one processor is configured to calculate, based on the offset, a transformation between the camera coordinate system and the second coordinate system.
 12. The camera system of claim 2, wherein the coordinate data is derived from radiographic images taken by the radiotherapy treatment system.
 13. The camera system of claim 2, wherein camera is a stereoscopic camera.
 14. A radiotherapy system comprising: a radiotherapy treatment system having a treatment coordinate system; and a camera system comprising: a camera configured to image a physical calibration phantom positioned in a camera coordinate system; and at least one processor configured to: receive image data of the physical calibration phantom from the camera assembly; receive coordinate data indicative of the physical calibration phantom's position and orientation with respect to the treatment coordinate system; determine, based on the image data and the coordinate data, an offset between the camera coordinate system and the treatment coordinate system; and compensate, based on the offset, the camera coordinate system for provision of an updated camera coordinate system.
 15. The radiotherapy system of claim 14, wherein camera is a stereoscopic camera.
 16. The radiotherapy system of claim 14, wherein the image data is optical image data.
 17. The radiotherapy system of claim 14, wherein the at least one processor is configured to receive the coordinate data from the radiotherapy treatment system.
 18. The radiotherapy system according to claim 14, wherein the at least one processor is configured to calculate, based on the offset, a transformation between the camera coordinate system and the second coordinate system.
 19. The radiotherapy system according to claim 18, wherein to compensate the camera coordinate system comprises to apply the transformation to the camera coordinate system.
 20. The radiotherapy system of claim 14, wherein the at least one processor is configured to generate a 3D model based on the image data.
 21. The radiotherapy system of claim 14, wherein the coordinate data is derived from radiographic images taken by the radiotherapy treatment system.
-