DX-D 100	Type 5410	SB No. 32
DX-D 100 Wireless	Type 5411	SB No. 29
DX-D 300	Type 8207	SB No. 38
DX-D 400	Type 5420	SB No. 10
DX-D 600	Type 5430	SB No. 23

DD+DIS146.13E

Service Information Bulletin

This Bulletin is for information only.

JPI Grid Test Film replaced by DICOM CD

Task

Timing		Category
Next service	○	**Apply at all sites**
	◉	**Apply at affected sites as listed below**
	○	**Optional to improve functionality of product**

Task Tracking

After completion of your task the following entry in your Service Report is required:	DD+DIS146.13E	*
* Insert the document number into the field "Comment" (SMS form).		

Purpose of this document:

This document communicates that the X-ray film delivered with JPI grids is replaced by a DICOM image on a CD.

Affected serial number(s) / batch: Systems delivered as from August 2013.

1   Introduction/purpose

All DX-D systems except DX-D 500 are delivered with grids manufactured by JPI.

To prove uniformity and quality at the time of shipment, each grid is inspected by the manufacturer and an X-ray is taken to capture the results.

Up to now, an X-ray film is delivered with each grid as proof of quality and stored on site.

From August 2013, this film (35 x 43 cm) is replaced by a DICOM image on a CD.
This CD is stored in the same way as the film.

Figure 1: Grid and X-ray film in envelope

2   Prerequisites

None

3   Instructions

Not applicable.

4   Verification

Not required.

5   Keywords

JPI; DICOM; test film; grid

6   Version history

Version	Change	Date
1.0	Initial version	10-2013