

Northbridge District Hospital

Infusion Pump Checklist

240 Pacific Highway, Northbridge NSW 2063
Tel: (02) 9405 0000 | Fax: (02) 9405 0001
IV Infusion Pump Safety Check

SYNTHETIC TRAINING DOCUMENT

Patient Name	Isabella Lisa Romano	Medical Record No.	137689
Date of Birth	09/03/1980 (Age 46)	Medicare No.	2485 11776 5
Allergies	Penicillin (anaphylaxis)	Reference No.	IPC-2026-41159
Drug Infusing	0.9% Sodium chloride 1L over 8 hours	Date	05/09/2026

CHECK DETAILS

Reference Number	IPC-2026-41159
Date / Time of Check	05/09/2026 12:39
Pump Make / Model	Baxter Spectrum IQ
Asset / Serial Number	BAXSPEC-62090
Drug Being Infused	0.9% Sodium chloride 1L over 8 hours
Reviewing Nurse	RN Michael Tran

PUMP SETUP

#	Item	Status	Initials
1	Pump checked for visible damage	Yes	KM
2	Pump cleaned with detergent / disinfectant wipes	Yes	JL
3	Mains power and battery backup verified	Yes	JL
4	Pump self-test passed	Yes	AT
5	Drug library updated and date-of-last-update confirmed	Yes	JL
6	Battery alarm threshold checked	Yes	JL

LINE AND ACCESS

#	Item	Status	Initials
1	IV access patent (no infiltration / phlebitis)	Yes	KM
2	Cannula/PICC site assessed and dressing intact	Yes	AT
3	Line traced from solution to patient (line-tracing safety)	No	MK
4	Correct line for the prescribed fluid (giving set type)	N/A	MK
5	Air eliminated from giving set (primed correctly)	Yes	JL
6	Filter present if required for the medication	No	LH

MEDICATION RIGHTS (5 RIGHTS)

#	Item	Status	Initials
1	Right Patient verified (3-point check)	Yes	SP
2	Right Drug verified against prescription	Yes	SR
3	Right Dose calculated and verified	Yes	AT
4	Right Route verified	Yes	JL

5	Right Time verified	Yes	DC
6	Right Documentation completed	N/A	DC
7	Two-nurse check completed for high-risk medications	Yes	DC

PUMP SETTINGS

#	Item	Status	Initials
1	Rate (mL/hr) programmed correctly	Yes	MK
2	Volume to be infused (VTBI) set correctly	Yes	SR
3	Concentration matches drug label	Yes	KM
4	Bolus / loading dose programmed if indicated	N/A	KM
5	Maximum dose safety limits not exceeded	Yes	DC

ONGOING MONITORING

#	Item	Status	Initials
1	Patient observations every 30-60 minutes	Yes	RB
2	Pump display visible and accessible to staff	Yes	SP
3	Alarms audible at nursing station	Yes	KM
4	Documented at start, hourly, and end of infusion	Yes	JL
5	Patient and family education provided	No	SP

ISSUES / ESCALATION

Nil. Pump operating as expected, patient comfortable.

REVIEWING NURSE

RN Michael Tran

RN

AHPRA: NMW62178533

Date / Time: 05/09/2026 10:04