

Test Report

Patient ID: Spencer,Brandon
Patient ID 2: sars-cov2,rt-pcr
Sample ID: mrn#
Test Type: Specimen
Sample Type:

Assay Information

Assay Name	Assay Version	Assay Type
Xpert Xpress CoV-2 plus	1	In Vitro Diagnostic

Test Result:

SARS-CoV-2 NEGATIVE

Analyte Result

Analyte Name	Ct	EndPt	Analyte Result	Probe Check Result
E	0.0	0	NEG	PASS
N2	0.0	0	NEG	PASS
RdRP	0.0	1	NEG	PASS
SPC	33.8	132	PASS	PASS

User: KATY EMERGENCY ROOM
Status: Completed Start Time: 10/30/22 12:07:08
S/W Version: 6.4 End Time: 10/30/22 12:37:06
Expiration Date*: 02/05/23 Instrument S/N: 845192
Cartridge S/N*: 6026031705 Module S/N: 905917
Reagent Lot*: 02925 Module Name: A1
Notes:
Error Status: OK

Errors

<None>

Test Methodology: RT-PCR
For In Vitro Diagnostic Use Only.
For use under the Emergency Use Authorization (US).

Test Report

SM 10/30/22

Operator Initial/Date

SM 10/30/22

Supervisor Initial/Date

* indicates that a particular field is entered using a barcode scanner

Test Methodology: RT-PCR

For In Vitro Diagnostic Use Only.

For use under the Emergency Use Authorization (US).