# Design of the ISS-Bioreactor

Aykut C. Satici,\* Omor Khan,<sup>†</sup> and Gunes Uzer<sup>‡</sup>
Boise State University Mechanical and Biomedical Engineering
(Dated: April 7, 2023)

This is a technical report for ISS-Bioreactor project at Boise State University.

### 1. INTRODUCTION

We design an input PWM filter to create the desired  $0-10\mathrm{V}$  sine wave at 90Hz that the controller expects from a carrier PWM signal at a carrier frequency  $36.6\mathrm{kHz}$ .

### 2. THEORETICAL ANALYSIS

We perform basic PWM filter design to generate the driving 0 - 10V sine wave for the Physik Instrumente (PI)'s piezoelectric actuator [1].

## 2.1. Basic Lowpass Filter Design

Consider the first-order filter sketched in Figure 1 with a sinusoidal driving voltage. The current i over the capacitor is  $i = C \frac{\mathrm{d}v_c}{\mathrm{d}t}$ . KVL around the circuit gives

$$RC\frac{\mathrm{d}v_c}{\mathrm{d}t} + v_c = v_{\mathrm{sig}} = A\cos\left(\omega t\right) \tag{1}$$

This differential equation has the transfer function

$$G(s) = \frac{1}{RCs + 1}.$$

The steady-state solution of the differential equation (1) is obtained as

$$v_c(t) = A |G(j\omega)| \cos(\omega t + \angle G(j\omega))$$

$$= \frac{A}{\sqrt{1 + \omega^2 R^2 C^2}} \cos(\omega t - \arctan(\omega RC))$$
(2)




FIG. 1: An first-order low-pass filter circuit.

Since we do not want our signal to be attenuated by the low-pass filter, we must choose the values of R and C such that  $\omega RC \ll 1$  or  $2\pi RC \ll 1/f$ .

Unfortunately, our actual input from the microcontroller is not a pure sine wave, rather a PWM signal. Therefore, we also need the value of  $2\pi RC$  to be large so that it attenuates the high frequencies present in the PWM signal. This is difficult to achieve with just a single low-pass filter. Hence, we design a second-order low-pass filter in the next subsection.

### 2.2. Second-Order LPF Design

The second-order filter is implemented in the first part of the circuit presented in Figure 2. The second part of this circuit amplifies the sine wave extracted from its PWM modulation from  $0-3.3\mathrm{V}$  to  $0-10\mathrm{V}$ . We analyze the circuit so as to figure out the values of the various resistances and capacitances.

We start by analyzing how the input voltage,  $v_i$ , to the non-inverting amplifier responds to Teensy-generated PWM voltage,  $v_t$ . This is a 0-3.3V PWM signal to be filtered to extract the modulated sine wave. Thanks to the buffer op-amp, the transfer function from the Teensy input  $v_t$  to the input  $v_i$  to the non-inverting amplifier op-amp is given by

$$\frac{V_i(s)}{V_t(s)} = H(s) = \frac{1}{R^2 C^2 s^2 + 2RCs + 1}.$$

This is a fully-damped transfer function with poles at  $s_{1,2} = {}^1\!/RC$ . In other words, the cut-off frequency of this filter is at  $f = \frac{1}{2\pi RC}$  with a roll-off of 40dB per decade. Contrast this to the first-order filter of the previous subsection where the roll-off was 20decibel per decade. The greater roll-off rate allows us to can select  $2\pi RC \ll {}^1\!/f$  while simultaneously achieving excellent high-frequency attenuation.

Our desired signal is a 90Hz sinusoidal. We do not want to lose this signal, i.e., we want our transfer function H(s) to approximately have a unity gain at this frequency. We set  $2\pi RC \leq 1/900$  or  $RC \leq 1.768 \times 10^{-4} \Omega \, \mathrm{F}$  to satisfy this requirement. Using some standard values of the resistance  $R=1\mathrm{k}\Omega$  and the capacitance  $C=104\mathrm{nF}$ , we obtain  $RC=1.04\times 10^{-4}$ , achieving the desired goal. The attenuation at Teensy's PWM career frequency of 36.6kHz is found by

$$20\log_{10}\{|H(j2\pi 36600)|\} = -55.163 \,\mathrm{dB}.$$

We want to amplify the input voltage  $v_i$  thrice in order to hit the 0-10V mark. The gain of the non-inverting

<sup>\*</sup>Electronic address: aykutsatici@boisestate.edu; URL: https://symplectomorphism.github.io/

<sup>†</sup>Electronic address: omorkhan@u.boisestate.edu

<sup>&</sup>lt;sup>‡</sup>Electronic address: gunesuzer@boisestate.edu




FIG. 2: The signal generator circuit.

amplifier is  $k=1+R_f/R_{in}$ . We choose  $R_f=2k\Omega$  and  $R_{in}=1k\Omega$  to achieve k=3. The high gain bandwidth product of the OP292 op-amp is read from its datasheet to be GBP = 4MHz. Hence the transfer function from the input voltage  $v_i$  to the output voltage  $v_o$  that will be applied to the PI controller is approximately given by

$$\frac{V_o(s)}{V_i(s)} = \frac{k}{\frac{k}{2\pi \text{GBP}} s + 1} \approx \frac{3}{1.194 \times 10^{-7} s + 1},$$

which will have a firm unity gain at our desired oscillation frequency of 90Hz.

## 3. RESULTS

## 3.1. Simulation

We perform a realistic simulation with a second-order low-pass filter we came up with in Section 2.2. One of the important aspects of this design is the selection of the op-amp. In order to keep the common-mode voltage at 0V we choose the op-amps as CMOS type. One such op-amp is 0P292 [2], which is used in this simulation.

The PWM signal generated by Teensy [3] is simulated exactly with a carrier frequency of 36.6kHz modulating the signal

$$V_{\text{pwm}} = \frac{3.3}{2} + \frac{3.3}{2} \sin(2\pi 90t).$$

Finally, the impedance of the load (PI's controller) is read off from its datasheet and inserted as a  $100 \mathrm{k}\Omega$  resistance. The circuit that is simulated using LTSpice [4] is presented in Figure 3.

The simulation generates the relevant voltage responses, provided in Figure 4. The top plot shows the




FIG. 3: The signal generator circuit in LTSpice  $\,$ 




FIG. 4: The response from the simulation for one full period.

PWM signal generated by Teensy modulating a sine-wave at 90Hz frequency. The individual plots in the middle show the output of the first (cyan) and the second (purple) RC low-pass filters ( $v_f$  and  $v_i$ , respectively) that extract the modulated signal from its PWM representation. Finally, the last plot shows the amplified signal (gain: 3) through the op-amp 0P292. This signal is ready to be sent to the PI controller.

### 3.2. Experiment

### 4. DISCUSSION AND CONCLUSIONS

Let's go to ISS!

## 5. ACKNOWLEDGMENT

The first author thanks Drs. John Chiasson and Vishal Saxena of Electrical and Computer Engineering at Boise State University and University of Delaware, respectively, for their invaluable insights into the practical implementation of the signal generator.

<sup>[1]</sup> Physik Instrumente, E-610, URL https://www.physikinstrumente.com/en/products/controllers-and-drivers/nanopositioning-piezo-controllers/e-610-piezo-amplifier-controller-601000.
[2] Analog Devices, OP 292, URL https://www.analog.com/

<sup>[2]</sup> Analog Devices, OP 292, URL https://www.analog.com/ en/products/op292.html.

<sup>[3]</sup> PJRC, Teensy, URL https://www.pjrc.com/teensy/.

<sup>[4]</sup> Analog Devices, LTSpice, URL https://www.analog.com/en/design-center/design-tools-and-calculators/ltspice-simulator.html.