7th International Youth Summit (IYS) 2026 - International Essay Competition

Sub-theme: Health

Spiro: A Low-Cost 3D-Printed IoT Spirometer with a Disposable Mouthpiece Made from Burned-Field Crop Residue

Bridging air pollution, lung health and circular materials for communities that cannot reach a hospital spirometer

Team MEDTHAI

Team leader: Puttitada Pruetchayakul (Bodindecha (Sing Singhaseni) School)

Members: Pimnara Roongsetthavitr (International Community School); Suchanan Kittikraisak (The Newton Sixth Form); Natnicha Taepucharoen (Samsenwittayalai School); Nabhat Sethaudom (International Community School)

Institutions: Bodindecha (Sing Singhaseni) School; International Community School; The Newton Sixth Form; Samsenwittayalai School

Advisor: Tanachat S. (teacher / project advisor, not a competitor)

Country: Thailand

Theme: "Bridging Ideas and Youth Creativity to Shape Solutions for Tomorrow's Challenges"

1. Introduction

Every dry season, smoke from burning rice straw, stubble and sugarcane leaves settles over northern and central Thailand. The fine particles (PM2.5) in that smoke reach the deepest part of the lung. Breathing in them for years is linked to chronic cough, asthma and chronic obstructive pulmonary disease (COPD). A province-level modelling study of five residue crops estimated that open burning of crop residue costs Thailand roughly 524 billion baht in health and environmental damage, while the alternatives it tested (animal feed, fertiliser, leaving residue in the field, power generation) gave a net benefit [1]. Another Thai study found that burning all rice straw would raise health impacts by 81.1% compared with the business-as-usual case [2].

Yet the people who breathe this smoke are often far from a hospital that owns a spirometer, the standard breathing test that measures how much air, and how fast, a person can blow out. Without that test, a smoker's cough, a farmer's wheeze or a child's asthma can go unmeasured for years. At the same time, the same fields that create the smoke produce a material that can be turned into something useful: fibre.

This essay presents Spiro, a student-designed prototype that tries to connect these two problems. Spiro is a handheld spirometer whose measuring body is 3D-printed and cheap to build, whose results are sent to a phone or website, and whose single-use mouthpiece is planned to be moulded from the same crop residue that would otherwise be burned. We describe the design, the reasoning behind it, an online system for patients, doctors and administrators, and - just as importantly - what has not yet been proven. Spiro is a learning and screening prototype. It is not a certified medical device and must not be used for diagnosis.

2. Body

2.1 The problem and the gap

Spirometry is the test used to detect and follow obstructive lung disease; international standards from the American Thoracic Society and European Respiratory Society (ATS/ERS) define how it should be performed, including at least three acceptable blows that repeat within 0.15 L [3][4]. Commercial spirometers meet these standards but are designed for clinics. Our goal is more modest: a device that a school, a community health volunteer or a village health post could build and use for screening, and that passes results to a doctor. Research groups have already shown that printed venturi spirometers are feasible; one 2025 study reported about 1.5% error against a 2 L calibration pump for a 3D-printed venturi with an ESP32 [5]. What is missing in the literature we found is a low-cost design that also deals with infection control and waste, and that is paired with an accounts-based service connecting patient and doctor.

2.2 How Spiro works

Spiro uses a venturi: a tube that narrows to a throat. When air is blown through it, the air speeds up at the throat and its pressure drops. A Sensirion SDP810-500Pa differential-

pressure sensor measures the pressure difference between the wide part (28 mm) and the throat (23 mm). Flow follows Q = K * sqrt(dP), and integrating flow over time gives volume. From one blow the firmware reports forced vital capacity (FVC), forced expiratory volume in one second (FEV1), their ratio and peak expiratory flow (PEF). A XIAO ESP32-S3 reads the sensor at 400 Hz, shows the result on a 1.28-inch round display with a speaker prompt, and sends it by Bluetooth Low Energy or Wi-Fi.



Figure 1. CAD render of the handheld unit on its magnetic charging dock (mouthpiece shown in the planned crop-residue colour).

The body is printed in PETG as ten parts held by M3 screws into brass heat-set inserts. The venturi tube lifts out and can be washed. Even small assembly details were designed to be buildable: for example, the push-button and status LED are soldered to a small daughter board first, the board is dropped onto two wall posts, and two M2 screws lock it, so no part is trapped behind printed features. The estimated material cost of the whole kit is about 3,100 baht (about USD 85) in our bill of materials; this is a design estimate, not a purchase receipt.



Figure 2. Transparent view of the internal parts: sensor, round display, battery, microcontroller and wiring.

2.3 From burned fields to the mouthpiece

Each person must blow into a clean mouthpiece, so a spirometer generates a stream of single-use parts. Our design makes the mouthpiece a separate, replaceable piece that the device counts (a reminder appears after 20 blows or 7 days). The prototype mouthpiece is printed in PETG to check size and fit, but it is not biodegradable. Our proposal for production is to replace it with moulded pulp made from field-burned residue.

We deliberately chose rice straw and sugarcane leaves rather than bagasse. Bagasse, the fibre left after cane is crushed, is mostly burned inside the sugar mills as fuel to generate electricity (about 30.7 million tonnes in 2016 [6]), so it is not the waste that pollutes the air. The residue that is still burned in the open is the field residue: cane leaves and tops are estimated at about 20 million tonnes per year, of which about 12 million tonnes were open-burned [7], and rice straw makes up most of Thailand's openly burned crop residue [2]. Evidence that such fibre can be moulded is encouraging: a University of Bangor study found trays with up to 80% straw were stronger in tension than expanded polystyrene and lost about 20% of their mass in four weeks in unsterilised soil [8]; a Thai study moulded rice-straw and bagasse pulp with tapioca starch and reduced water absorption with a chitosan coating [9]. Disposable air tubes for spirometers made of cardboard, paper or other biodegradable materials are also described in patents [10], and a university student project used cardboard mouthpieces on a reusable body, reporting that 28 of 30 volume measurements met ATS accuracy limits after calibration [11].

Stage	Material	Purpose	Status
Prototype	PETG (3D-printed)	Check size, fit and airflow resistance	Used in the current design
Production	Moulded pulp from rice straw / cane leaves	Compostable, uses field-burned residue	Proposal; supplier and hygiene testing needed
Fallback	Standard cardboard spirometer mouthpiece	Available on the market today	Needs socket size check (ours 30.4 mm)

Two honest cautions apply. First, a mouthpiece is not a bacterial filter, so single-person use and hygiene rules still matter. Second, we found no peer-reviewed measurement of airflow resistance or accuracy for crop-residue pulp mouthpieces in a spirometer; the studies above support the material, not the medical use. That measurement is part of our plan.

2.4 An online system for patients, doctors and administrators

Data are only useful if the right person can see them. We built a web service with three roles. A patient registers, blows into the device or follows the on-screen coaching, and sees a history and trend. A doctor, after approval by an administrator, uses a clinic mode that guides a visit (select a patient, prepare, at least three blows, repeatability check), shows a live flow-volume curve, and prints an A4 report. A patient can share results with one doctor by entering a doctor code, and can unlink, export or delete their data at any time. An administrator approves doctors, manages accounts and settings, and can review an audit log. The site runs on GitHub Pages with Firebase Authentication and Cloud Firestore. Passwords are held only inside Firebase Authentication; the application never stores, shows or exports them. Database

rules enforce that a patient sees only their own results, a doctor sees only linked patients, and no one can promote themselves to administrator. We tested these rules with live accounts, including attempts at privilege escalation, which were rejected. Health data are sensitive personal data, so real use would also require consent, a retention policy and independent review of the rules.

2.5 What is not proven yet

Because the device has not been built and measured, the following limits are important to state clearly:

- The device has been designed, simulated and the firmware compiled, but it has not yet been printed, assembled or measured. The constant K must be calibrated with a 3 L calibration syringe, and accuracy and repeatability must be checked against a reference spirometer.
- To stop at the end of a blow, the firmware ends the test when flow stays under 0.25 L/s. A model calculation showed this cuts the slow tail of a healthy blow, so FVC may read about 12% low and the FEV1/FVC ratio may read too high, which could hide mild obstruction. The firmware now estimates the missing tail, but this correction is unvalidated.
- Our acceptance rule requires at least 2 s of blowing, looser than the 6 s in ATS/ERS guidance, and the interpretation uses a fixed FEV1/FVC cut-off of 0.70 with predicted values typed in by the user because we did not include reference equations whose coefficients we could not verify.
- No clinical study has been done, and no regulatory approval exists. Spiro is for education and screening only.

Our validation plan is straightforward: calibrate with the 3 L syringe at several flow rates, compare 20 or more volunteers against a hospital spirometer with consent and teacher/physician supervision, measure the resistance of the pulp mouthpiece, and report the results whether or not they are favourable.

3. Conclusion

Spiro joins three ideas that are usually treated separately: a low-cost printed spirometer, an accounts-based service that connects patients with doctors, and a mouthpiece that turns burned crop residue into a useful, compostable product. The first design choice - a washable measuring body with a replaceable mouthpiece - makes the third idea possible. If calibration and validation confirm acceptable accuracy, schools and health posts in smoke-affected areas could screen lungs early, and farming communities would have a reason to collect straw instead of burning it. If validation fails, the open design, bill of materials and software still allow other young engineers to find and fix the problem, which is also a way of bridging ideas and youth creativity. Our next steps are printing and calibration, supplier testing of straw-pulp mouthpieces, and a supervised comparison with a reference spirometer.

References

[1] Mahidol University Repository. Health-damage assessment of crop-residue open burning in Thailand (rice, sugarcane, maize, cassava, oil palm). https://repository.li.mahidol.ac.th/handle/123456789/98397

[2] Thai Environmental Engineering Journal (TCI-Thaijo). Health impacts of rice-straw burning scenarios. https://so05.tci-thaijo.org/index.php/teej/article/download/256260/174186/958025

[3] Graham BL, et al. Standardization of spirometry 2019 update: an official ATS/ERS technical statement. Am J Respir Crit Care Med. 2019;200(8):e70-e88. doi:10.1164/rccm.201908-1590ST

[4] Miller MR, et al. Standardisation of spirometry. Eur Respir J. 2005;26(2):319-338. doi:10.1183/09031936.05.00034805

[5] 3D-printed Venturi spirometer with ESP32. Frontiers in Bioengineering and Biotechnology. 2025. doi:10.3389/fbioe.2025.1664127

[6] Sugar Tech (Springer). Sugar industry by-products in Thailand (bagasse 2016 figure). 2018. https://link.springer.com/10.1007/s12355-018-0599-x

[7] Thaksin University e-book. Sugarcane field residues and open burning in Thailand. https://ebook.tsu.ac.th/store/book/research/full2013_e/full2013_e/assets/basic-html/page13.html

[8] Bangor University. Feasibility of using straw in a strong, thin, pulp moulded packaging material. Industrial Crops and Products. https://research.bangor.ac.uk/portal/en/researchoutputs/feasibility-of-using-straw-in-a-strong-thin-pulp-moulded-packaging-material(d81fbefc-e0b5-45b1-9006-a1bcc9c9c66d)/export.html

[9] Production of molded pulp packaging from rice straw and bagasse coated by chitosan. Burapha Science Journal. https://li05.tci-thaijo.org/index.php/buuscij/article/view/1311

[10] US Patents 5,980,466 and 6,176,833. Biodegradable air tube and spirometer employing same. https://patents.google.com/patent/US6176833

[11] University of Wisconsin-Madison Biomedical Engineering Design. Spirometer final poster (student project, not peer reviewed). https://bmedesign.engr.wisc.edu/public/projects/f09/spirometer/file/view/408e0168-4392-4fac-9b9c-8246935e60cf/Final_Poster_12_3.pdf

Attachments

A. Product page and 3D viewer: https://spiro-medthai.web.app/

B. Clinic and patient web application (sign-in): https://spiro-medthai.web.app/login.html

C. Open design files (CAD scripts, STL, firmware, bill of materials): https://github.com/TANACHAT007/diy-spirometer

Note: figures are CAD renders of the design, not photographs of a built device.