DFM
Danish National Metrology Institute

DFM A/S
Kogle Allé 5, 2970 Hørsholm, Denmark
Tlf. +45 7730 5800 | www.dfm.dk

Certificate of Calibration No XXX-01234

Object	**Description of the test item, clear identification**
Order	Brief description of the order (If the calibration scope is clear from the calibration item (e.g. resistance, gauge block), this section can be omitted in exceptional cases.)
Applicant	**Name und Adresse des Auftraggebers** Firma, [Abteilung], Adresse PLZ Ort
Traceability	The reported measurement values are traceable to national standards and thus to internationally supported realisations of the SI units.
Date of Calibration	dd.mm.yyyy
Marking	Calibration label DFM mm.yyyy

2970 Hørsholm, 25 October 2021

For the Measurements	First Name Last Name
Approved by	Dr First Name Last Name, Head of sector Sector xxx

Mutual recognition
This certificate is consistent with Calibration and Measurement Capabilities (CMCs) that are included in Appendix C of the Mutual Recognition Arrangement (MRA) drawn up by the International Committee for Weights and Measures. Under the MRA, all participating institutes recognize the validity of each other's calibration certificates and measurement reports for the quantities, ranges and measurement uncertainties specified in Appendix C (for details see www.bipm.org).

This document is only valid and reviewable in its electronic form.
Please observe the information given on www.daniamet.dk/dcc.

DFM
Kogle Allé 5, 2970 Hørsholm, Denmark, phone +45 7730 5800, www.dfm.dk

Extent of the Calibration

bla

Measurement Procedure

bla

Measurement Conditions

bla

Measurement Results

bla

Uncertainty of Measurement

The reported uncertainty of measurement is stated as the combined standard uncertainty multiplied by a coverage factor $k = 2$. The measured value (y) and the associated expanded uncertainty (U) represent the interval ($y \pm U$) which contains the value of the measured quantity with a probability of approximately 95 %. The uncertainty was estimated following the guidelines of the ISO (GUM:2008).
The measurement uncertainty contains contributions originating from the measurement standard, from the calibration method, from the environmental conditions and from the object being calibrated. The longterm characteristic of the object being calibrated is not included.