Name	Toto
Age	0
Gender	female
Address	
Doctor	
User remark	
Acquisition System	NeuroScan
Sampling Rate (Hz)	250
Number of Channels	9
Monitor Selected	Monitor 0 (Main window)
Run Mode	USE
Synchronous Mode	Normal
Feature Extraction	TRCA
Number of Blocks	1
Data Saved Path	D:\zhengjiahan\uniBrain-Speller\Result
Accuracy (%)	N/A
ITR (bits/min)	N/A
Accuracy (%) and ITR (bits/min) over window length	Image not found
Commenced Date and Time	May 14, 2023, 08:18 PM
Total Run Time	1H 30M 54S
Keyboard Name	Keyboard 43 keys HS

Keyboard Processing Table:

Key Name	Frequency (Hz)	Phase (rad)	Training Accuracy (%)	Testing Accuracy (%)	No of Outputs
1	7.4	0.0	73.3	33.3	4
2	8.2	0.0	100.0	100.0	5
3	9.0	0.0	100.0	100.0	4
4	9.8	0.0	93.3	100.0	7
5	10.6	0.0	80.0	100.0	3
6	11.4	0.0	93.3	100.0	4
7	12.2	0.0	100.0	100.0	3
8	13.0	0.0	100.0	100.0	5
9	13.8	0.0	100.0	33.3	2
0	14.6	0.0	86.7	100.0	5
←	15.6	1.571	100.0	100.0	31
Q	7.6	1.571	100.0	100.0	6
W	8.4	1.571	100.0	100.0	5
E	9.2	1.571	100.0	100.0	19
R	10.0	1.571	100.0	100.0	9
T	10.8	1.571	80.0	66.7	11
Y	11.6	1.571	40.0	100.0	13
U	12.4	1.571	80.0	62.5	7
I	13.2	1.571	73.3	100.0	19
O	14.0	1.571	93.3	100.0	12
P	14.8	1.571	100.0	100.0	10
A	7.8	3.142	100.0	100.0	22
S	8.6	3.142	100.0	100.0	16
D	9.4	3.142	100.0	90.0	12
F	10.2	3.142	86.7	100.0	4
G	11.0	3.142	80.0	100.0	12
H	11.8	3.142	93.3	100.0	10
J	12.6	3.142	86.7	100.0	5
K	13.4	3.142	100.0	100.0	5
L	14.2	3.142	93.3	100.0	16
↵	15.0	3.142	93.3	100.0	11
■	8.0	4.712	100.0	85.7	12
Z	8.8	4.712	86.7	66.7	3
X	9.6	4.712	100.0	100.0	5

C	10.4	4.712	100.0	100.0	4
V	11.2	4.712	100.0	50.0	1
B	12.0	4.712	73.3	50.0	5
N	12.8	4.712	66.7	66.7	10
M	13.6	4.712	80.0	88.9	15
,	14.4	4.712	86.7	91.7	15
.	15.2	4.712	66.7	85.7	8
/	15.8	3.142	86.7	50.0	3
	15.4	0.0	73.3	87.2	49