



Conventional Calibration - Homopolymers : Results

Peak RV - (ml)	16.013
Mn - (Daltons)	12,356
Mw - (Daltons)	20,523
Mz - (Daltons)	31,822
Mp - (Daltons)	17,743
Mw / Mn	1.661
Percent Above Mw:	0
Percent Below Mw:	0
Mw 10.0% Low	4,606
Mw 10.0% High	54,593
Wt Fr (Peak)	1.000
RI Area - (mvml)	478.84
UV Area - (mvml)	0.00

Annotation	
Method File	PS CONV-0000.vcm
Limits File	
Date Acquired	Sep 07, 2022 - 22:42:14
Solvent	THF
Acquisition Operator	admin : Administrator
Calculation Operator	admin : Administrator
Column Set	2 x PLGel Mixed-B
System	THF
Flow Rate - (ml/min)	1.000
lnj Volume - (ul)	100.0
Volume Increment - (ml)	0.00333
Detector Temp. - (deg C)	40.0
Column Temp. - (deg C)	40.0
OmniSEC Build Number	406

