



Conventional Calibration - Homopolymers : Results

Peak RV - (ml)	16.177
Mn - (Daltons)	9,339
Mw - (Daltons)	14,995
Mz - (Daltons)	22,236
Mp - (Daltons)	14,146
Mw / Mn	1.606
Percent Above Mw: 0	100.000
Percent Below Mw: 0	0.000
Mw 10.0% Low	3,447
Mw 10.0% High	38,095
Wt Fr (Peak)	1.000
RI Area - (mvml)	304.64
UV Area - (mvml)	0.00

Annotation	
Method File	ps conv-0004.vcm
Limits File	
Date Acquired	Jan 18, 2023 - 00:14:21
Solvent	THF
Acquisition Operator	admin : Administrator
Calculation Operator	admin : Administrator
Column Set	2 x PLGel Mixed-B
System	THF
Flow Rate - (ml/min)	1.000
lnj Volume - (ul)	100.0
Volume Increment - (ml)	0.00333
Detector Temp. - (deg C)	40.0
Column Temp. - (deg C)	40.0
OmniSEC Build Number	406

