



Conventional Calibration - Homopolymers : Results

Peak RV - (ml)	15.933
Mn - (Daltons)	12,377
Mw - (Daltons)	18,265
Mz - (Daltons)	25,339
Mp - (Daltons)	17,877
Mw / Mn	1.476
Percent Above Mw: 0	100.000
Percent Below Mw: 0	0.000
Mw 10.0% Low	4,918
Mw 10.0% High	43,062
Wt Fr (Peak)	1.000
RI Area - (mvml)	99.49
UV Area - (mvml)	0.00

Annotation	
Method File	ps conv-0003.vcm
Limits File	
Date Acquired	Oct 04, 2022 - 22:34:49
Solvent	THF
Acquisition Operator	admin : Administrator
Calculation Operator	admin : Administrator
Column Set	2 x PLGel Mixed-B
System	THF
Flow Rate - (ml/min)	1.000
lnj Volume - (ul)	100.0
Volume Increment - (ml)	0.00333
Detector Temp. - (deg C)	40.0
Column Temp. - (deg C)	40.0
OmniSEC Build Number	406

