



Conventional Calibration - Homopolymers : Results

Peak RV - (ml)	15.783
Mn - (Daltons)	19,866
Mw - (Daltons)	29,029
Mz - (Daltons)	40,915
Mp - (Daltons)	24,356
Mw / Mn	1.461
Percent Above Mw: 0	100.000
Percent Below Mw: 0	0.000
Mw 10.0% Low	8,222
Mw 10.0% High	70,786
Wt Fr (Peak)	1.000
RI Area - (mvml)	113.37
UV Area - (mvml)	0.00

Annotation	
Method File	PS CONV-0003.vcm
Limits File	
Date Acquired	Feb 28, 2023 - 14:26:31
Solvent	THF
Acquisition Operator	admin : Administrator
Calculation Operator	admin : Administrator
Column Set	2 x PLGel Mixed-B
System	THF
Flow Rate - (ml/min)	1.000
lnj Volume - (ul)	100.0
Volume Increment - (ml)	0.00333
Detector Temp. - (deg C)	40.0
Column Temp. - (deg C)	40.0
OmniSEC Build Number	406

