



Conventional Calibration - Homopolymers : Results

Peak RV - (ml)	17.050
Mn - (Daltons)	4,344
Mw - (Daltons)	5,629
Mz - (Daltons)	7,346
Mp - (Daltons)	4,780
Mw / Mn	1.296
Percent Above Mw: 0	100.000
Percent Below Mw: 0	0.000
Mw 10.0% Low	2,107
Mw 10.0% High	12,526
Wt Fr (Peak)	1.000
RI Area - (mvml)	658.57
UV Area - (mvml)	0.00

Annotation	
Method File	ps conv-0003.vcm
Limits File	
Date Acquired	Oct 04, 2022 - 17:23:49
Solvent	THF
Acquisition Operator	admin : Administrator
Calculation Operator	admin : Administrator
Column Set	2 x PLGel Mixed-B
System	THF
Flow Rate - (ml/min)	1.000
lnj Volume - (ul)	100.0
Volume Increment - (ml)	0.00333
Detector Temp. - (deg C)	40.0
Column Temp. - (deg C)	40.0
OmniSEC Build Number	406

