



#### Conventional Calibration - Homopolymers : Results

Peak RV - (ml)	16.097
Mn - (Daltons)	11,692
Mw - (Daltons)	18,534
Mz - (Daltons)	27,614
Mp - (Daltons)	15,750
Mw / Mn	1.585
Percent Above Mw:	0
Percent Below Mw:	0
Mw 10.0% Low	4,486
Mw 10.0% High	47,720
Wt Fr (Peak)	1.000
RI Area - (mvml)	558.92
UV Area - (mvml)	0.00

Annotation	
Method File	ps conv-0004.vcm
Limits File	
Date Acquired	Jan 17, 2023 - 23:43:15
Solvent	THF
Acquisition Operator	admin : Administrator
Calculation Operator	admin : Administrator
Column Set	2 x PLGel Mixed-B
System	THF
Flow Rate - (ml/min)	1.000
lnj Volume - (ul)	100.0
Volume Increment - (ml)	0.00333
Detector Temp. - (deg C)	40.0
Column Temp. - (deg C)	40.0
OmniSEC Build Number	406

