



Conventional Calibration - Homopolymers : Results

Peak RV - (ml)	15.683
Mn - (Daltons)	24,034
Mw - (Daltons)	31,506
Mz - (Daltons)	40,968
Mp - (Daltons)	27,466
Mw / Mn	1.311
Percent Above Mw: 0	100.000
Percent Below Mw: 0	0.000
Mw 10.0% Low	11,133
Mw 10.0% High	69,558
Wt Fr (Peak)	1.000
RI Area - (mvml)	67.39
UV Area - (mvml)	0.00

Annotation	
Method File	ps conv-0004.vcm
Limits File	
Date Acquired	Jan 18, 2023 - 06:58:41
Solvent	THF
Acquisition Operator	admin : Administrator
Calculation Operator	admin : Administrator
Column Set	2 x PLGel Mixed-B
System	THF
Flow Rate - (ml/min)	1.000
lnj Volume - (ul)	100.0
Volume Increment - (ml)	0.00333
Detector Temp. - (deg C)	40.0
Column Temp. - (deg C)	40.0
OmniSEC Build Number	406

