
TP58R Printed: Friday, April 14, 2023 10:57:57 C:\Users\user\Documents\OmniSEC Data\THF\2023-04-13\2023-04-13_17;07;18_TP58R_01.vdt



Conventional	Calibration .	Homonolymers · Results	

Peak RV - (ml)	16.153
Mn - (Daltons)	9,959
Mw - (Daltons)	16,107
Mz - (Daltons)	23,519
Mp - (Daltons)	15,305
Mw / Mn	1.617
Percent Above Mw: 0	100.000
Percent Below Mw: 0	0.000
Mw 10.0% Low	3,613
Mw 10.0% High	40,283
Wt Fr (Peak)	1.000
RI Area - (mvml)	253.02
UV Area - (mvml)	0.00

Annotation	
Method File	PS CONV-0003.vcm
Limits File	
Date Acquired	Apr 13, 2023 - 17:07:18
Solvent	THF
Acquisition Operator	admin : Administrator
Calculation Operator	admin : Administrator
Column Set	2 x PLGel Mixed-B
System	THF
Flow Rate - (ml/min)	1.000
Inj Volume - (ul)	100.0
Volume Increment - (ml)	0.00333
Detector Temp (deg C)	40.0
Column Temp (deg C)	40.0
OmniSEC Build Number	406

