



Conventional Calibration - Homopolymers : Results

Peak RV - (ml)	16.543
Mn - (Daltons)	6,739
Mw - (Daltons)	11,128
Mz - (Daltons)	16,974
Mp - (Daltons)	9,182
Mw / Mn	1.651
Percent Above Mw:	0
Percent Below Mw:	0
Mw 10.0% Low	2,524
Mw 10.0% High	29,222
Wt Fr (Peak)	1.000
RI Area - (mvml)	352.04
UV Area - (mvml)	0.00

Annotation	
Method File	ps conv-0004.vcm
Limits File	
Date Acquired	Dec 06, 2022 - 17:38:31
Solvent	THF
Acquisition Operator	ADMIN : Administrator
Calculation Operator	ADMIN : Administrator
Column Set	2 x PLGel Mixed-B
System	THF
Flow Rate - (ml/min)	1.000
lnj Volume - (ul)	100.0
Volume Increment - (ml)	0.00333
Detector Temp. - (deg C)	40.0
Column Temp. - (deg C)	40.0
OmniSEC Build Number	406

