



Conventional Calibration - Homopolymers : Results

Peak RV - (ml)	15.847
Mn - (Daltons)	17,188
Mw - (Daltons)	25,585
Mz - (Daltons)	36,212
Mp - (Daltons)	22,248
Mw / Mn	1.488
Percent Above Mw:	0
Percent Below Mw:	0
Mw 10.0% Low	6,980
Mw 10.0% High	62,372
Wt Fr (Peak)	1.000
RI Area - (mvml)	110.44
UV Area - (mvml)	0.00

Annotation	
Method File	PS CONV-0000.vcm
Limits File	
Date Acquired	Sep 07, 2022 - 18:33:48
Solvent	THF
Acquisition Operator	admin : Administrator
Calculation Operator	admin : Administrator
Column Set	2 x PLGel Mixed-B
System	THF
Flow Rate - (ml/min)	1.000
lnj Volume - (ul)	100.0
Volume Increment - (ml)	0.00333
Detector Temp. - (deg C)	40.0
Column Temp. - (deg C)	40.0
OmniSEC Build Number	406

