Refractive Index

Conventional Calibration - Homopolymers : Results

Peak RV - (ml)	15.890
Mn - (Daltons)	13,788
Mw - (Daltons)	20,955
Mz - (Daltons)	29,752
Mp - (Daltons)	18,853
Mw / Mn	1.520
Percent Above Mw: 0	100.000
Percent Below Mw: 0	0.000
Mw 10.0% Low	5,371
Mw 10.0% High	51,013
Wt Fr (Peak)	1.000
RI Area - (mvml)	58.13
UV Area - (mvml)	0.00

Annotation	
Method File	ps conv-0000.vcm
Limits File	
Date Acquired	Oct 20, 2022 - 22:07:33
Solvent	THF
Acquisition Operator	admin : Administrator
Calculation Operator	admin : Administrator
Column Set	2 x PLGel Mixed-B
System	THF
Flow Rate - (ml/min)	1.000
Inj Volume - (ul)	100.0
Volume Increment - (ml)	0.00333
Detector Temp. - (deg C)	40.0
Column Temp. - (deg C)	40.0
OmniSEC Build Number	406

Molecular Weight vs. Retention Volume

WF / dLog M vs. Log Molecular Weight