

CALIBRA - Standardized Test Report

Session ID: 1

Date: 2026-09-09 15:32:09

1. Instrument Profile

Manufacturer: Mettler Toledo

Model: MS205DU

Accuracy Class: III

Max Capacity: 30.0

Min Capacity: 0.2

Verification Interval (e): 0.01

2. Test Results

Test: Weighing Performance

Status: REVIEW

Reason: Load exceeds Maximum Capacity for Class III (10000 e)