

CALIBRA - Standardized Test Report

Session ID: 4

Date: 2026-09-09 15:42:22

1. Instrument Profile

Manufacturer: Mettler Toledo
Model: MS205DU
Accuracy Class: Class III
Max Capacity: 30.0 kg
Min Capacity: 0.2 kg
Verification Interval (e): 10.0 g

2. Test Results

Test: Weighing Performance	Status: PASS
Reason: Automated MPE check	
Test: Weighing Performance	Status: FAIL
Reason: Automated MPE check	
Test: Weighing Performance	Status: REVIEW
Reason: Load exceeds Maximum Capacity for Class III (10000 e)	