

## EU DECLARATION OF CONFORMITY

This is a declaration made in accordance with the requirements of the following relevant Union harmonisation legislation. The manufacturer assures that the device that is covered by the present declaration is in conformity with this Regulation (EU) 2017/745 for Medical Devices and, if applicable, with any other relevant Union legislation that provides for the issuing of an EU declaration of conformity. The declaration of conformity is issued under the sole responsibility of the manufacturer.



**Manufacturer's Name:** SHANGHAI KOHDEN MEDICAL ELECTRONIC INSTRUMENT CORP.

**Address:** No. 567 Huancheng Bei Road, Shanghai Comprehensive Industrial Development Zone, Fengxian District, Shanghai, 201401, China

**SRN:** CN-MF-000013567

**European**

**Representative:** NIHON KOHDEN EUROPE GmbH

**Address:** Raiffeisenstrasse 10, 61191 Rosbach, Germany

**SRN:** DE-AR-000010740

■ **Regulation (EU) 2017/745(MDR)**

**Classification/Risk Class:** IIa

**Conformity assessment procedure:**

Annex IX Chapter I and III

**Notified Body**

**Name and No. :** BSI Group The Netherlands B.V.; 2797

**EC Certificate:** MDR 748567

■ **Directive 2011/65/EU and 2015/863/EU**

**Standard Applied:** EN 50581:2012

■ **Directive 2014/53/EU (RED)**

**Notified Body**

**Name and No. :** UL Japan, Inc., No.1731(Module B)

**EU-Type Examination**

**Certificate No. :** ULAR2107111

**Standard Applied:**

IEC 60601-1:2005, IEC 60601-1 Amendment 1:2012  
IEC 60601-1-2:2014  
IEC 60601-2-25:2011  
EN 60601-2-25:1995+A1:1999  
IEC 60601-2-25:1933+A1:1999  
EN 300 328 V2.2.2  
EN 301 893 V2.1.1  
EN 62311:2008

Declaration No.: SKXY-10005A

**Product Name, Model Number and Basic UDI-DI :**

Product Name	Model Number	Basic UDI-DI	MDR	RoHS	RED
Electrocardiograph	ECG-3350	697001088ECG-33506V	√	√	√
Software kit	QS-335E	697001088QS-335EA5	√	√	×
Accessory kit	YD-321D	697001088YD-321D7P	√	√	×
Accessory kit	YD-322D	697001088YD-322D7S	√	√	×
Accessory kit	YD-323D	697001088YD-323D7V	√	√	×
Accessory kit	YD-324D	697001088YD-324D7Y	√	√	×


**Intended purpose:** The Shanghai Kohden ECG-3350 electrocardiograph is intended for medical use to process the electrical signals produced by the heart, which are acquired through two or more electrodes, and to display waveforms and/or prepare a record of these electrical signals. This device is a portable ECG acquisition terminal which measures up to 12-lead ECG waveforms. The intended use is only for diagnosis, not for monitoring of vital physiological parameters.  
ECG-3350 is intended to be used by qualified medical personnel within a medical facility, such as hospital or clinic and children and adults can be diagnosed with this electrocardiograph.  
The software kit QS-335E is the software for ECG-3350. It is intended for upgrading.

**Additional Information:**

**Authorized Signatory:**

Shanghai, China/ 2023/6/1

Place and date of issue



General Manager

Quality Management Division