InstruCal Pro

DEMONSTRATION DOCUMENT

Full calibration certificate

Pressure transmitter PT-101 | 0-10 bar | 4-20 mA

Test information

Fictional standard PR-01 | Conditions: 22 °C, 48% RH

Fictional client: Demo Workshop

Recorded measurements

Point (%)	Reference (bar)	Increasing (bar)	Decreasing (bar)
0	0.00	0.00	0.00
25	2.50	2.49	2.50
50	5.00	5.01	5.02
75	7.50	7.52	7.51
100	10.00	10.01	10.01

Illustrative result

Illustrative expanded uncertainty: 0.04 bar (k = 2).

Fictional data, illustrative example.

No actual measurements or declaration of conformity.