InstruCal Pro

DEMONSTRATION DOCUMENT

Quick calibration certificate

Pressure transmitter PT-101 | 0-10 bar | 4-20 mA

Test information

Fictional reference PR-01

Fictional client: Demo Workshop

Recorded measurements

Point (%)	Reference (bar)	Reading (bar)	Error (bar)
0	0.00	0.00	0.00
25	2.50	2.49	-0.01
50	5.00	5.01	+0.01
75	7.50	7.52	+0.02
100	10.00	10.01	+0.01

Illustrative result

Maximum absolute error: 0.02 bar.

Fictional data, illustrative example.

No actual measurements or declaration of conformity.