

# US Patent & Trademark Office

## Patent Public Search | Text View

---

United States Patent	12383276
Kind Code	B2
Date of Patent	August 12, 2025
Inventor(s)	Plaza; Claudio et al.

---

### Systems and methods for delivery of stents and stent-like devices

---

#### Abstract

Systems for completely or partially excluding an aneurysm from circulation of blood are described. In one embodiment, the system includes a microcatheter, a fully or partially self-expandable stent, and a delivery device configured to be deliverable together with the stent through a lumen of the microcatheter. The delivery device includes an elongate support member coupled to a self-expandable portion, which includes a tubular mesh structure having a compressed state and an expanded state. A distal portion of the self-expandable portion extends proximally from a distal end of the self-expandable portion and has a length having an expanded outer diameter that is equal to or greater than the self-expanded inner diameter of the stent. In some embodiments, the proximal end of the self-expandable portion is substantially non-expanded where it is coupled to the elongate support member.

---

**Inventors:** Plaza; Claudio (Rancho Santa Margarita, CA), Thompson; James M. (Lake Forest, CA), Tran; Hung P. (Midway City, CA)

**Applicant:** MICROVENTION, INC. (Aliso Viejo, CA)

**Family ID:** 59787614

**Assignee:** MICROVENTION, INC. (Aliso Viejo, CA)

**Appl. No.:** 18/197814

**Filed:** May 16, 2023

#### Prior Publication Data

Document Identifier	Publication Date
US 20230285028 A1	Sep. 14, 2023

#### Related U.S. Application Data

continuation parent-doc US 17176744 20210216 US 11690629 child-doc US 18197814  
continuation parent-doc US 15453581 20170308 US 10952739 20210323 child-doc US 17176744

**Publication Classification**

**Int. Cl.:** **A61B17/12** (20060101); **A61F2/90** (20130101); **A61F2/966** (20130101); A61B17/00 (20060101); A61F2/82 (20130101)

**U.S. Cl.:**

**CPC** **A61B17/12118** (20130101); **A61B17/12109** (20130101); **A61B17/12172** (20130101); **A61F2/90** (20130101); A61B2017/00836 (20130101); A61B2017/1205 (20130101); A61F2002/823 (20130101); A61F2210/0009 (20130101); A61F2310/00029 (20130101)

**Field of Classification Search**

**CPC:** A61F (2/86); A61F (2/90); A61F (2/91); A61F (2/95); A61F (2/962); A61F (2/966); A61F (2002/9505); A61F (2002/9511); A61F (2002/9528); A61F (2002/9534); A61F (2002/9665)

---

**References Cited****U.S. PATENT DOCUMENTS**

<b>Patent No.</b>	<b>Issued Date</b>	<b>Patentee Name</b>	<b>U.S. Cl.</b>	<b>CPC</b>
5951569	12/1998	Tuckey et al.	N/A	N/A
6080177	12/1999	Igaki et al.	N/A	N/A
8043353	12/2010	Kaufmann et al.	N/A	N/A
9078658	12/2014	Hewitt et al.	N/A	N/A
9597087	12/2016	Marchand et al.	N/A	N/A
10952739	12/2020	Plaza	N/A	A61B 17/12118
11690629	12/2022	Plaza	606/198	A61B 17/12118
2005/0125052	12/2004	Iwata et al.	N/A	N/A
2005/0278011	12/2004	Peckham	N/A	N/A
2006/0184226	12/2005	Austin	N/A	N/A
2006/0190070	12/2005	Dieck et al.	N/A	N/A
2006/0200234	12/2005	Hines	N/A	N/A
2007/0055351	12/2006	Eidenschink et al.	N/A	N/A
2008/0082165	12/2007	Wilson et al.	N/A	N/A
2009/0216309	12/2008	Granada et al.	N/A	N/A
2011/0082490	12/2010	Connelly et al.	N/A	N/A
2011/0276121	12/2010	Levine	N/A	N/A
2012/0203325	12/2011	Weisman et al.	N/A	N/A
2012/0271407	12/2011	Jones et al.	N/A	N/A
2012/0316632	12/2011	Gao	N/A	N/A
2013/0018452	12/2012	Weitzner et al.	N/A	N/A
2013/0297003	12/2012	Pinchuk	N/A	N/A
2014/0088689	12/2013	Bales, Jr. et al.	N/A	N/A
2014/0180387	12/2013	Khenansho et al.	N/A	N/A
2014/0303543	12/2013	Meade et al.	N/A	N/A
2014/0358178	12/2013	Hewitt et al.	N/A	N/A

**OTHER PUBLICATIONS**

WO, PCT/US17/21577 ISR and Written Opinion, May 25, 2017. cited by applicant

---

*Primary Examiner:* Scherbel; Todd J

*Attorney, Agent or Firm:* ONE LLP

---

**Background/Summary**

**RELATED APPLICATIONS** (1) This application is a continuation of U.S. application Ser. No. 17/176,744, filed Feb. 16, 2021, which is a continuation of U.S. application Ser. No. 15/453,581, filed Mar. 8, 2017, now issued as U.S. Pat. No. 10,952,739, which claims priority under 35 U.S.C. § 119(e) to U.S. Provisional Application Ser. No. 62/307,123, filed Mar. 11, 2016, entitled “Systems and Methods for Delivery of Stents and Stent-like Devices,” all of which are herein expressly incorporated by reference in their entireties for all purposes.

**FIELD OF THE INVENTION**

(1) Embodiments of devices and methods herein are directed to blocking a flow of fluid through a tubular vessel or into a small interior chamber of a saccular cavity or vascular defect within a mammalian body. More specifically, embodiments herein are directed to devices and methods for treatment of a vascular defect of a patient including some embodiments directed specifically to the treatment of cerebral aneurysms of patients.

**BACKGROUND**

(2) The mammalian circulatory system is comprised of a heart, which acts as a pump, and a system of blood vessels which transport the blood to various points in the body. Due to the force exerted by the flowing blood on the blood vessel, the blood vessels may develop a variety of vascular defects. One common vascular defect known as an aneurysm results from the abnormal widening of the blood vessel. Typically, vascular aneurysms are formed as a result of the weakening of the wall of a blood vessel and subsequent ballooning and expansion of the vessel wall. If, for example, an aneurysm is present within an artery of the brain, and the aneurysm should burst with resulting cranial hemorrhaging, death could occur.

(3) Surgical techniques for the treatment of cerebral aneurysms typically involve a craniotomy requiring creation of an opening in the skull of the patient through which the surgeon can insert instruments to operate directly on the patient's brain. For some surgical approaches, the brain must be retracted to expose the parent blood vessel from which the aneurysm arises. Once access to the aneurysm is gained, the surgeon places a clip across the neck of the aneurysm thereby preventing arterial blood from entering the aneurysm. Upon correct placement of the clip the aneurysm will be obliterated in a matter of minutes. Surgical techniques may be effective treatment for many aneurysms. Unfortunately, surgical techniques for treating these types of conditions include major invasive surgical procedures which often require extended periods of time under anesthesia involving high risk to the patient. Such procedures thus require that the patient be in generally good physical condition in order to be a candidate for such procedures.

(4) Various alternative and less invasive procedures have been used to treat cerebral aneurysms without resorting to major surgery. Some such procedures involve the delivery of embolic or filling materials into an aneurysm. The delivery of such vaso-occlusion devices or materials may be used to promote hemostasis or fill an aneurysm cavity entirely. Vaso-occlusion devices may be placed within the vasculature of the human body, typically via a catheter, either to block the flow of blood

through a vessel with an aneurysm through the formation of an embolus or to form such an embolus within an aneurysm stemming from the vessel. A variety of implantable, coil-type vaso-occlusion devices are known. The coils of such devices may themselves be formed into a secondary coil shape, or any of a variety of more complex secondary shapes. Vaso-occlusive coils are commonly used to treat cerebral aneurysms but suffer from several limitations including poor packing density, compaction due to hydrodynamic pressure from blood flow, poor stability in wide-necked aneurysms, and complexity and difficulty in the deployment thereof as most aneurysm treatments with this approach require the deployment of multiple coils. Recently, expandable mesh devices have been used to fill the volume of an aneurysm and to disrupt normal blood flow in the aneurysm.

(5) Another approach to treating aneurysms without the need for invasive surgery involves the placement of sleeves or stents into the vessel and across the region where the aneurysm occurs. Such devices maintain blood flow through the vessel while reducing blood pressure applied to the interior of the aneurysm. Certain types of stents are expanded to the proper size by inflating a balloon catheter, referred to as balloon expandable stents, while other stents are designed to elastically expand in a self-expanding manner. Certain subsets of these devices are sometimes referred to as flow diversion devices. Some stents are covered typically with a sleeve of polymeric material called a graft to form a stent-graft. Stents and stent-grafts are generally delivered to a preselected position adjacent a vascular defect through a delivery catheter. In the treatment of cerebral aneurysms, covered stents or stent-grafts have seen very limited use due to the likelihood of inadvertent occlusion of small perforator vessels that may be near the vascular defect being treated.

(6) In addition, current uncovered stents are generally not sufficient as a stand-alone treatment. In order for stents to fit through the microcatheters used in small cerebral blood vessels, their density is usually reduced such that when expanded there is only a small amount of stent structure bridging the aneurysm neck. Thus, they do not block enough flow to cause clotting of the blood in the aneurysm and are thus generally used in combination with vaso-occlusive devices, such as the coils discussed above, to achieve aneurysm occlusion.

(7) A number of aneurysm neck bridging devices with defect spanning portions or regions have been attempted; however, none of these devices have had a significant measure of clinical success or usage. A major limitation in their adoption and clinical usefulness is the inability to position the defect spanning portion to assure coverage of the neck. Existing stent delivery systems that are neurovascular compatible (i.e., deliverable through a microcatheter and highly flexible) do not have the necessary rotational positioning capability. Another limitation of many aneurysm bridging devices described in the prior art is poor flexibility. Cerebral blood vessels are tortuous and a high degree of flexibility is required for effective delivery to most aneurysm locations in the brain.

## SUMMARY

(8) An embodiment of the invention includes a system for treating an aneurysm in a cerebral vessel. The system includes an elongate tubular member having a proximal end, a distal end, and a lumen therebetween having an inner diameter  $d$ . The system also includes an expandable stent having a proximal end, a distal end, and a lumen therebetween. The expandable stent has a constrained state with an outer diameter  $od.sub.1$  that is configured for delivery through the lumen of the elongate tubular member, and an expanded state having an inside diameter  $id.sub.2$  and an outer diameter  $od.sub.2$ ;  $od.sub.2$  is greater than  $od.sub.1$ . The expanded state of the stent is configured for placement within the cerebral vessel adjacent the aneurysm. The system also includes a delivery device that includes an elongate member having proximal and distal ends, and a self-expandable portion having proximal and distal ends. The proximal end of the self-expandable portion may be coupled to the elongate member at or near the distal end of the elongate member. The self-expandable portion includes a tubular mesh structure having a constrained state with an outer diameter  $OD.sub.1$  that is configured for delivery through the lumen of the elongate tubular

member, and an expanded state having an outer diameter OD.sub.2. The stent is engaged with the self-expandable portion of the delivery device.

(9) In alternative embodiments of the inventions, the stent may have different dimensions. In one embodiment, the inside diameter id.sub.2 of the expanded state of the stent decreases between the proximal and distal ends of the stent. Alternatively, the inside diameter id.sub.2 of the expanded state of the stent increases between the proximal and distal ends of the stent. Alternatively, the inside diameter id.sub.2 of the expanded state of the stent is substantially the same between the proximal and distal ends of the stent. Alternatively, the inside diameter id.sub.2 of the expanded state of the stent is between about 2 mm and about 5 mm. Alternatively, the inside diameter id.sub.2 of the expanded state of the stent increases between about 0.5 mm and about 2.0 mm between the proximal and distal ends of the stent.

(10) In alternative embodiments of the inventions, the stent may be made from a tubular mesh. Alternatively, the stent may be made from a slotted tube having a plurality of struts. The stent may be completely self-expanding or partially self-expanding. In an alternative embodiment, for the tubular mesh embodiment of the stent, a braid angle of the stent may be different than a braid angle of the self-expandable portion. Alternatively, a braid angle of the stent is substantially similar to a braid angle of the self-expandable portion.

(11) In alternative embodiments of the inventions, the self-expandable portion of the delivery device may have different dimensions. The outer diameter OD.sub.2 of the expanded state of the self-expandable portion of the delivery device may decrease between the distal and proximal ends. Alternatively, the expanded state of the self-expandable portion has a tubular portion having the outer diameter OD.sub.2 over a length of at least 2 mm. Alternatively, the expanded state of the self-expandable portion has a tubular portion having the outer diameter OD.sub.2 over a length of about 2 mm to about 15 mm. Alternatively, the expanded state of the self-expandable portion has a tubular portion having the outer diameter OD.sub.2 over a length of about 3 mm to about 10 mm.

(12) In alternative embodiments, at least part of the self-expandable portion of the delivery device is disposed within the lumen of the stent. In one embodiment, the at least part of the self-expandable portion mechanically engages at least a portion of the stent. In an alternative embodiment, the at least part of the self-expandable portion frictionally engages at least a portion of the stent. In an alternative embodiment, the at least part of the self-expandable portion mechanically engages at least a portion of the stent through intermeshing. In an alternative embodiment, the at least part of the self-expandable portion engages the stent substantially from the proximal end to the distal end of the stent. In an alternative embodiment, the at least part of the self-expandable portion mechanically engages the stent substantially from the proximal end to the distal end of the stent. In an alternative embodiment, the at least part of the self-expandable portion frictionally engages the stent substantially from the proximal end to the distal end of the stent.

(13) In other embodiments of the invention, the tubular mesh of the self-expandable portion of the delivery device includes a plurality of filaments. In alternative embodiments, the plurality of filaments comprises a cobalt-chromium alloy. In alternative embodiments, at least a portion of the plurality of filaments comprises a cobalt-chromium alloy. In alternative embodiments, the plurality of filaments each have a diameter of between about 0.0010 inches and 0.0020 inches. In alternative embodiments, the plurality of filaments each has a diameter of between about 0.0013 inches and 0.0017 inches. In alternative embodiments, the plurality of filaments each have a diameter of about 0.0015 inches. In alternative embodiments, the plurality of filaments are braided. In alternative embodiments, the braided filaments have a braid angle of between about 75° and about 80°.

(14) In another embodiment, a system for completely or partially excluding an aneurysm from circulation of blood is described. The system includes a microcatheter having a lumen, a fully or partially self-expandable stent, and a delivery device. The fully or partially self-expandable stent has a first end and a second end, and is configured to be deliverable through the lumen of the microcatheter. The stent also has a self-expanded inner diameter. The delivery device is configured

to be deliverable together with the stent through the lumen of the microcatheter. The delivery device includes an elongate support member having a proximal end, a distal end, and a self-expandable portion having a proximal end and a distal end. The proximal end of the self-expandable portion is coupled at or near the distal end of the elongate support member. The self-expandable portion of the delivery device includes a tubular mesh structure having a compressed state and an expanded state. A distal portion of the self-expandable portion extends proximally from the distal end of the self-expandable portion and has a length having an expanded outer diameter. The proximal end of the self-expandable portion is substantially non-expanded where it is coupled to the elongate support member. The expanded outer diameter of the distal portion of the self-expandable portion is equal to or greater than the self-expanded inner diameter of the stent.

(15) In another embodiment, the invention includes a delivery device configured to deliver a medical implant through the lumen of a catheter. The delivery device includes an elongate support member having a proximal end and a distal end, and a self-expandable portion having a proximal end and a distal end. The proximal end of the self-expandable portion is coupled at or near the distal end of the elongate support member. The self-expandable portion includes a tubular mesh structure having a compressed state and an expanded state. A distal portion of the self-expandable portion extends proximally from the distal end of the self-expandable portion and has a length having an expanded outer diameter. The proximal end of the self-expandable portion is substantially non-expanded where it is coupled to the elongate support member. The expanded outer diameter of the distal portion of the self-expandable portion is equal to or greater than a maximum inner diameter of the medical implant.

(16) The invention also includes various methods of treatment. In alternative embodiments, methods for treating an aneurysm in a cerebral vessel using the stents and delivery devices described in the various embodiments above are also described. The method includes the step of providing a microcatheter having a proximal end, a distal end, a lumen extending therebetween, a distal opening communicating with the lumen, and an inner diameter  $d$ . A system comprising an expandable stent and a delivery device, as described in the various embodiments above, is provided. The distal end of the microcatheter is advanced to a position distal to a neck of the aneurysm. The system is advanced within the lumen of the microcatheter, wherein both the expandable stent and self-expandable portion are each in the constrained state within the lumen of the microcatheter. At least part of the stent and at least part of the self-expandable portion of the delivery device are then released through the distal opening of the microcatheter, wherein the distal end of the stent is located in the cerebral vessel distal of the aneurysm neck.

(17) In alternative embodiments, the step of releasing at least part of the stent and at least part of the self-expandable portion of the delivery device can be carried out in different ways. The releasing step may occur by advancing the distal end of the stent and the distal end of the self-expandable portion of the delivery device through the distal opening of the microcatheter. Alternatively, the releasing step may occur by withdrawing the microcatheter proximally while the elongate member of the delivery device is held substantially in place, wherein the stent and self-expandable portion of the delivery device expand to their expanded states.

(18) In alternative embodiments, a method for treating an aneurysm in a cerebral vessel using the stents described in the various embodiments above may include the additional step of assessing the placement of the stent relative to the aneurysm neck (e.g., using a contrast (dye) injection).

(19) In alternative embodiments, where the physician is dissatisfied with the placement of the device, the method can further include the step of recapturing the at least part of the stent and at least part of the self-expandable portion within the lumen of the microcatheter. The step of recapturing the at least part of the stent and at least part of the self-expandable portion of the delivery device may include the step of advancing the microcatheter distally while maintaining traction on the elongate member of the delivery device to recapture the stent and self-expandable portion within the lumen of the microcatheter. Additionally, the distal end of the microcatheter can

then be repositioned to a position distal of a neck of the aneurysm. The stent and at least part of the self-expandable portion of the delivery device can then be released through the distal opening of the microcatheter, wherein the stent in its expanded state is positioned in the cerebral vessel such that the distal end of the stent is located distal of the aneurysm neck and the proximal end of the stent is located proximal of the aneurysm neck.

(20) In alternative embodiments, the step of recapturing the at least part of the stent and at least part of the self-expandable portion includes the step of advancing the microcatheter distally while maintaining traction on the elongate member of the delivery device to recapture the stent and self-expandable portion within the lumen of the microcatheter.

(21) In alternative embodiments, the method includes the step of advancing the microcatheter distally while applying traction to the proximal end of the elongate member of the delivery device, thereby disengaging the stent and the self-expandable portion, and withdrawing the self-expandable portion of the delivery device into the lumen of the microcatheter.

(22) In alternative embodiments, methods for treating an aneurysm in a cerebral vessel using the stents described in the various embodiments above are also described. The method includes the step of providing a microcatheter having a proximal end, a distal end, a lumen extending therebetween, a distal opening communicating with the lumen, and an inner diameter  $d$ . A system comprising an expandable stent and a delivery device, as described in the various embodiments above, is provided. The distal end of the microcatheter is advanced to a position distal to a neck of the aneurysm. The system is advanced within the lumen of the microcatheter, wherein both the expandable stent and self-expandable portion are each in the constrained state within the lumen. At least part of the stent and at least part of the self-expandable portion of the delivery device are then released through the distal opening of the microcatheter, wherein the distal end of the stent is located in the cerebral vessel distal of the aneurysm neck. At least part of the stent and at least part of the self-expandable portion is then recaptured within the lumen of the microcatheter. The distal end of the microcatheter is repositioned to a position distal of a neck of the aneurysm. The at least part of the stent and at least part of the self-expandable portion of the delivery device are then released through the distal opening of the microcatheter, wherein the stent and self-expandable portion of the delivery device expand to their expanded states, wherein the stent in its expanded state is positioned in the cerebral vessel such that the distal end of the stent is located distal of the aneurysm neck and the proximal end of the stent is located proximal of the aneurysm neck.

---

## Description

### BRIEF DESCRIPTION OF THE DRAWINGS

- (1) FIG. 1 is a perspective view of an embodiment of a delivery system for use with devices for treatment of a patient's vasculature.
- (2) FIG. 1A is a perspective view of an embodiment of a delivery system and a mesh therapeutic device (stent or flow diversion device) within an elongate tubular member.
- (3) FIG. 1B is a sectional view of FIG. 1A along plane 1B.
- (4) FIG. 1C is a sectional view of FIG. 1B along plane 1C.
- (5) FIG. 2 is an elevation view of a mesh therapeutic device (stent or flow diversion device) for treatment of a patient's vasculature.
- (6) FIG. 3 is a sectional view of a delivery device in its expanded state prior to compression within an introducer sheath.
- (7) FIG. 4 is a sectional view of the delivery device of FIG. 3 in an at least partially compressed state within an introducer sheath.
- (8) FIG. 5 is a sectional view of the delivery device in its compressed state within the introducer sheath and a mesh therapeutic device in its expanded state.

- (9) FIG. 6 is a sectional view of the delivery device and the mesh therapeutic device being coupled together in longitudinal alignment.
- (10) FIG. 7 is a sectional view of the delivery device and the mesh therapeutic device fully coupled together in longitudinal alignment.
- (11) FIG. 8 is a sectional view of the delivery device and the mesh therapeutic device being inserted within the introducer sheath.
- (12) FIG. 8A is a sectional view showing an alternative insertion technique to that of FIG. 8.
- (13) FIG. 8B is a sectional view showing an alternative insertion technique to that of FIG. 8.
- (14) FIG. 9 is a sectional view of the delivery device and the mesh therapeutic device positioned together completely within the introducer sheath.
- (15) FIG. 10 is a sectional view of a microcatheter in position in proximity to a cerebral aneurysm.
- (16) FIG. 11 is a sectional view of the delivery device and the mesh therapeutic device being delivered from the introducer sheath into a hub at the proximal end of a microcatheter.
- (17) FIG. 12 is a sectional view of the mesh therapeutic device being delivered by the delivery device from the distal end of a microcatheter adjacent the cerebral aneurysm.
- (18) FIG. 13 is a sectional view of the mesh therapeutic device partially delivered in an undesired position or configuration.
- (19) FIG. 14 is a sectional view of the mesh therapeutic device being retracted by the delivery device into the microcatheter.
- (20) FIG. 15 is a sectional view of the mesh therapeutic device fully delivered in a desired position and configuration spanning the cerebral aneurysm.
- (21) FIG. 16 is a sectional view of the delivery device compressed by the microcatheter, with the mesh therapeutic device remaining in position spanning the cerebral aneurysm.
- (22) FIG. 17 is a sectional view of the mesh therapeutic device in position spanning the cerebral aneurysm, with the delivery device and microcatheter retracted or removed.
- (23) FIG. 18 is a side view of a mesh therapeutic device in a curved condition.
- (24) FIG. 19 is a detailed view of FIG. 18 taken within circle 19.
- (25) FIG. 20 is a graph of braid diameter vs. radial hoop force.
- (26) FIG. 21 is a graph of braid diameter vs. bending stiffness.

#### DETAILED DESCRIPTION

(27) Discussed herein are devices and methods for the treatment of vascular defects that are suitable for minimally invasive deployment within a patient's vasculature, and particularly, within the cerebral vasculature of a patient. For such embodiments to be safely and effectively delivered to a desired treatment site and effectively deployed, some device embodiments may be configured for collapse to a low profile constrained state with a transverse dimension suitable for delivery through an inner lumen of a microcatheter and deployment from a distal end thereof. Unless otherwise stated, one or more of the features, dimensions, or materials of the various embodiments may be used in other similar embodiments discussed herein.

(28) FIG. 1 illustrates a delivery system 2 comprising a microcatheter 4 and a delivery device 6. FIG. 1A illustrates a delivery system 2 comprising a microcatheter 4, a delivery device 6, and a mesh therapeutic device 34, such as a stent or a flow diversion device. The microcatheter 4 comprises an elongate tubular structure 12 having a wall 14, and having a lumen 16 extending therethrough. The microcatheter 4 may comprise a single extrusion, or may be of a composite construction comprising a plurality of layers. For example, a braided and/or coil-reinforced or otherwise reinforced construction may be sandwiched between an outer polymeric layer and an inner polymeric layer. The construction of the microcatheter 4 may change along its length, such that the distal end has a greater degree of flexibility than the proximal end. A middle portion may have an intermediate degree of flexibility relative to the proximal and distal ends. The inner surface 18 of the microcatheter 4 may be constructed from a lubricious material, such as polyethylene, ETFE (ethylene tetrafluoroethylene) or PTFE (polytetrafluoro ethylene). The inner surface 18 of



the microcatheter may also, or alternatively, be coated with a lubricious material such as silicone to increase lubricity.

(29) The delivery device **6** may comprise a self-expandable portion **8** and an elongate support member **10**. The self-expandable portion **8** has a proximal end **20** and a distal end **22** and the elongate support member **10** has a proximal end (not shown) and a distal end **24**. The proximal end **20** of the self-expandable portion **8** may be coupled at or near the distal end **24** of the elongate support member **10**. The self-expandable portion **8** may comprise a tubular mesh structure (e.g., braided, woven, etc.) having a distal opening **26**. In some embodiments, the tubular mesh may be a braided tube which is formed from filaments **30** comprising shape memory alloy, such as Nitinol (nickel-titanium). The filaments of the self-expandable portion **8** may have a diameter of between about 0.0010 inches and 0.0030 inches, alternatively between about 0.0010 inches and 0.0020 inches, alternatively between about 0.0013 inches and 0.0020 inches, alternatively between about 0.0013 inches and 0.0017 inches, alternatively between about 0.0015 inches and 0.0020 inches. The filaments may have a diameter of about 0.0010 inches, alternatively about 0.0013 inches, alternatively about 0.0015 inches, alternatively about 0.0017 inches, alternatively about 0.0018 inches, alternatively about 0.0020 inches. The self-expandable portion **8** of the delivery device **6** may be made from between 12 and 96 filaments, alternatively between 18 and 72 filaments, alternatively between 36 and 48 filaments, alternatively between 12 and 1, alternatively between 72 and 180, alternatively greater than 96 filaments. The self-expandable portion **8** of the delivery device **6** may be made from about 12, 18, 36, 48, 72, 96, 120, or 180 filaments.

(30) The tubular mesh of self-expandable portion **8** may be heat-formed to create an expanded state (FIG. 3). In the expanded state, a distal portion **28** has an expanded outer diameter  $D$  over a length  $L$ . The expanded outer diameter  $D$  of the self-expandable portion **8** may be about 2.0 mm to about 6.0 mm, alternatively about 3.0 mm to about 6.0 mm, alternatively about 3.5 mm to about 6.0 mm, alternatively about 4.0 mm to about 6.0 mm, alternatively about 4.5 mm to about 6.0 mm, alternatively about 5.0 mm to about 6.0 mm, alternatively about 3.5 mm to about 5.5 mm, alternatively about 4.0 mm to about 5.0 mm. The expanded outer diameter  $D$  of the self-expandable portion **8** may be about 2.0 mm, alternatively about 2.5 mm, alternatively about 3.0 mm, alternatively about 3.5 mm, alternatively about 4.0 mm, alternatively about 4.5 mm, alternatively about 5.0 mm, alternatively about 5.5 mm, alternatively about 6.0 mm. The length  $L$  of the section of the delivery device **6** having the expanded diameter may be about 2-15 mm, alternatively about 3-10 mm, alternatively about 3-5 mm. The length  $L$  of the section of the delivery device **6** having the expanded diameter may be about 2 mm, alternatively about 3 mm, alternatively about 4 mm, alternatively about 7 mm, alternatively about 10 mm. When the self-expandable portion **8** which has been heat-formed is allowed to self-expand (unconstrained) to its maximum diameter (e.g., expanded outer diameter  $D$ , as in FIG. 3) the expanded outer diameter  $D$  may be configured to have a substantially constant diameter along its length  $L$  (as shown), or may have a diameter that tapers along its length. For example, it may taper to an increasing diameter from proximal to distal, or may taper to an increasing diameter from distal to proximal. In some embodiments, it may have a maximum diameter in a center portion of its length, and taper down in diameter toward the distal end and toward the proximal end. In FIG. 1, the distal end **22** and portion A of the self-expandable portion **8** are able to self-expand more than portion B, which is constrained within the lumen **16** of the microcatheter **4**. However, portion A is still not fully expanded, as the constriction caused by the microcatheter on portion B also affects the expansion of portion A, more so at a point near the distal end **32** of the microcatheter **4**, than at the distal end **22** of the self-expandable portion **8**.

(31) Turning to FIG. 2, a stent **34** having a proximal end **36** and a distal end **38** is shown. The term “stent” is intended to broadly describe a generally tubular or partially tubular structure which may be configured to engage the wall of a body lumen or duct of a patient, such as an artery or vein. “Stent” may describe a flow diversion device, such as a flow diversion device configured to extend within an artery at the site of an aneurysm. “Stent” may also describe a support lattice, such as

stents used in angioplasty procedures. The stent **34** may be completely self-expanding (reach its desired target diameter without aid), partially self-expanding (reach an increased diameter by self-expanding, but then reach its desired target diameter by forced expansion, such as expansion from the delivery device **6** described herein or a balloon placed within its interior), or may be configured to be expandable (e.g., expandable using the delivery device **6** described herein or a balloon). The stent **34** may comprise a tubular mesh structure (e.g., braided, woven, etc.) formed of a filaments **44** and having a distal opening **40**, a proximal opening **42**, and a lumen **43** extending therebetween. In some embodiments, the tubular mesh may be a braided tube which is formed from filaments **44** comprising a shape memory alloy, such as Nitinol (nickel-titanium). The stent **34** in FIG. 2 is shown in its self-expanded condition (unconstrained), and in this self-expanded condition has a self-expanded inner diameter  $d$ . The self-expandable portion **8** of delivery device **6** is configured to engage the stent from within the lumen **43**. For purposes of delivering the stent **34** into a vessel (e.g., artery) using the delivery device **6**, the expanded outer diameter  $D$  of the self-expandable portion **8** should be greater than or equal to the self-expanded inner diameter  $d$  of the stent **34**, stated by the equation:

$$D \geq d$$

(32) In some embodiments, the self-expanded inner diameter  $d$  of the stent **34** may be between about 2 mm and about 5 mm, or between about 3 mm and about 4 mm. The self-expanded inner diameter  $d$  of the stent **34** may also be about 2.0 mm, alternatively about 2.5 mm, alternatively about 3.0 mm, alternatively about 3.5 mm, alternatively about 4.0 mm, alternatively about 4.5 mm, alternatively about 5.0 mm, alternatively about 5.5 mm, alternatively about 6.0 mm. Stent **34** may have a total length of approximately 4 mm, alternatively about 5 mm, alternatively about 6 mm, alternatively about 10 mm, alternatively about 15 mm. In some cases, the self-expanded inner diameter  $d$  of the stent **34** may be tapered. For example, the stent **34** may be configured to extend in an artery that itself tapers, e.g., decreasing in diameter from proximal to distal. In some embodiments, the self-expanded inner diameter  $d$  may increase a total of between about 0.5 mm and about 2.0 mm over the length  $l$  of the stent **34**. In some embodiments, the self-expanded inner diameter  $d$  may increase a total of between about 1.0 mm and about 1.5 mm over the length  $l$  of the stent **34**. The expanded outer diameter  $D$  of the self-expandable portion **8** of delivery device **6** may also taper over its length in order to better mechanically engage with a tapered self-expanded inner diameter  $d$  of the stent **34**. In some embodiments, the expanded outer diameter  $D$  of the self-expandable portion **8** of delivery device **6** may stay generally constant, even though it is configured to mechanically engage with a tapered self-expanded inner diameter  $d$  of the stent **34**. For example, the self-expanded inner diameter  $d$  of the stent **34** may taper between 3.0 mm and 3.5 mm, and the expanded outer diameter  $D$  of the self-expandable portion **8** of delivery device **6** may stay constant at 3.75 mm. When it is engaged with the 3.0 mm to 3.5 mm inner diameter of the stent **34**, the 3.75 mm expanded outer diameter  $D$  may be semi-constricted, so that it conforms to the tapered inner diameter of the stent **34**.

(33) The filaments of the stent **34** may have a diameter of between about 0.0010 inches and 0.0030 inches, alternatively between about 0.0010 inches and 0.0020 inches, alternatively between about 0.0013 inches and 0.0020 inches, alternatively between about 0.0013 inches and 0.0017 inches, alternatively between about 0.0015 inches and 0.0020 inches. The filaments may have a diameter of about 0.0010 inches, alternatively about 0.0013 inches, alternatively about 0.0015 inches, alternatively about 0.0017 inches, alternatively about 0.0018 inches, alternatively about 0.0020 inches. The self-expandable portion **8** of the delivery device **6** may be made from between **36** and **144** filaments, alternatively between **48** and **108** filaments, alternatively between 48 and 96 filaments, alternatively between 72 and 96 filaments. The self-expandable portion **8** of the delivery device **6** may be made from about 36, 48, 72, 96, or 108 filaments. The total number of filaments in the stent may be about 2.0 to about 5 times, alternatively about 2.5 to about 4.25 times, alternatively about 2.6 to about 4.05 times, alternatively about 2.5 times to about 4.0 times,

alternatively about 2.0 times to about 4.0 times, alternatively about 2.0 times to about 3.5 times more than the number of filaments in the self-expandable portion **8** of the delivery device **6**.

(34) As illustrated in FIGS. **1A-1C**, the stent **34** is engaged with the self-expandable portion **8** of the delivery device **6**. The inner surface of the stent **34** is in contact with the outer surface of the self-expandable portion **8** of the delivery device **6**. The inner surface of the stent **34** may be engaged with the outer surface of the self-expandable portion **8** of the delivery device **6**, for example, through mechanical engagement, frictional engagement, or intermeshing. The inner surface of the stent **34** may be engaged to at least a portion of the outer surface of the self-expandable portion **8** of the delivery device **6** substantially from the proximal end to the distal end of the stent, which allows for improved control, placement, and delivery of the stent at the target site.

(35) Prior to loading the delivery device **6** and stent **34** together in an introducer sheath **46** (FIGS. **3-9**), the delivery device **6** and stent **34** may be mechanically engaged to each other. Mechanical engagement may in some cases be due predominantly to friction between the stent **34** and the self-expandable portion **8** of delivery device **6**, or may be due predominantly to intermeshing between the stent **34** and the self-expandable portion **8** of delivery device **6**. Mechanical engagement may also in some cases be due to a combination of friction and intermeshing between the stent **34** and the self-expandable portion **8** of delivery device **6**. Intermeshing may include “keying” between filaments **30** of the self-expandable portion **8** of delivery device **6** and filaments **44** (or struts) of the stent **34**. The intermeshing may be predominantly along a particular axis or may be along several axes. In some embodiments, the filaments **30** of the self-expandable portion **8** of delivery device **6** and filaments **44** of the stent **34** may be configured to match or register with each other, either by having similar or identical dimensions (diameters, thicknesses) or by being braided or woven with similar or identical dimensions (pics or crossings per unit dimension) or patterns (one filament over one filament, two filaments over two filaments, two filaments over one filament, etc.). In other embodiments, it may be desired to provide a stent **34** and a self-expandable portion **8** that do not match each other, for example, wherein they are mechanically different from each other. In cases wherein it is desired to have a self-expandable portion **8** that is relatively stiffer than the stent **34** (e.g., axial stiffness and/or radial stiffness), the stent **34** and the self-expandable portion **8** may be constructed differently from each other. For example, the braid angle of the stent may be different from the braid angle of the self-expandable portion **8** and/or the filament diameter of the stent **34** may be different from the filament diameter of the self-expandable portion **8**. The braid angle of the self-expandable portion **8** of the delivery device **6** may be between about 70° and about 85°, alternatively between about 73° and about 82°, alternatively between about 75° and about 80°. The braid angle of the stent **34** may be between about 70° and about 85°, alternatively between about 73° and about 82°, alternatively between about 75° and about 80°. The degree of mechanical engagement may be controlled by configuring the self-expandable portion **8** and the stent **34** such that an axial force *F* (FIG. **6**) applied to an external portion of the stent **34** can be allowed to reach a critical level without any disruption (e.g., relative axial displacement) between the self-expandable portion **8** and the stent **34**. This critical level may be about 227 grams.sub.f (0.5 pounds.sub.f) or greater, 454 grams.sub.f (1.0 pounds.sub.f) or greater, or 2,270 grams.sub.f (5.0 pounds.sub.f) or greater. This critical level may also be about 227 grams.sub.f (0.5 pounds.sub.f) to about 4,540 grams.sub.f (10.0 pounds.sub.f), alternatively about 454 grams.sub.f (1.0 pounds.sub.f) to about 2,270 grams.sub.f (5.0 pounds.sub.f). The axial force *F* at which the stent **34** becomes separated from the self-expandable portion **8** may be referred to as the separation force.

(36) FIGS. **3-9** illustrate a method for preparing a stent **34** for delivery via a delivery device **6**, from within an introducer sheath **46**. In FIG. **3**, the elongate support member **10** is backloaded (arrow **48**) through lumen **50** of the introducer **46** sheath. When the proximal end (not shown) of the elongate support member **10** exits the proximal end (not shown) of the introducer sheath **46**, elongate support member **10** exits the proximal end (not shown) of the introducer sheath **46**, a

tension T is placed on the elongate support member (at its proximal end), while also holding the introducer sheath **46**, and the self-expandable portion **8** is pulled into the lumen **50** of the introducer sheath **46** and is constricted by the inner surface **52** of the introducer sheath **46** (FIG. 4). It may be desired to leave a gap g between the distal opening **54** of the introducer sheath **46** and the distal end **22** of the self-expandable portion **8**. The dimension of the gap g may be chosen in order to aid longitudinal alignment with the self-expandable portion **8** with the stent **34**.

(37) Prior to loading the stent and delivery device into the introducer sheath, the stent and the delivery device can be aligned in different ways. As seen in FIG. 5, by placing the stent **34** over the introducer sheath **46** at a particular longitudinal position, the stent **34** may be longitudinally aligned, for example, either with the self-expandable portion **8**, or with the distal opening **54** (i.e., distal end) of the introducer sheath (or both). The distal end **38** of the sheath **34** may be the portion aligned, or the proximal end **36** of the sheath **34** may be the portion aligned. Alternatively, the proximal end **36** of the sheath **36** may be aligned with the proximal end **20** of the self-expandable portion **8**. In some embodiments, it may be desired to longitudinally center the length 1 of the stent **34** along the length L of the distal portion **28** of the self-expandable portion **8**. In some embodiments, it may be desired to skew or shift the length 1 of the stent **34** in relation to the length L of the distal portion **28** of the self-expandable portion **8**, either skewed distally or skewed proximally.

(38) The self-expandable portion **8** of the delivery device **6** must engage the stent **34** before loading both the delivery device and stent into the introducer sheath **46**. In FIG. 6, the self-expandable portion **8** is shown being extended from the introducer sheath (e.g., by placing a compressive force on the proximal end of the elongate support member **10**) such that the filaments **30** of the self-expandable portion **8** mechanically engage the filaments **44** of the stent **34** over an engagement length L.sub.E. As the self-expandable portion **8** is further extended, the engagement length L.sub.E increases, and the mechanical engagement (e.g., axial decoupling force) increases accordingly. In FIG. 7, the engagement length L.sub.E is at its maximum, as the stent **34** and the distal portion **28** of the self-expandable portion **8** are fully outside the lumen **50** of the introducer sheath **46**.

(39) There are numerous possible ways to load the stent **34** and delivery device **6** into the introducer sheath **46**. In order to load the stent **34** and the self-expandable portion **8** of delivery device **6** together into the lumen **50** of the introducer sheath **46**, an assembler may manipulate the stent **34** and the self-expandable portion **8** using fingers **56**, **58**, for example, first finger **56** and thumb **58** (FIG. 8). In some cases, the fingers **56**, **58** are used to pinch or squeeze the stent **34** and the self-expandable portion **8** to decrease their total diameter at the distal opening **54** of the introducer sheath **46** so that they are able to be inserted within the lumen **50**. Tension T (traction) may be applied on the proximal end of the elongate support member **10** while holding the proximal end of the introducer sheath **46** to further aid the insertion of the stent **34** and the self-expandable portion **8** within the lumen **50**.

(40) An alternative technique to that of FIG. 8 is shown in FIG. 8A. An insertion funnel **60** may be constructed by flaring a tube or by shrinking a shrink tube, in order to provide a small inner diameter **62** and an opposing large inner diameter **64**, with a tapered inner diameter **66** communicating therebetween. The dimension of the small inner diameter **62** can be made to be equal to or slightly less than the inner diameter of the lumen **50** of the introducer sheath **46**, thus allowing sliding insertion of the step-up **68** in outer diameter caused by the proximal end **36** of the stent **34**. A joint member **70** serves to align the lumen **50** of the introducer sheath **46** with the small inner diameter **62** of the insertion funnel **60**, and may be configured by shrinking a shrink tube over the introducer sheath **46** and the insertion funnel **60**. The large inner diameter **64** can be made larger than the outer diameter of the stent **34**, when it is engaged with the self-expandable portion **8**, thus allowing for smooth insertion. This diameter may be larger than the unconstrained outer diameter of the stent **34**. Once the entirety of the stent **34** and the self-expandable portion **8** are

inserted within the lumen 50, the insertion funnel 60 and the joint member 70 may be removed. Yet another alternative to the technique of FIG. 8 is shown in FIG. 8B. The introducer sheath 46 includes a flare 72 having a large inner diameter 74. The dimension of the large inner diameter 74 can be larger than the outer diameter of the stent 34, when it is engaged with the self-expandable portion 8, thus allowing for smooth insertion. Once the entirety of the stent 34 and the self-expandable portion 8 are inserted within the lumen 50 and pulled back so that the most distal end of the stent 34 and/or self-expandable portion are sufficiently out of the way, the flare 72 may be cut from the introducer sheath, for example at a cutting point 76. Thus, this cutting point will create the new distal opening 54 of the introducer sheath 46. FIG. 9 illustrates the stent 34 and self-expandable portion 8 fully retracted into the lumen 50 of the introducer sheath 46. The configuration of FIG. 9 may represent the packaging configuration of the stent 34 and the delivery device 6. Alternatively, the configuration of FIG. 9 may represent the configuration of the stent 34 and the delivery device 6, after preparation by a user. FIGS. 3-9 may represent the steps performed by a supplier. Or FIGS. 3-9 may represent the steps performed by a user.

(41) FIGS. 10-17 illustrate a delivery system 2 comprising a microcatheter 4 and a delivery device 6 in use for delivering a stent 34 (e.g., flow diversion device) across the neck 82 of an aneurysm 78 of an artery 80. In FIG. 10 a user inserts a microcatheter 4 into the vascular system of a patient, and tracks the microcatheter 4 such that the distal tip 32 of the microcatheter 4 is located in a position adjacent the neck 82 of the aneurysm 78. The microcatheter 4 may be inserted through a guiding catheter, which has been inserted through an introducer (e.g., valved introducer), which itself has been inserted into a peripheral access artery (femoral artery, brachial artery, radial artery, etc.), or even directly into a carotid artery. In FIG. 10, the distal tip 32 of the microcatheter 4 has been placed (or tracked) to a position just distal to the neck 82 of the aneurysm 78. With the microcatheter 4 in a desired position adjacent the neck 82 of the aneurysm 78, the stent 34 and the self-expandable portion 8 of the delivery device 6 are inserted together into the lumen 16 of the microcatheter 4, through the introducer sheath 46 (FIG. 11). With the stent 34 and the self-expandable portion 8 still completely retracted within the lumen 50 of the introducer sheath 46, the distal end 55 of the introducer sheath 46 is inserted into a cavity 86 within a hub 84 which is coupled to the microcatheter 4. The cavity 86 is coextensive with the lumen 16 of the microcatheter 4. The cavity 86 may comprise a luer taper and may include an interface (e.g., bump, groove) for interfacing with the distal end 55 of the introducer sheath 46. With the distal end 55 of the introducer sheath in position within the cavity 86, the distal opening 54 of the introducer sheath 46 empties into the cavity 86, thus allowing smooth, direct insertion of the stent 34 and the self-expandable portion 8 of delivery device 6 into the cavity 86 and into the lumen 16 of the microcatheter 4. The hub 84 may include a luer lock 88, for coupling to a syringe or to other luer connectors. In FIG. 11, the stent 34 and the self-expandable portion 8 are shown being inserted into the cavity 86, as the user applied an insertion force (compression) on the proximal end of the support member 10. After the stent 34 and the self-expandable portion 8 have been completely inserted into the lumen 16 of the microcatheter 4, and in some cases, after a proximal, stiff portion of the support member 10 has also been inserted into the lumen of the microcatheter 4, the introducer sheath 46 may be removed proximally, while holding the support member 10 distal of the introducer sheath 46.

(42) Now, the user is able to continue inserting the delivery device 6 by pushing on the proximal end of the support member 10 until, as shown in FIG. 12, the stent 34 begins to self-expand against the inner wall 90 of the artery 80. In addition, the self-expansive action of the self-expandable portion 8 of the delivery device 6 may further expand the stent 34 (e.g., increase its diameter further and/or increase the radial force it applies against the inner wall 90 of the artery 80). The microcatheter 4 may be pulled proximally (arrow 92) while the support member 10 of the delivery device 6 is held substantially static (via a compressive force applied to its proximal end) so as to deliver the stent 34 in place across the neck 82 of the aneurysm 78.

(43) Prior to and during the delivery of the devices the user may view the delivery of the stent **34** and the self-expandable portion by use of angiography or fluoroscopy or other imaging modalities. Radiopaque marker bands may be located on the distal end **32** of the microcatheter **4** and at one or more locations on the stent **34** and/or the self-expandable portion **8**. In addition, or alternatively, the materials (e.g., filaments) of the stent **34** and/or the self-expandable portion **8** may include radiopaque materials (platinum, platinum/tungsten, platinum/iridium, gold, etc.) so that they may be visualized in angiography or fluoroscopy. In some embodiments, the filaments may comprise drawn filled tubes (DFT), such as those supplied by Fort Wayne Metals, Fort Wayne, Indiana, USA. Such DFT filaments may comprise an outer shell of nickel-titanium and an inner core of platinum. In other embodiments, the filaments may comprise cobalt-chromium alloys or platinum-tungsten alloys. The outer diameter may range from about 0.0004 inches to 0.005 inches or from about 0.00075 inches to about 0.003 inches. The filaments may comprise a mixture of multiple different filament diameters or thicknesses and/or a mixture of different filament materials.

(44) As shown in FIG. **13**, in some cases, a user may discover, while delivering the stent **34**, that the stent **34** is about to be delivered in an undesirable location or position, for example, in relation to the neck **82** of the aneurysm **78**. For example, in FIG. **13**, the longitudinal distance between the distal end **38** of the stent **34** and a distal extremity **81** of the neck **82** may appear to be significantly greater than the longitudinal distance between the proximal end **36** of the stent **34** and a proximal extremity **83** of the neck **82**, if the stent **34** gets fully delivered. The user may determine that this is undesired, and thus has the ability to advance the microcatheter **4** (FIG. **14**, arrow **94**) while maintaining traction on the support member **10**, in order to recapture or recompress the stent **34** and the self-expandable portion **8** back within the lumen **16** of the microcatheter **4**. The mechanical engagement between the stent **34** and the self-expandable portion **8**, as described herein, allows the stent **34** to be delivered and to be retracted one-to-one (1:1) with the delivery device **6**, without any significant kinking, buckling, or stretching. The frictional and/or longitudinal engagement forces are distributed along the length of the stent **34**, both in the constricted configuration (inside the lumen **16**) and in the expanded configuration (e.g., FIG. **13**), thus avoiding kink points or stress concentrations that may cause kinking, buckling, or stretching. The distribution of the frictional and/or longitudinal engagement forces along the length of the stent **34** also allows for one-to-one (1:1) movement of the stent with the delivery device, i.e., for every 1.0 mm that the delivery device is advanced, the stent will also be advanced 1.0 mm.

(45) After retracting the stent **34** and the self-expandable portion **8** back into the lumen **16** of the microcatheter **4**, the microcatheter **4** can be repositioned, and the stent **34** can be redelivered, as shown in FIG. **15**. The user may prefer the positioning depicted in FIG. **15** because the stent **34** is relatively evenly positioned longitudinally in relation to the neck **82** of the aneurysm **78**, with the longitudinal distance between the distal end **38** of the stent **34** and a distal extremity **81** of the neck **82** generally comparable to the longitudinal distance between the proximal end **36** of the stent **34** and a proximal extremity **83** of the neck **82**. Once the user is satisfied with the position of the stent **34**, the user may then advance the microcatheter **4** (FIG. **16**, arrow **96**) while applying traction to the proximal end of the support member **10**. Because the end **36** of the stent **34** is outside of the lumen **16** of the microcatheter **4** (FIG. **15**) when the stent is fully deployed, the advancement of the microcatheter **4** as in FIG. **16** causes only the self-expandable portion **8** to be constrained within the lumen **16** of the microcatheter, leaving the stent **34** in place within the artery **80**, across the neck **82** of the aneurysm. The microcatheter **4** and the delivery device **6** may now be removed, leaving the stent **34** in place (FIG. **17**). The inner lumen **43** of the stent **34** may now allow blood flow through the artery **80**, and the filaments **44** of the stent **34** may limit a significant amount of blood from entering the aneurysm **78**, thus allowing stagnation of blood flow within the aneurysm **78** and thrombosis.

(46) FIG. **18** illustrates a stent **134** (or flow diversion device) having a proximal end **136** and a distal end **138**, and a distal opening **140** and a proximal opening **142**. The stent **134** is in a curved

condition, so that it has a convex side **131** and a concave side **133**. The woven or braided mesh may comprise a plurality of diamond-shaped modules **1008** (FIG. **19**). The diamond-shaped modules **1008** include diamond-shaped openings **1011**, bounded by filaments **1013**, **1015**, **1017**, **1019**. The diamond-shaped openings **1011** may each have low flow zones **1088**, especially as the angles **1078** are decreased. The flow of blood through the diamond-shaped openings **1011** varies depending upon the total cross-sectional area of the diamond-shaped openings **1011** and the perimeter of the diamond-shaped openings **1011**. The diamond-shaped openings **1011** may be considered to act as if they have a theoretical hydraulic diameter **1033**. Longitudinal axis Z.sub.L is shown in FIG. **19** and extends from a proximal end **136** to a distal end **138** of the stent **134**.

(47) In an embodiment of a stent **134** having a fixed diameter, fixed circumference, and a fixed number of filaments, the number of diamond-shaped modules **1008** fitting within the fixed circumference will not change, regardless of how sparsely or densely the braid is formed. Therefore, the module width **1084** (tangent to the diameter) can remain the same dimension, regardless of how sparsely or densely the braid is formed. However, the module length **1086** will be shorter on the concave side **133**, and the module length **1086** will be longer on the convex side **131**. This can occur when the stent **134** is placed in the curved condition as filament **1015** and filament **1017** slide over one another at crossing **1025** and filament **1013** and filament **1019** slide over one another at crossing **1029**, while angle **1082** (and angle across from angle **1082**) change. In conjunction, filament **1013** and filament **1015** will swivel in relation to one another at crossing **1023**, and filament **1017** and filament **1019** will swivel in relation to one another at crossing **1027** while angle **1078** (and the angle across from angle **1078**) changes. For example, along the convex side **131**, angle **1078** may be increased and angle **1082** may be reduced. And, along the concave side **133**, angle **1078** may be reduced and angle **1082** may be increased. It should be noted that angle **1082** in braiding nomenclature would be two times the “braid angle.”

$$(48) \text{BraidAngle}(\alpha) = \frac{(\text{angle}1082)}{2}$$

(49) Thus along the convex side **131**, the braid angle ( $\alpha$ ) is reduced and along the concave side **133**, the braid angle ( $\alpha$ ) is increased. There is also a mechanical difference between the convex side **131** and the concave side **133**, as the convex side **131** may become stiffer than the concave side.

(50) Additionally, average braid angle ( $\alpha$ ) may be changed at one location along the length of the stent **134** (between the proximal end **136** and a distal end **138**) in comparison to another location along the length. This may be controlled by varying the braiding process. For example, the average braid angle ( $\alpha$ ) at an intermediate portion **135** of the stent **134** may be made larger than at the braid angle at portions near the proximal end **136** or the distal end **138** of the stent **134**. In this manner, a higher filament density portion (having a smaller diamond-shaped opening **1011**) may be located at the intermediate portion **135**, which may be placed adjacent the neck **82** of the aneurysm **78**, thus better inhibiting blood flow into the aneurysm **78**. Traditionally, the formula for hydraulic diameter D.sub.H is:

$D_{\text{sub.H}} = (4 \times A_{\text{sub.O}}) / P_{\text{sub.O}}$  Where A.sub.O is the area of the diamond-shaped opening, and P.sub.O is the perimeter of the diamond-shaped opening.

(51) In anatomies wherein the artery **80** is curved, and the neck **82** of an aneurysm **78** is located on a convex portion of the curve of the artery **80**, the flow through the diamond-shaped opening **1011** may be increased above a desirable level, and may therefore also have an undesirable increased flow into the aneurysm **78**.

(52) In some cases, it may be preferred to construct the stent **134** with relatively large braid angles ( $\alpha$ ) (thus a larger angle **1082** and a smaller angle **1078**) in order to produce a smaller diamond-shaped opening **1011** (and smaller hydraulic diameter H.sub.D) in the stent **134** towards the neck **82** of the aneurysm, and to assure that the stent **134** is not unacceptably stiff. This way, there is greater chance of causing stagnation of the blood at the neck **82** to promote aneurysm embolization and occlusion. Additionally, there will be less possibility of causing damage to the artery **80** or aneurysm **78** because the more flexible stent **134** is also more atraumatic. Additionally, a stiff stent

might reside in the artery in a manner that it could actually buckle or project into the aneurysm. This may be particularly problematic in fusiform aneurysms. Although a more flexible stent **134** as described may minimize these problems, it may have reduced column stiffness, and thus may be very difficult to deliver on its own with greater tendency to buckle, kink, stretch, or otherwise deform when being delivered. Thus, the delivery device **6** comprising a self-expandable portion **8** and an elongate support member **10** can improve the safety and reliability of delivering a stent **134** that is more flexible and/or has a smaller braid angle ( $\alpha$ ) because the self-expandable portion **8** may be constructed to be relatively stiffer, which can to fully support and/or protect the stent **134** during delivery. The stent **34**, **134** characteristics and self-expandable portion **8** characteristics may be matched in order to produce desired mechanical characteristics of the pair during delivery and/or during retrieval, and so that the stent **134** mechanical characteristics and flow (hydraulic diameter of diamond-shaped opening **1011**, etc.) characteristics are also desired.

(53) In one embodiment of a system for completely or partially excluding an aneurysm from circulation of blood, the stent **34**, **134** and the self-expandable portion **8** have the following characteristics:

(54) TABLE-US-00001 Self-Expandable Portion 8 Stent 34, 134 Material Cobalt- Drawn Filled Tube Chromium (Nitinol outer shell/ (CoCr) platinum inner core) # of Filaments 18 72 Filament diameter 0.0015 inches 0.0013 inches Device Diameter 4.5 mm 4.5 mm Braid Angle ( $\alpha$ ) 80° 65° (Braid Angle = one-half of angle 1082) Collapsed profile 0.0093 inches 0.0161 inches

(55) In another embodiment of a system for completely or partially excluding an aneurysm from circulation of blood, the stent **34**, **134** and the self-expandable portion **8** have the following characteristics:

(56) TABLE-US-00002 Self-Expandable Portion 8 Stent 34, 134 Material Cobalt- Drawn Filled Tube Chromium (Nitinol outer shell/ (CoCr) platinum inner core) # of Filaments 18 48 Filament diameter 0.0015 inches 0.0015 inches Device Diameter 4.5 mm 4.5 mm Braid Angle ( $\alpha$ ) 75° 75° (one-half of angle 1082) Collapsed profile 0.0093 inches 0.0152 inches

(57) FIGS. **20** and **21** illustrate the changes in mechanical characteristics of four different braided tubes as their diameters are changed (e.g., by compression, tension, etc.). FIG. **20** illustrates the changes in radial hoop force (N) at different diameters. FIG. **21** illustrates changes in bending stiffness (N/mm) at different diameters. In the graphs of FIGS. **20** and **21**, the radial hoop force and the bending stiffness are normalized over a single unit length (total length of the braided tube), thus producing the displayed units of N/mm and N/mm/mm, respectively.

(58) In any of the embodiments, it is possible to include bioresorbable filaments, for example, filaments comprising (PGLA), (PGA), or (PLLA). It is even possible to make a fully bioresorbable mesh device. Bioresorbable metals such as magnesium, magnesium alloys, iron, or iron alloy may also be used to make bioresorbable filaments. In any of the embodiments, it is possible to coat at least some of the permeable shell or filaments with a growth factor, for example a CE34 antibody, in order to encourage the growth of endothelial cells to form a healing cap on an occluded aneurysm. The action of the CE34 antibody is to bind to an endothelial-derived growth factor.

(59) Although the foregoing invention has, for the purposes of clarity and understanding, been described in some detail by way of illustration and example, it will be obvious that certain changes and modifications may be practiced which will still fall within the scope of the appended claims.

## Claims

1. A system for treating an aneurysm, comprising: a microcatheter comprising a proximal end, a distal end, and a lumen therebetween, a fully or partially self-expandable stent having a first end and a second end deliverable through the lumen of the microcatheter, the stent having a self-expanded inner diameter; and a delivery device deliverable together with the stent through the lumen of the microcatheter, the delivery device comprising an elongate support member having a



proximal end and a distal end and a self-expandable portion having a proximal end, a distal end, an outer surface, and an inner surface, the proximal end of the self-expandable portion coupled at or near the distal end of the elongate support member; wherein the self-expandable portion of the delivery device comprises a tubular mesh structure having a compressed state and an expanded state, wherein the self-expandable portion has an open distal end, wherein a distal portion of the self-expandable portion extending proximally from the distal end of the self-expandable portion has a length having an expanded outer diameter, and the proximal end of the self-expandable portion is substantially non-expanded where it is coupled to the elongate support member; wherein the expanded outer diameter of the distal portion of the self-expandable portion is equal to or greater than the self-expanded inner diameter of the stent, and wherein in a configuration in the lumen of the microcatheter, the fully or partially self-expandable stent is engaged with the outer surface of the self-expandable portion of the delivery device along a length of the self-expandable stent from the first end to the second end, wherein in a delivery configuration, at least a portion of the delivery device is positioned distal to the distal end of the microcatheter, at least a portion of the fully or partially self-expandable stent is positioned distal to the distal end of the microcatheter, and the fully or partially self-expandable stent is engaged with the outer surface of the self-expandable portion of the delivery device.

2. The system of claim 1, wherein the self-expanded inner diameter of the stent tapers in diameter over a taper length positioned between the first end and the second end of the stent.
  3. The system of claim 1, wherein the expanded outer diameter of the distal portion of the self-expandable portion of the delivery device tapers in diameter.
  4. The system of claim 1, wherein the stent comprises a tubular mesh.
  5. The system of claim 1, wherein the stent comprises a slotted tube having a plurality of struts.
  6. The system of claim 1, wherein mechanical engagement between the expanded outer diameter of the distal portion of the self-expandable portion of the delivery device and the self-expanded inner diameter of the stent maintains the stent on the self-expandable portion.
  7. The system of claim 6, wherein the mechanical engagement between the expanded outer diameter of the distal portion of the self-expandable portion of the delivery device and the self-expanded inner diameter of the stent maintains the stent on the self-expandable portion with an axial separation force of at least 227 grams.
  8. The system of claim 7, wherein the mechanical engagement between the expanded outer diameter of the distal portion of the self-expandable portion of the delivery device and the self-expanded inner diameter of the stent maintains the stent on the self-expandable portion with an axial separation force of at least 454 grams.
  9. The system of claim 6, wherein the mechanical engagement is predominantly due to frictional engagement.
  10. The system of claim 6, wherein the mechanical engagement is predominantly due to intermeshing.
  11. The system of claim 1, wherein in the delivery configuration, the fully or partially self-expandable stent is positioned entirely distal to the distal end of the microcatheter.
  12. The system of claim 1, wherein in the delivery configuration, the self-expandable portion of the delivery device is positioned entirely distal to the distal end of the microcatheter.
-