

# US Patent & Trademark Office

## Patent Public Search | Text View

---

United States Patent	12390665
Kind Code	B1
Date of Patent	August 19, 2025
Inventor(s)	Duryea; Alexander P. et al.

---

### Histotripsy systems and methods

---

#### Abstract

A histotripsy therapy system configured for the treatment of tissue is provided, which may include any number of features. Provided herein are systems and methods that provide efficacious non-invasive and minimally invasive therapeutic, diagnostic and research procedures. In particular, provided herein are optimized systems and methods that provide targeted, efficacious histotripsy in a variety of different regions and under a variety of different conditions without causing undesired tissue damage to intervening/non-target tissues or structures.

---

**Inventors:** Duryea; Alexander P. (Ann Arbor, MI), Edsall; Connor (Ann Arbor, MI), Gerhardson; Tyler I. (Ann Arbor, MI)

**Applicant:** HistoSonics, Inc. (Plymouth, MN)

**Family ID:** 1000007618757

**Assignee:** HistoSonics, Inc. (Plymouth, MN)

**Appl. No.:** 18/498979

**Filed:** October 31, 2023

#### Related U.S. Application Data

continuation parent-doc US 18497856 20231030 PENDING child-doc US 18498979  
us-provisional-application US 63381401 20221028

---

#### Publication Classification

**Int. Cl.:** A61N7/00 (20060101)

**U.S. Cl.:**

**CPC** A61N7/00 (20130101); A61N2007/0039 (20130101); A61N2007/0095 (20130101)

## Field of Classification Search

**CPC:** A61N (7/00); A61N (2007/0039); A61N (2007/0095)

---

### References Cited

#### U.S. PATENT DOCUMENTS

Patent No.	Issued Date	Patentee Name	U.S. Cl.	CPC
3243497	12/1965	Kendall et al.	N/A	N/A
3679021	12/1971	Goldberg et al.	N/A	N/A
3693415	12/1971	Whittington	367/105	A61B 8/12
3879699	12/1974	Pepper	N/A	N/A
4016749	12/1976	Wachter	N/A	N/A
4024501	12/1976	Herring et al.	N/A	N/A
4051394	12/1976	Tieden	N/A	N/A
4117446	12/1977	Alais	N/A	N/A
4266747	12/1980	Souder, Jr. et al.	N/A	N/A
4269174	12/1980	Adair	N/A	N/A
4277367	12/1980	Madsen et al.	N/A	N/A
4351038	12/1981	Alais	N/A	N/A
4406153	12/1982	Ophir et al.	N/A	N/A
4440025	12/1983	Hayakawa et al.	N/A	N/A
4447031	12/1983	Souder, Jr. et al.	N/A	N/A
4453408	12/1983	Clayman	N/A	N/A
4483343	12/1983	Beyer et al.	N/A	N/A
4483345	12/1983	Miwa	N/A	N/A
4548374	12/1984	Thompson et al.	N/A	N/A
4549533	12/1984	Cain et al.	N/A	N/A
4550606	12/1984	Drost	N/A	N/A
4551794	12/1984	Sandell	N/A	N/A
4575330	12/1985	Hull	N/A	N/A
4622972	12/1985	Giebelier, Jr.	N/A	N/A
4625731	12/1985	Quedens et al.	N/A	N/A
4641378	12/1986	McConnell et al.	N/A	N/A
4669483	12/1986	Hepp et al.	N/A	N/A
4689986	12/1986	Carson et al.	N/A	N/A
4757820	12/1987	Itoh	N/A	N/A
4791915	12/1987	Barsotti et al.	N/A	N/A
4819621	12/1988	Ueberle et al.	N/A	N/A
4829491	12/1988	Saugeon et al.	N/A	N/A
4856107	12/1988	Dory	N/A	N/A
4865042	12/1988	Umemura et al.	N/A	N/A
4888746	12/1988	Wurster et al.	N/A	N/A
4890267	12/1988	Rudolph	N/A	N/A
4922917	12/1989	Dory	N/A	N/A
4928672	12/1989	Grasser	601/4	A61B 17/2258
4938217	12/1989	Lele	N/A	N/A

4957099	12/1989	Hassler	N/A	N/A
4973980	12/1989	Howkins et al.	N/A	N/A
4984575	12/1990	Uchiyama et al.	N/A	N/A
4991151	12/1990	Dory	N/A	N/A
4995012	12/1990	Dory	N/A	N/A
RE33590	12/1990	Dory	N/A	N/A
5014686	12/1990	Schafer	N/A	N/A
5065751	12/1990	Wolf	N/A	N/A
5078140	12/1991	Kwoh	N/A	N/A
5080101	12/1991	Dory	N/A	N/A
5080102	12/1991	Dory	N/A	N/A
5091893	12/1991	Smith et al.	N/A	N/A
5092336	12/1991	Fink	N/A	N/A
5097709	12/1991	Masuzawa et al.	N/A	N/A
5111822	12/1991	Dory	N/A	N/A
5143073	12/1991	Dory	N/A	N/A
5143074	12/1991	Dory	N/A	N/A
5150711	12/1991	Dory	N/A	N/A
5158070	12/1991	Dory	N/A	N/A
5158071	12/1991	Umemura et al.	N/A	N/A
5163421	12/1991	Bernstein et al.	N/A	N/A
5165412	12/1991	Okazaki	N/A	N/A
5174294	12/1991	Saito et al.	N/A	N/A
5195509	12/1992	Rentschler et al.	N/A	N/A
5209221	12/1992	Riedlinger	N/A	N/A
5215680	12/1992	D'Arrigo	N/A	N/A
5219401	12/1992	Cathignol et al.	N/A	N/A
5222806	12/1992	Roberts	N/A	N/A
5230340	12/1992	Rhyne	N/A	N/A
5295484	12/1993	Marcus et al.	N/A	N/A
5316000	12/1993	Chapelon et al.	N/A	N/A
5354258	12/1993	Dory	N/A	N/A
5380411	12/1994	Schlieff	N/A	N/A
5393296	12/1994	Rattner	N/A	N/A
5409002	12/1994	Pell	N/A	N/A
5431621	12/1994	Dory	N/A	N/A
5435311	12/1994	Umemura et al.	N/A	N/A
5443069	12/1994	Schaetzle	N/A	N/A
5450305	12/1994	Boys et al.	N/A	N/A
5469852	12/1994	Nakamura et al.	N/A	N/A
5474071	12/1994	Chapelon et al.	N/A	N/A
5474531	12/1994	Carter	N/A	N/A
5490051	12/1995	Messana	N/A	N/A
5492126	12/1995	Hennige et al.	N/A	N/A
5501655	12/1995	Rolt et al.	N/A	N/A
5520188	12/1995	Hennige et al.	N/A	N/A
5523058	12/1995	Umemura et al.	N/A	N/A
5524620	12/1995	Rosenschein	N/A	N/A
5540909	12/1995	Schutt	N/A	N/A
5542935	12/1995	Unger et al.	N/A	N/A

5558092	12/1995	Unger et al.	N/A	N/A
5563346	12/1995	Bartelt et al.	N/A	N/A
5566675	12/1995	Li et al.	N/A	N/A
5573497	12/1995	Chapelon	N/A	N/A
5580575	12/1995	Unger et al.	N/A	N/A
5582578	12/1995	Zhong et al.	N/A	N/A
5590657	12/1996	Cain et al.	N/A	N/A
5601526	12/1996	Chapelon et al.	N/A	N/A
5617862	12/1996	Cole et al.	N/A	N/A
5648098	12/1996	Porter	N/A	N/A
5665054	12/1996	Dory	601/3	A61N 7/02
5666954	12/1996	Chapelon et al.	N/A	N/A
5676452	12/1996	Scholz	N/A	N/A
5676692	12/1996	Sanghvi et al.	N/A	N/A
5678554	12/1996	Hossack et al.	N/A	N/A
5683064	12/1996	Copeland et al.	N/A	N/A
5694936	12/1996	Fujimoto et al.	N/A	N/A
5695460	12/1996	Siegel et al.	N/A	N/A
5717657	12/1997	Ruffa	N/A	N/A
5720287	12/1997	Chapelon et al.	N/A	N/A
5724972	12/1997	Petrofsky	N/A	N/A
5743863	12/1997	Chapelon	N/A	N/A
5753929	12/1997	Bliss	N/A	N/A
5759162	12/1997	Oppelt et al.	N/A	N/A
5766138	12/1997	Rattner	N/A	N/A
5769790	12/1997	Watkins et al.	N/A	N/A
5797848	12/1997	Marian et al.	N/A	N/A
5800365	12/1997	Zhong et al.	N/A	N/A
5820623	12/1997	Ng	N/A	N/A
5823962	12/1997	Schaetzle et al.	N/A	N/A
5827204	12/1997	Grandia et al.	N/A	N/A
5836896	12/1997	Rosenschein	N/A	N/A
5849727	12/1997	Porter et al.	N/A	N/A
5873902	12/1998	Sanghvi et al.	N/A	N/A
5879314	12/1998	Peterson et al.	N/A	N/A
5928169	12/1998	Schitzle et al.	N/A	N/A
5932807	12/1998	Mallart	N/A	N/A
5947904	12/1998	Hossack et al.	N/A	N/A
6001069	12/1998	Tachibana et al.	N/A	N/A
6007499	12/1998	Martin et al.	N/A	N/A
6022309	12/1999	Celliers et al.	N/A	N/A
6036667	12/1999	Manna et al.	N/A	N/A
6088613	12/1999	Unger	N/A	N/A
6093883	12/1999	Sanghvi et al.	N/A	N/A
6113558	12/1999	Rosenschein et al.	N/A	N/A
6126607	12/1999	Whitmore, III et al.	N/A	N/A
6128958	12/1999	Cain	N/A	N/A
6143018	12/1999	Beuthan et al.	N/A	N/A
6165144	12/1999	Talish et al.	N/A	N/A
6176842	12/2000	Tachibana et al.	N/A	N/A

6296619	12/2000	Brisken et al.	N/A	N/A
6308585	12/2000	Nilsson et al.	N/A	N/A
6308710	12/2000	Silva	N/A	N/A
6309355	12/2000	Cain et al.	N/A	N/A
6318146	12/2000	Madsen et al.	N/A	N/A
6321109	12/2000	Ben-Haim et al.	N/A	N/A
6338566	12/2001	Verdier	N/A	N/A
6344489	12/2001	Spears	N/A	N/A
6391020	12/2001	Kurtz et al.	N/A	N/A
6413216	12/2001	Cain et al.	N/A	N/A
6419648	12/2001	Vitek et al.	N/A	N/A
6470204	12/2001	Uzgiris et al.	N/A	N/A
6488639	12/2001	Ribault et al.	N/A	N/A
6490469	12/2001	Candy	N/A	N/A
6500141	12/2001	Irion et al.	N/A	N/A
6506154	12/2002	Ezion et al.	N/A	N/A
6506171	12/2002	Vitek et al.	N/A	N/A
6508774	12/2002	Acker et al.	N/A	N/A
6511428	12/2002	Azuma et al.	N/A	N/A
6511444	12/2002	Hynynen et al.	N/A	N/A
6522142	12/2002	Freundlich	N/A	N/A
6524251	12/2002	Rabiner et al.	N/A	N/A
6536553	12/2002	Scanlon	N/A	N/A
6543272	12/2002	Vitek	N/A	N/A
6556750	12/2002	Constantino et al.	N/A	N/A
6559644	12/2002	Froundlich et al.	N/A	N/A
6576220	12/2002	Unger	N/A	N/A
6599288	12/2002	Maguire et al.	N/A	N/A
6607498	12/2002	Eshel	N/A	N/A
6612988	12/2002	Maor et al.	N/A	N/A
6613004	12/2002	Vitek et al.	N/A	N/A
6613005	12/2002	Friedman et al.	N/A	N/A
6626854	12/2002	Friedman et al.	N/A	N/A
6626855	12/2002	Weng et al.	N/A	N/A
6645162	12/2002	Friedman et al.	N/A	N/A
6648839	12/2002	Manna et al.	N/A	N/A
6666833	12/2002	Friedman et al.	N/A	N/A
6685640	12/2003	Fry et al.	N/A	N/A
6685657	12/2003	Jones	N/A	N/A
6705994	12/2003	Vortman et al.	N/A	N/A
6719449	12/2003	Laugharn, Jr. et al.	N/A	N/A
6719694	12/2003	Weng et al.	N/A	N/A
6735461	12/2003	Vitek et al.	N/A	N/A
6736814	12/2003	Manna et al.	N/A	N/A
6750463	12/2003	Riley	N/A	N/A
6770031	12/2003	Hynynen et al.	N/A	N/A
6775438	12/2003	Gaedke et al.	N/A	N/A
6788977	12/2003	Fenn et al.	N/A	N/A
6790180	12/2003	Vitek	N/A	N/A
6820160	12/2003	Allman	N/A	N/A

6852082	12/2004	Strickberger et al.	N/A	N/A
6869439	12/2004	White et al.	N/A	N/A
6890332	12/2004	Truckal et al.	N/A	N/A
6929609	12/2004	Asafusa	N/A	N/A
7004282	12/2005	Manna et al.	N/A	N/A
7059168	12/2005	Hibi et al.	N/A	N/A
7128711	12/2005	Medan et al.	N/A	N/A
7128719	12/2005	Rosenberg	N/A	N/A
7175596	12/2006	Vitek et al.	N/A	N/A
7175599	12/2006	Hynynen et al.	N/A	N/A
7196313	12/2006	Quinones	N/A	N/A
7223239	12/2006	Schulze et al.	N/A	N/A
7258674	12/2006	Cribbs et al.	N/A	N/A
7273458	12/2006	Prausnitz et al.	N/A	N/A
7273459	12/2006	Desilets et al.	N/A	N/A
7300414	12/2006	Holland et al.	N/A	N/A
7311679	12/2006	Desilets et al.	N/A	N/A
7331951	12/2007	Eshel et al.	N/A	N/A
7341569	12/2007	Soltani et al.	N/A	N/A
7347855	12/2007	Eshel et al.	N/A	N/A
7358226	12/2007	Dayton et al.	N/A	N/A
7359640	12/2007	Onde et al.	N/A	N/A
7367948	12/2007	O'Donnell et al.	N/A	N/A
7374551	12/2007	Liang et al.	N/A	N/A
7377900	12/2007	Vitek et al.	N/A	N/A
7429249	12/2007	Winder et al.	N/A	N/A
7431704	12/2007	Babaev	N/A	N/A
7442168	12/2007	Novak et al.	N/A	N/A
7462488	12/2007	Madsen et al.	N/A	N/A
7559905	12/2008	Kagosaki et al.	N/A	N/A
7656638	12/2009	Laakso et al.	N/A	N/A
7695437	12/2009	Quistgaard et al.	N/A	N/A
7714481	12/2009	Sakai	N/A	N/A
7771359	12/2009	Adam	N/A	N/A
7967763	12/2010	Deem et al.	N/A	N/A
8057408	12/2010	Cain et al.	N/A	N/A
8295912	12/2011	Gertner	N/A	N/A
8333115	12/2011	Garvey et al.	N/A	N/A
8337407	12/2011	Quistgaard et al.	N/A	N/A
8342467	12/2012	Stachowski et al.	N/A	N/A
8376970	12/2012	Babaev	N/A	N/A
8539813	12/2012	Cain et al.	N/A	N/A
8568339	12/2012	Rybyanets	N/A	N/A
8636664	12/2013	Brannan	N/A	N/A
8715187	12/2013	Landberg Davis et al.	N/A	N/A
8845537	12/2013	Tanaka et al.	N/A	N/A
8932239	12/2014	Sokka et al.	N/A	N/A
9028434	12/2014	Tanaka	N/A	N/A
9049783	12/2014	Teofilovic	N/A	N/A

9061131	12/2014	Jahnke et al.	N/A	N/A
9144694	12/2014	Cain	N/A	N/A
9220476	12/2014	Coussios et al.	N/A	N/A
9228730	12/2015	Inbody	N/A	N/A
9302124	12/2015	Konofagou et al.	N/A	N/A
9457201	12/2015	Hoelscher et al.	N/A	N/A
9526923	12/2015	Jahnke et al.	N/A	N/A
9636133	12/2016	Hall et al.	N/A	N/A
9642634	12/2016	Cain et al.	N/A	N/A
9763688	12/2016	Stulen et al.	N/A	N/A
9901753	12/2017	Cain et al.	N/A	N/A
9943708	12/2017	Roberts et al.	N/A	N/A
10022107	12/2017	Thornton et al.	N/A	N/A
10046179	12/2017	Oskar-Kohler	N/A	N/A
10046181	12/2017	Barthe et al.	N/A	N/A
10058352	12/2017	Carvell et al.	N/A	N/A
10071266	12/2017	Cain	N/A	N/A
10130828	12/2017	Vortman et al.	N/A	N/A
10219815	12/2018	Maxwell et al.	N/A	N/A
10293187	12/2018	Cannata et al.	N/A	N/A
10751015	12/2019	Anderson et al.	N/A	N/A
10751125	12/2019	Levy et al.	N/A	N/A
10765892	12/2019	Vitek et al.	N/A	N/A
10772646	12/2019	Lu et al.	N/A	N/A
10780298	12/2019	Cain et al.	N/A	N/A
10791991	12/2019	Burkett et al.	N/A	N/A
10799209	12/2019	Lahti et al.	N/A	N/A
10806421	12/2019	Keller	N/A	N/A
10820813	12/2019	Alpert	N/A	N/A
10847264	12/2019	Mansker et al.	N/A	N/A
10849511	12/2019	Tochterman et al.	N/A	N/A
10869603	12/2019	Millett et al.	N/A	N/A
10869633	12/2019	Burkett	N/A	N/A
10869648	12/2019	Hubbard et al.	N/A	N/A
10874353	12/2019	Assif	N/A	N/A
10874409	12/2019	Matsubara et al.	N/A	N/A
10878586	12/2019	Brokman et al.	N/A	N/A
10888232	12/2020	Anderson et al.	N/A	N/A
10893808	12/2020	Dorando	N/A	N/A
10900933	12/2020	Prus et al.	N/A	N/A
10905394	12/2020	Stigall et al.	N/A	N/A
10912463	12/2020	Davies et al.	N/A	N/A
10925688	12/2020	Millett et al.	N/A	N/A
10927003	12/2020	Millett et al.	N/A	N/A
10932678	12/2020	Burkett	N/A	N/A
10939826	12/2020	Glynn et al.	N/A	N/A
10942022	12/2020	Johansson et al.	N/A	N/A
10973419	12/2020	Corl	N/A	N/A
10993618	12/2020	Mansker et al.	N/A	N/A
10993628	12/2020	Tochterman	N/A	N/A

10993694	12/2020	Meyer et al.	N/A	N/A
11000185	12/2020	Stigall et al.	N/A	N/A
11006840	12/2020	Stigall	N/A	N/A
11013491	12/2020	Rice et al.	N/A	N/A
11020087	12/2020	Hoffman	N/A	N/A
11020089	12/2020	Corl	N/A	N/A
11026591	12/2020	Burkett et al.	N/A	N/A
11040140	12/2020	Unser et al.	N/A	N/A
11058399	12/2020	Xu et al.	N/A	N/A
11071522	12/2020	Hynynen et al.	N/A	N/A
11103731	12/2020	Vortman et al.	N/A	N/A
11112473	12/2020	Assif	N/A	N/A
11119552	12/2020	Spencer et al.	N/A	N/A
11120896	12/2020	Balignasay et al.	N/A	N/A
11123019	12/2020	Merritt et al.	N/A	N/A
11123575	12/2020	Vortman et al.	N/A	N/A
11135454	12/2020	Xu et al.	N/A	N/A
11141063	12/2020	Kemp et al.	N/A	N/A
11141131	12/2020	Stigall et al.	N/A	N/A
11160513	12/2020	Anderson et al.	N/A	N/A
11205507	12/2020	Anderson et al.	N/A	N/A
11219748	12/2021	Burkett et al.	N/A	N/A
11224349	12/2021	Davies et al.	N/A	N/A
11224403	12/2021	Corl	N/A	N/A
11224407	12/2021	Wrolstad et al.	N/A	N/A
11234649	12/2021	Matsubara et al.	N/A	N/A
11246533	12/2021	Henderson et al.	N/A	N/A
11246565	12/2021	Corl	N/A	N/A
11253225	12/2021	Hancock et al.	N/A	N/A
11260160	12/2021	Matsubara et al.	N/A	N/A
11272845	12/2021	Cheline et al.	N/A	N/A
11272904	12/2021	Vortman et al.	N/A	N/A
11291866	12/2021	Levy et al.	N/A	N/A
11298030	12/2021	Davies et al.	N/A	N/A
11309071	12/2021	Anderson	N/A	N/A
11311271	12/2021	Stigall et al.	N/A	N/A
11324410	12/2021	Burkett	N/A	N/A
11350906	12/2021	Castella et al.	N/A	N/A
11350954	12/2021	De Cicco et al.	N/A	N/A
11364042	12/2021	Maxwell et al.	N/A	N/A
11369346	12/2021	De Cicco et al.	N/A	N/A
11369994	12/2021	Greenberg et al.	N/A	N/A
11395638	12/2021	Shin et al.	N/A	N/A
11406334	12/2021	Merritt	N/A	N/A
11406355	12/2021	Hoffman et al.	N/A	N/A
11406498	12/2021	Stigall et al.	N/A	N/A
11408987	12/2021	Vignon et al.	N/A	N/A
11413017	12/2021	Stigall et al.	N/A	N/A
11419580	12/2021	Stigall et al.	N/A	N/A
11426140	12/2021	Sudol et al.	N/A	N/A



11432795	12/2021	Merritt	N/A	N/A
11432900	12/2021	Rakic et al.	N/A	N/A
11446000	12/2021	Minas et al.	N/A	N/A
11452496	12/2021	Minas et al.	N/A	N/A
11452506	12/2021	Perez et al.	N/A	N/A
11471215	12/2021	Stigall et al.	N/A	N/A
11484294	12/2021	Hancock et al.	N/A	N/A
11510632	12/2021	Begin et al.	N/A	N/A
11517291	12/2021	Kantor et al.	N/A	N/A
11520874	12/2021	Kennedy et al.	N/A	N/A
11524183	12/2021	Garcia Gutierrez et al.	N/A	N/A
11527001	12/2021	Brokman et al.	N/A	N/A
11547389	12/2022	Shin et al.	N/A	N/A
11553889	12/2022	Spencer et al.	N/A	N/A
11554386	12/2022	Pernot et al.	N/A	N/A
11559207	12/2022	Stigall et al.	N/A	N/A
11567153	12/2022	Stormont et al.	N/A	N/A
11576649	12/2022	Corl	N/A	N/A
11576652	12/2022	De Cicco et al.	N/A	N/A
11583193	12/2022	Groenland et al.	N/A	N/A
11589835	12/2022	Stigall et al.	N/A	N/A
11596351	12/2022	Nair	N/A	N/A
11596384	12/2022	Nair et al.	N/A	N/A
11596385	12/2022	Stigall et al.	N/A	N/A
11596387	12/2022	Song	N/A	N/A
11596389	12/2022	Nair	N/A	N/A
11596469	12/2022	Nair	N/A	N/A
11622746	12/2022	Minas et al.	N/A	N/A
11638576	12/2022	Groenland et al.	N/A	N/A
11647989	12/2022	Hope Simpson et al.	N/A	N/A
11648424	12/2022	Cannata	601/2	A61B 34/30
11653895	12/2022	Stigall et al.	N/A	N/A
11660070	12/2022	Stigall et al.	N/A	N/A
11666245	12/2022	Rajguru et al.	N/A	N/A
11666307	12/2022	Unser	N/A	N/A
11672433	12/2022	Park et al.	N/A	N/A
11672552	12/2022	Pasquino et al.	N/A	N/A
11672953	12/2022	May	N/A	N/A
11684342	12/2022	Groenland et al.	N/A	N/A
11684807	12/2022	Vortman et al.	N/A	N/A
11701134	12/2022	Maxwell et al.	N/A	N/A
11707207	12/2022	Stigall et al.	N/A	N/A
11707254	12/2022	Di Tullio et al.	N/A	N/A
11733881	12/2022	Perez	N/A	N/A
11737728	12/2022	Davies et al.	N/A	N/A
11744527	12/2022	Scott et al.	N/A	N/A
11744547	12/2022	Hynynen	N/A	N/A
11759169	12/2022	Corl	N/A	N/A

11759174	12/2022	Sarooha et al.	N/A	N/A
11766237	12/2022	Anderson	N/A	N/A
11771370	12/2022	Hynynen	N/A	N/A
11771405	12/2022	Rhodes	N/A	N/A
11771869	12/2022	Cicco	N/A	N/A
11779307	12/2022	Norris et al.	N/A	N/A
11806167	12/2022	Burkett	N/A	N/A
11813484	12/2022	Cannata et al.	N/A	N/A
11813485	12/2022	Xu et al.	N/A	N/A
11819712	12/2022	Cain et al.	N/A	N/A
11854687	12/2022	Keller	N/A	N/A
11857362	12/2023	Wrolstad et al.	N/A	N/A
11857807	12/2023	Levy et al.	N/A	N/A
11864918	12/2023	Burkett et al.	N/A	N/A
11872412	12/2023	Vortman et al.	N/A	N/A
11879973	12/2023	Prus et al.	N/A	N/A
11883235	12/2023	Stigall et al.	N/A	N/A
11890025	12/2023	Stigall et al.	N/A	N/A
11890136	12/2023	Stigall et al.	N/A	N/A
11890137	12/2023	Jenkins et al.	N/A	N/A
11950954	12/2023	Hyun et al.	N/A	N/A
11963822	12/2023	Wrolstad	N/A	N/A
11986682	12/2023	Prus et al.	N/A	N/A
11992366	12/2023	Stigall et al.	N/A	N/A
12017013	12/2023	Sasamine et al.	N/A	N/A
12035919	12/2023	Unser	N/A	N/A
12036066	12/2023	De Cicco et al.	N/A	N/A
12053194	12/2023	Goertz et al.	N/A	N/A
12220259	12/2024	Burkett et al.	N/A	N/A
12232907	12/2024	Chao et al.	N/A	N/A
12246195	12/2024	Levy et al.	N/A	N/A
2001/0039420	12/2000	Burbank et al.	N/A	N/A
2001/0041163	12/2000	Sugita	N/A	N/A
2002/0045890	12/2001	Celliers et al.	N/A	N/A
2002/0078964	12/2001	Kovac et al.	N/A	N/A
2002/0099356	12/2001	Unger et al.	N/A	N/A
2002/0145091	12/2001	Talish et al.	N/A	N/A
2003/0092982	12/2002	Eppstein	N/A	N/A
2003/0112922	12/2002	Burdette et al.	N/A	N/A
2003/0149352	12/2002	Liang et al.	N/A	N/A
2003/0157025	12/2002	Unger et al.	N/A	N/A
2003/0169591	12/2002	Cochran	N/A	N/A
2003/0181833	12/2002	Faragalla et al.	N/A	N/A
2003/0199857	12/2002	Eizenhofer	N/A	N/A
2003/0221561	12/2002	Milo	N/A	N/A
2003/0236539	12/2002	Rabiner et al.	N/A	N/A
2004/0127815	12/2003	Marchitto et al.	N/A	N/A
2004/0138563	12/2003	Moehring et al.	N/A	N/A
2004/0162571	12/2003	Rabiner et al.	N/A	N/A
2004/0164213	12/2003	Stephan	N/A	N/A

2004/0236248	12/2003	Svedman	N/A	N/A
2004/0243021	12/2003	Murphy et al.	N/A	N/A
2004/0249509	12/2003	Rogers et al.	N/A	N/A
2004/0260214	12/2003	Echt et al.	N/A	N/A
2005/0011296	12/2004	Koseki	N/A	N/A
2005/0020945	12/2004	Tosaya et al.	N/A	N/A
2005/0038339	12/2004	Chauhan et al.	N/A	N/A
2005/0038361	12/2004	Zhong et al.	N/A	N/A
2005/0152561	12/2004	Spencer	N/A	N/A
2005/0154314	12/2004	Quistgaard	N/A	N/A
2005/0154431	12/2004	Quistgaard et al.	N/A	N/A
2005/0203399	12/2004	Vaezy et al.	N/A	N/A
2005/0215901	12/2004	Anderson et al.	N/A	N/A
2005/0234438	12/2004	Mast et al.	N/A	N/A
2005/0283098	12/2004	Conston et al.	N/A	N/A
2006/0058678	12/2005	Vitek et al.	N/A	N/A
2006/0060991	12/2005	Holsteyns et al.	N/A	N/A
2006/0074303	12/2005	Chornenky et al.	N/A	N/A
2006/0089636	12/2005	Christopherson et al.	N/A	N/A
2006/0173321	12/2005	Kubota et al.	N/A	N/A
2006/0173387	12/2005	Hansmann et al.	N/A	N/A
2006/0184166	12/2005	Valle et al.	N/A	N/A
2006/0206028	12/2005	Lee et al.	N/A	N/A
2006/0229659	12/2005	Gifford et al.	N/A	N/A
2006/0241466	12/2005	Ottoboni et al.	N/A	N/A
2006/0241523	12/2005	Sinelnikov et al.	N/A	N/A
2006/0241533	12/2005	Geller	N/A	N/A
2006/0264760	12/2005	Liu et al.	N/A	N/A
2006/0293598	12/2005	Fraser	N/A	N/A
2006/0293630	12/2005	Manna et al.	N/A	N/A
2007/0010805	12/2006	Fedewa et al.	N/A	N/A
2007/0016039	12/2006	Vortman et al.	N/A	N/A
2007/0044562	12/2006	Sarr	N/A	N/A
2007/0065420	12/2006	Johnson	N/A	N/A
2007/0083120	12/2006	Cain et al.	N/A	N/A
2007/0161902	12/2006	Dan	N/A	N/A
2007/0167764	12/2006	Hynynen	N/A	N/A
2007/0205785	12/2006	Nilsson	N/A	N/A
2007/0219448	12/2006	Seip et al.	N/A	N/A
2007/0239001	12/2006	Mehl et al.	N/A	N/A
2008/0013593	12/2007	Kawabata	N/A	N/A
2008/0033297	12/2007	Sliwa	N/A	N/A
2008/0033417	12/2007	Nields et al.	N/A	N/A
2008/0051656	12/2007	Vaezy et al.	N/A	N/A
2008/0055003	12/2007	Unnikrishnan et al.	N/A	N/A
2008/0082026	12/2007	Schmidt et al.	N/A	N/A
2008/0091125	12/2007	Owen et al.	N/A	N/A
2008/0126665	12/2007	Burr et al.	N/A	N/A
2008/0154132	12/2007	Hall et al.	N/A	N/A

2008/0167555	12/2007	Qian et al.	N/A	N/A
2008/0177180	12/2007	Azhari et al.	N/A	N/A
2008/0194965	12/2007	Sliwa et al.	N/A	N/A
2008/0214964	12/2007	Chapelon et al.	N/A	N/A
2008/0262345	12/2007	Fichtinger et al.	N/A	N/A
2008/0262486	12/2007	Zvuloni et al.	N/A	N/A
2008/0300485	12/2007	Liu et al.	N/A	N/A
2008/0312561	12/2007	Chauhan	N/A	N/A
2008/0319356	12/2007	Cain	600/300	A61B 17/22004
2008/0319376	12/2007	Wilcox et al.	N/A	N/A
2009/0012514	12/2008	Moonen et al.	N/A	N/A
2009/0030339	12/2008	Cheng et al.	N/A	N/A
2009/0036773	12/2008	Lau et al.	N/A	N/A
2009/0112098	12/2008	Vaezy et al.	N/A	N/A
2009/0198094	12/2008	Fenster et al.	N/A	N/A
2009/0211587	12/2008	Lawrentschuk	N/A	N/A
2009/0227874	12/2008	Suri et al.	N/A	N/A
2009/0230822	12/2008	Kushculey et al.	N/A	N/A
2009/0254008	12/2008	Shields, Jr.	N/A	N/A
2009/0287083	12/2008	Kushculey et al.	N/A	N/A
2009/0306502	12/2008	Lacoste	N/A	N/A
2010/0011845	12/2009	Laugharn et al.	N/A	N/A
2010/0042020	12/2009	Ben-Ezra	N/A	N/A
2010/0056924	12/2009	Powers	N/A	N/A
2010/0059264	12/2009	Hasegawa et al.	N/A	N/A
2010/0069797	12/2009	Cain et al.	N/A	N/A
2010/0125225	12/2009	Gelbart et al.	N/A	N/A
2010/0152624	12/2009	Tanis et al.	N/A	N/A
2010/0163694	12/2009	Fadler et al.	N/A	N/A
2010/0204578	12/2009	Schmidt et al.	N/A	N/A
2010/0261994	12/2009	Davalos et al.	N/A	N/A
2010/0274136	12/2009	Cerofolini	N/A	N/A
2010/0286519	12/2009	Lee et al.	N/A	N/A
2010/0298744	12/2009	Altshuler et al.	N/A	N/A
2010/0305432	12/2009	Duhay et al.	N/A	N/A
2010/0317971	12/2009	Fan et al.	N/A	N/A
2010/0318002	12/2009	Prus et al.	N/A	N/A
2011/0072970	12/2010	Slobodzian et al.	N/A	N/A
2011/0112400	12/2010	Emery et al.	N/A	N/A
2011/0118600	12/2010	Gertner	N/A	N/A
2011/0118602	12/2010	Weng et al.	N/A	N/A
2011/0144490	12/2010	Davis et al.	N/A	N/A
2011/0144545	12/2010	Fan et al.	N/A	N/A
2011/0172529	12/2010	Gertner	N/A	N/A
2011/0178444	12/2010	Slayton et al.	N/A	N/A
2011/0245671	12/2010	Sato	N/A	N/A
2011/0251528	12/2010	Canney et al.	N/A	N/A
2011/0257524	12/2010	Gertner	N/A	N/A
2011/0263967	12/2010	Bailey et al.	N/A	N/A

2011/0270136	12/2010	Vitek et al.	N/A	N/A
2011/0319927	12/2010	Nita	N/A	N/A
2012/0029337	12/2011	Aoyagi	N/A	N/A
2012/0029353	12/2011	Slayton et al.	N/A	N/A
2012/0029393	12/2011	Lee	N/A	N/A
2012/0059264	12/2011	Hope Simpson et al.	N/A	N/A
2012/0059285	12/2011	Soltani et al.	N/A	N/A
2012/0092724	12/2011	Pettis	N/A	N/A
2012/0111339	12/2011	Barthe et al.	N/A	N/A
2012/0130288	12/2011	Holland et al.	N/A	N/A
2012/0136279	12/2011	Tanaka et al.	N/A	N/A
2012/0158013	12/2011	Stefanchik et al.	N/A	N/A
2012/0172720	12/2011	Asami et al.	N/A	N/A
2012/0189998	12/2011	Kruecker et al.	N/A	N/A
2012/0215157	12/2011	Berryman et al.	N/A	N/A
2012/0232388	12/2011	Curra et al.	N/A	N/A
2012/0259250	12/2011	Sapozhnikov et al.	N/A	N/A
2012/0271167	12/2011	Holland et al.	N/A	N/A
2012/0271223	12/2011	Khanna	N/A	N/A
2012/0281902	12/2011	Oikawa	382/131	G01S 7/52034
2013/0051178	12/2012	Rybyanets	N/A	N/A
2013/0053691	12/2012	Kawabata et al.	N/A	N/A
2013/0090579	12/2012	Cain et al.	N/A	N/A
2013/0102932	12/2012	Cain et al.	N/A	N/A
2013/0144165	12/2012	Ebbini et al.	N/A	N/A
2013/0172739	12/2012	Paladini	N/A	N/A
2013/0190623	12/2012	Bertolina et al.	N/A	N/A
2013/0190661	12/2012	Wing et al.	N/A	N/A
2013/0255426	12/2012	Kassow et al.	N/A	N/A
2013/0303906	12/2012	Cain et al.	N/A	N/A
2014/0030806	12/2013	Dudley et al.	N/A	N/A
2014/0039358	12/2013	Zhou et al.	N/A	N/A
2014/0046181	12/2013	Benchimol et al.	N/A	N/A
2014/0058293	12/2013	Hynynen et al.	N/A	N/A
2014/0058294	12/2013	Gross et al.	N/A	N/A
2014/0073995	12/2013	Teofilovic et al.	N/A	N/A
2014/0074076	12/2013	Gertner	N/A	N/A
2014/0088613	12/2013	Seo et al.	N/A	N/A
2014/0100459	12/2013	Xu et al.	N/A	N/A
2014/0112107	12/2013	Guo et al.	N/A	N/A
2014/0128734	12/2013	Genstler et al.	N/A	N/A
2014/0200489	12/2013	Behar et al.	N/A	N/A
2014/0243664	12/2013	El-Sayed et al.	N/A	N/A
2014/0316269	12/2013	Zhang et al.	N/A	N/A
2014/0324034	12/2013	Assaf et al.	N/A	N/A
2014/0330124	12/2013	Carol	N/A	N/A
2014/0378832	12/2013	Sanghvi et al.	N/A	N/A
2014/0378964	12/2013	Pearson	N/A	N/A

2015/0011875	12/2014	Noordhoek et al.	N/A	N/A
2015/0063668	12/2014	You et al.	N/A	N/A
2015/0073261	12/2014	Kohler et al.	N/A	N/A
2015/0080926	12/2014	Emery	N/A	N/A
2015/0148659	12/2014	Vahala	N/A	N/A
2015/0151141	12/2014	Arnal et al.	N/A	N/A
2015/0190121	12/2014	Slayton et al.	N/A	N/A
2015/0190659	12/2014	Kolher	N/A	N/A
2015/0196239	12/2014	Meehan et al.	N/A	N/A
2015/0224347	12/2014	Barthe et al.	N/A	N/A
2015/0257779	12/2014	Sinelnikov et al.	N/A	N/A
2015/0258352	12/2014	Lin et al.	N/A	N/A
2015/0273246	12/2014	Darlington et al.	N/A	N/A
2015/0297177	12/2014	Boctor et al.	N/A	N/A
2016/0004933	12/2015	Hu et al.	N/A	N/A
2016/0114194	12/2015	Gertner	N/A	N/A
2016/0120572	12/2015	Lee	N/A	N/A
2016/0135782	12/2015	Chen et al.	N/A	N/A
2016/0151618	12/2015	Powers et al.	N/A	N/A
2016/0184614	12/2015	Matula et al.	N/A	N/A
2016/0206341	12/2015	Slayton	N/A	N/A
2016/0206867	12/2015	Hossack et al.	N/A	N/A
2016/0249859	12/2015	Babkin et al.	N/A	N/A
2016/0287909	12/2015	Maxwell et al.	N/A	N/A
2016/0303166	12/2015	Katz et al.	N/A	N/A
2016/0331583	12/2015	Geringer	N/A	N/A
2016/0331585	12/2015	Kim	N/A	N/A
2016/0339273	12/2015	Al Mayiah	N/A	N/A
2016/0345938	12/2015	Tanter	N/A	A61B 8/485
2016/0354087	12/2015	Noonan et al.	N/A	N/A
2016/0361574	12/2015	Barthe et al.	N/A	N/A
2017/0000376	12/2016	Partanen et al.	N/A	N/A
2017/0049463	12/2016	Popovic et al.	N/A	N/A
2017/0071515	12/2016	Chevillet et al.	N/A	N/A
2017/0072227	12/2016	Khokhlova et al.	N/A	N/A
2017/0072228	12/2016	Wang et al.	N/A	N/A
2017/0100145	12/2016	Khoklova et al.	N/A	N/A
2017/0120080	12/2016	Phillips et al.	N/A	N/A
2017/0165046	12/2016	Johnson et al.	N/A	N/A
2017/0183062	12/2016	Lee	N/A	N/A
2017/0232277	12/2016	Hall et al.	N/A	N/A
2017/0281983	12/2016	Marquet et al.	N/A	N/A
2018/0000444	12/2017	Dayton et al.	N/A	N/A
2018/0028841	12/2017	Konofagou et al.	N/A	N/A
2018/0064412	12/2017	Messas et al.	N/A	N/A
2018/0161086	12/2017	Davalos et al.	N/A	N/A
2018/0169444	12/2017	Averkioiu et al.	N/A	N/A
2018/0206816	12/2017	Prus et al.	N/A	N/A
2018/0236271	12/2017	Tanter et al.	N/A	N/A
2018/0317884	12/2017	Chapelon et al.	N/A	N/A

2019/0000422	12/2018	West et al.	N/A	N/A
2019/0023804	12/2018	Onik et al.	N/A	N/A
2019/0275353	12/2018	Cannata et al.	N/A	N/A
2019/0282294	12/2018	Davalos et al.	N/A	N/A
2019/0314045	12/2018	Cunitz et al.	N/A	N/A
2019/0320904	12/2018	Mei	N/A	N/A
2019/0323086	12/2018	Leuthardt et al.	N/A	N/A
2019/0328500	12/2018	Cichon et al.	N/A	N/A
2020/0010575	12/2019	Hode et al.	N/A	N/A
2020/0055085	12/2019	Taffier	N/A	N/A
2020/0078608	12/2019	Maxwell et al.	N/A	N/A
2020/0107843	12/2019	Goertz et al.	N/A	N/A
2020/0182989	12/2019	Freeman et al.	N/A	N/A
2020/0194117	12/2019	Krieger et al.	N/A	N/A
2020/0253550	12/2019	Nair	N/A	N/A
2020/0254285	12/2019	Jang et al.	N/A	N/A
2020/0260964	12/2019	Collins et al.	N/A	N/A
2020/0282239	12/2019	Beder et al.	N/A	N/A
2020/0305842	12/2019	Rosenzweig et al.	N/A	N/A
2020/0323515	12/2019	Levy	N/A	N/A
2020/0330039	12/2019	Burkett et al.	N/A	N/A
2020/0330075	12/2019	O'Reilly et al.	N/A	N/A
2020/0353293	12/2019	Khokhlova et al.	N/A	N/A
2020/0367835	12/2019	Anderson	N/A	N/A
2020/0375576	12/2019	Moulton	N/A	N/A
2020/0405258	12/2019	Dayton et al.	N/A	N/A
2020/0405259	12/2019	Merritt	N/A	N/A
2021/0000541	12/2020	Levy et al.	N/A	N/A
2021/0022703	12/2020	Nair	N/A	N/A
2021/0022714	12/2020	Zagrodsky et al.	N/A	N/A
2021/0100527	12/2020	Martin	N/A	N/A
2021/0108866	12/2020	Lucht et al.	N/A	N/A
2021/0161398	12/2020	Millett et al.	N/A	N/A
2021/0169515	12/2020	Pahk et al.	N/A	N/A
2021/0170204	12/2020	Vortman et al.	N/A	N/A
2021/0170205	12/2020	Vortman et al.	N/A	N/A
2021/0187331	12/2020	Zadicario et al.	N/A	N/A
2021/0196295	12/2020	Goudot et al.	N/A	N/A
2021/0220607	12/2020	Sasamine et al.	N/A	N/A
2021/0330294	12/2020	Hynynen et al.	N/A	N/A
2021/0353161	12/2020	Merritt et al.	N/A	N/A
2021/0386451	12/2020	Escudero et al.	N/A	N/A
2021/0401400	12/2020	Sheehan et al.	N/A	N/A
2022/0008036	12/2021	Gulsen et al.	N/A	N/A
2022/0022845	12/2021	Corl	N/A	N/A
2022/0043143	12/2021	Prus et al.	N/A	N/A
2022/0079563	12/2021	Kemp	N/A	N/A
2022/0087640	12/2021	Minas et al.	N/A	N/A
2022/0166462	12/2021	Deurenberg et al.	N/A	N/A
2022/0167920	12/2021	Margolis	N/A	N/A

2022/0168470	12/2021	Ricotti et al.	N/A	N/A
2022/0196771	12/2021	Zur et al.	N/A	N/A
2022/0203139	12/2021	Shapira	N/A	N/A
2022/0219019	12/2021	Xu et al.	N/A	N/A
2022/0233890	12/2021	Hynynen et al.	N/A	N/A
2022/0240890	12/2021	Hancock et al.	N/A	N/A
2022/0257329	12/2021	Stigall et al.	N/A	N/A
2022/0280233	12/2021	Park et al.	N/A	N/A
2022/0280367	12/2021	Diodato et al.	N/A	N/A
2022/0296211	12/2021	Saroha et al.	N/A	N/A
2022/0338750	12/2021	Allen et al.	N/A	N/A
2022/0346756	12/2021	Chen	N/A	N/A
2022/0370025	12/2021	Regensburger et al.	N/A	N/A
2022/0386970	12/2021	Merritt	N/A	N/A
2022/0395333	12/2021	Merritt et al.	N/A	N/A
2022/0409171	12/2021	Sudol et al.	N/A	N/A
2022/0409858	12/2021	Lin	N/A	N/A
2023/0000466	12/2022	Levy et al.	N/A	N/A
2023/0000469	12/2022	Prus et al.	N/A	N/A
2023/0008714	12/2022	Rajguru et al.	N/A	N/A
2023/0012365	12/2022	Alpert et al.	N/A	N/A
2023/0024998	12/2022	Greenberg	N/A	N/A
2023/0031859	12/2022	Davies et al.	N/A	N/A
2023/0037603	12/2022	Pombo et al.	N/A	N/A
2023/0038498	12/2022	Xu et al.	N/A	N/A
2023/0038543	12/2022	Minas et al.	N/A	N/A
2023/0042834	12/2022	Henderson et al.	N/A	N/A
2023/0045488	12/2022	Rajguru et al.	N/A	N/A
2023/0050732	12/2022	Hancock et al.	N/A	N/A
2023/0061534	12/2022	Stopek	N/A	N/A
2023/0073447	12/2022	Minas et al.	N/A	N/A
2023/0100912	12/2022	Amar et al.	N/A	N/A
2023/0112722	12/2022	Hoffman et al.	N/A	N/A
2023/0114972	12/2022	Bigham et al.	N/A	N/A
2023/0121688	12/2022	Begin et al.	N/A	N/A
2023/0145064	12/2022	Vortman et al.	N/A	N/A
2023/0181140	12/2022	Cohen et al.	N/A	N/A
2023/0181156	12/2022	Cohen et al.	N/A	N/A
2023/0190119	12/2022	Groenland et al.	N/A	N/A
2023/0190215	12/2022	Nachtomy et al.	N/A	N/A
2023/0190224	12/2022	Rajguru et al.	N/A	N/A
2023/0190225	12/2022	Cohen et al.	N/A	N/A
2023/0190226	12/2022	Rajguru et al.	N/A	N/A
2023/0190227	12/2022	Cohen et al.	N/A	N/A
2023/0190228	12/2022	Cohen et al.	N/A	N/A
2023/0190229	12/2022	Cohen et al.	N/A	N/A
2023/0190230	12/2022	Yang et al.	N/A	N/A
2023/0191162	12/2022	Yau et al.	N/A	N/A
2023/0196569	12/2022	Cohen et al.	N/A	N/A
2023/0200899	12/2022	Nair	N/A	N/A



2023/0201553	12/2022	Levy et al.	N/A	N/A
2023/0218230	12/2022	Wu et al.	N/A	N/A
2023/0218262	12/2022	Boutelle et al.	N/A	N/A
2023/0218266	12/2022	Stigall et al.	N/A	N/A
2023/0218269	12/2022	Alpert et al.	N/A	N/A
2023/0218930	12/2022	Stopek et al.	N/A	N/A
2023/0240615	12/2022	May et al.	N/A	N/A
2023/0240647	12/2022	Minas et al.	N/A	N/A
2023/0240792	12/2022	Rakic et al.	N/A	N/A
2023/0255597	12/2022	O'Reilly et al.	N/A	N/A
2023/0260601	12/2022	Abel et al.	N/A	N/A
2023/0263507	12/2022	Groenland et al.	N/A	N/A
2023/0270388	12/2022	Richardson et al.	N/A	N/A
2023/0293148	12/2022	Stigall et al.	N/A	N/A
2023/0293149	12/2022	Stigall et al.	N/A	N/A
2023/0309859	12/2022	Sreedhar et al.	N/A	N/A
2023/0310899	12/2022	Hall et al.	N/A	N/A
2023/0310900	12/2022	Cannata et al.	N/A	N/A
2023/0310901	12/2022	Cannata et al.	N/A	N/A
2023/0320600	12/2022	Tochterman et al.	N/A	N/A
2023/0321327	12/2022	Maxwell	424/423	A61L 31/145
2023/0321398	12/2022	May	N/A	N/A
2023/0329559	12/2022	Xu et al.	N/A	N/A
2023/0333617	12/2022	Spencer et al.	N/A	N/A
2023/0334659	12/2022	Marama et al.	N/A	N/A
2023/0334677	12/2022	Sturm	N/A	N/A
2023/0338010	12/2022	Sturm	N/A	N/A
2023/0372025	12/2022	Van der Zaag et al.	N/A	N/A
2023/0381544	12/2022	Penot et al.	N/A	N/A
2023/0389891	12/2022	Cohen et al.	N/A	N/A
2023/0398381	12/2022	Vitek et al.	N/A	N/A
2024/0000422	12/2023	Corl	N/A	N/A
2024/0000426	12/2023	Davies et al.	N/A	N/A
2024/0001157	12/2023	Cannata et al.	N/A	N/A
2024/0001158	12/2023	Cannata et al.	N/A	N/A
2024/0023928	12/2023	Di Tullio et al.	N/A	N/A
2024/0023930	12/2023	Anderson	N/A	N/A
2024/0023941	12/2023	Rhodes	N/A	N/A
2024/0065632	12/2023	Burkett	N/A	N/A
2024/0138807	12/2023	Minas	N/A	N/A
2024/0139553	12/2023	Miller et al.	N/A	N/A
2024/0165666	12/2023	Hynynen et al.	N/A	N/A
2024/0188929	12/2023	Minas et al.	N/A	N/A
2024/0188931	12/2023	Ossmann et al.	N/A	N/A
2024/0189267	12/2023	Bogott et al.	N/A	N/A
2024/0189628	12/2023	Grumbir et al.	N/A	N/A
2024/0225592	12/2023	May et al.	N/A	N/A
2024/0245374	12/2023	Jenkins et al.	N/A	N/A
2024/0245390	12/2023	Winkler Brown et al.	N/A	N/A

2024/0245465	12/2023	Jenkins et al.	N/A	N/A
2024/0269491	12/2023	Xu et al.	N/A	N/A
2024/0285249	12/2023	May	N/A	N/A
2024/0408416	12/2023	Cannata et al.	N/A	N/A
2025/0018227	12/2024	Son et al.	N/A	N/A
2025/0040912	12/2024	Levy et al.	N/A	N/A
2025/0041577	12/2024	Shapira et al.	N/A	N/A
2025/0072872	12/2024	Nachtomy et al.	N/A	N/A
2025/0090130	12/2024	Schell et al.	N/A	N/A
2025/0090871	12/2024	Snell et al.	N/A	N/A

## FOREIGN PATENT DOCUMENTS

Patent No.	Application Date	Country	CPC
2017222925	12/2020	AU	N/A
112018017326	12/2021	BR	N/A
102023017634	12/2024	BR	N/A
3073552	12/2018	CA	N/A
3101381	12/2018	CA	N/A
3055856	12/2019	CA	N/A
3153080	12/2020	CA	N/A
2910561	12/2020	CA	N/A
2908740	12/2020	CA	N/A
2980976	12/2022	CA	N/A
2840014	12/2022	CA	N/A
1669672	12/2004	CN	N/A
1732031	12/2005	CN	N/A
201197744	12/2008	CN	N/A
102292123	12/2010	CN	N/A
102481164	12/2011	CN	N/A
102665585	12/2011	CN	N/A
103537016	12/2013	CN	N/A
103648361	12/2013	CN	N/A
103812477	12/2013	CN	N/A
104013444	12/2013	CN	N/A
104135938	12/2013	CN	N/A
104208822	12/2013	CN	N/A
106999076	12/2016	CN	N/A
109185113	12/2018	CN	N/A
109219415	12/2018	CN	N/A
109689160	12/2018	CN	N/A
208725992	12/2018	CN	N/A
111565642	12/2019	CN	N/A
111655337	12/2019	CN	N/A
111699022	12/2019	CN	N/A
111712300	12/2019	CN	N/A
111712301	12/2019	CN	N/A
106999053	12/2019	CN	N/A
107660137	12/2019	CN	N/A
111757769	12/2019	CN	N/A

112204412	12/2020	CN	N/A
112236195	12/2020	CN	N/A
106661535	12/2020	CN	N/A
112533673	12/2020	CN	N/A
112566694	12/2020	CN	N/A
106999054	12/2020	CN	N/A
106793997	12/2020	CN	N/A
107530049	12/2020	CN	N/A
112912011	12/2020	CN	N/A
112912012	12/2020	CN	N/A
112912013	12/2020	CN	N/A
112969413	12/2020	CN	N/A
112996445	12/2020	CN	N/A
113167877	12/2020	CN	N/A
113196080	12/2020	CN	N/A
109196369	12/2020	CN	N/A
109200484	12/2020	CN	N/A
113316419	12/2020	CN	N/A
113329788	12/2020	CN	N/A
109640830	12/2020	CN	N/A
113473917	12/2020	CN	N/A
113507946	12/2020	CN	N/A
113518588	12/2020	CN	N/A
113705586	12/2020	CN	N/A
110662575	12/2020	CN	N/A
113905666	12/2021	CN	N/A
114222536	12/2021	CN	N/A
114366154	12/2021	CN	N/A
114423362	12/2021	CN	N/A
110248606	12/2021	CN	N/A
115227992	12/2021	CN	N/A
109843181	12/2021	CN	N/A
115461000	12/2021	CN	N/A
115515504	12/2021	CN	N/A
109091768	12/2022	CN	N/A
115779285	12/2022	CN	N/A
115779287	12/2022	CN	N/A
115813438	12/2022	CN	N/A
111032157	12/2022	CN	N/A
115916035	12/2022	CN	N/A
110958858	12/2022	CN	N/A
116172611	12/2022	CN	N/A
111655337	12/2022	CN	N/A
109416908	12/2022	CN	N/A
116507295	12/2022	CN	N/A
107529989	12/2022	CN	N/A
111372522	12/2022	CN	N/A
116617589	12/2022	CN	N/A
112236195	12/2022	CN	N/A
113615098	12/2022	CN	N/A

114555247	12/2022	CN	N/A
116744856	12/2022	CN	N/A
116761554	12/2022	CN	N/A
109416907	12/2022	CN	N/A
117321444	12/2022	CN	N/A
117500437	12/2023	CN	N/A
117580499	12/2023	CN	N/A
111212606	12/2023	CN	N/A
113490459	12/2023	CN	N/A
118042992	12/2023	CN	N/A
118414127	12/2023	CN	N/A
119367006	12/2024	CN	N/A
112704620	12/2024	CN	N/A
114287963	12/2024	CN	N/A
3220751	12/1982	DE	N/A
3544628	12/1986	DE	N/A
3817094	12/1988	DE	N/A
4012760	12/1991	DE	N/A
0017382	12/1979	EP	N/A
0320303	12/1988	EP	N/A
0332871	12/1988	EP	N/A
0384831	12/1989	EP	N/A
0755653	12/1996	EP	N/A
1374785	12/2003	EP	N/A
1504713	12/2004	EP	N/A
1566201	12/2004	EP	N/A
2397188	12/2010	EP	N/A
2934308	12/2014	EP	N/A
2934309	12/2014	EP	N/A
3097180	12/2015	EP	N/A
2759003	12/2019	EP	N/A
3558457	12/2019	EP	N/A
3700629	12/2019	EP	N/A
3218829	12/2019	EP	N/A
3229688	12/2019	EP	N/A
3723857	12/2019	EP	N/A
2887989	12/2020	EP	N/A
3777689	12/2020	EP	N/A
2938253	12/2020	EP	N/A
3076864	12/2020	EP	N/A
2802276	12/2020	EP	N/A
2809221	12/2020	EP	N/A
3801761	12/2020	EP	N/A
3801762	12/2020	EP	N/A
3801763	12/2020	EP	N/A
2967369	12/2020	EP	N/A
2967488	12/2020	EP	N/A
3184048	12/2020	EP	N/A
2967370	12/2020	EP	N/A
3482390	12/2020	EP	N/A

3870067	12/2020	EP	N/A
3870069	12/2020	EP	N/A
3876843	12/2020	EP	N/A
2931130	12/2020	EP	N/A
2934304	12/2020	EP	N/A
3887843	12/2020	EP	N/A
3888534	12/2020	EP	N/A
3895604	12/2020	EP	N/A
3897391	12/2020	EP	N/A
3229672	12/2020	EP	N/A
3902603	12/2020	EP	N/A
3903672	12/2020	EP	N/A
2964096	12/2020	EP	N/A
3930776	12/2021	EP	N/A
3545829	12/2021	EP	N/A
3959530	12/2021	EP	N/A
3060129	12/2021	EP	N/A
3986296	12/2021	EP	N/A
3988167	12/2021	EP	N/A
2914166	12/2021	EP	N/A
3229674	12/2021	EP	N/A
2779907	12/2021	EP	N/A
3102098	12/2021	EP	N/A
2965263	12/2021	EP	N/A
2726152	12/2021	EP	N/A
4041387	12/2021	EP	N/A
4042936	12/2021	EP	N/A
3298959	12/2021	EP	N/A
2931131	12/2021	EP	N/A
2938268	12/2021	EP	N/A
3581103	12/2021	EP	N/A
4087492	12/2021	EP	N/A
4093470	12/2021	EP	N/A
3091905	12/2021	EP	N/A
4098203	12/2021	EP	N/A
2950737	12/2022	EP	N/A
3057496	12/2022	EP	N/A
4114274	12/2022	EP	N/A
4117534	12/2022	EP	N/A
2869804	12/2022	EP	N/A
2938265	12/2022	EP	N/A
3024403	12/2022	EP	N/A
4138672	12/2022	EP	N/A
4151156	12/2022	EP	N/A
2938271	12/2022	EP	N/A
4161360	12/2022	EP	N/A
4179995	12/2022	EP	N/A
3171764	12/2022	EP	N/A
4201342	12/2022	EP	N/A
2931132	12/2022	EP	N/A

3229695	12/2022	EP	N/A
4209178	12/2022	EP	N/A
4209179	12/2022	EP	N/A
4226864	12/2022	EP	N/A
4230121	12/2022	EP	N/A
4230146	12/2022	EP	N/A
4233972	12/2022	EP	N/A
2866733	12/2022	EP	N/A
3870069	12/2022	EP	N/A
4247489	12/2022	EP	N/A
3484371	12/2022	EP	N/A
3658037	12/2022	EP	N/A
3685874	12/2022	EP	N/A
3870070	12/2022	EP	N/A
2938255	12/2022	EP	N/A
3229906	12/2022	EP	N/A
3764914	12/2022	EP	N/A
3903672	12/2022	EP	N/A
4272654	12/2022	EP	N/A
4275609	12/2022	EP	N/A
3316804	12/2022	EP	N/A
3519109	12/2022	EP	N/A
3166479	12/2023	EP	N/A
3537984	12/2023	EP	N/A
3908195	12/2023	EP	N/A
3182920	12/2023	EP	N/A
3174643	12/2023	EP	N/A
3814917	12/2023	EP	N/A
4349283	12/2023	EP	N/A
3681419	12/2023	EP	N/A
4368118	12/2023	EP	N/A
2804525	12/2023	EP	N/A
4380667	12/2023	EP	N/A
4385428	12/2023	EP	N/A
4289415	12/2024	EP	N/A
4406484	12/2024	EP	N/A
3190958	12/2024	EP	N/A
2774069	12/2019	ES	N/A
2819552	12/2020	ES	N/A
2829822	12/2020	ES	N/A
2998435	12/2024	ES	N/A
2099582	12/1981	GB	N/A
1245715	12/2020	HK	N/A
254768	12/2020	IL	N/A
261285	12/2021	IL	N/A
202117039853	12/2020	IN	N/A
387413	12/2021	IN	N/A
445766	12/2022	IN	N/A
60-80779	12/1984	JP	N/A
61-196718	12/1985	JP	N/A

02-215451	12/1989	JP	N/A
H0422351	12/1991	JP	N/A
06-197907	12/1993	JP	N/A
07-504339	12/1994	JP	N/A
08-84740	12/1995	JP	N/A
06-304178	12/1995	JP	N/A
08-131454	12/1995	JP	N/A
09-55571	12/1996	JP	N/A
10-512477	12/1997	JP	N/A
2000300559	12/1999	JP	N/A
2003510159	12/2002	JP	N/A
2004505660	12/2003	JP	N/A
2004249106	12/2003	JP	N/A
2005167058	12/2004	JP	N/A
2006511265	12/2005	JP	N/A
2007144225	12/2006	JP	N/A
2007520307	12/2006	JP	N/A
2010019554	12/2009	JP	N/A
2010029650	12/2009	JP	N/A
2010204068	12/2009	JP	N/A
2013538097	12/2012	JP	N/A
2004512502	12/2013	JP	N/A
2015519970	12/2014	JP	N/A
2016508808	12/2015	JP	N/A
2017/506542	12/2016	JP	N/A
2017506538	12/2016	JP	N/A
2019051295	12/2018	JP	N/A
2020525167	12/2019	JP	N/A
2020525168	12/2019	JP	N/A
2020525169	12/2019	JP	N/A
6785554	12/2019	JP	N/A
6789944	12/2019	JP	N/A
2020534077	12/2019	JP	N/A
2020195788	12/2019	JP	N/A
2020535895	12/2019	JP	N/A
6832958	12/2020	JP	N/A
6835719	12/2020	JP	N/A
6838057	12/2020	JP	N/A
6849592	12/2020	JP	N/A
2021510104	12/2020	JP	N/A
6896719	12/2020	JP	N/A
6934933	12/2020	JP	N/A
6951560	12/2020	JP	N/A
6979633	12/2020	JP	N/A
6980696	12/2020	JP	N/A
7012726	12/2021	JP	N/A
2022500092	12/2021	JP	N/A
2022500093	12/2021	JP	N/A
2022501080	12/2021	JP	N/A
2022504159	12/2021	JP	N/A

2022509389	12/2021	JP	N/A
2022509391	12/2021	JP	N/A
2022509392	12/2021	JP	N/A
2022509393	12/2021	JP	N/A
2022509395	12/2021	JP	N/A
2022509401	12/2021	JP	N/A
2022509453	12/2021	JP	N/A
2022510217	12/2021	JP	N/A
7019679	12/2021	JP	N/A
7026118	12/2021	JP	N/A
2022514272	12/2021	JP	N/A
2022515488	12/2021	JP	N/A
2022516078	12/2021	JP	N/A
7053500	12/2021	JP	N/A
2022526104	12/2021	JP	N/A
2022527043	12/2021	JP	N/A
2022095785	12/2021	JP	N/A
7171645	12/2021	JP	N/A
7171663	12/2021	JP	N/A
7175640	12/2021	JP	N/A
2022546288	12/2021	JP	N/A
7187715	12/2021	JP	N/A
2022551875	12/2021	JP	N/A
2022552229	12/2021	JP	N/A
7201819	12/2022	JP	N/A
7232204	12/2022	JP	N/A
7239466	12/2022	JP	N/A
7265525	12/2022	JP	N/A
2023071859	12/2022	JP	N/A
7292448	12/2022	JP	N/A
7299992	12/2022	JP	N/A
2023085350	12/2022	JP	N/A
7302936	12/2022	JP	N/A
7304344	12/2022	JP	N/A
7321162	12/2022	JP	N/A
7325430	12/2022	JP	N/A
7335367	12/2022	JP	N/A
2023116673	12/2022	JP	N/A
7340594	12/2022	JP	N/A
7346293	12/2022	JP	N/A
7351972	12/2022	JP	N/A
7352561	12/2022	JP	N/A
2023123676	12/2022	JP	N/A
2023134811	12/2022	JP	N/A
7358391	12/2022	JP	N/A
7359765	12/2022	JP	N/A
7370386	12/2022	JP	N/A
2023162327	12/2022	JP	N/A
2024010135	12/2023	JP	N/A
2024020483	12/2023	JP	N/A



7479288	12/2023	JP	N/A
7479351	12/2023	JP	N/A
7485383	12/2023	JP	N/A
7530561	12/2023	JP	N/A
7542708	12/2023	JP	N/A
7612816	12/2024	JP	N/A
2025013082	12/2024	JP	N/A
7641600	12/2024	JP	N/A
7643694	12/2024	JP	N/A
2025031814	12/2024	JP	N/A
102574559	12/2022	KR	N/A
10276498281	12/2024	KR	N/A
20250019597	12/2024	KR	N/A
WO94/06355	12/1993	WO	N/A
WO02/32506	12/2001	WO	N/A
WO2005/018469	12/2004	WO	N/A
WO2008/051484	12/2007	WO	N/A
WO2011/040054	12/2010	WO	N/A
WO2011/092683	12/2010	WO	N/A
WO2011/154654	12/2010	WO	N/A
WO2014/008594	12/2013	WO	N/A
WO2014/071386	12/2013	WO	N/A
WO2015/000953	12/2014	WO	N/A
WO2015/031532	12/2014	WO	N/A
WO2015/153909	12/2014	WO	N/A
WO2018/149671	12/2017	WO	N/A
WO2019/081329	12/2018	WO	N/A
WO2019/122941	12/2018	WO	N/A
WO2020/074615	12/2019	WO	N/A
WO2020/087049	12/2019	WO	N/A
WO2020/217098	12/2019	WO	N/A
WO2020/237382	12/2019	WO	N/A
WO2020/245660	12/2019	WO	N/A
WO2021/014221	12/2020	WO	N/A
WO2021/032887	12/2020	WO	N/A
WO2021/069216	12/2020	WO	N/A
WO2021/069971	12/2020	WO	N/A
WO2021/089810	12/2020	WO	N/A
WO2021/105358	12/2020	WO	N/A
WO2021/115958	12/2020	WO	N/A
WO2021/116763	12/2020	WO	N/A
WO2021/122253	12/2020	WO	N/A
WO2021/123905	12/2020	WO	N/A
WO2021/123906	12/2020	WO	N/A
WO2021/140042	12/2020	WO	N/A
WO2021/142090	12/2020	WO	N/A
WO2021/170510	12/2020	WO	N/A
WO2021/175626	12/2020	WO	N/A
WO2021/176275	12/2020	WO	N/A
WO2021/178961	12/2020	WO	N/A

WO2021/180501	12/2020	WO	N/A
WO2021/180550	12/2020	WO	N/A
WO2021/213927	12/2020	WO	N/A
WO2021/249936	12/2020	WO	N/A
WO2021/258007	12/2020	WO	N/A
WO2022/013266	12/2021	WO	N/A
WO2022/040493	12/2021	WO	N/A
WO2022/047193	12/2021	WO	N/A
WO2022/056394	12/2021	WO	N/A
WO2022/069254	12/2021	WO	N/A
WO2022/069303	12/2021	WO	N/A
WO2022/069327	12/2021	WO	N/A
WO2022/078744	12/2021	WO	N/A
WO2022/097138	12/2021	WO	N/A
WO2022/106891	12/2021	WO	N/A
WO2022/152724	12/2021	WO	N/A
WO2022/152827	12/2021	WO	N/A
WO2022/152828	12/2021	WO	N/A
WO2022/228922	12/2021	WO	N/A
WO2022/238058	12/2021	WO	N/A
WO2022/238092	12/2021	WO	N/A
WO2022/238229	12/2021	WO	N/A
WO2022/238274	12/2021	WO	N/A
WO2022/238276	12/2021	WO	N/A
WO2022/238392	12/2021	WO	N/A
WO2022/247242	12/2021	WO	N/A
WO2022/258561	12/2021	WO	N/A
WO2022/260746	12/2021	WO	N/A
WO2022/260747	12/2021	WO	N/A
WO2023/274899	12/2022	WO	N/A
WO2023/275617	12/2022	WO	N/A
WO2023/275771	12/2022	WO	N/A
WO2023/012516	12/2022	WO	N/A
WO2023/036742	12/2022	WO	N/A
WO2023/052278	12/2022	WO	N/A
WO2023/084307	12/2022	WO	N/A
WO2023/104599	12/2022	WO	N/A
WO2023/104841	12/2022	WO	N/A
WO2023/110555	12/2022	WO	N/A
WO2023/110556	12/2022	WO	N/A
WO2023/110594	12/2022	WO	N/A
WO2023/110607	12/2022	WO	N/A
WO2023/117721	12/2022	WO	N/A
WO2023/117821	12/2022	WO	N/A
WO2023/117822	12/2022	WO	N/A
WO2023/118080	12/2022	WO	N/A
WO2023105288	12/2022	WO	N/A
WO2023105290	12/2022	WO	N/A
WO2023/131566	12/2022	WO	N/A
WO2023/131574	12/2022	WO	N/A

WO2023/135024	12/2022	WO	N/A
WO2023/141653	12/2022	WO	N/A
WO2023/152639	12/2022	WO	N/A
WO2023/169967	12/2022	WO	N/A
WO2023/180811	12/2022	WO	N/A
WO2023/218428	12/2022	WO	N/A
WO2023/230053	12/2022	WO	N/A
WO2023/230054	12/2022	WO	N/A
WO2024/009143	12/2023	WO	N/A
WO2024/016088	12/2023	WO	N/A
WO2024/040185	12/2023	WO	N/A
WO2024/047580	12/2023	WO	N/A
WO2024/092272	12/2023	WO	N/A
WO2024/120659	12/2023	WO	N/A
WO2024/125872	12/2023	WO	N/A
WO2024/130252	12/2023	WO	N/A
WO2024/144897	12/2023	WO	N/A
WO2025/038127	12/2024	WO	N/A
WO2025/059671	12/2024	WO	N/A

## OTHER PUBLICATIONS

Feasibility of Using Ultrasound Phased Arrays for MRI Monitored Noninvasive Surgery (Year: 1996). cited by examiner

Concentric-Ring and Sector-Vortex Phased-Array Applicators for Ultrasound Hyperthermia (Year: 1986). cited by examiner

Akiyama et al.; Elliptically curved acoustic lens for emitting strongly focused finite-amplitude beams: Application of the spheroidal beam equation model to the theoretical prediction; Acoustical Science and Technology, vol. 26, pp. 279-284, May 2005. cited by applicant

Appel et al.; Stereoscopic highspped recording of bubble filaments; Ultrasonics Sonochemistry; vol. 11(1); pp. 39-42; Jan. 2004. cited by applicant

Arani et al.; Transurethral prostate magnetic resonance elastography; prospective imaging requirements; Magn. Reson. Med.; 65(2); pp. 340-349; Feb. 2011. cited by applicant

Aschoff et al.; How does alteration of hepatic blood flow affect liver perfusion and radiofrequency-induced thermal lesion size in rabbit liver?; J Magn Reson Imaging; 13(1); pp. 57-63; Jan. 2001. cited by applicant

Atchley et al.; Thresholds for cavitation produced in water by pulsed ultrasound; Ultrasonics.; vol. 26(5); pp. 280-285; Sep. 1988. cited by applicant

Avago Technologies; ACNV2601 High Insulation Voltage 10 MBd Digital Opotcoupler. Avago Technologies Data Sheet; pp. 1-11; Jul. 29, 2010. cited by applicant

Avago Technologies; Avago's ACNV2601 optocoupler is an optically coupled logic gate; Data Sheet; 2 pages; Jul. 29, 2010. cited by applicant

Avtech; AVR-8 Data sheet; May 23, 2004; 3 pages; retrieved from the internet (<http://www.avtechpulse.com>). cited by applicant

BAK; Rapid prototyping or rapid production? 3D printing processes move industry towards the latter; Assembly Automation; 23(4); pp. 340-345; Dec. 1, 2003. cited by applicant

Billson et al.; Rapid prototyping technologies for ultrasonic beam focussing in NDE; IEEE International Ultrasonic Symposium Proceedings; pp. 2472-2474; Oct. 2011. cited by applicant

Bjoerk et al.; Cool/MOS CP—How to make most beneficial use of the generation of super junction technology devices. Infineon Technologies AG. [retrieved Feb. 4, 2014] from the internet (<http://www.infineon.com/dgdl/Infineon+-+Application+Note+-+PowerMOSFETs+->

+600V+CoolMOS%E284%A2+-

+CP+Most+beneficial+use+of+superjunction+technologie+devices.pdf?

folderid=db3a304412b407950112b408e8090004&fileid=db3a304412b407950112b40ac9a40688>pp.

1, 4, 14; Feb. 2007. cited by applicant

Bland et al.; Surgical Oncology; McGraw Hill; Chap. 5 (Cavitron Ultrasonic Aspirator); pp. 461-462; Jan. 29, 2001. cited by applicant

Burdin et al.; Implementation of the laser diffraction technique for cavitation bubble investigations; Particle & Particle Systems Characterization; vol. 19; pp. 73-83; May 2002. cited by applicant

Cain, Charles A.; Histotripsy: controlled mechanical sub-division of soft tissues by high intensity pulsed ultrasound (conference presentation); American Institute of Physics (AIP) Therapeutic Ultrasound: 5th International Symposium on Therapeutic Ultrasound; 44 pgs.; Oct. 27-29, 2005. cited by applicant

Canney et al.; Shock-Induced Heating and Millisecond Boiling in Gels and Tissue Due to High Intensity Focused Ultrasound; Ultrasound in Medicine & Biology, vol. 36, pp. 250-267; Feb. 2010 (author manuscript). cited by applicant

Chan et al.; An image-guided high intensity focused ultrasound device for uterine fibroids treatment; Medical Physics, vol. 29, pp. 2611-2620, Nov. 2002. cited by applicant

Clasen et al.; MR-guided radiofrequency ablation of hepatocellular carcinoma: Long-term effectiveness; J Vase Interv Radiol; 22(6); pp. 762-770; Jun. 2011. cited by applicant

Clement et al.; A hemisphere array for non-invasive ultrasound brain therapy and surgery; Physics in Medicine and Biology, vol. 45, p. 3707-3719, Dec. 2000. cited by applicant

Cline et al.; Magnetic resonance-guided thermal surgery; Magnetic Resonance in Medicine; 30(1); pp. 98-106; Jul. 1993. cited by applicant

Curiel et al.; Elastography for the follow-up of high-intensity focused ultrasound prostate cancer treatment: Initial comparison with MRI; Ultrasound Med. Biol; 31(11); pp. 1461-1468; Nov. 2005. cited by applicant

Desilets et al.; The Design of Efficient Broad-Band Piezoelectric Transducers; Sonics and Ultrasonics, IEEE Transactions on, vol. 25, pp. 115-125, May 1978. cited by applicant

Dovedi et al.; Acquired Resistance to Fractionated Radiotherapy Can Be Overcome by Concurrent PD-L1 Blockade; Cancer Research; 74(19); pp. 5458-5468; Oct. 1, 2014. cited by applicant

Emelianov et al.; Triplex ultrasound: Elasticity imaging to age deep venous thrombosis; Ultrasound Med Biol; 28(6); pp. 757-767; Jun. 2002. cited by applicant

Giannatsis et al.; Additive fabrication technologies applied to medicine and health care: a review; The International Journal of Advanced Manufacturing Technology; 40(1-2); pp. 116-127; Jan. 2009. cited by applicant

Gudra et al.; Influence of acoustic impedance of multilayer acoustic systems on the transfer function of ultrasonic airborne transducers; Ultrasonics, vol. 40, pp. 457-463, May 2002. cited by applicant

Hall et al.; A Low Cost Compact 512 Channel Therapeutic Ultrasound System For Transcutaneous Ultrasound Surgery; AIP Conference Proceedings, Boston, MA; vol. 829, pp. 445-449, Oct. 27-29, 2005. cited by applicant

Hall et al.; Acoustic Access to the Prostate for Extracorporeal Ultrasound Ablation; Journal of Endourology, vol. 24, pp. 1875-1881, Nov. 2010. cited by applicant

Hall et al.; Histotripsy of the prostate: dose effects in a chronic canine model; Urology; 74(4); pp. 932-937; Oct. 2009 (author manuscript). cited by applicant

Hall et al.; Imaging feedback of tissue liquefaction (histotripsy) in ultrasound surgery; IEEE Ultrasonic Symposium, Sep. 18-21, 2005, pp. 1732-1734. cited by applicant

Haller et al.; Determination of acoustic cavitation probabilities and thresholds using a single focusing transducer to induce and detect acoustic cavitation events: I. Method and terminology; Ultrasound in Medicine & Biology; 44(2); pp. 377-396; Feb. 1, 2018. cited by applicant

Hartmann; Ultrasonic properties of poly(4-methyl pentene-1), Journal of Applied Physics, vol. 51,

pp. 310-314, Jan. 1980. cited by applicant

Hobarth et al.; Color flow doppler sonography for extracorporeal shock wave lithotripsy; Journal of Urology; 150(6); pp. 1768-1770; Dec. 1, 1993. cited by applicant

Holland et al.; Thresholds for transient cavitation produced by pulsed ultrasound in a controlled nuclei environment; J. Acoust. Soc. Am.; vol. 88(5); pp. 2059-2069; Nov. 1990. cited by applicant

Huber et al.; Influence of shock wave pressure amplitude and pulse repetition frequency on the lifespan, size and number of transient cavities in the field of an electromagnetic lithotripter; Physics in Medicine and Biology; vol. 43(10); pp. 3113-3128; Oct. 1998. cited by applicant

Hynynen et al.; Tissue thermometry during ultrasound exposure; European Urology; 23(Suppl 1); pp. 12-16; (the year of publication is sufficiently earlier than the effective U.S. filing date and any foreign priority date so that the particular month of publication is not in issue)1993. cited by applicant

Kallel et al.; The feasibility of elastographic visualization of HIFU-induced thermal lesions in soft tissues: Image-guided high-intensity focused ultrasound; Ultrasound Med. Biol; 25(4); pp. 641-647; May 1999. cited by applicant

Khokhlova et al.; Controlled tissue emulsification produced by high intensity focused ultrasound shock waves and millisecond boiling; J. Acoust. Soc. Am.; 130(5), pt. 2; pp. 3498-3510; Nov. 2011. cited by applicant

Kim et al.; Dependence of particle volume fraction on sound velocity and attenuation of EPDM composites; Ultrasonics, vol. 46, pp. 177-183, Feb. 2007. cited by applicant

Kim et al.; Development of a wearable robotic positioning system for noninvasive transcranial focused ultrasound stimulation. IEEE/ASME Transactions on Mechatronics; 21(5); pp. 2284-2293; Jun. 13, 2016. cited by applicant

Konofagou; Quo vadis elasticity imaging?; Ultrasonics; 42(1-9); pp. 331-336; Apr. 2004. cited by applicant

Krimholtz et al.; New equivalent circuits for elementary piezoelectric transducers; Electronics Letters, vol. 6, pp. 398-399, Jun. 1970. cited by applicant

Kruse et al.; Tissue characterization using magnetic resonance elastography: Preliminary results; Phys. Med. Biol; 45(6); pp. 1579-1590; Jun. 2000. cited by applicant

Lake et al.; Histotripsy: minimally invasive technology for prostatic tissue ablation in an in vivo canine model; Urology; 72(3); pp. 682-686; Sep. 2008. cited by applicant

Lauterborn et al.; Cavitation bubble dynamics studied by high speed photography and holography: part one; Ultrasonics; vol. 23; pp. 260-268; Nov. 1985. cited by applicant

Lensing et al.; Deep-vein thrombosis; The Lancet, vol. 353, pp. 479-485, Feb. 6, 1999. cited by applicant

Lin et al.; Dual-beam histotripsy: a low-frequency pump enabling a high-frequency probe for precise lesion formation; IEEE Trans. Ultrason. Ferroelectr. Control; 61(2); pp. 325-340; Feb. 2014; (Author Manuscript; 29 pages). cited by applicant

Liu et al.; Real-time 2-D temperature imaging using ultrasound; IEEE Trans Biomed Eng; 57(1); pp. 12-16; Jan. 2010 (author manuscript, 16 pgs.). cited by applicant

Liu et al.; Viscoelastic property measurement in thin tissue constructs using ultrasound; IEEE Trans Ultrason Ferroelectr Freq Control; 55(2); pp. 368-383; Feb. 2008 (author manuscript, 37 pgs.). cited by applicant

Macoskey; Acoustic methods for histotripsy feedback; (Dissertation); Biomedical Engineering and Science Computing; University of Michigan 2019; 207 pages; retrieved from the internet (<https://deepblue.lib.umich.edu/handle/2027.42/149988>) on Feb. 2022. cited by applicant

Manes et al.; Design of a Simplified Delay System for Ultrasound Phased Array Imaging; Sonics and Ultrasonics, IEEE Transactions on, vol. 30, pp. 350-354, Nov. 1983. cited by applicant

Maréchal et al; Effect of Radial Displacement of Lens on Response of Focused Ultrasonic Transducer; Japanese Journal of Applied Physics, vol. 46, p. 3077-3085; May 15, 2007. cited by

applicant

Maréchal et al.; Lens-focused transducer modeling using an extended KLM model; *Ultrasonics*, vol. 46, pp. 155-167, May 2007. cited by applicant

Martin et al.; Water-cooled, high-intensity ultrasound surgical applicators with frequency tracking; *Ultrasonics, Ferroelectrics and Frequency Control, IEEE Transactions on*, vol. 50, pp. 1305-1317, Oct. 2003. cited by applicant

Maxwell et al.; Cavitation clouds created by shock scattering from bubbles during histotripsy; *J. Acoust. Soc. Am.*; 130(4); pp. 1888-1898; Oct. 2011. cited by applicant

Maxwell et al.; Noninvasive Thrombolysis Using Pulsed Ultrasound Cavitation Therapy—Histotripsy; *Ultrasound in Medicine & Biology*, vol. 35, pp. 1982-1994, Dec. 2009 (author manuscript). cited by applicant

Maxwell; Noninvasive thrombolysis using histotripsy pulsed ultrasound cavitation therapy; PhD Dissertation. University of Michigan, Ann Arbor, Michigan. Jun. 2012. cited by applicant

Maxwell et al.; In-vivo study of non-invasive thrombolysis by histotripsy in a porcine model; *IEEE international Ultrasonics Symposium; IEEE*; p. 220-223; Sep. 20, 2009. cited by applicant

Maxwell et al.; The role of compressional pressure in the formation of dense bubble clouds in histotripsy, 2009 IEEE International Ultrasonics Symposium; pp. 81-84, Sep. 20, 2009. cited by applicant

Miller et al.; A review of in vitro bioeffects of inertial ultrasonic cavitation from a mechanistic perspective; *Ultrasound in Medicine and Biology*; vol. 22; pp. 1131-1154; (the year of publication is sufficiently earlier than the effective U.S. filing date and any foreign priority date so that the particular month of publication is not in issue) 1996. cited by applicant

Miller et al.; Investigation of the mechanism of ARFI-based color doppler feedback of histotripsy tissue fractionation; *Ultrasonic Symposium (IUS); 2013 IEEE International*; 4 pages; Jul. 21-25, 2013. cited by applicant

Miller et al.; Real-time elastography-based monitoring of histotripsy tissue fractionation using color doppler; *Ultrasonics Symposium (IUS); 2012 IEEE International*; 8 pages; Oct. 7-10, 2012. cited by applicant

Nightingale et al.; Analysis of contrast in images generated with transient acoustic radiation force; *Ultrasound Med Biol*; 32(1); pp. 61-72; Jan. 2006. cited by applicant

Ohl et al.; Bubble dynamics, shock waves and sonoluminescence; *Phil. Trans. R. Soc. Lond. A*; vol. 357; pp. 269-294; (the year of publication is sufficiently earlier than the effective U.S. filing date and any foreign priority date so that the particular month of publication is not in issue) 1999. cited by applicant

Okada et al.; A case of hepatocellular carcinoma treated by MR-guided focused ultrasound ablation with respiratory gating; *Magn Reson Med Sci*; 5(3); pp. 167-171; Oct. 2006. cited by applicant

Palmeri et al.; Acoustic radiation force-based elasticity imaging methods; *Interface Focus*; 1; pp. 553-564; Aug. 2011. cited by applicant

Parsons et al.; Cost-effective assembly of a basic fiber-optic hydrophone for measurement of high-amplitude therapeutic ultrasound fields; *The Journal of the Acoustical Society of America*, vol. 119, pp. 1432-1440, Mar. 2006. cited by applicant

Parsons et al.; Pulsed cavitation ultrasound therapy for controlled tissue homogenization; *Ultrasound in Med. & Biol.*; vol. 32(1); pp. 115-129; Jan. 2006. cited by applicant

Pishchalnikov et al.; Cavitation Bubble Cluster Activity in the Breakage of Kidney Stones by Lithotripter Shock Waves; *J Endourol.*; 17(7): 435-446; Sep. 2003. cited by applicant

Porter et al.; Reduction in left ventricular cavitory attenuation and improvement in posterior myocardial contrast . . . ; *J Am Soc Echocardiography*; pp. 437-441; Jul.-Aug. 1996. cited by applicant

Qu et al.; Non-thermal histotripsy tumor ablation promotes abscopal immune responses that enhance cancer immunotherapy; *Journal for immunotherapy of cancer*; 8(1); Jan. 15, 2020. cited by applicant

Roberts et al.; Pulsed cavitational ultrasound: a noninvasive technology for controlled tissue ablation (histotripsy) in the rabbit kidney; *Journal of Urology*; vol. 175(2); pp. 734-738; Feb. 2006. cited by applicant

Rosenschein et al.; Ultrasound Imaging-Guided Noninvasive Ultrasound Thrombolysis: Preclinical Results; *Circulation*; vol. 102; pp. 238-245, Jul. 11, 2000. cited by applicant

Rowland et al.; MRI study of hepatic tumours following high intensity focused ultrasound surgery; *British Journal of Radiology*; 70; pp. 144-153; Feb. 1997. cited by applicant

Roy et al.; A precise technique for the measurement of acoustic cavitation thresholds and some preliminary results; *Journal of the Acoustical Society of America*; vol. 78(5); pp. 1799-1805; Nov. 1985. cited by applicant

Sapareto et al.; Thermal dose determination in cancer therapy; *Int J Radiat Oncol Biol Phys*; 10(6); pp. 787-800; Apr. 1984. cited by applicant

Sapozhnikov et al.; Ultrasound-Guided Localized Detection of Cavitation During Lithotripsy in Pig Kidney in Vivo; *IEEE Ultrasonics Symposium*, vol. 2; pp. 1347-1350; Oct. 7-10, 2001. cited by applicant

Sato et al.; Experimental Investigation of Phased Array Using Tapered Matching Layers. 2002 *IEEE Ultrasound Symposium*. vol. 2; pp. 1235-1238, Oct. 2002. cited by applicant

Sferruzza et al.; Generation of high power unipolar pulse with a piezocomposite transducer; In 1999 *IEEE Ultrasonics Symposium Proceedings; International Symposium (Cat. No. 99CH37027)*; vol. 2; pp. 1125-1128; Oct. 17, 1999. cited by applicant

Shung; *Diagnostic Ultrasound: Imaging and Blood Flow Measurements*; Taylor and Francis Group, LLC; Boca Raton, FL; 207 pages; (the year of publication is sufficiently earlier than the effective U.S. filing date and any foreign priority date so that the particular month of publication is not in issue) 2006. cited by applicant

Simonin et al.; Characterization of heterogeneous structure in a polymer object manufactured by stereolithography with low-frequency microechography; *Journal of Materials Chemistry*; vol. 6, pp. 1595-1599, Sep. 1996. cited by applicant

Sokolov et al.; Use of a dual-pulse lithotripter to generate a localized and intensified cavitation field; *Journal of the Acoustical Society of America*; vol. 110(3); pp. 1685-1695; Sep. 2001. cited by applicant

Song et al.; Feasibility of Using Lateral Mode Coupling Method for a Large Scale Ultrasound Phased Array for Noninvasive Transcranial Therapy; *Biomedical Engineering; IEEE Transactions on*, vol. 57, pp. 124-133; Jan. 2010 (author manuscript). cited by applicant

Souchon et al.; Visualisation of HIFU lesions using elastography of the human prostate in vivo: Preliminary results; *Ultrasound Med. Biol.*; 29(7); pp. 1007-1015; Jul. 2003. cited by applicant

Souquet et al.; Design of Low-Loss Wide-Band Ultrasonic Transducers for Noninvasive Medical Application; *Sonics and Ultrasonics, IEEE Transactions on*, vol. 26, pp. 75-80, Mar. 1979. cited by applicant

Therapeutic Ultrasound Group. Non-invasive Ultrasonic Tissue Fraction for Treatment of Benign Disease and Cancer—"Histotripsy" University research [online]. Biomedical Engineering Department, University of Michigan. Jul. 2011[retrieved on Jan. 28, 2014] from: (<http://web.archive.org/web/20110720091822/http://www.histotripsy.umich.edu/index.html>> entire document) Jul. 2011. cited by applicant

TODA; Narrowband impedance matching layer for high efficiency thickness mode ultrasonic transducers; *Ultrasonics, Ferroelectrics and Frequency Control, IEEE Transactions on*, vol. 49, pp. 299-306, Mar. 2002. cited by applicant

Urban et al.; Measurement of prostate viscoelasticity using shearwave dispersion ultrasound vibrometry (SDUV): an in vitro study; *IEEE International Ultrasonics Symposium Proceedings (IUS)*; pp. 1141-1144; Oct. 11, 2010. cited by applicant

Van Kervel et al.; A calculation scheme for the optimum design of ultrasonic transducers;

Ultrasonics, vol. 21, pp. 134-140, May 1983. cited by applicant

Wang et al.; Quantitative ultrasound backscatter for pulsed cavitation ultrasound therapy-histotripsy; Ultrasonics, Ferroelectrics and Frequency Control, IEEE Transactions on, vol. 56, pp. 995-1005, May 2009. cited by applicant

Wikipedia; Medical ultrasound; 15 pages; retrieved from the internet ([https://en.wikipedia.org/w/index.php?title=Medical\\_ultrasound&oldid=515340960](https://en.wikipedia.org/w/index.php?title=Medical_ultrasound&oldid=515340960)) on Jan. 12, 2018. cited by applicant

Wu et al.; Mechanism and dynamics of hydrodynamic-acoustic cavitation (HAC); Ultrasonics sonochemistry; vol. 49., pp. 89-96; Dec. 1, 2018. cited by applicant

Xie et al.; Correspondence of ultrasound elasticity imaging to direct mechanical measurement in aging DVT in rats; Ultrasound Med Biol; 31(10); pp. 1351-1359; Oct. 2005 (author manuscript, 20 pgs.). cited by applicant

Xu et al.; A new strategy to enhance cavitation tissue erosion by using a high intensity initiating sequence; IEEE Trans Ultrasonics Ferroelectrics and Freq Control; vol. 53(8); pp. 1412-1424; Aug. 2006. cited by applicant

Xu et al.; Controlled ultrasound tissue erosion: the role of dynamic interaction between insonation and microbubble activity; Journal of the Acoustical Society of America; vol. 117(1); pp. 424-435; Jan. 2005. cited by applicant

Xu et al.; Controlled ultrasound tissue erosion; IEEE Transaction on Ultrasonics, Ferroelectrics, and Frequency Control; vol. 51 (6); pp. 726-736; Jun. 2004. cited by applicant

Xu et al.; Effects of acoustic parameters on bubble cloud dynamics in ultrasound tissue erosion (histotripsy); Journal of the Acoustical Society of America; vol. 122(1); pp. 229-236; Jul. 2007. cited by applicant

Xu et al.; High Speed Imaging of Bubble Clouds Generated in Pulsed Ultrasound Cavitation Therapy Histotripsy; IEEE Trans Ultrason Ferroelectr Freq Control; ; vol. 54; No. 10; pp. 2091R2101; Oct. 2007. cited by applicant

Xu et al.; Investigation of intensity threshold for ultrasound tissue erosion; Ultrasound in Med. & Biol.; vol. 31(12); pp. 1673-1682; Dec. 2005. cited by applicant

Xu et al.; Optical and acoustic monitoring of bubble cloud dynamics at a tissue-fluid interface in ultrasound tissue erosion; Journal of the Acoustical Society of America; vol. 121(4); pp. 2421-2430; Apr. 2007. cited by applicant

Yan et al.; A review of rapid prototyping technologies and systems; Computer-Aided Design, vol. 28, pp. 307-318, Apr. 1996. cited by applicant

Zhang et al.; A fast tissue stiffness-dependent elastography for HIFU-induced lesions inspection; Ultrasonics; 51(8); pp. 857-869; Dec. 2011. cited by applicant

Zheng et al.; An acoustic backscatter-based method for localization of lesions induced by high-intensity focused ultrasound; Ultrasound Med Biol; 36(4); pp. 610-622; Apr. 2010. cited by applicant

Gateau et al.; Transcranial ultrasonic therapy based on time reversal of acoustically induced cavitation bubble signature. IEEE Transactions on Biomedical Engineering; 57(1); pp. 134-144; Sep. 18, 2009. cited by applicant

Cannata et al.; U.S. Appl. No. 18/311,050 entitled "Histotripsy systems and methods," filed May 2, 2023. cited by applicant

Maxwell et al.; U.S. Appl. No. 18/329,459 entitled "Histotripsy for thrombolysis," filed Jun. 5, 2023. cited by applicant

Xu et al.; U.S. Appl. No. 18/478,342 entitled "Systems and methods for histotripsy immunosensitization," filed Sep. 29, 2023. cited by applicant

Duryea et al.; U.S. Appl. No. 18/497,856 entitled "Histotripsy systems and methods," filed Oct. 31, 2023. cited by applicant

Duryea et al.; U.S. Appl. No. 18/498,966 entitled "Histotripsy systems and methods," filed Oct. 31, 2023. cited by applicant



Cain et al.; U.S. Appl. No. 18/485,904 entitled "Histotripsy using very short ultrasound pulses," filed Oct. 12, 2023. cited by applicant

Xu et al.; U.S. Appl. No. 18/555,683 entitled "Design and fabrication of therapeutic ultrasound transducer with arbitrarily shaped, densely packing, removable modular elements," filed Oct. 16, 2023. cited by applicant

International Society for Magnetic Resonance in Medicine (ISMRM); No. 105; XP040714022; I Jul. 24, 2020. cited by applicant

Hoogenboom et al.; Mechanical high-intensity focused ultrasound destruction of soft tissue: working mechanisms and physiologic effects; Ultrasound in medicine & biology; 41(6); pp. 1500-1517; Jun. 1, 2015. cited by applicant

Sukovich et al.; Real-time transcranial histotripsy treatment localization and mapping using acoustic cavitation emission feedback; IEEE transactions on ultrasonics, ferroelectrics, and frequency control; 67(6); pp. 1178-1791; Jan. 17, 2020. cited by applicant

Bader et al.; For whom the bubble grows: physical principles of bubble nucleation and dynamics in histotripsy ultrasound therapy; Ultrasound in medicine & biology; 45(5); pp. 1056-1080, May 1, 2019. cited by applicant

Xu et al.; U.S. Appl. No. 18/568,038 entitled "Minimally invasive histotripsy systems and methods," filed Dec. 7, 2023. cited by applicant

Cannata et al.; U.S. Appl. No. 18/594,843 entitled "Histotripsy systems and methods," filed Mar. 4, 2024. cited by applicant

Cannata et al.; U.S. Appl. No. 18/630,758 entitled "Histotripsy systems and methods," filed Apr. 9, 2024. cited by applicant

Cannata et al.; U.S. Appl. No. 18/642,529 entitled "Histotripsy systems and associated methods including user interfaces and workflows for treatment planning and therapy," filed Apr. 22, 2024. cited by applicant

Maxwell et al.; U.S. Appl. No. 18/737,731 entitled "Histotripsy for thrombolysis," filed Jun. 7, 2024. cited by applicant

Cannata et al.; U.S. Appl. No. 18/737,746 entitled "Histotripsy excitation sequences optimized for bubble cloud formation using shock scattering," filed Jun. 7, 2024. cited by applicant

Stopek.; U.S. Appl. No. 18/761,937 entitled "Minimally invasive histotripsy systems and methods," filed Jul. 2, 2024. cited by applicant

Shaffer et al.; U.S. Appl. No. 18/832,708 entitled "Histotripsy systems and methods," filed Jul. 24, 2024. cited by applicant

Miller et al.; U.S. Appl. No. 18/924,812 entitled "Histotripsy systems and methods," filed Oct. 23, 2024. cited by applicant

Stocker et al.; Endocavity histotripsy for efficient tissue ablation Transducer design and characterization. IEEE transactions on ultrasonics, ferroelectrics, and frequency control; 68(9); pp. 2896-2905; Jan. 28, 2021. cited by applicant

Woodacre et al.; A low-cost miniature histotripsy transducer for precision tissue ablation. IEEE transactions on ultrasonics, ferroelectrics, and frequency control; 65(11); pp. 2131-2140; Nov. 1, 2018. cited by applicant

---

*Primary Examiner:* Nguyen; Hien N

*Attorney, Agent or Firm:* Shay Glenn LLP

---

## **Background/Summary**

CROSS REFERENCE TO RELATED APPLICATIONS (1) This patent application is a continuation of U.S. application Ser. No. 18/497,856, filed Oct. 30, 2023, which claims priority to U.S. provisional patent application No. 63/381,401, titled “HISTOTRIPSY SYSTEMS AND METHODS” and filed on Oct. 28, 2022, which is herein incorporated by reference in its entirety.

## INCORPORATION BY REFERENCE

(1) All publications and patent applications mentioned in this specification are herein incorporated by reference to the same extent as if each individual publication or patent application was specifically and individually indicated to be incorporated by reference.

## FIELD

(2) The present disclosure details novel high intensity therapeutic ultrasound (HITU) systems configured to produce acoustic cavitation, methods, devices and procedures for the minimally and non-invasive treatment of healthy, diseased and/or injured tissue. The acoustic cavitation systems and methods described herein, also referred to Histotripsy, may include transducers, drive electronics, positioning robotics, imaging systems, and integrated treatment planning and control software to provide comprehensive treatment and therapy for soft tissues in a patient.

## BACKGROUND

(3) Histotripsy, or pulsed ultrasound cavitation therapy, is a technology where extremely short, intense bursts of acoustic energy induce controlled cavitation (microbubble formation) within the focal volume. The vigorous expansion and collapse of these microbubbles mechanically homogenizes cells and tissue structures within the focal volume. This is a very different end result than the coagulative necrosis characteristic of thermal ablation. To operate within a non-thermal, Histotripsy realm; it is necessary to deliver acoustic energy in the form of high amplitude acoustic pulses with low duty cycle.

(4) Compared with conventional focused ultrasound technologies, Histotripsy has important advantages: 1) the destructive process at the focus is mechanical, not thermal; 2) cavitation appears bright on ultrasound imaging thereby confirming correct targeting and localization of treatment; 3) treated tissue generally, but not always, appears darker (more hypoechoic) on ultrasound imaging, so that the operator knows what has been treated; and 4) Histotripsy produces lesions in a controlled and precise manner. It is important to emphasize that unlike thermal ablative technologies such as microwave, radiofrequency, high-intensity focused ultrasound (HIFU) cryo or radiation, Histotripsy relies on the mechanical action of cavitation for tissue destruction and not on heat, cold or ionizing energy.

---

## Description

### BRIEF DESCRIPTION OF THE DRAWINGS

(1) The novel features of the invention are set forth with particularity in the claims that follow. A better understanding of the features and advantages of the present invention will be obtained by reference to the following detailed description that sets forth illustrative embodiments, in which the principles of the invention are utilized, and the accompanying drawings of which:

(2) FIGS. 1A-1B illustrate an ultrasound imaging and therapy system.

(3) FIG. 2 is one embodiment of a histotripsy therapy and imaging system with a coupling system.

(4) FIG. 3 is one example of an ultrasound pulse for generating histotripsy via a shock scattering mechanism.

(5) FIGS. 4A-4B illustrate packing techniques for treating a target tissue volume.

(6) FIG. 5A illustrates one embodiment of creating a bubble cloud at a single focal location for a given discrete treatment location.

(7) FIG. 5B illustrates one example of creating a bubble cloud with a plurality of axially spaced

focal locations for a given discrete treatment location.

(8) FIG. 5C illustrates one example of creating a bubble cloud with a plurality of laterally spaced focal locations for a given discrete treatment location.

(9) FIGS. 6A-6B illustrate one technique for electronically beam-steering an ultrasound transducer laterally to form an enhanced excitation volume.

(10) FIGS. 7A-7B illustrate one technique for electronically beam-steering an ultrasound transducer in 3D space to form an enhanced excitation volume.

(11) FIGS. 8A-8B illustrate one technique for electronically beam-steering an ultrasound transducer in 3D space to form an enhanced excitation volume.

(12) FIGS. 9A-9D illustrate beam profiles for a plurality of axially spaced bubble cloud focal locations.

(13) FIGS. 10A-10C illustrate beam profiles for a plurality of laterally spaced bubble cloud focal locations.

(14) FIGS. 11A-11B illustrate the effect of spacing between adjacent focal locations.

(15) FIGS. 12A-12I illustrate heat maps, exposure levels, and lesion levels for a plurality of treatment protocols.

(16) FIG. 13 is a schematic drawing showing the spacing between focal locations relative to a therapy transducer.

(17) FIG. 14 is a chart showing the effect of spacing on bubble cloud size.

#### SUMMARY OF THE DISCLOSURE

(18) A method of delivering histotripsy therapy to a target tissue is provided, comprising: receiving a digital treatment plan that includes a target tissue volume of a subject divided into a plurality of discrete treatment locations; mechanically positioning a focus of an ultrasound therapy transducer at a first focal location in a first discrete treatment location within the target tissue volume; delivering a first histotripsy pulse to the first focal location to produce a first cavitation bubble cloud at the first focal location; electronically beam-steering the focus of the ultrasound therapy transducer to a second focal location in the first discrete treatment location that overlaps with the first focal location; and delivering a second histotripsy pulse to the second focal location to produce a second bubble cloud at the second focal location and to at least partially re-excite the first cavitation bubble cloud at the first focal location.

(19) In some aspects, the histotripsy pulse comprises a positive half-cycle followed by a peak negative half-cycle followed by a trailing positive half-cycle.

(20) In some aspects, the trailing positive half-cycle has an amplitude lower than that of the leading positive half-cycle

(21) In one aspect, delivering the second histotripsy pulse to the second focal location further comprises delivering the second histotripsy pulse to the second focal location to produce the second bubble cloud at the second focal location during a life-cycle of the first bubble cloud.

(22) In other aspects, delivering the second histotripsy pulse to the second focal location further comprises delivering the second histotripsy pulse to the second focal location to produce the second bubble cloud at the second focal location during a life-cycle of residual cavitation nuclei of the first bubble cloud.

(23) In one aspect, electronically beam-steering the focus further comprises electronically beam-steering the focus in an axial direction with respect to the ultrasound therapy transducer.

(24) In other aspects, electronically beam-steering the focus further comprises electronically beam-steering the focus in any lateral direction with respect to the ultrasound therapy transducer.

(25) In some aspects, electronically beam-steering the focus further comprises electronically beam-steering the focus in a lateral and an axial direction with respect to the ultrasound therapy transducer.

(26) In another aspect, the method includes repeating the electronic beam-steering and delivering steps for a third focal location that overlaps with the second focal location.

- (27) In some aspects, the method includes repeating the electronic beam-steering and delivering steps for a third focal location that overlaps with the first focal location.
- (28) In one aspect, after each delivering step, the method includes delivering at least one low amplitude waveform into the target tissue to spatially manipulate residual cavitation nuclei in the target tissue.
- (29) In other aspects, the method includes repeating the delivering and electronically beam-steering steps for the first and second focal locations until a desired dose is applied to the first discrete treatment location.
- (30) In some aspects, the method includes mechanically positioning the focus of the ultrasound therapy transducer at a first focal location in a second discrete treatment location within the target tissue volume after the desired dose has been applied to the first discrete treatment location.
- (31) In one aspect, the method includes mechanically positioning the focus further comprises mechanically positioning the focus with a robotic positioning system.
- (32) An ultrasound therapy system is provided, comprising; optionally a robotic positioning system; an ultrasound therapy transducer array connected to the robotic positioning system; a generator operatively coupled to the ultrasound therapy transducer array, the generator and ultrasound therapy transducer array being configured to deliver histotripsy pulses into a subject to generate a cavitation bubble cloud in the subject; at least one processor operatively coupled to the robotic positioning system and the generator, the at least one processor being configured to control the robotic positioning system and the generator to provide histotripsy therapy to the subject according to a treatment plan that includes a plurality of discrete treatment locations within a target tissue volume of the subject by: controlling the robotic positioning system to mechanically position a focus of the ultrasound therapy transducer array at a first focal location in a first discrete treatment location within the treatment plan; controlling the generator to deliver at least one histotripsy pulse with the ultrasound therapy transducer array to form a first cavitation bubble cloud at the first focal location; controlling the generator to electronically beam-steer the focus of the ultrasound therapy transducer to a second focal location in the first discrete treatment location that overlaps with the first focal location; and controlling the generator to deliver at least one histotripsy pulse to the second focal location to produce a second bubble cloud at the second focal location and to re-excite the first cavitation bubble cloud at the first focal location.
- (33) In some aspects, each of the histotripsy pulses comprises a positive half-cycle followed by a peak negative half-cycle followed by a trailing positive half-cycle,
- (34) In one aspect, the trailing positive half-cycle has an amplitude lower than that of the leading positive half-cycle.
- (35) In one aspect, controlling the generator to deliver the at least one histotripsy pulse to the second focal location further comprises controlling the generator to deliver the at least one histotripsy pulse to the second focal location to produce the second bubble cloud at the second focal location during a life-cycle of the first bubble cloud.
- (36) In other aspects, controlling the generator to deliver the at least one histotripsy pulse to the second focal location further comprises controlling the generator to deliver the at least one histotripsy pulse to the second focal location to produce the second bubble cloud at the second focal location during a life-cycle of residual cavitation nuclei of the first bubble cloud.
- (37) In some aspects, controlling the generator to electronically beam-steer the focus further comprises controlling the generator to electronically beam-steer the focus in an axial direction with respect to the ultrasound therapy transducer array.
- (38) In some aspects, controlling the generator to electronically beam-steer the focus further comprises controlling the generator to electronically beam-steer the focus in any lateral direction with respect to the ultrasound therapy transducer array.
- (39) In some aspects, controlling the generator to electronically beam-steer the focus further comprises controlling the generator to electronically beam-steer the focus in a lateral and an axial

direction with respect to the ultrasound therapy transducer array.

(40) In one aspect, the system is configured to repeat controlling the generator to electronically beam-steer and controlling the generator to deliver at least one histotripsy pulse for a third focal location that overlaps with the second focal location.

(41) In other aspects, the system is configured to repeat controlling the generator to electronically beam-steer and controlling the generator to deliver at least one histotripsy pulse for a third focal location that overlaps with the first focal location.

(42) In one aspect, after controlling the generator to deliver at least one histotripsy pulse the system is configured to control the generator to deliver at least one low amplitude waveform into the target tissue to spatially manipulate residual cavitation nuclei in the target tissue.

(43) In one aspect, the system is configured for repeating controlling the generator to electronically beam-steer and controlling the generator to deliver at least one histotripsy pulse for the first and second focal locations until a desired dose is applied to the first discrete treatment location.

(44) In other aspects, the system is configured for controlling the robotic positioning system to mechanically position the focus of the ultrasound therapy transducer at a first focal location in a second discrete treatment location within the target tissue volume after the desired dose has been applied to the first discrete treatment location.

(45) A method of delivering histotripsy therapy to a target tissue volume is provided, comprising the steps of: dividing the target tissue volume into a plurality of discrete treatment locations; determining a treatment plan that includes a pathway for providing therapy to each of the discrete treatment locations; positioning a focus of an ultrasound therapy transducer at a first focal location in a first discrete treatment location within the target tissue volume; delivering one or more histotripsy pulses to the first focal location to produce a first bubble cloud at the first focal location; electronically beam-steering the focus of the ultrasound therapy transducer to a second focal location in the first discrete treatment location that overlaps with the first focal location; delivering one or more histotripsy pulses to the second focal location to produce a second bubble cloud at the second focal location; and repeating the positioning and delivering steps for each of the plurality of discrete treatment locations according to the treatment plan and pathway.

(46) In one aspect, each histotripsy pulse comprises a positive half-cycle followed by a peak negative half-cycle followed by a trailing positive half-cycle.

(47) In other aspects, the trailing positive half-cycle has an amplitude lower than that of the leading positive half-cycle.

(48) In some aspects, the method includes delivering the one or more histotripsy pulses to the second focal location further comprises delivering the one or more histotripsy pulses to the second focal location to produce the second bubble cloud at the second focal location during a life-cycle of the first bubble cloud.

(49) In other aspects, the method includes delivering the one or more histotripsy pulses to the second focal location further comprises delivering the one or more histotripsy pulses to the second focal location to produce the second bubble cloud at the second focal location during a life-cycle of residual cavitation nuclei of the first bubble cloud.

(50) In some aspects, electronically beam-steering the focus further comprises electronically beam-steering the focus in an axial direction with respect to the ultrasound therapy transducer.

(51) In other aspects, electronically beam-steering the focus further comprises electronically beam-steering the focus in any lateral direction with respect to the ultrasound therapy transducer.

(52) In one aspect, electronically beam-steering the focus further comprises electronically beam-steering the focus in a lateral and an axial direction with respect to the ultrasound therapy transducer.

(53) In another aspect, the method includes repeating the electronic beam-steering and delivering steps for a third focal location that overlaps with the second focal location.

(54) In some aspects, the method includes repeating the electronic beam-steering and delivering

steps for a third focal location that overlaps with the first focal location.

(55) In another aspect, after each delivering step the method includes delivering at least one low amplitude waveform into the target tissue to spatially manipulate residual cavitation nuclei in the target tissue.

(56) In some aspects, the method includes repeating the delivering and electronically beam-steering steps for the first and second focal locations until a desired dose is applied to the first discrete treatment location.

(57) In another aspect, the method includes mechanically positioning the focus of the ultrasound therapy transducer at a first focal location in a second discrete treatment location within the target tissue volume after the desired dose has been applied to the first discrete treatment location.

(58) In some aspects, mechanically positioning the focus further comprises mechanically positioning the focus with a robotic positioning system.

(59) A method of delivering histotripsy therapy to a target tissue volume is provided, comprising: delivering first histotripsy therapy pulses to a first focal location to create a corresponding bubble cloud within a first focal zone; electronically beam-steering the focus a second focal location; and delivering second histotripsy therapy pulses to the second focal location to create a corresponding bubble cloud within a second focal zone that spatially overlaps with the first focal zone; wherein the second histotripsy pulses are timed to re-excite residual cavitation nuclei in the first focal zone to create an enhanced excitation volume comprising the first focal zone and second focal zone.

(60) In some aspects, the histotripsy pulse comprises a positive half-cycle followed by a peak negative half-cycle followed by a trailing positive half-cycle.

(61) In other aspects, the trailing positive half-cycle has an amplitude lower than that of the leading positive half-cycle.

(62) In some aspects, delivering the second histotripsy pulse to the second focal location further comprises delivering the second histotripsy pulse to the second focal location to produce the second bubble cloud at the second focal location during a life-cycle of the first bubble cloud.

(63) In another aspect, delivering the second histotripsy pulse to the second focal location further comprises delivering the second histotripsy pulse to the second focal location to produce the second bubble cloud at the second focal location during a life-cycle of residual cavitation nuclei of the first bubble cloud.

(64) In some aspects, electronically beam-steering the focus further comprises electronically beam-steering the focus in an axial direction with respect to the ultrasound therapy transducer.

(65) In one aspect, electronically beam-steering the focus further comprises electronically beam-steering the focus in any lateral direction with respect to the ultrasound therapy transducer.

(66) In another aspect, electronically beam-steering the focus further comprises electronically beam-steering the focus in a lateral and an axial direction with respect to the ultrasound therapy transducer.

(67) In some aspects, the method includes repeating the electronic beam-steering and delivering steps for a third focal location that overlaps with the second focal location.

(68) In other aspects, the method includes repeating the electronic beam-steering and delivering steps for a third focal location that overlaps with the first focal location.

(69) In one aspect, after each delivering step, the method includes delivering at least one low amplitude waveform into the target tissue to spatially manipulate residual cavitation nuclei in the target tissue.

(70) In some aspects, the method includes repeating the delivering and electronically beam-steering steps for the first and second focal locations until a desired dose is applied to the first discrete treatment location.

(71) In other aspects, the method includes mechanically positioning the focus of the ultrasound therapy transducer at a first focal location in a second discrete treatment location within the target tissue volume after the desired dose has been applied to the first discrete treatment location.

(72) In some aspects, mechanically positioning the focus further comprises mechanically positioning the focus with a robotic positioning system.

(73) A method of producing an enhanced excitation volume with histotripsy energy is provided, comprising positioning a focus of an ultrasound therapy transducer at a first focal location in a first discrete treatment location within a target tissue volume, the first focal location being positioned along a central axis of the ultrasound therapy transducer, delivering a first histotripsy pulse to the first focal location to produce a first cavitation bubble cloud at the first focal location, electronically beam-steering the focus of the ultrasound therapy transducer to a second focal location in the first discrete treatment location that is at least partially outside of the central axis, delivering a second histotripsy pulse to the second focal location to produce a second bubble cloud at the second focal location, electronically beam-steering the focus of the ultrasound therapy transducer to a focal location at least partially inside the central axis; and delivering a third histotripsy pulse to the focal location to produce a third bubble cloud at the focal location.

(74) In some aspects, a natural focus of the ultrasound therapy transducer array is located along the central axis.

(75) In other aspects, the first focal location comprises the natural focus.

(76) In some aspects, the second focal location partially overlaps with the central axis.

(77) In other aspects, the focal location comprises the first focal location.

(78) In additional aspects, the focal location comprises a third focal location.

(79) In some aspects, electronically beam-steering to the second focal location comprises steering laterally.

(80) In other aspects, electronically beam-steering to the second focal location comprises steering laterally and axially.

(81) An ultrasound therapy system is provided, comprising; a robotic positioning system; an ultrasound therapy transducer array connected to the robotic positioning system; a generator operatively coupled to the ultrasound therapy transducer array, the generator and ultrasound therapy transducer array being configured to deliver histotripsy pulses into a subject to generate a cavitation bubble cloud in the subject; at least one processor operatively coupled to the robotic positioning system and the generator, the at least one processor being configured to control the robotic positioning system and the generator to provide histotripsy therapy to the subject according to a treatment plan that includes a plurality of discrete treatment locations within a target tissue volume of the subject by: controlling the robotic positioning system to mechanically position a focus of the ultrasound therapy transducer array at a first focal location in a first discrete treatment location within the treatment plan, the first focal location being positioned along a central axis of the ultrasound therapy transducer array; controlling the generator to deliver at least one histotripsy pulse with the ultrasound therapy transducer array to form a first cavitation bubble cloud at the first focal location; controlling the generator to electronically beam-steer the focus of the ultrasound therapy transducer to a second focal location in the first discrete treatment location that is at least partially outside the central axis; controlling the generator to deliver at least one histotripsy pulse to the second focal location to produce a second bubble cloud at the second focal location; controlling the generator to electronically beam-steer the focus of the ultrasound therapy transducer to a focal location in the first discrete treatment location that is at least partially inside the central axis; and controlling the generator to deliver at least one histotripsy pulse to the focal location to produce a third bubble cloud at the focal location.

(82) In some aspects, a natural focus of the ultrasound therapy transducer array is located along the central axis.

(83) In other aspects, the first focal location comprises the natural focus.

(84) In some aspects, the second focal location partially overlaps with the central axis.

(85) In other aspects, the focal location comprises the first focal location.

(86) In some aspects, the focal location comprises a third focal location.

(87) In other aspects, the system is configured for controlling the generator to electronically beam-steer to the second focal location comprises steering laterally.

(88) In other aspects, the system is configured for controlling the generator to electronically beam-steer to the second focal location comprises steering laterally and axially.

(89) An ultrasound therapy system is provided, comprising; an ultrasound therapy transducer array; a generator operatively coupled to the ultrasound therapy transducer array, the generator and ultrasound therapy transducer array being configured to deliver histotripsy pulses into a subject to generate a cavitation bubble cloud in the subject; at least one processor operatively coupled to the generator, the at least one processor being configured to control the generator to provide histotripsy therapy to the subject according to a treatment plan that includes a plurality of discrete treatment locations within a target tissue volume of the subject by: controlling the generator to deliver at least one histotripsy pulse with the ultrasound therapy transducer array to form a first cavitation bubble cloud at a first focal location in a first discrete treatment location within the treatment plan; controlling the generator to electronically beam-steer the focus of the ultrasound therapy transducer to a second focal location in the first discrete treatment location that overlaps with the first focal location; and controlling the generator to deliver at least one histotripsy pulse to the second focal location to produce a second bubble cloud at the second focal location and to re-excite the first cavitation bubble cloud at the first focal location.

(90) An ultrasound therapy system is provided, comprising; an ultrasound therapy transducer array; a generator operatively coupled to the ultrasound therapy transducer array, the generator and ultrasound therapy transducer array being configured to deliver histotripsy pulses into a subject to generate a cavitation bubble cloud in the subject; at least one processor operatively coupled to the generator, the at least one processor being configured to control the generator to provide histotripsy therapy to the subject according to a treatment plan that includes a plurality of discrete treatment locations within a target tissue volume of the subject by: controlling the generator to deliver at least one histotripsy pulse with the ultrasound therapy transducer array to form a first cavitation bubble cloud at a first focal location in a first discrete treatment location within the treatment plan, the first focal location being positioned along a central axis of the ultrasound therapy transducer array; controlling the generator to electronically beam-steer the focus of the ultrasound therapy transducer to a second focal location in the first discrete treatment location that is at least partially outside the central axis; controlling the generator to deliver at least one histotripsy pulse to the second focal location to produce a second bubble cloud at the second focal location; controlling the generator to electronically beam-steer the focus of the ultrasound therapy transducer to a focal location in the first discrete treatment location that is at least partially inside the central axis; and controlling the generator to deliver at least one histotripsy pulse to the focal location to produce a third bubble cloud at the focal location.

#### DETAILED DESCRIPTION

(91) The system, methods and devices of the disclosure may be used for open surgical, minimally invasive surgical (laparoscopic and percutaneous), robotic surgical (integrated into a robotically-enabled medical system), endoscopic or completely transdermal extracorporeal non-invasive acoustic cavitation for the treatment of healthy, diseased and/or injured tissue including but not limited to tissue destruction, cutting, skeletonizing and ablation. Furthermore, due to tissue selective properties, histotripsy may be used to create a cytoskeleton that allows for subsequent tissue regeneration either de novo or through the application of stem cells and other adjuvants. Finally, histotripsy can be used to cause the release of delivered agents such as chemotherapy and immunotherapy by locally causing the release of these agents by the application of acoustic energy to the targets. As will be described below, the acoustic cavitation system may include various sub-systems, including a Cart, Therapy, Integrated Imaging,

(92) Robotics, Coupling and Software. The system also may comprise various Other Components, Ancillaries and Accessories, including but not limited to computers, cables and connectors,



networking devices, power supplies, displays, drawers/storage, doors, wheels, and various simulation and training tools, etc. All systems, methods and means creating/controlling/delivering histotripsy are considered to be a part of this disclosure, including new related inventions disclosed herein.

(93) FIG. 1A generally illustrates histotripsy system **100** according to the present disclosure, comprising a therapy transducer **102**, an imaging system **104**, a display and control panel **106**, a robotic positioning arm **108**, and a cart **110**. The system can further include an ultrasound coupling interface and a source of coupling medium, not shown.

(94) FIG. 1B is a bottom view of the therapy transducer **102** and the imaging system **104**. As shown, the imaging system can be positioned in the center of the therapy transducer. However, other embodiments can include the imaging system positioned in other locations within the therapy transducer, or even directly integrated into the therapy transducer. In some embodiments, the imaging system is configured to produce real-time imaging at a focal point of the therapy transducer. The system also allows for multiple imaging transducers to be located within the therapy transducer to provide multiple views of the target tissue simultaneously and to integrate these images into a single 3-D image.

(95) The histotripsy system may comprise one or more of various sub-systems, including a Therapy sub-system that can create, apply, focus and deliver acoustic cavitation/histotripsy through one or more therapy transducers, Integrated Imaging sub-system (or connectivity to) allowing real-time visualization of the treatment site and histotripsy effect through-out the procedure, a Robotics positioning sub-system to mechanically and/or electronically steer the therapy transducer, further enabled to connect/support or interact with a Coupling sub-system to allow acoustic coupling between the therapy transducer and the patient, and Software to communicate, control and interface with the system and computer-based control systems (and other external systems) and various Other Components, Ancillaries and Accessories, including one or more user interfaces and displays, and related guided work-flows, all working in part or together. The system may further comprise various fluidics and fluid management components, including but not limited to, pumps, valve and flow controls, temperature and degassing controls, and irrigation and aspiration capabilities, as well as providing and storing fluids. It may also contain various power supplies and protectors.

(96) As described above, the histotripsy system may include integrated imaging. However, in other embodiments, the histotripsy system can be configured to interface with separate imaging systems, such as C-arm, fluoroscope, cone beam CT, MRI, etc., to provide real-time imaging during histotripsy therapy. In some embodiments, the histotripsy system can be sized and configured to fit within a C-arm, fluoroscope, cone beam CT, MRI, etc.

(97) Cart

(98) The Cart **110** may be generally configured in a variety of ways and form factors based on the specific uses and procedures. In some cases, systems may comprise multiple Carts, configured with similar or different arrangements. In some embodiments, the cart may be configured and arranged to be used in a radiology environment and in some cases in concert with imaging (e.g., CT, cone beam CT and/or MRI scanning). In other embodiments, it may be arranged for use in an operating room and a sterile environment for open surgical or laparoscopic surgical and endoscopic application, or in a robotically enabled operating room, and used alone, or as part of a surgical robotics procedure wherein a surgical robot conducts specific tasks before, during or after use of the system and delivery of acoustic cavitation/histotripsy. As such and depending on the procedure environment based on the aforementioned embodiments, the cart may be positioned to provide sufficient work-space and access to various anatomical locations on the patient (e.g., torso, abdomen, flank, head and neck, etc.), as well as providing work-space for other systems (e.g., anesthesia cart, laparoscopic tower, surgical robot, endoscope tower, etc.).

(99) The Cart may also work with a patient surface (e.g., table or bed) to allow the patient to be

presented and repositioned in a plethora of positions, angles and orientations, including allowing changes to such to be made pre, peri and post-procedurally. It may further comprise the ability to interface and communicate with one or more external imaging or image data management and communication systems, not limited to ultrasound, CT, fluoroscopy, cone beam CT, PET, PET/CT, MRI, optical, ultrasound, and image fusion and or image flow, of one or more modalities, to support the procedures and/or environments of use, including physical/mechanical interoperability (e.g., compatible within cone beam CT work-space for collecting imaging data pre, peri and/or post histotripsy) and to provide access to and display of patient medical data including but not limited to laboratory and historical medical record data.

(100) In some embodiments one or more Carts may be configured to work together. As an example, one Cart may comprise a bedside mobile Cart equipped with one or more Robotic arms enabled with a Therapy transducer, and Therapy generator/amplifier, etc., while a companion cart working in concert and at a distance of the patient may comprise Integrated Imaging and a console/display for controlling the Robotic and Therapy facets, analogous to a surgical robot and master/slave configurations.

(101) In some embodiments, the system may comprise a plurality of Carts, all slave to one master Cart, equipped to conduct acoustic cavitation procedures. In some arrangements and cases, one Cart configuration may allow for storage of specific sub-systems at a distance reducing operating room clutter, while another in concert Cart may comprise essentially bedside sub-systems and componentry (e.g., delivery system and therapy).

(102) One can envision a plethora of permutations and configurations of Cart design, and these examples are in no way limiting the scope of the disclosure.

(103) Histotripsy

(104) Histotripsy comprises short, high amplitude, focused ultrasound pulses to generate a dense, energetic, “bubble cloud”, capable of the targeted fractionation and destruction of tissue.

Histotripsy is capable of creating controlled tissue erosion when directed at a tissue interface, including tissue/fluid interfaces, as well as well-demarcated tissue fractionation and destruction, at sub-cellular levels, when it is targeted at bulk tissue. Unlike other forms of ablation, including thermal and radiation-based modalities, histotripsy does not rely on heat cold or ionizing (high) energy to treat tissue. Instead, histotripsy uses acoustic cavitation generated at the focus to mechanically effect tissue structure, and in some cases liquefy, suspend, solubilize and/or destruct tissue into sub-cellular components.

(105) Histotripsy can be applied in various forms, including: 1) Intrinsic-Threshold Histotripsy: Delivers pulses typically with a 1-2 cycles of high amplitude negative/tensile phase pressure exceeding the intrinsic threshold to generate cavitation in the medium (e.g., ~24-28 MPa for water-based soft tissue), 2) Shock-Scattering Histotripsy: Delivers typically pulses 1-20 cycles in duration. The shockwave (positive/compressive phase) scattered from an initial individual microbubble generated forms inverted shockwave, which constructively interfere with the incoming negative/tensile phase to form high amplitude negative/rarefactional phase exceeding the intrinsic threshold. In this way, a cluster of cavitation microbubbles is generated. The amplitude of the tensile phases of the pulses is sufficient to cause bubble nuclei in the medium to undergo inertial cavitation within the focal zone throughout the duration of the pulse. These nuclei scatter the incident shockwaves, which invert and constructively interfere with the incident wave to exceed the threshold for intrinsic nucleation, and 3) Boiling Histotripsy: Employs pulses roughly 1-20 ms in duration. Absorption of the shocked pulse rapidly heats the medium, thereby reducing the threshold for intrinsic nuclei. Once this intrinsic threshold coincides with the peak negative pressure of the incident wave, boiling bubbles form at the focus.

(106) The large pressure generated at the focus causes a cloud of acoustic cavitation bubbles to form above certain thresholds, which creates localized stress and strain in the tissue and mechanical breakdown without significant heat deposition. At pressure levels where cavitation is not generated,

minimal effect is observed on the tissue at the focus. This cavitation effect is observed only at pressure levels significantly greater than those which define the inertial cavitation threshold in water for similar pulse durations, on the order of 10 to 30 MPa peak negative pressure.

(107) Histotripsy may be performed in multiple ways and under different parameters. It may be performed entirely or completely non-invasively by acoustically coupling a focused ultrasound transducer over the skin of a patient and transmitting acoustic pulses transcutaneously through overlying (and intervening) tissue to the focal zone (treatment zone and site). The application of histotripsy is not limited to a transdermal approach but can be applied through any means that allows contact of the transducer with tissue including open surgical laparoscopic surgical, percutaneous and robotically mediated surgical procedures. It may be further targeted, planned, directed and observed under direct visualization, via ultrasound imaging, given the bubble clouds generated by histotripsy may be visible as highly dynamic, echogenic regions on, for example, B Mode ultrasound images, allowing continuous visualization through its use (and related procedures). Likewise, the treated and fractionated tissue shows a dynamic change in echogenicity (typically a reduction), which can be used to evaluate, plan, observe and monitor treatment.

(108) Generally, in histotripsy treatments, ultrasound pulses with 1 or more acoustic cycles are applied, and the bubble cloud formation relies on the pressure release scattering of the positive shock fronts (sometimes exceeding 100 MPa, P+) from initially initiated, sparsely distributed bubbles (or a single bubble). This is referred to as the “shock scattering mechanism”.

(109) FIG. 3 illustrates an ultrasound pulse that can be used for shock scattering histotripsy. As shown the ultrasound pulse can include a leading negative half cycle, a peak positive half cycle, a peak negative half cycle, and a trailing positive half cycle (with the pulse traveling from right to left on the page). As shown, the trailing positive half cycle has a lower amplitude than the peak positive half cycle. This mechanism depends on one (or a few sparsely distributed) bubble(s) initiated with the initial negative half cycle(s) of the pulse at the focus of the transducer. A cloud of microbubbles then forms due to the pressure release backscattering of the high peak positive shock fronts from these sparsely initiated bubbles. These back-scattered high-amplitude rarefactional waves exceed the intrinsic threshold thus producing a localized dense bubble cloud. Each of the following acoustic cycles then induces further cavitation by the backscattering from the bubble cloud surface if the amplitude of those cycles is sufficient, which grows towards the transducer. As a result, an elongated dense bubble cloud growing along the acoustic axis opposite the ultrasound propagation direction is observed with the shock scattering mechanism. This shock scattering process makes the bubble cloud generation not only dependent on the peak negative pressure, but also the number of acoustic cycles and the amplitudes of the positive shocks. Without at least one intense shock front developed by nonlinear propagation, no dense bubble clouds are generated when the peak negative half-cycles are below the intrinsic threshold.

(110) When the amplitude(s) of positive half cycle(s) of each pulse are limited, shock scattering can be minimized, and the generation of a dense bubble cloud depends on the negative half cycle(s) of the applied ultrasound pulses exceeding an “intrinsic threshold” of the medium. This is referred to as the “intrinsic threshold mechanism”.

(111) This threshold can be in the range of 26-30 MPa for soft tissues with high water content, such as tissues in the human body. In some embodiments, using this intrinsic threshold mechanism, the spatial extent of the lesion may be well-defined and more predictable. With peak negative pressures (P-) not significantly higher than this threshold, sub-wavelength reproducible lesions as small as half of the -6 dB beam width of a transducer may be generated.

(112) With high-frequency Histotripsy pulses, the size of the smallest reproducible lesion becomes smaller, which is beneficial in applications that require precise lesion generation. However, high-frequency pulses are more susceptible to attenuation and aberration, rendering problematical treatments at a larger penetration depth (e.g., ablation deep in the body) or through a highly aberrative medium (e.g., transcranial procedures, or procedures in which the pulses are transmitted

through bone(s)). Histotripsy may further also be applied as a low-frequency “pump” pulse (typically <2 cycles and having a frequency between 100 kHz and 1 MHz) can be applied together with a high-frequency “probe” pulse (typically <2 cycles and having a frequency greater than 2 MHz, or ranging between 2 MHz and 10 MHz) wherein the peak negative pressures of the low and high-frequency pulses constructively interfere to exceed the intrinsic threshold in the target tissue or medium. The low-frequency pulse, which is more resistant to attenuation and aberration, can raise the peak negative pressure P-level for a region of interest (ROI), while the high-frequency pulse, which provides more precision, can pin-point a targeted location within the ROI and raise the peak negative pressure P-above the intrinsic threshold. This approach may be referred to as “dual frequency”, “dual beam histotripsy” or “parametric histotripsy.”

(113) Additional systems, methods and parameters to deliver optimized histotripsy, using shock scattering, intrinsic threshold, and various parameters enabling frequency compounding and bubble manipulation, are herein included as part of the system and methods disclosed herein, including additional means of controlling said histotripsy effect as pertains to steering and positioning the focus, and concurrently managing tissue effects (e.g., prefocal thermal collateral damage) at the treatment site or within intervening tissue. Further, it is disclosed that the various systems and methods, which may include a plurality of parameters, such as but not limited to, frequency, operating frequency, center frequency, pulse repetition frequency, pulses, bursts, number of pulses, cycles, length of pulses, amplitude of pulses, pulse period, delays, burst repetition frequency, sets of the former, loops of multiple sets, loops of multiple and/or different sets, sets of loops, and various combinations or permutations of, etc., are included as a part of this disclosure, including future envisioned embodiments of such.

(114) Challenges of Residual/Remnant Cavitation Nuclei

(115) The concept of residual/remnant cavitation nuclei is known. It relates to the phenomenon that after a bubble cloud is generated and rapidly collapses within a medium (e.g., tissue), bubbles may remain sparsely distributed throughout the medium where prior histotripsy therapy has been applied. In some situations, the residual cavitation nuclei can affect histotripsy treatment and can create undesirable local environments that can challenge the ability to deliver histotripsy, effectively. This includes their ability to seed cavitation by subsequent therapy pulses, influencing the spatial distribution of subsequent bubble clouds such that cavitation occurs at prior treated locations (i.e., cavitation memory effect), while other and/or adjacent sites may experience reduced bubble cloud exposure or undertreatment. In some scenarios and conditions, remnant nuclei may result in bubble clouds forming preferentially where the remnant nuclei are located and at a distance from the prescribed focal location, resulting in less uniform treatment and/or distribution of treatment through the larger planned volume. Furthermore, and potentially more problematic in some cases, remnant nuclei can cause shielding/attenuation of subsequent pulses if they exist along the acoustic propagation path such that they can attenuate the incident therapy waveforms and reduce the energy delivered to tissue. In some examples, they may partially or fully block therapy.

(116) Bubble Cloud Enhancement with Rapid Steering Between Overlapping Focal Zones

(117) As described above, histotripsy therapy bubble clouds can be generated in a number of ways, including methods based on shock scattering and intrinsic threshold pulses. These sequences may include, but are not limited to, various types of pulses, including therapy, bubble manipulation pulses, etc. In some cases, bubble manipulation pulses may be used to modify the focal and treatment zone in effort to move or manipulate residual bubbles/nuclei remaining from prior pulses, which may adversely affect histotripsy treatments (e.g., act as blockers to subsequent histotripsy pulses).

(118) This disclosure provides electronic beam-steering strategies configured to modify a single focal location environment to overcome the blockage, attenuation, and cavitation memory effect. This disclosure further provides additional systems, methods, and techniques to advantageously use adjacent residual cavitation nuclei to produce enhanced or enlarged cavitation bubble clouds in a

highly controlled manner through the use of specifically designed computer-controlled hardware/software systems.

(119) Provided herein are novel systems and methods for forming “enhanced excitation volumes” by enhancing, enlarging, and/or shaping a defined histotripsy “excitation volume”, leveraging unique approaches to re-excite remnant nuclei to create a larger bubble cloud than would be possible without remnant nuclei re-excitation. For purposes of this disclosure, an “excitation volume” can comprise a volume within which a cavitation bubble cloud generated by a single, discrete focal location with an ultrasound therapy transducer array is expected to be formed. An “enhanced excitation volume” can comprise an enhanced or enlarged volume within which a cavitation bubble cloud generated by overlapping focal locations with electronic beam-steering of the ultrasound therapy transducer array between the overlapping focal locations is expected to be formed. As described herein, the pulse sequences and electronic beam-steering for enhanced excitation volumes can be chosen to advantageously exploit residual cavitation nuclei to enhance the volume of cavitation formed in a highly controlled manner. The systems and methods that enable this enhanced excitation volume may generally include 1) initiation of a bubble cloud at a first focal location (e.g., a natural focus of the therapy transducer array) and creation of a corresponding volume referred to as the “focal zone”, 2) electronically beam-steering (e.g., phased array beam-steering) the focus to one or more additional focal locations in 2D (two-dimensional) or 3D (three-dimensional) space for creation of corresponding focal zones, 3) wherein the corresponding focal zone(s) are at least partially overlapping (with the first/prior focal zone volume(s)) and temporally and spatially positioned where pressures are within a threshold or range of the peak-pressure of the prior focal location/zone, 4) to re-excite remaining remnant nuclei (in the overlapping volume), 5) which together create the overall enhanced excitation volume or “larger bubble cloud” as shown in at least FIGS. 5B and 5C.

(120) The above-described enhanced excitation volumes may be created in unlimited permutations, based on the selection of, but not limited to, the spatial and temporal parameters for creating these 3D excitation volumes. They may be comprised of varied shape and size by in part modifying the position of the steered focal locations in 3D space (and corresponding focal zones), the specific range and spacing of overlap of focal zones, beam profile overlap, maximum distance from the first focal location to the most distant focal location, sequencing, timing and/or phasing of therapy and/or bubble manipulation pulses including any pauses or wait periods, and/or the number of pulses per focal location and total number of pulses applied to the defined 3D excitation volume to continuously grow and/or fill in the excitation volume.

(121) Further, in some embodiments, the design of the steering direction may be selected to allow the rapid steering and re-excitement of remnant nuclei, but to also minimize the need to steer directly distal to a focal location that was targeted immediately prior to aid in minimizing shielding or blocking effects not conducive to re-exciting the desired effect for creating larger excitation volumes. As will be described later, the order of the steered locations can be of importance in minimizing the shielding or blocking effects not conducive to the invention of the present disclosure.

#### (122) Spatial and Temporal Considerations

(123) The enhanced excitation volumes described above may be designed, configured, and optimized based on various spatial and temporal considerations of 1) the spacing of the selected focal locations/zones so that the focal zone of a subsequent pulse overlaps with the focal zone of the/a previous pulse and 2) the timing of the subsequent pulse within a time span sufficient to re-excite residual nuclei present in the overlapping zone from one or more previous pulses.

(124) The selection of spacing parameters may include selecting both the distance between adjacent focal locations, as well as the distance (range) between the most extreme focal locations. In some embodiments, there may be a plurality of focal location positions at uniformly and/or non-uniformly spaced coordinates (x, y, z), which may be selected/defined based on desired target

excitation volume size (e.g., specific spacing values may result in varied per pulse bubble cloud area). The selection of these parameters may also be based on the specific histotripsy sequence, acoustic field properties and corresponding bubble cloud size. In some examples, the distance between steered focal locations may be based on the wavelength of the therapy pulses. In particular, the beam profile can be proportional to the wavelength of the ultrasound signal. In some implementations, the longer the wavelength of the ultrasound signal, the larger the beam profile. Similarly, the shorter the wavelength of the ultrasound signal, the smaller the beam profile. Furthermore, the beam profile in each axis (x, y, z) is also affected by the specific geometry of the therapy transducer. For example, transducers with f-numbers between 0.5 and 1 may result in pulses that generate “elongate” bubble clouds that have longer z dimensions than dimensions in x or y. Following this same principle, this may allow for the ability to electronically steer to a greater distance in the z direction than in x and y.

(125) In terms of additional temporal aspects, the timing of electronic steering to adjacent or distant focal locations (and creation of overlapping focal zones), may use similar or varied pulse repetition frequency (PRF) and steering periods, and may be selected based on desired dose delivered. In one example, an enhanced excitation volume is created using one representative histotripsy pulse delivered to a first focal location at a selected PRF, and a following pulse is electronically beam-steered and delivered to a subsequent (or second) focal location that overlaps with the first focal location, wherein the steering period (amount of time from one electronic steering event to the next) is equivalent to the therapy PRF (amount of time from one therapy pulse to the next). This may be repeated until the total desired number of therapy pulses are delivered to the enhanced excitation volume. In other examples, based on the number and position of steered focal locations, some steered locations may have the same or different pulse(s), steering period or PRF, and/or other therapy parameters. For example, the electronic steering period may be shorter (faster) than the PRF. The systems described herein may also have embedded logic and/or algorithms to account and/or adjust for the spatial/temporal relationships.

(126) Volumes may be created based on the selection of spacing distances, various shapes, and morphologies of excitation volumes. These may include continuous or discontinuous volumes and the selection of these parameters may be based on application (e.g., complete tumor destruction versus gross tissue debulking, etc.). As a representative example, some excitation volumes may be ellipsoidal in shape, whereas others may be more bulb-like. Other shapes and geometries of excitation volumes are envisioned and within the scope of this disclosure.

(127) In some embodiments, these may be created by using rapid axial beam-steering (z axis) over short distances between focal locations (<5 mm) of overlapping focal zones to aid in creating a longer continuous enhanced excitation volume but using approximately the same number of pulses for a single corresponding natural focus (only) bubble cloud. In some embodiments, lateral beam-steering (x-y axis) overlapping focal locations may be used. In other embodiments, a combination of axial and lateral electronic beam-steering between focal locations (and overlapping focal locations) may be used to create more spherical enhanced excitation volumes, in part to enable more efficient/easier packing of such volumes into a larger grid pattern of planned treatment volume. In some examples, the distribution of focal points within an electronic steering embodiment can be adjusted to alter an aspect ratio of the ellipsoidal cloud specific to best fit desired treatment patterns and plan geometries.

(128) Planned Treatment Volumes Based on Patterned/Packed “Enhanced Excitation Volumes”

(129) A plurality of enhanced excitation volumes (or larger bubble clouds) may also be arranged and packed into subsequent larger structures comprised of defined 3D patterns and shapes, of which may be used to create a defined planned treatment volume. For example, if the target tissue to be treated is a tumor, then the treatment volume can encompass the tumor (and any desired margins) and be filled with a plurality of enhanced excitation volumes. The plurality of enhanced excitation volumes may comprise one or more specific shapes or sizes (based on spatial selection

of focal locations and how pulses are applied to those positions). The system and control software delivering therapy may also comprise logic to populate a desired and/or user selected planned treatment volume with enhanced excitation volumes using packing rules for the placement, position and numbers required to fill that desired volume/shape.

(130) The further arrangement and packing of the enhanced excitation volumes may include the selection of their axial and radial spacing, and degree of overlap. Planned treatment volumes may comprise the packing of one iteration of an enhanced excitation volume or more than one variation of them. For example, larger enhanced excitation volumes may be packed more centrally within a planned treatment volume, with smaller excitation volumes positioned along the perimeter of the planned treatment volume. Rules and logic including overlapping of focal zones and temporal and distance spacing apply to both of the above examples and in other implementations discussed herein. In an analogous example, varied shapes may be used to pack planned treatment volumes as well. In this use case, the use of more eccentric ellipsoidal shapes may be positioned more centrally, with surrounding/adjacent enhanced treatment volumes being more spherical in shape to better enable packing the outer perimeter of a treatment volume (and plan).

(131) In particular, the enhanced excitation volumes described are generally larger than an excitation volume formed by a single, discreet focal zone volume, such as an excitation volume formed at the natural focus of the therapy transducer array. Therefore, fewer enhanced excitation volumes are required to pack a given treatment volume, such as a 3 cm sphere, compared to the number of discreet excitation volumes that would be required to pack the same volume if the enhanced excitation volume techniques were not used. In some embodiments, for a given treatment plan, a specified number of pulses are calculated for each excitation volume or enhanced excitation volume comprising the treatment volume to achieve complete treatment of the treatment plan while managing the thermal dose delivered to the patient. The same number of pulses are employed for either discreet focal zones (excitation volumes) or enhanced excitation volumes. Because fewer enhanced excitation volumes are required to populate the planned treatment volume, faster treatment times are enabled with using enhanced excitation volumes.

(132) FIGS. 4A-4B show a comparison of packing a treatment volume with a plurality of excitation volumes vs. enhanced excitation volumes. FIGS. 4A-4B show a cross-sectional view of the treatment volumes in the z-x plane of the therapy transducer array. In this example, the therapy transducer array is positioned at the top of the page and directs treatment along the z-axis towards the bottom of the page.

(133) FIG. 4A shows a treatment volume **401** packed with a plurality of excitation volumes **404**. In this implementation, the excitation volumes are each discrete cavitation bubble cloud locations within the treatment volume, wherein each excitation volume is formed at the natural focus **405** of the therapy transducer array. To navigate through the treatment volume of FIG. 4A, the histotripsy system can use mechanical movement of the ultrasound therapy transducer array between excitation volumes as controlled by the robotic positioning system (e.g., mechanically moving the transducer array to the location for a given excitation volume), followed by delivering histotripsy therapy to generate the bubble cloud/excitation volume. This process can then be repeated for the remainder of the treatment volume by mechanically moving the transducer array to the next location for the next excitation volume, and so forth.

(134) In contrast, FIG. 4B shows a treatment volume **401** packed with a plurality of enhanced excitation volumes **407**. In this implementation, the enhanced excitation volumes are formed by rapidly electronically beam-steering the ultrasound therapy transducer array to a plurality of overlapping focal locations within the enhanced excitation volume to form an enhanced or larger bubble cloud by re-exciting residual cavitation nuclei within the enhanced excitation volume. As shown in FIG. 4B, the enhanced excitation volumes **407** are larger than the excitation volumes **403** of FIG. 4A. This allows for packing of the treatment volume **401** with a smaller number of enhanced excitation volumes **407**, compared to the number of excitation volumes **404** required to

pack the volume of FIG. 4A. To navigate through the treatment volume of FIG. 4B, the histotripsy system can use mechanical movement of the ultrasound therapy transducer array between enhanced excitation volumes as controlled by the robotic positioning system (e.g., mechanically moving the transducer array to a given enhanced excitation volume, and then electronically beam-steering the ultrasound transducer array within the enhanced excitation volume and delivering histotripsy therapy to generate the enhance bubble cloud/excitation volume. This process can then be repeated for the remainder of the treatment volume by mechanically moving the transducer array to the next enhanced excitation volume, electronically beam-steering the transducer and delivering histotripsy to form subsequent enhanced excitation volumes, and so forth.

(135) In one specific implementation, an excitation volume formed at a natural focus of the ultrasound therapy transducer array, without rapid electronic beam-steering, can have volume dimensions of approximately 3 mm×3 mm×7 mm. In contrast, an enhanced excitation volume formed by rapid electronic beam-steering of the ultrasound transducer array can have volume dimensions of approximately 3 mm×3 mm×11 mm. With a treatment volume size of a 3 cm sphere in FIG. 4A, packing of the treatment volume with the natural focus excitation volumes requires approximately 1049 excitation volumes. However, implementing the enhanced excitation volumes, with the larger volume size, requires only 605 enhanced excitation volumes to pack the same 3 cm treatment volume sphere. It can therefore be seen that the enhanced excitation volumes of FIG. 4B require fewer volumes to pack the same sized treatment volume, which therefore increases the efficiency of treatment and reduces treatment times.

(136) Additionally, some permutations of parameters selected for the design of enhanced excitation volumes may preferentially assist in breaking through the biomechanical interfaces of adjacent focal zones (e.g., between natural focus and subsequent beam-steered location(s)), and/or of adjacent excitation volumes (additional discrete positions of volumes placed adjacently as part of a larger 3D planned treatment volume).

(137) During therapy, the histotripsy therapy transducer and robotic positioning system can then be configured to automatically traverse a focus of the therapy transducer sequentially through each of the discrete therapy locations of the planned treatment volume. This traversal can be, for example, mechanical movement of the histotripsy therapy transducer by physically changing a position/orientation of the histotripsy therapy transducer with the robotic positioning system.

(138) Implementation of Enhanced Excitation Volumes

(139) As described above, with a discrete excitation volume, the transducer focus is positioned at each discrete therapy location (e.g., planned focal locations) and a cavitation bubble cloud or focal zone is formed at each location for a pre-determined time based on the dose of therapy desired for each treatment location. FIG. 5A illustrates an example of an excitation volume **504** defined by the histotripsy bubble cloud or cavitation **502** formed at a discrete focal location. The dashed ellipses of the excitation volume **504** are intended to depict the general location of the bubble cloud but do not depict a specific size with respect to the acoustic field. In the illustrated example, the therapy transducer is assumed to be on the right side of the page, with the histotripsy pulses being delivered from right to left on the page along the z-axis of the therapy transducer. In this example, the excitation volume **504** can be placed on a discrete treatment location of a treatment plan within a target tissue volume (e.g., as depicted in FIG. 4A). As shown, the bubble cloud **502** can include bubbles, bubble clouds, or cavitation outside of the excitation volume **504**.

(140) According to one aspect of this disclosure, as shown in FIG. 5B, an enhanced excitation volume **507** can comprise a plurality of focal zones (e.g., **504a**, **504b**, and **504c**), and the therapy transducer can be configured and controlled to rapidly electronically beam-steer between the plurality of closely spaced focal zones **504a**, **504b**, and **504c** to produce the cavitation or bubble cloud(s) which collectively form the enhanced excitation volume **507**. In some embodiments, as shown in FIG. 5B, the steering is accomplished with electronic beam-steering of the therapy transducer array. In some examples, axial steering along the z-axis of the transducer array can be



used, starting with the first focal zone **504a** furthest away from the transducer. Electronic beam-steering through the focal zones can be in any direction, including proximal-most to distal-most, or starting at the central focal location and traversing either distally or proximally to the next focal zone. However, transitioning between focal zones from distal-most to proximal-most (with respect to the transducer) can advantageously limit the shielding effects of residual cavitation nuclei for the first axial pass through the focal zones.

(141) In the embodiment shown in FIG. 5B, the focal zones **504a**, **504b**, and **504c** are partially overlapping so as to produce an elongated, continuous enhanced excitation volume **507** across the focal zones **504a**, **504b**, and **504c**. It is an object of this disclosure that lower amplitude regions of the histotripsy beam sufficiently overlap with previously targeted focal locations to re-excite residual nuclei. However, it should be understood that in other embodiments, the focal zones need not overlap. Generally, according to this disclosure, the system can be configured to position an enhanced excitation volume at a first discrete treatment location (e.g., by mechanically moving or manipulating the robotic positioning system), deliver a first histotripsy pulse to the first treatment location comprising the enhanced excitation volume to create a bubble cloud(s)/cavitation, rapidly electronically beam-steer the transducer array focus to a second focal location comprising the enhanced excitation volume, deliver a second histotripsy pulse to the second focal location to create a second bubble cloud(s)/cavitation, and repeat for as many desired focal locations/focal zones for the given enhanced excitation volume positioned at the discrete treatment location.

(142) While the embodiment of FIG. 5B shows three overlapping focal zones (e.g., **504a**, **504b**, **504c**) for each enhanced excitation volume **507**, any number of focal zones can be created depending on the treatment plan, target tissue volume size, and dose of therapy required or prescribed. For example, two overlapping focal locations per enhanced excitation volume are possible, as are four or more focal locations per enhanced excitation volume, providing there is sufficient beam overlap of the lower amplitude portions of the beam re-exciting residual cavitation nuclei in previously treated focal locations.

(143) The distance of electronic beam-steering between focal locations/zones can vary, but in general the distance between the focal locations can be optimized based on the size of the bubble cloud generated by a chosen pulse. In some examples, the distance of steering between focal locations can be based on the wavelength of the ultrasound pulses and whether the steering is axial or lateral. For a histotripsy system using a selected pulse generating the bubble cloud sizes described and depicted in this disclosure, it has been found that optimal axial spacing between focal locations can be in the range of 1 mm to 5 mm between subsequent focal locations. In one embodiment, 2.5 mm spacing between focal locations advantageously provides sufficient overlap between bubble clouds at adjacent focal locations and optimally takes advantage over lower amplitude portions of the beam re-exciting residual cavitation nuclei in previously treated focal locations to improve the consistency and continuity of the bubble cloud across all the focal locations. In embodiments where the focal locations are laterally spaced from another, the range of movement/steering can differ from that of the axially spaced embodiment. Generally, lateral spacing can be within a 1 mm radius from the focus of the transducer. In one embodiment, 0.5 mm spacing between laterally spaced focal locations advantageously provides overlap between bubble clouds at adjacent focal locations and optimally takes advantage over lower amplitude portions of the beam re-exciting residual cavitation nuclei in previously treated focal locations to improve the consistency and continuity of the bubble cloud across all the focal locations.

(144) The histotripsy pulse delivered to each focal location can comprise, for example, a shocked scattering histotripsy pulse, such as the one illustrated in FIG. 3. For example, in FIG. 5B, the histotripsy system can be configured to position the therapy transducer focus at the first focal location **504a** and deliver the histotripsy pulse of FIG. 3 to that first focal location. More specifically, the pulse can include a peak positive half cycle, followed by a peak negative half cycle, followed by a trailing positive half cycle. As shown in FIG. 3, the trailing positive half cycle

can have a smaller amplitude than the peak positive half cycle. After the pulse is delivered to the first focal location, the transducer array can be rapidly electronically beam-steered to the second focal location and the pulse can be delivered to the second focal location. This process can be repeated for the third focal location. In some embodiments, the focus of the transducer is steered so rapidly, and subsequent pulses are delivered to subsequent focal locations such that a bubble cloud can be generated at a second or subsequent focal location during the life-cycle of the previous bubble cloud. In some examples, the “life-cycle” of the bubble cloud is considered to be the time during which the bubble cloud is formed within the focal location. In other embodiments, the “life-cycle” of the bubble cloud can include the presence of residual nuclei in the focal location even after the initial bubble cloud has expanded and collapsed. For example, a first bubble cloud can be generated at a first focal location with therapy pulses, and the transducer can then be rapidly electronically beam-steered to a second focal location and deliver therapy pulses to the second focal location to generate a second bubble cloud during the life-cycle of the first bubble cloud (e.g., before the first bubble cloud collapses). The overlapping focal zones combined with the rapid electronic beam-steering and repetition of this sequence creates a larger enhanced excitation volume.

(145) In one specific embodiment, the therapy pulses can have a repetition rate of ~600 Hz, and timing of the rapid electronic beam-steering can be as follows: 1) One iteration of the therapy pulse is delivered to the first focal location (e.g., **504a** in FIG. 5B). 2) The next iteration of the therapy pulse is delivered to the second focal location (e.g., **504b** in FIG. 5B) via electronic beam-steering. In this example, the period of steering is equal to the period of the therapy sequence ( $1/600\text{Hz} \sim 1.7$  ms). 3) The next iteration of the therapy pulse is delivered to the third focal location (e.g., **504c** in FIG. 5B) via electronic beam-steering; the period of steering is equal to the period of the therapy sequence ( $1/600\text{ Hz} \sim 1.7$  ms). 4) The next iteration of the therapy pulse is delivered at the first focal location (e.g., **504a** in FIG. 5B) via electronic beam-steering; the period of steering is equal to the period of the therapy sequence ( $1/600\text{ Hz} \sim 1.7$  ms). And the process described above can be repeated until the target dose, which may include a target number of pulses, is delivered to the focal locations at that treatment location. While the embodiments described above and in FIG. 5B describe rapidly steering between three distinct focal locations for each enhanced excitation volume, it should be understood that other embodiments can include any number of rapidly beam-steerable focal locations (e.g., two, four, five, six, or more) and variably positioned in 3D space.

(146) Depending on the dose desired or prescribed for a given discrete treatment location, doses of therapy can be delivered multiple times to each focal location. For example, therapy can be provided to the first focal location, then the second focal location, then the third focal location, before moving back to the first focal location and repeating the process. In other embodiments, only a single dose at each focal location may be sufficient to achieve the desired dose for that treatment location. In some embodiments, the dose is varied spatially and temporally non-uniformly across a planned treatment volume comprised of a plurality of enhanced excitation volumes. For example, the extent of therapy in each iteration can also be configurable (e.g.,  $n$  pulses delivered to a first focal location followed by  $n$  pulses delivered to a second focal location followed by  $n$  pulses delivered to a third focal location, where  $n$  could be equal to 1, 2, . . . , 100 or more, etc. prior to iterating through the focal locations again).

(147) In the embodiment of FIG. 5B, the therapy transducer can be electronically beam-steered axially along its z-axis in between the discrete focal locations. For example, the first focal location **504a** can be the natural focus of the therapy transducer, the second focal location **504b** can be axially displaced towards the transducer by a predetermined (overlapping) distance from the first focal location, and the third focal location can be axially displaced towards the transducer from the second focal location.

(148) It should be understood that in other embodiments, the electronic beam-steering can be implemented laterally in the x-y direction. One example of lateral electronic beam-steering is

shown in FIG. 5C, in which the first focal zone **504a** is displaced laterally in the x-y direction from the second focal zone **504b** to form an enhanced excitation volume **507**. As shown in this embodiment, the focal zones partially overlap, but as described above, while there are advantages to overlapping focal zones this is not necessary. It should also be understood that any of the steering described in FIGS. 5B and 5C can be implemented via electronic beam-steering.

(149) While FIG. 5B shows an embodiment where multiple focal zones are spaced axially apart in the z-direction, and FIG. 5C shows an embodiment where multiple focal zones are spaced laterally apart in the x-y direction, it should be understood that in other embodiments, the rapid steering can include a combination of movement/steering in both the x-y direction and the z-direction.

(150) FIGS. 6A-6B illustrate one embodiment of an electronic beam-steering pattern used to generate an enhanced excitation volume. In this example, the focus of the ultrasound transducer array can be steered laterally (e.g., in the x-y axis) from the natural focus **604a** of the transducer array. The specific sequence and timing of the electronic beam-steering can be selected and chosen to optimally maintain cavitation and/or advantageously use or re-excite residual cavitation. In one specific embodiment, the histotripsy therapy can be initiated at the natural focus **604a** of the transducer array. The focus can then be sequentially electronically beam-steered through the remaining focal locations of the enhanced excitation volume, moving from natural focus **604a** to focal locations **604b**, **604c**, **604d**, **604e**, **604f**, **604g**, **604h**, and **604i**. The individual points marked as focal locations in FIGS. 6A-6B represent the center of the focal locations. It should be understood that the focal zones themselves corresponding to the bubble cloud or cavitation created when histotripsy therapy is delivered can be overlapping in the illustrated embodiments.

(151) While the beam-steering pattern of FIG. 6A is shown generally as a wheel or ring around the central natural focus, it should be understood that the pattern can be traversed either clockwise or counter-clockwise in direction.

(152) Furthermore, according to some embodiments, the beam-steering pattern may revisit or repeat therapy delivery to some or all of the focal locations as the transducer array is beam-steered through the locations. The parameters for revisiting or repeating therapy delivery to focal locations can be predetermined, or customized by a user. In one embodiment, an important element of the electronically steered patterns is that in succession at the steered focal location within the enhanced volume there are both instances of new cavitation events and the restimulation of residual nuclei. In addition to these events at the current focal location, previously sonicated locations are being treated by the re-excitation of their residual nuclei. This process of periodically revisiting points can greatly assist in the sustained cavitation across the enhanced excitation volume, especially along a central axis of the transducer. In practice, it has been determined that periodically revisiting the central axis of the enhanced excitation volume during traversal of the electronic beam-steering pattern is beneficial to maintaining cavitation throughout the enhanced excitation volume.

(153) Therefore, according to one embodiment and further depicted in FIG. 6B, the natural focus **604a** of the enhanced excitation volume can be re-excited after a predetermined number of focal locations (e.g., after every new focal location, every 2 focal locations, every 3 focal locations, etc.). Alternatively, the steering logic can require that 2 or fewer, 3 or fewer, 4 or fewer, or 5 or fewer locations are treated before returning to a focal location along the central axis of the transducer to re-excite that focal location along the center axis of the transducer.

(154) For example, one specific beam-steering pattern may require revisiting and re-exciting the natural focus (or a focal location central to the beam steering pattern) at least every 3 focal locations. Referring to FIG. 6A, this beam-steering pattern may start at the natural focus **604a**, electronically beam-steer to focal location **604b** and then to focal location **604c**, before returning to natural focus **604a** maintaining re-excitation of that entire steered sector throughout the raster of the pattern. The pattern can then continue to focal location **604c**, move to focal location **604d**, and then return to natural focus **604a**. From there, the pattern can continue back to focal location **604d**, to focal location **604e**, to natural focus **604a**, back to focal location **604e**, to focal location **604f**, to

natural focus **604a**, back to focal location **604f**, to focal location **604g**, to natural focus **604a**, back to focal location **604g**, to focal location **604h**, to natural focus **604a**, back to focal location **604h**, and finally to focal location **604i**. In the pattern described above, at least a portion of the natural focus **604a** is re-excited every 3 focal locations. This ensures that the residual nuclei in the central portion of the enhanced excitation volume (e.g., in an axial plane that intersects at least partially with the natural focus of the transducer) is periodically being re-excited (e.g., after a set number of focal locations), and furthermore assists with cavitation formation at the peripheral/perimeter focal locations. While the electronic beam-steering pattern of FIGS. **6A-6B** steers between focal locations only laterally from the natural focus (e.g., in the x-y plane), it should be understood that the enhanced excitation volume exists in three-dimensions. The result of this lateral steering is to increase the lateral dimensions of the enhanced excitation volume, which serves to form a more spherical excitation volume compared to the generally column shaped excitation volume that is formed at a single focal location.

(155) FIGS. **7A-7B** describe another embodiment of an electronic beam-steering pattern that includes beam-steering in 3D space (e.g., a combination of steering in the x, y, and z directions). Similar to the pattern described in FIGS. **6A-6B**, the pattern in FIGS. **7A-7B** can include an electronic beam-steering pathway of plurality of equilateral triangles. However, these triangles have been stretched over the axial (z) dimension resulting in curved arcs from the natural focus (0,0,0) to (0,0,-Z) with the two intervening points located at  $\frac{1}{4}$  and  $\frac{1}{2}(-Z)$ . A single arc moves from the natural focus of the transducer **704a**=(0,0,0) to **704b**=( $(\sqrt{2}) \cdot R/2$ ),  $-(\sqrt{2}) \cdot R/2$ ,  $-Z/4$ ) to **704c**=(R,0,-Z/2) and returns to the central axis at **704d**=(0,0,-Z). The entire arc can be rotated 45° every fifth pulse when the pattern again returns to **704a** at the natural focus. In some embodiments, the rotation could be finer (e.g., less than) 45° or more coarse (e.g., up to 90°) depending on the implementation. In the example of FIGS. **7A-7B**, the residual nuclei in a central axis of the enhanced excitation volume (e.g., an axial plane that intersects with the natural focus of the transducer) is at least partially re-excited after a set number of focal locations to maintain cavitation in the central axis and assist with cavitation formation at focal locations on the periphery, lateral to the central axis, or outside of the central axis of the enhanced excitation volume. The pattern iterates through an ordered list of the planned focal locations with **704a** and **704d** bookending each series of four points resulting in 18 unique focal locations for a pattern with 32 planned focal locations. The 32-point rotated pattern can then be repeated until reaching the proscribed dose for each target location within a given volumetric plan.

(156) FIGS. **8A-8B** illustrate yet another embodiment of an electronic beam-steering pattern that includes beam-steering in 3D space between overlapping focal locations. In the embodiment of FIGS. **8A-8B**, the beam-steering pattern comprises a central column **809a** of focal locations positioned within a plurality of vertical or axial peripheral or outer columns **809b/c/d/e** of focal locations. It should be understood that in FIGS. **8A-8B**, the therapy transducer is positioned at the top of the page and directing ultrasound energy towards the bottom of the page. In some examples, the central column can contain the natural focus **804a** or at least partially overlap with the natural focus of the therapy transducer array or be positioned along a central axis of the transducer array. Bubble clouds formed at focal locations in the central column **809a** may partially overlap with bubble clouds formed at focal locations in the peripheral columns **809b/c/d/e**.

(157) In some embodiments, the beam-steering pattern of FIGS. **8A-8B** may revisit or repeat therapy delivery to some or all of the focal locations within the central column **809a** or within a central axis of the transducer array as the transducer array is beam-steered through the focal locations in columns **809b/c/d/e**. The parameters for revisiting or repeating therapy delivery to focal locations within the central column or central axis can be predetermined, or customized by a user. In practice, it has been determined that at least partially revisiting the central column or central axis focal locations of the enhanced excitation volume periodically during traversal of the electronic beam-steering pattern can advantageously re-excite residual cavitation in the center of

the volume, which then assists with enhanced cavitation at the peripheral or perimeter focal locations. Therefore, according to one embodiment and further depicted in FIG. 8B, a focal location in the central column **809a** of focal locations of the enhanced excitation volume can be re-excited after a predetermined number of focal locations (e.g., after every new focal location, every 2 focal locations, every 3 focal locations, etc.). In practice, it has been determined that revisiting at least one focal location in the central column of the enhanced excitation volume periodically during traversal of the electronic beam-steering pattern is beneficial to maintaining cavitation throughout the enhanced excitation volume.

(158) In one specific embodiment, the steering pattern is initiated by forming cavitation at the natural focus **804a** of the central column **809a**. Electronic beam-steering can then shift the focus of the transducer array to focal location **804b** in column **809b** (the central focal location in that column), then returning the focus to focal location **804c** in column **809a**. To complete excitation in both the central column and column **809b**, the electronic beam-steering can continue by exciting focal location **804d** in column **809b**, then focal location **804e** in central column **809a**, then focal location **804f** in column **809b**. The steering pattern can then initiate cavitation in a new column (e.g., **809c**), and continue the pattern of exciting a focal location in an outer or peripheral column followed by exciting a focal location in the central column. It is noted that in this specific implementation, the order of excitation/steering within the columns remains consistent. For example, the order of excitation of focal locations within the central column can start with the natural focus **804a**, move to an outer or peripheral column, return to focal location **804c**, move to the outer or peripheral portion, and then return to the central focal location **804e** within the central column. Similarly, the pattern within outer or peripheral columns can also follow a specific sequence, e.g., starting with a central focal location within a given outer or peripheral column (e.g., focal location **804b** in column **809b**), returning to the central column, steering back to the lower or distal focal location in the outer or peripheral column (e.g., focal location **804d** in column **809b**), returning to the central column, and then moving to the top or proximal focal location in that column (e.g., focal location **804f** in column **809b**). In some embodiments, steering can be performed from the most distal point to the most proximal point in any series of three focal locations before returning to the most distal end of the pattern to begin a new series of three points.

(159) In some aspects, the specific order of excitation within a column can be changed from as described above, however the overall concept of alternating between the central column and a peripheral column with each subsequent therapy pulse sequence delivery can be maintained. It is also desirable to re-excite each focal location within the central column once before repeating the process. For example, focal locations **804a**, **804c**, and **804e** should all be re-excited in some predetermined order, instead of re-exciting focal location **804a** twice before re-exciting focal location **804c** once.

(160) Generally, it is desirable to alternate between a single peripheral or outer column and the central column, until all points in both columns have been excited/re-excited, before moving on to a different peripheral column/central column pair. However, it should be understood that other embodiments can include electronically steering between columns without exciting/re-exciting all focal locations in each peripheral column.

(161) FIG. 8B shows the general pathway between focal locations of all the columns. the result is a 3D steering pattern that includes electronic steering between the central column and peripheral columns, with control logic that requires re-excitation of central column focal locations to assist in maintaining cavitation throughout the enhanced excitation volume.

(162) The rapid electronic beam-steering concepts described and shown herein serve to advantageously account for residual cavitation nuclei to enhance treatment of a discrete target location by creating an enhanced excitation volume with overlapping or adjacent focal zones.

(163) In the example of FIG. 5B in which the rapid steering is in the z-direction towards the transducer for each subsequent focal location, residual cavitation nuclei tend to reside in the prior

focal location. For example, in FIG. 5B, after the first pulse is delivered to the first focal location to form a focal zone **504a**, the bubble cloud rapidly expands and collapses, and residual cavitation nuclei remain in the vicinity of the first focal zone. Instead of transmitting another pulse to the first focal location, the system instead rapidly steers to the second focal location and delivers another pulse to that location, forming another and overlapping focal zone **504b** that is substantially free from the interference or negative effects of the residual cavitation nuclei remaining at the first focal zone **504a**. It should be noted that this benefit is achieved because the focus of the transducer is steered in the z-direction towards the therapy transducer, as opposed to being steered away from the therapy transducer. This allows the residual cavitation nuclei to remain distal to the current bubble cloud. If instead, the focus were steered to a new focal location away from the transducer (e.g., distal to earlier bubble clouds), then subsequent pulses would have to travel through residual cavitation nuclei, which may cause issues with bubble cloud formation and/or dispersion or uniformity of the subsequent bubble clouds. While the second focal zone **504b** is free from residual cavitation nuclei, the nuclei that remain in the vicinity of the first focal zone **504a** can be re-excited by the pulse delivered to the second focal location effectively enlarging the bubble cloud at that focal zone (e.g., forming cavitation at the first focal zone **504a** and the second focal zone **504b** with pulse(s) delivered to the second focal location). Similarly, the transducer can then be rapidly steered to the third focal location, and another pulse can be delivered to generate cavitation at the third focal zone **504c**. During this first pass, the third focal zone **504c** is relatively free of residual cavitation nuclei, maximizing the efficiency of pulses delivered to the third focal location. Additionally, residual cavitation nuclei in the first focal zone **504a** and the second focal zone **504b** can also be re-excited by the pulse(s) delivered to the third focal location, effectively increasing the size of the enhanced excitation volume. During this first pass, the third focal zone **504c** is free from remnant nuclei. Given the desire to rapidly steer and re-excite remnant nuclei, the choice of electronic steering direction minimizes the number of times that the system must steer distal to a point at which a cloud was generated immediately prior (thus minimizing the effects of shielding to the best extent possible, although not eliminating them). As described, the position of the focal locations/zones and the steered direction and pathway to the next and/or adjacent focal location can be varied based on desired effects (e.g., more advantageous to re-initiate versus potential to act as a blocker) and may contemplate how therapy versus bubble manipulation are phased in context to one another.

(164) The time to return to a given focal location for additional therapy pulses can also be adjusted or controlled for to account for residual cavitation nuclei. In one embodiment, the following equation can be used to determine the timing of steering the beam between focal locations and/or returning to previously treated focal locations:

time to return to given focal location = number of unique focal locations / PRF

(165) As described in the equation, for some embodiments and focal locations, the time to return to a given focal location can equal the number of unique focal locations (e.g., three focal locations) divided by the pulse repetition frequency (PRF) of the pulse.

(166) As described above, a therapy pulse of a given pulse sequence delivered to each focal location can be a shocked scattering pulse that can include a peak positive half cycle, followed by a peak negative half cycle, followed by a trailing peak positive half cycle. This pulse is generally responsible for forming cavitation at each focal location.

(167) In some embodiments, the specific pulse can include one or more low amplitude pulses following the therapy pulses that are designed and configured to manipulate, push, interact with, or suppress residual cavitation nuclei. In one embodiment, these additional pulses can comprise one, two, or more low amplitude pulses. These low amplitude pulses can have an amplitude that is substantially smaller than the amplitude of the therapy pulse(s). In some embodiments, the low amplitude pulses are designed and configured to push or flush out residual cavitation nuclei from previously treated focal locations. In other embodiments, the low amplitude pulses are designed

and configured to spatially modulate cavitation nuclei within the previously treated focal location. For example, referring to FIG. 5B, the focus of the transducer can be positioned on first focal location **504a**, and a shocked scattering pulse can be delivered to the first focal location to create a bubble cloud on the first focal location, as previously described. Immediately following the shocked scattering pulse, the transducer array can deliver one or more low amplitude pulses into the focal location to spatially modulate cavitation nuclei within the first focal location. Next, the transducer array can be rapidly steered to the second focal location, and treatment can continue as previously described. Simply put, the pulse can be modified to include a plurality of low amplitude pulses immediately following the therapy or shocked scattering pulses so as to advantageously spatially modulate or manipulate residual cavitation nuclei in a favorable way.

(168) FIG. 9A illustrates an axial beam profile for axially spaced focal locations illustrated in FIGS. 9B, 9C, and 9D. These focal locations can correspond to focal zones **504a**, **504b**, and **504c** and focal locations described above in reference to FIG. 5B. In this embodiment, the first focal location, shown in FIG. 9B, corresponds to the natural focus of the therapy transducer. FIG. 9C depicts the second focal location steered 2.5 mm in the z-direction towards the transducer from the first focal location. FIG. 9D depicts the third focal location steered 2.5 mm in the z-direction towards the transducer from the second focal location.

(169) FIG. 10A illustrates an axial beam profile for laterally spaced focal locations illustrated in FIGS. 10B, and 10C. These focal locations can correspond to focal zones **504a** and **504b** described above in reference to FIG. 5C. In this embodiment, the first focal location, shown in FIG. 10B, corresponds to a location steered 0.5 mm in the y direction from the natural focus of the therapy transducer. FIG. 10C depicts the second focal location steered 0.5 mm in the opposite direction from the natural focus of the therapy transducer (e.g., -0.5 mm from the natural focus).

(170) FIGS. 11A and 11B illustrate how the spacing between focal locations can affect overlap between adjacent focal zones and determine whether or not residual cavitation nuclei are re-excited by subsequent therapy pulses. In the graph of FIG. 11A, spacing between the three focal locations comprises 7 mm in the axial direction between each adjacent focal location, resulting in a first focal location at the natural focus of the transducer, a second focal location 7 mm axially spaced from the first focal location, and a third focal location 7 mm axially spaced from the second focal location. In this example, the spacing results in focal zones of insufficient overlap between adjacent beam profiles, causing the overall treatment at that treatment location (i.e., within the enhanced excitation volume) to be non-uniform. In contrast, the example of FIG. 9A, with 2.5 mm spacing between adjacent focal locations, results in sufficient overlap between adjacent focal zones to facilitate re-excitation of residual cavitation nuclei and more uniform cavitation distributions in the treatment location.

(171) FIGS. 12A, 12B, and 12C illustrate heat maps for histotripsy therapy provided to 1) a single focal location at a treatment location with 1000 pulses delivered (FIG. 12A), 2) three axially spaced focal locations at a treatment location with a total of 1000 pulses delivered amongst them (FIG. 12B), and 3) a single focal location at a treatment location with 580 pulses delivered (FIG. 12C). FIGS. 12D, 12E, and 12F illustrate exposure levels for histotripsy therapy provided to 1) a single focal location at a treatment location with 1000 pulses delivered (FIG. 12D), 2) three axially spaced focal locations at a treatment location with a total of 1000 pulses delivered amongst them (FIG. 12E), and 3) a single focal location at a treatment location with 580 pulses delivered (FIG. 12F). FIGS. 12G, 12H, and 12I illustrate lesion and exposure levels for histotripsy therapy provided to 1) a single focal location at a treatment location with 1000 pulses delivered (FIG. 12G), 2) three axially spaced focal locations at a treatment location with a total of 1000 pulses delivered amongst them (FIG. 12H), and 3) a single focal location at a treatment location with 580 pulses delivered (FIG. 12I). Compared to treatment at a single focal location, the treatment that includes three axially spaced focal locations (FIGS. 12B, 12E, 12H) distributes a sufficiently destructive # of exposures over a greater area for same number of pulses. In one experiment, treatment of a 3 cm

sphere of tissue with a plurality of discrete treatment locations, wherein each treatment location is treated with a single focal location (e.g., the treatment protocol of FIGS. 12A, 12D, 12G) resulted in “therapy on-times” of 21 minutes and 26 seconds. In contrast, treatment of a 3 cm sphere of tissue with a plurality of discrete treatment locations, wherein each treatment location is treated with three axially spaced focal locations (e.g., the treatment protocol of FIGS. 12B, 12E, 12H) resulted in “therapy on-times” of only 12 minutes and 21 seconds, a major reduction in therapy on-times.

(172) The embodiment of FIGS. 12C, 12F, and 12I shows what treatment would look like at a single focal location for each discrete treatment location of a 3 cm sphere of tissue with a therapy on-time matching that of the embodiment of FIGS. 12B, 12E, and 12H (e.g., 12 minutes and 21 seconds of on-time). As can be seen, the heat map, exposure level, and lesion generation lag far behind the other two embodiments. This experiment shows how rapid steering of the transducer between multiple focal locations for a given treatment location can improve the quality of therapy and rapidly reduce the amount of on-time (e.g., dosage of ultrasound energy) delivered for a given treatment. Rapid steering between multiple focal locations for a given discrete treatment location uses the same number of pulses more efficiently than single focal location treatment by distributing a sufficiently destructive number of exposures over a larger volume.

(173) FIG. 13 is a schematic drawing showing the spatial relationship between a therapy transducer 1300, a first focal location 1304a, a second focal location 1304b, and a third focal location 1304c. As described above, in some embodiments, the first focal location can comprise the natural focus of the transducer, and the second and third focal locations can be achieved with electronic steering of the transducer array. In this example, a distance dz separates adjacent focal locations.

(174) FIG. 14 illustrates the relationship between bubble cloud size (area) and the distance between adjacent focal locations. As shown in FIG. 14, in one embodiment, 2.5 mm spacing between focal locations also produces the largest cloud per pulse, highlighting how previously excited zones can be re-excited by targeting adjacent zones, enlarging the cloud on a per pulse basis.

(175) Therapy Components

(176) The Therapy sub-system may work with other sub-systems to create, optimize, deliver, visualize, monitor and control acoustic cavitation, also referred to herein and in following as “histotripsy”, and its derivatives of, including boiling histotripsy and other thermal high frequency ultrasound approaches. It is noted that the disclosed inventions may also further benefit other acoustic therapies that do not comprise a cavitation, mechanical or histotripsy component. The therapy sub-system can include, among other features, an ultrasound therapy transducer and a pulse generator system configured to deliver ultrasound pulses into tissue.

(177) In order to create and deliver histotripsy and derivatives of histotripsy, the therapy sub-system may also comprise components, including but not limited to, one or more function generators, amplifiers, therapy transducers and power supplies.

(178) The therapy transducer can comprise a single element or multiple elements configured to be excited with high amplitude electric pulses (>1000V or any other voltage that can cause harm to living organisms). The amplitude necessary to drive the therapy transducers for Histotripsy vary depending on the design of the transducer and the materials used (e.g., solid or polymer/piezoelectric composite including ceramic or single crystal) and the transducer center frequency which is directly proportional to the thickness of the piezo-electric material. Transducers therefore operating at a high frequency require lower voltage to produce a given surface pressure than is required by low frequency therapy transducers. In some embodiments, the transducer elements are formed using a piezoelectric-polymer composite material or a solid piezoelectric material. Further, the piezoelectric material can be of polycrystalline/ceramic or single crystalline formulation. In some embodiments the transducer elements can be formed using silicon using MEMs technology, including CMUT and PMUT designs.

(179) In some embodiments, the function generator may comprise a field programmable gate array



(FPGA) or other suitable function generator. The FPGA may be configured with parameters disclosed previously herein, including but not limited to frequency, pulse repetition frequency, bursts, burst numbers, where bursts may comprise pulses, numbers of pulses, length of pulses, pulse period, delays, burst repetition frequency or period, where sets of bursts may comprise a parameter set, where loop sets may comprise various parameter sets, with or without delays, or varied delays, where multiple loop sets may be repeated and/or new loop sets introduced, of varied time delay and independently controlled, and of various combinations and permutations of such, overall and throughout.

(180) In some embodiments, the generator or amplifier may be configured to be a universal single-cycle or multi-cycle pulse generator, and to support driving via Class D or inductive driving, as well as across all envisioned clinical applications, use environments, also discussed in part later in this disclosure. In other embodiments, the class D or inductive current driver may be configured to comprise transformer and/or auto-transformer driving circuits to further provide step up/down components, and in some cases, to preferably allow a step up in the amplitude. They may also comprise specific protective features, to further support the system, and provide capability to protect other parts of the system (e.g., therapy transducer and/or amplifier circuit components) and/or the user, from various hazards, including but not limited to, electrical safety hazards, which may potentially lead to use environment, system and therapy system, and user harms, damage or issues.

(181) Disclosed generators may allow and support the ability of the system to select, vary and control various parameters (through enabled software tools), including, but not limited to those previously disclosed, as well as the ability to start/stop therapy, set and read voltage level, pulse and/or burst repetition frequency, number of cycles, duty ratio, channel enabled and delay, etc., modulate pulse amplitude on a fast time-scale independent of a high voltage supply, and/or other service, diagnostic or treatment features.

(182) In some embodiments, the Therapy sub-system and/or components of, such as the amplifier, may comprise further integrated computer processing capability and may be networked, connected, accessed, and/or be removable/portable, modular, and/or exchangeable between systems, and/or driven/commanded from/by other systems, or in various combinations. Other systems may include other acoustic cavitation/histotripsy, HIFU, HITU, radiation therapy, radiofrequency, microwave, and cryoablation systems, navigation and localization systems, open surgical, laparoscopic, single incision/single port, endoscopic and non-invasive surgical robots, laparoscopic or surgical towers comprising other energy-based or vision systems, surgical system racks or booms, imaging carts, etc.

(183) In some embodiments, one or more amplifiers may comprise a Class D amplifier and related drive circuitry including matching network components. Depending on the transducer element electric impedance and choice of the matching network components (e.g., an LC circuit made of an inductor L1 in series and the capacitor C1 in parallel), the combined impedance can be aggressively set low in order to have high amplitude electric waveform necessary to drive the transducer element. The maximum amplitude that Class D amplifiers is dependent on the circuit components used, including the driving MOSFET/IGBT transistors, matching network components or inductor, and transformer or autotransformer, and of which may be typically in the low kV (e.g., 1-3 kV) range.

(184) Therapy transducer element(s) are excited with an electrical waveform with an amplitude (voltage) to produce a pressure output sufficient for Histotripsy therapy. The excitation electric field can be defined as the necessary waveform voltage per thickness of the piezoelectric element. For example, because a piezoelectric element operating at 1 MHz transducer is half the thickness of an equivalent 500 kHz element, it will require half the voltage to achieve the same electric field and surface pressure.

(185) The Therapy sub-system may also comprise therapy transducers of various designs and

working parameters, supporting use in various procedures (and procedure settings). Systems may be configured with one or more therapy transducers, that may be further interchangeable, and work with various aspects of the system in similar or different ways (e.g., may interface to a robotic arm using a common interface and exchange feature, or conversely, may adapt to work differently with application specific imaging probes, where different imaging probes may interface and integrate with a therapy transducer in specifically different ways).

(186) Therapy transducers may be configured of various parameters that may include size, shape (e.g., rectangular or round; anatomically curved housings, etc.), geometry, focal length, number of elements, size of elements, distribution of elements (e.g., number of rings, size of rings for annular patterned transducers), frequency, enabling electronic beam steering, etc. Transducers may be composed of various materials (e.g., piezoelectric, silicon, etc.), form factors and types (e.g., machined elements, chip-based, etc.) and/or by various methods of fabrication of.

(187) Transducers may be designed and optimized for clinical applications (e.g., abdominal tumors, peripheral vascular disease, fat ablation, etc.) and desired outcomes (e.g., acoustic cavitation/histotripsy without thermal injury to intervening tissue), and affording a breadth of working ranges, including relatively shallow and superficial targets (e.g., thyroid or breast nodules), versus, deeper or harder to reach targets, such as central liver or brain tumors. They may be configured to enable acoustic cavitation/histotripsy under various parameters and sets of, as enabled by the aforementioned system components (e.g., function generator and amplifier, etc.), including but not limited to frequency, pulse repetition rate, pulses, number of pulses, pulse length, pulse period, delays, repetitions, sync delays, sync period, sync pulses, sync pulse delays, various loop sets, others, and permutations of. The transducer may also be designed to allow for the activation of a drug payload either deposited in tissue through various means including injection, placement or delivery in micelle or nanostructures.

(188) Integrated Imaging

(189) The disclosed system may comprise various imaging modalities to allow users to visualize, monitor and collect/use feedback of the patient's anatomy, related regions of interest and treatment/procedure sites, as well as surrounding and intervening tissues to assess, plan and conduct procedures, and adjust treatment parameters as needed. Imaging modalities may comprise various ultrasound, x-ray, CT, MRI, PET, fluoroscopy, optical, contrast or agent enhanced versions, and/or various combinations of. It is further disclosed that various image processing and characterization technologies may also be utilized to afford enhanced visualization and user decision making. These may be selected or commanded manually by the user or in an automated fashion by the system. The system may be configured to allow side by side, toggling, overlays, 3D reconstruction, segmentation, registration, multi-modal image fusion, image flow, and/or any methodology affording the user to identify, define and inform various aspects of using imaging during the procedure, as displayed in the various system user interfaces and displays. Examples may include locating, displaying and characterizing regions of interest, organ systems, potential treatment sites within, with on and/or surrounding organs or tissues, identifying critical structures such as ducts, vessels, nerves, ureters, fissures, capsules, tumors, tissue trauma/injury/disease, other organs, connective tissues, etc., and/or in context to one another, of one or more (e.g., tumor draining lymphatics or vasculature; or tumor proximity to organ capsule or underlying other organ), as unlimited examples.

(190) Systems may be configured to include onboard integrated imaging hardware, software, sensors, probes and wetware, and/or may be configured to communicate and interface with external imaging and image processing systems. The aforementioned components may be also integrated into the system's Therapy sub-system components wherein probes, imaging arrays, or the like, and electrically, mechanically or electromechanically integrated into therapy transducers. This may afford, in part, the ability to have geometrically aligned imaging and therapy, with the therapy directly within the field of view, and in some cases in line, with imaging. In some embodiments,

this integration may comprise a fixed orientation of the imaging capability (e.g., imaging probe) in context to the therapy transducer. In other embodiments, the imaging solution may be able to move or adjust its position, including modifying angle, extension (e.g., distance from therapy transducer or patient), rotation (e.g., imaging plane in example of an ultrasound probe) and/or other parameters, including moving/adjusting dynamically while actively imaging. The imaging component or probe may be encoded so its orientation and position relative to another aspect of the system, such as the therapy transducer, and/or robotically-enabled positioning component may be determined.

(191) In one embodiment, the system may comprise onboard ultrasound, further configured to allow users to visualize, monitor and receive feedback for procedure sites through the system displays and software, including allowing ultrasound imaging and characterization (and various forms of), ultrasound guided planning and ultrasound guided treatment, all in real-time. The system may be configured to allow users to manually, semi-automated or in fully automated means image the patient (e.g., by hand or using a robotically-enabled imager).

(192) In some embodiments, imaging feedback and monitoring can include monitoring changes in: backscatter from bubble clouds; speckle reduction in backscatter; backscatter speckle statistics; mechanical properties of tissue (i.e., elastography); tissue perfusion (i.e., ultrasound contrast); shear wave propagation; acoustic emissions, electrical impedance tomography, and/or various combinations of, including as displayed or integrated with other forms of imaging (e.g., CT or MRI).

(193) In some embodiments, imaging including feedback and monitoring from backscatter from bubble clouds, may be used as a method to determine immediately if the histotripsy process has been initiated, is being properly maintained, or even if it has been extinguished. For example, this method enables continuously monitored in real time drug delivery, tissue erosion, and the like. The method also can provide feedback permitting the histotripsy process to be initiated at a higher intensity and maintained at a much lower intensity. For example, backscatter feedback can be monitored by any transducer or ultrasonic imager. By measuring feedback for the therapy transducer, an accessory transducer can send out interrogation pulses or be configured to passively detect cavitation. Moreover, the nature of the feedback received can be used to adjust acoustic parameters (and associated system parameters) to optimize the drug delivery and/or tissue erosion process.

(194) In some embodiments, imaging including feedback and monitoring from backscatter, and speckle reduction, may be configured in the system.

(195) For systems comprising feedback and monitoring via backscattering, and as means of background, as tissue is progressively mechanically subdivided, in other words homogenized, disrupted, or eroded tissue, this process results in changes in the size and distribution of acoustic scatter. At some point in the process, the scattering particle size and density is reduced to levels where little ultrasound is scattered, or the amount scattered is reduced significantly. This results in a significant reduction in speckle, which is the coherent constructive and destructive interference patterns of light and dark spots seen on images when coherent sources of illumination are used; in this case, ultrasound. After some treatment time, the speckle reduction results in a dark area in the therapy volume. Since the amount of speckle reduction is related to the amount of tissue subdivision, it can be related to the size of the remaining tissue fragments. When this size is reduced to sub-cellular levels, no cells are assumed to have survived. So, treatment can proceed until a desired speckle reduction level has been reached. Speckle is easily seen and evaluated on standard ultrasound imaging systems. Specialized transducers and systems, including those disclosed herein, may also be used to evaluate the backscatter changes.

(196) Further, systems comprising feedback and monitoring via speckle, and as means of background, an image may persist from frame to frame and change very little as long as the scatter distribution does not change and there is no movement of the imaged object. However, long before

the scatters are reduced enough in size to cause speckle reduction, they may be changed sufficiently to be detected by signal processing and other means. This family of techniques can operate as detectors of speckle statistics changes. For example, the size and position of one or more speckles in an image will begin to decorrelate before observable speckle reduction occurs. Speckle decorrelation, after appropriate motion compensation, can be a sensitive measure of the mechanical disruption of the tissues, and thus a measure of therapeutic efficacy. This feedback and monitoring technique may permit early observation of changes resulting from the acoustic cavitation/histotripsy process and can identify changes in tissue before substantial or complete tissue effect (e.g., erosion occurs). In one embodiment, this method may be used to monitor the acoustic cavitation/histotripsy process for enhanced drug delivery where treatment sites/tissue is temporally disrupted, and tissue damage/erosion is not desired. In other embodiments, this may comprise speckle decorrelation by movement of scatters in an increasingly fluidized therapy volume. For example, in the case where partial or complete tissue erosion is desired.

(197) For systems comprising feedback and monitoring via elastography, and as means of background, as treatment sites/tissue are further subdivided per an acoustic cavitation/histotripsy effect (homogenized, disrupted, or eroded), its mechanical properties change from a soft but interconnected solid to a viscous fluid or paste with few long-range interactions. These changes in mechanical properties can be measured by various imaging modalities including MRI and ultrasound imaging systems. For example, an ultrasound pulse can be used to produce a force (i.e., a radiation force) on a localized volume of tissue. The tissue response (displacements, strains, and velocities) can change significantly during histotripsy treatment allowing the state of tissue disruption to be determined by imaging or other quantitative means.

(198) Systems may also comprise feedback and monitoring via shear wave propagation changes. As means of background, the subdivision of tissues makes the tissue more fluid and less solid and fluid systems generally do not propagate shear waves. Thus, the extent of tissue fluidization provides opportunities for feedback and monitoring of the histotripsy process. For example, ultrasound and MRI imaging systems can be used to observe the propagation of shear waves. The extinction of such waves in a treated volume is used as a measure of tissue destruction or disruption. In one system embodiment, the system and supporting sub-systems may be used to generate and measure the interacting shear waves. For example, two adjacent ultrasound foci might perturb tissue by pushing it in certain ways. If adjacent foci are in a fluid, no shear waves propagate to interact with each other. If the tissue is not fluidized, the interaction would be detected with external means, for example, by a difference frequency only detected when two shear waves interact nonlinearly, with their disappearance correlated to tissue damage. As such, the system may be configured to use this modality to enhance feedback and monitoring of the acoustic cavitation/histotripsy procedure.

(199) For systems comprising feedback and monitoring via acoustic emission, and as means of background, as a tissue volume is subdivided, its effect on acoustic cavitation/histotripsy (e.g., the bubble cloud here) is changed. For example, bubbles may grow larger and have a different lifetime and collapse changing characteristics in intact versus fluidized tissue. Bubbles may also move and interact after tissue is subdivided producing larger bubbles or cooperative interaction among bubbles, all of which can result in changes in acoustic emission. These emissions can be heard during treatment and they change during treatment. Analysis of these changes, and their correlation to therapeutic efficacy, enables monitoring of the progress of therapy, and may be configured as a feature of the system.

(200) For systems comprising feedback and monitoring via electrical impedance tomography, and as means of background, an impedance map of a therapy site can be produced based upon the spatial electrical characteristics throughout the therapy site. Imaging of the conductivity or permittivity of the therapy site of a patient can be inferred from taking skin surface electrical measurements. Conducting electrodes are attached to a patient's skin and small alternating currents

are applied to some or all of the electrodes. One or more known currents are injected into the surface and the voltage is measured at a number of points using the electrodes. The process can be repeated for different configurations of applied current. The resolution of the resultant image can be adjusted by changing the number of electrodes employed. A measure of the electrical properties of the therapy site within the skin surface can be obtained from the impedance map, and changes in and location of the acoustic cavitation/histotripsy (e.g., bubble cloud, specifically) and histotripsy process can be monitored using this as configured in the system and supporting sub-systems.

(201) The user may be allowed to further select, annotate, mark, highlight, and/or contour, various regions of interest or treatment sites, and defined treatment targets (on the image(s)), of which may be used to command and direct the system where to image, test and/or treat, through the system software and user interfaces and displays. In some arrangements, the user may use a manual ultrasound probe (e.g., diagnostic hand-held probe) to conduct the procedure. In another arrangement, the system may use a robot and/or electromechanical positioning system to conduct the procedure, as directed and/or automated by the system, or conversely, the system can enable combinations of manual and automated uses.

(202) The system may further include the ability to conduct image registration, including imaging and image data set registration to allow navigation and localization of the system to the patient, including the treatment site (e.g., tumor, critical structure, bony anatomy, anatomy and identifying features of, etc.). In one embodiment, the system allows the user to image and identify a region of interest, for example the liver, using integrated ultrasound, and to select and mark a tumor (or surrogate marker of) comprised within the liver through/displayed in the system software, and wherein said system registers the image data to a coordinate system defined by the system, that further allows the system's Therapy and Robotics sub-systems to deliver synchronized acoustic cavitation/histotripsy to said marked tumor. The system may comprise the ability to register various image sets, including those previously disclosed, to one another, as well as to afford navigation and localization (e.g., of a therapy transducer to a CT or MRI/ultrasound fusion image with the therapy transducer and Robotics sub-system tracking to said image).

(203) The system may also comprise the ability to work in a variety of interventional, endoscopic and surgical environments, including alone and with other systems (surgical/laparoscopic towers, vision systems, endoscope systems and towers, ultrasound enabled endoscopic ultrasound (flexible and rigid), percutaneous/endoscopic/laparoscopic and minimally invasive navigation systems (e.g., optical, electromagnetic, shape-sensing, ultrasound-enabled, etc.), of also which may work with, or comprise various optical imaging capabilities (e.g., fiber and or digital). The disclosed system may be configured to work with these systems, in some embodiments working alongside them in concert, or in other embodiments where all or some of the system may be integrated into the above systems/platforms (e.g., acoustic cavitation/histotripsy-enabled endoscope system or laparoscopic surgical robot). In many of these environments, a therapy transducer may be utilized at or around the time of use, for example, of an optically guided endoscope/bronchoscope, or as another example, at the time a laparoscopic robot (e.g., Intuitive Da Vinci\*Xi system) is viewing/manipulating a tissue/treatment site. Further, these embodiments and examples may include where said other systems/platforms are used to deliver (locally) fluid to enable the creation of a man-made acoustic window, where on under normal circumstances may not exist (e.g., fluidizing a segment or lobe of the lung in preparation for acoustic cavitation/histotripsy via non-invasive transthoracic treatment (e.g., transducer externally placed on/around patient). Systems disclosed herein may also comprise all or some of their sub-system hardware packaged within the other system cart/console/systems described here (e.g., acoustic cavitation/histotripsy system and/or sub-systems integrated and operated from said navigation or laparoscopic system).

(204) The system may also be configured, through various aforementioned parameters and other parameters, to display real-time visualization of a bubble cloud in a spatial-temporal manner, including the resulting tissue effect peri/post-treatment from tissue/bubble cloud interaction,

wherein the system can dynamically image and visualize, and display, the bubble cloud, and any changes to it (e.g., decreasing or increasing echogenicity), which may include intensity, shape, size, location, morphology, persistence, etc. These features may allow users to continuously track and follow the treatment in real-time in one integrated procedure and interface/system, and confirm treatment safety and efficacy on the fly (versus other interventional or surgical modalities, which either require multiple procedures to achieve the same, or where the treatment effect is not visible in real-time (e.g., radiation therapy), or where it is not possible to achieve such (e.g., real-time visualization of local tissue during thermal ablation), and/or where the other procedure further require invasive approaches (e.g., incisions or punctures) and iterative imaging in a scanner between procedure steps (e.g., CT or MRI scanning). The above disclosed systems, sub-systems, components, modalities, features and work-flows/methods of use may be implemented in an unlimited fashion through enabling hardware, software, user interfaces and use environments, and future improvements, enhancements and inventions in this area are considered as included in the scope of this disclosure, as well as any of the resulting data and means of using said data for analytics, artificial intelligence or digital health applications and systems.

(205) Robotics

(206) They system may comprise various Robotic sub-systems and components, including but not limited to, one or more robotic arms and controllers, which may further work with other sub-systems or components of the system to deliver and monitor acoustic cavitation/histotripsy. As previously discussed herein, robotic arms and control systems may be integrated into one or more Cart configurations.

(207) For example, one system embodiment may comprise a Cart with an integrated robotic arm and control system, and Therapy, Integrated Imaging and Software, where the robotic arm and other listed sub-systems are controlled by the user through the form factor of a single bedside Cart.

(208) In other embodiments, the Robotic sub-system may be configured in one or more separate Carts, that may be a driven in a master/slave configuration from a separate master or Cart, wherein the robotically-enabled Cart is positioned bed/patient-side, and the Master is at a distance from said Cart.

(209) Disclosed robotic arms may be comprised of a plurality of joints, segments, and degrees of freedom and may also include various integrated sensor types and encoders, implemented for various use and safety features. Sensing technologies and data may comprise, as an example, vision, potentiometers, position/localization, kinematics, force, torque, speed, acceleration, dynamic loading, and/or others. In some cases, sensors may be used for users to direct robot commands (e.g., hand gesture the robot into a preferred set up position, or to dock home). Additional details on robotic arms can be found in US Patent Pub. No. 2013/0255426 to Kassow et al. which is disclosed herein by reference in its entirety.

(210) The robotic arm receives control signals and commands from the robotic control system, which may be housed in a Cart. The system may be configured to provide various functionalities, including but not limited to, position, tracking, patterns, triggering, and events/actions.

(211) Position may be configured to comprise fixed positions, pallet positions, time-controlled positions, distance-controlled positions, variable-time controlled positions, variable-distance controlled positions.

(212) Tracking may be configured to comprise time-controlled tracking and/or distance-controlled tracking.

(213) The patterns of movement may be configured to comprise intermediate positions or waypoints, as well as sequence of positions, through a defined path in space.

(214) Triggers may be configured to comprise distance measuring means, time, and/or various sensor means including those disclosed herein, and not limited to, visual/imaging-based, force, torque, localization, energy/power feedback and/or others.

(215) Events/actions may be configured to comprise various examples, including proximity-based

(approaching/departing a target object), activation or de-activation of various end-effectors (e.g., therapy transducers), starting/stopping/pausing sequences of said events, triggering or switching between triggers of events/actions, initiating patterns of movement and changing/toggling between patterns of movement, and/or time-based and temporal over the defined work and time-space.

(216) In one embodiment, the system comprises a three degree of freedom robotic positioning system, enabled to allow the user (through the software of the system and related user interfaces), to micro-position a therapy transducer through X, Y, and Z coordinate system, and where gross macro-positioning of the transducer (e.g., aligning the transducer on the patient's body) is completed manually. In some embodiments, the robot may comprise 6 degrees of freedom including X, Y, Z, and pitch, roll and yaw. In other embodiments, the Robotic sub-system may comprise further degrees of freedom, that allow the robot arm supporting base to be positioned along a linear axis running parallel to the general direction of the patient surface, and/or the supporting base height to be adjusted up or down, allowing the position of the robotic arm to be modified relative to the patient, patient surface, Cart, Coupling sub-system, additional robots/robotic arms and/or additional surgical systems, including but not limited to, surgical towers, imaging systems, endoscopic/laparoscopic systems, and/or other.

(217) One or more robotic arms may also comprise various features to assist in maneuvering and modifying the arm position, manually or semi-manually, and of which said features may interface on or between the therapy transducer and the most distal joint of the robotic arm. In some embodiments, the feature is configured to comprise a handle allowing maneuvering and manual control with one or more hands. The handle may also be configured to include user input and electronic control features of the robotic arm, to command various drive capabilities or modes, to actuate the robot to assist in gross or fine positioning of the arm (e.g., activating or deactivating free drive mode). The work-flow for the initial positioning of the robotic arm and therapy head can be configured to allow either first positioning the therapy transducer/head in the coupling solution, with the therapy transducer directly interfaced to the arm, or in a different work-flow, allowing the user to set up the coupling solution first, and enabling the robot arm to be interfaced to the therapy transducer/coupling solution as a later/terminal set up step.

(218) In some embodiments, the robotic arm may comprise a robotic arm on a laparoscopic, single port, endoscopic, hybrid or combination of, and/or other robot, wherein said robot of the system may be a slave to a master that controls said arm, as well as potentially a plurality of other arms, equipped to concurrently execute other tasks (vision, imaging, grasping, cutting, ligating, sealing, closing, stapling, ablating, suturing, marking, etc.), including actuating one or more laparoscopic arms (and instruments) and various histotripsy system components. For example, a laparoscopic robot may be utilized to prepare the surgical site, including manipulating organ position to provide more ideal acoustic access and further stabilizing said organ in some cases to minimize respiratory motion. In conjunction and parallel to this, a second robotic arm may be used to deliver non-invasive acoustic cavitation through a body cavity, as observed under real-time imaging from the therapy transducer (e.g., ultrasound) and with concurrent visualization via a laparoscopic camera. In other related aspects, a similar approach may be utilized with a combination of an endoscopic and non-invasive approach, and further, with a combination of an endoscopic, laparoscopic and non-invasive approach.

(219) Software

(220) The system may comprise various software applications, features and components which allow the user to interact, control and use the system for a plethora of clinical applications. The Software may communicate and work with one or more of the sub-systems, including but not limited to Therapy, Integrated Imaging, Robotics and Other Components, Ancillaries and Accessories of the system.

(221) Overall, in no specific order of importance, the software may provide features and support to initialize and set up the system, service the system, communicate and import/export/store data,

modify/manipulate/configure/control/command various settings and parameters by the user, mitigate safety and use-related risks, plan procedures, provide support to various configurations of transducers, robotic arms and drive systems, function generators and amplifier circuits/slaves, test and treatment ultrasound sequences, transducer steering and positioning (electromechanical and electronic beam steering, etc.), treatment patterns, support for imaging and imaging probes, manual and electromechanical/robotically-enabling movement of, imaging support for measuring/characterizing various dimensions within or around procedure and treatment sites (e.g., depth from one anatomical location to another, etc., pre-treatment assessments and protocols for measuring/characterizing in situ treatment site properties and conditions (e.g., acoustic cavitation/histotripsy thresholds and heterogeneity of), targeting and target alignment, calibration, marking/annotating, localizing/navigating, registering, guiding, providing and guiding through work-flows, procedure steps, executing treatment plans and protocols autonomously, autonomously and while under direct observation and viewing with real-time imaging as displayed through the software, including various views and viewports for viewing, communication tools (video, audio, sharing, etc.), troubleshooting, providing directions, warnings, alerts, and/or allowing communication through various networking devices and protocols. It is further envisioned that the software user interfaces and supporting displays may comprise various buttons, commands, icons, graphics, text, etc., that allow the user to interact with the system in a user-friendly and effective manner, and these may be presented in an unlimited number of permutations, layouts and designs, and displayed in similar or different manners or feature sets for systems that may comprise more than one display (e.g., touch screen monitor and touch pad), and/or may network to one or more external displays or systems (e.g., another robot, navigation system, system tower, console, monitor, touch display, mobile device, tablet, etc.).

(222) The software, as a part of a representative system, including one or more computer processors, may support the various aforementioned function generators (e.g., FPGA), amplifiers, power supplies and therapy transducers. The software may be configured to allow users to select, determine and monitor various parameters and settings for acoustic cavitation/histotripsy, and upon observing/receiving feedback on performance and conditions, may allow the user to stop/start/modify said parameters and settings.

(223) The software may be configured to allow users to select from a list or menu of multiple transducers and support the auto-detection of said transducers upon connection to the system (and verification of the appropriate sequence and parameter settings based on selected application). In other embodiments, the software may update the targeting and amplifier settings (e.g., channels) based on the specific transducer selection. The software may also provide transducer recommendations based on pre-treatment and planning inputs. Conversely, the software may provide error messages or warnings to the user if said therapy transducer, amplifier and/or function generator selections or parameters are erroneous, yield a fault or failure. This may further comprise reporting the details and location of such.

(224) In addition to above, the software may be configured to allow users to select treatment sequences and protocols from a list or menu, and to store selected and/or previous selected sequences and protocols as associated with specific clinical uses or patient profiles. Related profiles may comprise any associated patient, procedure, clinical and/or engineering data, and maybe used to inform, modify and/or guide current or future treatments or procedures/interventions, whether as decision support or an active part of a procedure itself (e.g., using serial data sets to build and guide new treatments).

(225) As a part of planning or during the treatment, the software (and in working with other components of the system) may allow the user to evaluate and test acoustic cavitation/histotripsy thresholds at various locations in a user-selected region of interest or defined treatment area/volume, to determine the minimum cavitation thresholds throughout said region or area/volume, to ensure treatment parameters are optimized to achieve, maintain and dynamically



control acoustic cavitation/histotripsy. In one embodiment, the system allows a user to manually evaluate and test threshold parameters at various points. Said points may include those at defined boundary, interior to the boundary and center locations/positions, of the selected region of interest and treatment area/volume, and where resulting threshold measurements may be reported/displayed to the user, as well as utilized to update therapy parameters before treatment. In another embodiment, the system may be configured to allow automated threshold measurements and updates, as enabled by the aforementioned Robotics sub-system, wherein the user may direct the robot, or the robot may be commanded to execute the measurements autonomously.

(226) Software may also be configured, by working with computer processors and one or more function generators, amplifiers and therapy transducers, to allow various permutations of delivering and positioning optimized acoustic cavitation/histotripsy in and through a selected area/volume. This may include, but not limited to, systems configured with a fixed/natural focus arrangement using purely electromechanical positioning configuration(s), electronic beam steering (with or without electromechanical positioning), electronic beam steering to a new selected fixed focus with further electromechanical positioning, axial (Z axis) electronic beam steering with lateral (X and Y) electromechanical positioning, high speed axial electronic beam steering with lateral electromechanical positioning, high speed beam steering in 3D space, various combinations of including with dynamically varying one or more acoustic cavitation/histotripsy parameters based on the aforementioned ability to update treatment parameters based on threshold measurements (e.g., dynamically adjusting amplitude across the treatment area/volume).

#### OTHER COMPONENTS, ANCILLARIES AND ACCESSORIES

(227) The system may comprise various other components, ancillaries and accessories, including but not limited to computers, computer processors, power supplies including high voltage power supplies, controllers, cables, connectors, networking devices, software applications for security, communication, integration into information systems including hospital information systems, cellular communication devices and modems, handheld wired or wireless controllers, goggles or glasses for advanced visualization, augmented or virtual reality applications, cameras, sensors, tablets, smart devices, phones, internet of things enabling capabilities, specialized use “apps” or user training materials and applications (software or paper based), virtual proctors or trainers and/or other enabling features, devices, systems or applications, and/or methods of using the above.

#### (228) System Variations and Methods/Applications

(229) In addition to performing a breadth of procedures, the system may allow additional benefits, such as enhanced planning, imaging and guidance to assist the user. In one embodiment, the system may allow a user to create a patient, target and application specific treatment plan, wherein the system may be configured to optimize treatment parameters based on feedback to the system during planning, and where planning may further comprise the ability to run various test protocols to gather specific inputs to the system and plan.

(230) Feedback may include various energy, power, location, position, tissue and/or other parameters.

(231) The system, and the above feedback, may also be further configured and used to autonomously (and robotically) execute the delivery of the optimized treatment plan and protocol, as visualized under real-time imaging during the procedure, allowing the user to directly observe the local treatment tissue effect, as it progresses through treatment, and start/stop/modify treatment at their discretion. Both test and treatment protocols may be updated over the course of the procedure at the direction of the user, or in some embodiments, based on logic embedded within the system.

(232) It is also recognized that many of these benefits may further improve other forms of acoustic therapy, including thermal ablation with high intensity focused ultrasound (HIFU), high intensity therapeutic ultrasound (HITU) including boiling histotripsy (thermal cavitation), and are considered as part of this disclosure. The disclosure also considers the application of histotripsy as

a means to activate previously delivered in active drug payloads whose activity is inert due to protection in a micelle, nanostructure or similar protective structure or through molecular arrangement that allows activation only when struck with acoustic energy.

(233) In another aspect, the Therapy sub-system, comprising in part, one or more amplifiers, transducers and power supplies, may be configured to allow multiple acoustic cavitation and histotripsy driving capabilities, affording specific benefits based on application, method and/or patient specific use. These benefits may include, but are not limited to, the ability to better optimize and control treatment parameters, which may allow delivery of more energy, with more desirable thermal profiles, increased treatment speed and reduced procedure times, enable electronic beam steering and/or other features.

(234) This disclosure also includes novel systems and concepts as related to systems and sub-systems comprising new and “universal” amplifiers, which may allow multiple driving approaches (e.g., single and multi-cycle pulsing). In some embodiments, this may include various novel features to further protect the system and user, in terms of electrical safety or other hazards (e.g., damage to transducer and/or amplifier circuitry).

(235) In another aspect, the system, and Therapy sub-system, may include a plethora of therapy transducers, where said therapy transducers are configured for specific applications and uses and may accommodate treating over a wide range of working parameters (target size, depth, location, etc.) and may comprise a wide range of working specifications (detailed below). Transducers may further adapt, interface and connect to a robotically-enabled system, as well as the Coupling sub-system, allowing the transducer to be positioned within, or along with, an acoustic coupling device allowing, in many embodiments, concurrent imaging and histotripsy treatments through an acceptable acoustic window. The therapy transducer may also comprise an integrated imaging probe or localization sensors, capable of displaying and determining transducer position within the treatment site and affording a direct field of view (or representation of) the treatment site, and as the acoustic cavitation/histotripsy tissue effect and bubble cloud may or may not change in appearance and intensity, throughout the treatment, and as a function of its location within said treatment (e.g., tumor, healthy tissue surrounding, critical structures, adipose tissue, etc.).

(236) The systems, methods and use of the system disclosed herein, may be beneficial to overcoming significant unmet needs in the areas of soft tissue ablation, oncology, immuno-oncology, advanced image guided procedures, surgical procedures including but not limited to open, laparoscopic, single incision, natural orifice, endoscopic, non-invasive, various combination of, various interventional spaces for catheter-based procedures of the vascular, cardiovascular pulmonary and/or neurocranial-related spaces, cosmetics/aesthetics, metabolic (e.g., type 2 diabetes), plastic and reconstructive, ocular and ophthalmology, orthopedic, gynecology and men's health, and other systems, devices and methods of treating diseased, injured, undesired, or healthy tissues, organs or cells.

(237) Systems and methods are also provided for improving treatment patterns within tissue that can reduce treatment time, improve efficacy, and reduce the amount of energy and prefocal tissue heating delivered to patients.

(238) Use Environments

(239) The disclosed system, methods of use, and use of the system, may be conducted in a plethora of environments and settings, with or without various support systems such as anesthesia, including but not limited to, procedure suites, operating rooms, hybrid rooms, in and out-patient settings, ambulatory settings, imaging centers, radiology, radiation therapy, oncology, surgical and/or any medical center, as well as physician offices, mobile healthcare centers or systems, automobiles and related vehicles (e.g., van), aero and marine transportation vehicles such as planes and ships, and/or any structure capable of providing temporary procedure support (e.g., tent). In some cases, systems and/or sub-systems disclosed herein may also be provided as integrated features into other environments, for example, the direct integration of the histotripsy Therapy sub-system into a MRI

scanner or patient surface/bed, wherein at a minimum the therapy generator and transducer are integral to such, and in other cases wherein the histotripsy configuration further includes a robotic positioning system, which also may be integral to a scanner or bed centered design.

#### (240) Coupling

(241) Systems may comprise a variety of Coupling sub-system embodiments, of which are enabled and configured to allow acoustic coupling to the patient to afford effective acoustic access for ultrasound visualization and acoustic cavitation/histotripsy (e.g., provide acoustic window and medium between the transducer(s) and patient, and support of). These may include different form factors of such, including open and enclosed device solutions, and some arrangements which may be configured to allow dynamic control over the acoustic medium (e.g., temperature, dissolved gas content, level of particulate filtration, sterility, volume, composition, etc.). Such dynamic control components may be directly integrated to the system (within the Cart), or may be in temporary/intermittent or continuous communication with the system, but externally situated in a separate device and/or cart.

(242) The Coupling sub-system typically comprises, at a minimum, coupling medium (e.g., degassed water or water solutions), a reservoir/container to contain said coupling medium, and a support structure (including interfaces to other surfaces or devices). In most embodiments, the coupling medium is water, and wherein the water may be conditioned before or during the procedure (e.g., chilled, degassed, filtered, etc.). Various conditioning parameters may be employed based on the configuration of the system and its intended use/application.

(243) The reservoir or medium container may be formed and shaped to various sizes and shapes, and to adapt/conform to the patient, allow the therapy transducer to engage/access and work within the acoustic medium, per defined and required working space (minimum volume of medium to allow the therapy transducer to be positioned and/or move through one or more treatment positions or patterns, and at various standoffs or depths from the patient, etc.), and wherein said reservoir or medium container may also mechanically support the load, and distribution of the load, through the use of a mechanical and/or electromechanical support structure. As a representative example, this may include a support frame. The container may be of various shapes, sizes, curvatures, and dimensions, and may be comprised of a variety of materials compositions (single, multiple, composites, etc.), of which may vary throughout. In some embodiments, it may comprise features such as films, drapes, membranes, bellows, etc. that may be insertable and removable, and/or fabricated within, of which may be used to conform to the patient and assist in confining/containing the medium within the container. It may further contain various sensors (e.g., volume/fill level), drains (e.g., inlet/outlet), lighting (e.g., LEDs), markings (e.g., fill lines, set up orientations, etc.), text (e.g., labeling), etc.

(244) In one embodiment, the reservoir or medium container contains a sealable frame, of which a membrane and/or film may be positioned within, to afford a conformable means of contacting the reservoir (later comprising the treatment head/therapy transducer) as an interface to the patient, that further provides a barrier to the medium (e.g., water) between the patient and therapy transducer). In other embodiments, the membrane and/or film may comprise an opening, the patient contacting edge of which affords a fluid/mechanical seal to the patient, but in contrast allows medium communication directly with the patient (e.g., direct degassed water interface with patient). The superstructure of the reservoir or medium container in both these examples may further afford the proximal portion of the structure (e.g., top) to be open or enclosed (e.g., to prevent spillage or afford additional features).

(245) Disclosed membranes may be comprised of various elastomers, viscoelastic polymers, thermoplastics, thermoplastic elastomers, thermoset polymers, silicones, urethanes, rigid/flexible co-polymers, block co-polymers, random block co-polymers, etc. Materials may be hydrophilic, hydrophobic, surface modified, coated, extracted, etc., and may also contain various additives to enhance performance, appearance or stability. In some embodiments, the thermoplastic elastomer

may be styrene-ethylene-butylene-styrene (SEBS), or other like strong and flexible elastomers. The membrane form factor can be flat or pre-shaped prior to use. In other embodiments, the membrane could be inelastic (i.e., a convex shape) and pressed against the patient's skin to acoustically couple the transducer to the tissue. Systems and methods are further disclosed to control the level of contaminants (e.g., particulates, etc.) on the membrane to maintain the proper level of ultrasound coupling. Too many particulates or contaminants can cause scattering of the ultrasound waves. This can be achieved with removable films or coatings on the outer surfaces of the membrane to protect against contamination.

(246) Said materials may be formed into useful membranes through molding, casting, spraying, ultrasonic spraying, extruding, and/or any other processing methodology that produces useful embodiments. They may be single use or reusable/reusable. They may be provided non-sterile, aseptically cleaned or sterile, where sterilization may comprise any known method, including but not limited to ethylene oxide, gamma, e-beam, autoclaving, steam, peroxide, plasma, chemical, etc. Membranes can be further configured with an outer molded or over molded frame to provide mechanical stability to the membrane during handling including assembly, set up and take down of the coupling sub-system. Various parameters of the membrane can be optimized for this method of use, including thickness, thickness profile, density, formulation (e.g., polymer molecular weight and copolymer ratios, additives, plasticizers, etc.), including optimizing specifically to maximize acoustic transmission properties, including minimizing impact to cavitation initiation threshold values, and/or ultrasound imaging artifacts, including but not limited to membrane reflections, as representative examples.

(247) Open reservoirs or medium containers may comprise various methods of filling, including using pre-prepared medium or water, that may be delivered into the containers, in some cases to a defined specification of water (level of temperature, gas saturation, etc.), or they may comprise additional features integral to the design that allow filling and draining (e.g., ports, valves, hoses, tubing, fittings, bags, pumps, etc.). These features may be further configured into or to interface to other devices, including for example, a fluidics system. In some cases, the fluidics system may be an in-house medium preparation system in a hospital or care setting room, or conversely, a mobile cart-based system which can prepare and transport medium to and from the cart to the medium container, etc.

(248) Enclosed iterations of the reservoir or medium container may comprise various features for sealing, in some embodiments sealing to a proximal/top portion or structure of a reservoir/container, or in other cases where sealing may comprise embodiments that seal to the transducer, or a feature on the transducer housings. Further, some embodiments may comprise the dynamic ability to control the volume of fluid within these designs, to minimize the potential for air bubbles or turbulence in said fluid and to allow for changes in the focal length to the target area without moving the transducer. As such, integrated features allowing fluid communication, and control of, may be provided (ability to provide/remove fluid on demand), including the ability to monitor and control various fluid parameters, some disclosed above. In order to provide this functionality, the overall system, and as part, the Coupling sub-system, may comprise a fluid conditioning system, which may contain various electromechanical devices, systems, power, sensing, computing, pumping, filtering and control systems, etc. The reservoir may also be configured to receive signals that cause it to deform or change shape in a specific and controlled manner to allow the target point to be adjusted without moving the transducer.

(249) Coupling support systems may include various mechanical support devices to interface the reservoir/container and medium to the patient, and the workspace (e.g., bed, floor, etc.). In some embodiments, the support system comprises a mechanical arm with 3 or more degrees of freedom. Said arm may have a proximal interface with one or more locations (and features) of the bed, including but not limited to, the frame, rails, customized rails or inserts, as well as one or more distal locations of the reservoir or container. The arm may also be a feature implemented on one or

more Carts, wherein Carts may be configured in various unlimited permutations, in some cases where a Cart only comprises the role of supporting and providing the disclosed support structure. (250) In some embodiments, the support structure and arm may be a robotically-enabled arm, implemented as a stand-alone Cart, or integrated into a Cart further comprising two or more system sub-systems, or where in the robotically-enabled arm is an arm of another robot, of interventional, surgical or other type, and may further comprise various user input features to actuate/control the robotic arm (e.g., positioning into/within coupling medium) and/or Coupling solution features (e.g., filling, draining, etc.). In some examples, the support structure robotic arm positional encoders may be used to coordinate the manipulation of the second arm (e.g. comprising the therapy transducer/treatment head), such as to position the therapy transducer to a desired/known location and pose within the coupling support structure.

(251) Overall, significant unmet needs exist in interventional and surgical medical procedures today, including those procedures utilizing minimally invasive devices and approaches to treat disease and/or injury, and across various types of procedures where the unmet needs may be solved with entirely new medical procedures. Today's medical system capabilities are often limited by access, wherein a less or non-invasive approach would be preferred, or wherein today's tools aren't capable to deliver preferred/required tissue effects (e.g., operate around/through critical structures without serious injury), or where the physical set up of the systems makes certain procedure approaches less desirable or not possible, and where a combination of approaches, along with enhanced tissue effecting treatments, may enable entirely new procedures and approaches, not possible today.

(252) In addition, specific needs exist for enabling histotripsy delivery, including robotic histotripsy delivery, wherein one or more histotripsy therapy transducers may be configured to acoustically couple to a patient, using a completely sealed approach (e.g., no acoustic medium communication with the patient's skin) and allowing the one or more histotripsy transducers to be moved within the coupling solution without impeding the motion/movement of the robotic arm or interfering/disturbing the coupling interface, which could affect the intended treatment and/or target location.

(253) Disclosed herein are histotripsy acoustic and patient coupling systems and methods, to enable histotripsy therapy/treatment, as envisioned in any setting, from interventional suite, operating room, hybrid suites, imaging centers, medical centers, office settings, mobile treatment centers, and/or others, as non-limiting examples. The following disclosure further describes novel systems used to create, control, maintain, modify/enhance, monitor and setup/takedown acoustic and patient coupling systems, in a variety of approaches, methods, environments, architectures and work-flows. In general, the disclosed novel systems may allow for a coupling medium, in some examples degassed water, to be interfaced between a histotripsy therapy transducer and a patient, wherein the acoustic medium provides sufficient acoustic coupling to said patient, allowing the delivery of histotripsy pulses through a user desired treatment location (and volume), where the delivery may require physically moving the histotripsy therapy transducer within a defined work-space comprising the coupling medium, and also where the coupling system is configured to allow said movement of the therapy transducer (and positioning system, e.g., robot) freely and unencumbered from by the coupling support system (e.g., a frame or manifold holding the coupling medium).

(254) Coupling System and Sub-Systems/Components

(255) The disclosed histotripsy acoustic and patient coupling systems, in general, may comprise one or more of the following sub-systems and components, an example of which is depicted in FIG. 2, including but not limited to 1) a membrane/barrier film to provide an enclosed, sealed and conformal patient coupling and histotripsy system interface, 2) a frame and assembly to retain the membrane and provide sufficient work and head space for a histotripsy therapy transducers required range of motion (x, y and z, pitch, roll and yaw), 3) a sufficient volume of ultrasound medium to afford acoustic coupling and interfaces to a histotripsy therapy transducer and robotic

arm, 4) one or more mechanical support arms to allow placement, positioning and load support of the frame, assembly and medium and 5) a fluidics system to prepare, provide and remove ultrasound medium(s) from the frame and assembly.

(256) In some embodiments, the coupling system may be fully sealed, and in other embodiments and configurations, it may be partially open to afford immediate access (physical and/or visual).

(257) The acoustic and patient coupling systems and sub-systems may further comprise various features and functionality, and associated work-flows, and may also be configured in a variety of ways to enable histotripsy procedures as detailed below.

(258) FIG. 2 illustrates one embodiment of a histotripsy therapy and imaging system **200**, including a coupling assembly **212**. As described above, a histotripsy therapy and imaging system can include a therapy transducer **202**, an imaging system **204**, a robotic positioning arm **208**, and a cart **210**.

(259) The therapy and/or imaging transducers can be housed in a coupling assembly **212** which can further include a coupling membrane **214** and a membrane constraint **216** configured to prevent the membrane from expanding too far from the transducer. The coupling membrane can be filled with an acoustic coupling medium such as a fluid or a gel. The membrane constraint can be, for example, a semi-rigid or rigid material configured to restrict expansion/movement of the membrane. In some embodiments, the membrane constraint is not used, and the elasticity and tensile strength of the membrane prevent over expansion. The coupling membrane can be a mineral-oil infused SEBS membrane to prevent direct fluid contact with the patient's skin. In the illustrated embodiment, the coupling assembly **212** is supported by a mechanical support arm **218** which can be load bearing in the x-y plane but allow for manual or automated z-axis adjustment. The mechanical support arm can be attached to the floor, the patient table, or the cart **210**. The mechanical support is designed and configured to conform and hold the coupling membrane **214** in place against the patient's skin while still allowing movement of the therapy/imaging transducer relative to the patient and also relative to the coupling membrane **214** with the robotic positioning arm **208**.

(260) The system can further include a fluidics system **220** that can include a fluid source, a cooling and degassing system, and a programmable control system. The fluidics system is configured for external loading of the coupling membrane with automated control of fluidic sequences. Further details on the fluidics system **220** are provided below.

(261) Membranes/Barrier Films and Related Architectures

(262) Membranes and barrier films may be composed of various biocompatible materials which allow conformal coupling to patient anatomy with minimal or no entrapped bubbles capable of interfering with ultrasound imaging and histotripsy therapy, and that are capable of providing a sealed barrier layer between said patient anatomy and the ultrasound medium, of which is contained within the work-space provided by the frame and assembly.

(263) Membrane and barrier film materials may comprise flexible and elastomeric biocompatible materials/polymers, such as various thermoplastic and thermoset materials, as well as permanent or bioresorbable polymers. Additionally, the frame of the UMC can also comprise the same materials. In some examples, the membrane may be rigid or semi-rigid polymers which are pre-shaped or flat.

(264) Ultrasound Medium

(265) As previously described, the ultrasound medium may comprise any applicable medium capable of providing sufficient and useful acoustic coupling to allow histotripsy treatments and enable sufficient clinical imaging (e.g., ultrasound). Ultrasound mediums, as a part of this disclosure and system, may comprise, but are not limited to, various aqueous solutions/mediums, including mixtures with other co-soluble fluids, of which may have preferred or more preferred acoustic qualities, including ability to match speed of sound, etc. Example mediums may comprise degassed water and/or mixtures/co-solutions of degassed water and various alcohols, such as ethanol.

(266) Mechanical Support Arms and Arm Architectures

(267) In order to support the acoustic and patient coupling system, including providing efficient and ergonomic work-flows for users, various designs and configurations of mechanical support arms (and arm architectures) may be employed. Support arms may be configured with a range of degrees of freedom, including but not limited to allowing, x, y, z, pitch, roll and yaw, as well additional interfacing features that may allow additional height adjustment or translation.

(268) Arms may comprise a varied number and type of joints and segments. Typically, arms may comprise a minimum of 2 segments. In some configurations, arms may comprise 3 to 5 segments.

(269) Arms are also be configured to interface proximally to a main support base or base interface (e.g., robot, table, table/bed rail, cart, floor mount, etc.) and distally to the frame/assembly and overall “UMC” or “coupling solution”. This specific distal interface may further include features for controlling position/orientation of the frame/assembly, at the frame/assembly interface.

(270) For example, in some embodiments, the arm/frame interface may comprise a ball joint wrist. In another example, the interface may include use of a gimbal wrist or an adjustable pitch and roll controlled wrist. These interfaces may be further employed with specific user interfaces and inputs, to assist with interacting with the various wrists, of which may include additional handles or knobs (as an unlimited example), to further enable positioning the UMC/coupling solution. For example, a gimbal wrist may benefit from allowing the frame/assembly to have 3 degrees of freedom (independent of the arm degrees of freedom), including pitch, roll and yaw adjustments.

(271) Support arms, configured with arm wrists, further interfaced with frames/assemblies, may comprise features such as brakes, including cable or electronic actuated brakes, and quick releases, which may interact with one or more axis, individually, or in groupings. They may also include electronic lift systems and base supports. In some embodiments, these lift systems/base supports are co-located with robot arm bases, wherein said robot arm is equipped with the histotripsy therapy transducer configured to fit/work within the enclosed coupling solution. In other embodiments, the support arm is located on a separate cart. In some cases, the separate cart may comprise a fluidics system or user console. In other embodiments, it is interfaced to a bed/table, including but not limited to a rail, side surface, and/or bed/table base. In other examples/embodiments, it's interfaced to a floor-based structure/footing, capable of managing weight and tipping requirements.

(272) Fluidics Systems, Control Systems and System Architectures

(273) As a part of overall fluidics management, histotripsy systems including acoustic/patient coupling systems, may be configured to include an automated fluidics system, which primarily is responsible for providing a reservoir for preparation and use of coupling medium, where preparation may include the ability to degas, chill, monitor, adjust, dispense/fill, and retrieve/drain coupling medium to/from the frame/assembly. The fluidics system may include an emergency high flow rate system for rapid draining of the coupling medium from the UMC. In some embodiments, the fluidics system can be configured for a single use of the coupling medium, or alternatively, for re-use of the medium. In some embodiments, the fluidics system can implement positive air pressure or vacuum to carry out leak tests of the UMC and membrane prior to filling with a coupling medium. Vacuum assist can also be used for removal of air from the UMC during the filling process. The fluidics system can further include filters configured to prevent particulate contamination from reaching the UMC.

(274) The fluidics system may be implemented in the form of a mobile fluidics cart. The cart may comprise an input tank, drain tank, degassing module, fill pump, drain pump, inert gas tank, air compressor, tubing/connectors/lines, electronic and manual controls systems and input devices, power supplies and one or more batteries. The cart in some cases may also comprise a system check vessel/reservoir for evaluating histotripsy system performance and related system diagnostics (configured to accommodate a required water volume and work-space for a therapy transducer).

(275) It should be understood that any feature described herein with respect to one embodiment can

be substituted for or combined with any feature described with respect to another embodiment.

(276) When a feature or element is herein referred to as being “on” another feature or element, it can be directly on the other feature or element or intervening features and/or elements may also be present. In contrast, when a feature or element is referred to as being “directly on” another feature or element, there are no intervening features or elements present. It will also be understood that, when a feature or element is referred to as being “connected”, “attached” or “coupled” to another feature or element, it can be directly connected, attached or coupled to the other feature or element or intervening features or elements may be present. In contrast, when a feature or element is referred to as being “directly connected”, “directly attached” or “directly coupled” to another feature or element, there are no intervening features or elements present. Although described or shown with respect to one embodiment, the features and elements so described or shown can apply to other embodiments. It will also be appreciated by those of skill in the art that references to a structure or feature that is disposed “adjacent” another feature may have portions that overlap or underlie the adjacent feature.

(277) Terminology used herein is for the purpose of describing particular embodiments only and is not intended to be limiting of the invention. For example, as used herein, the singular forms “a”, “an” and “the” are intended to include the plural forms as well, unless the context clearly indicates otherwise. It will be further understood that the terms “comprises” and/or “comprising,” when used in this specification, specify the presence of stated features, steps, operations, elements, and/or components, but do not preclude the presence or addition of one or more other features, steps, operations, elements, components, and/or groups thereof. As used herein, the term “and/or” includes any and all combinations of one or more of the associated listed items and may be abbreviated as “/”.

(278) Spatially relative terms, such as “under”, “below”, “lower”, “over”, “upper” and the like, may be used herein for ease of description to describe one element or feature's relationship to another element(s) or feature(s) as illustrated in the figures. It will be understood that the spatially relative terms are intended to encompass different orientations of the device in use or operation in addition to the orientation depicted in the figures. For example, if a device in the figures is inverted, elements described as “under” or “beneath” other elements or features would then be oriented “over” the other elements or features. Thus, the exemplary term “under” can encompass both an orientation of over and under. The device may be otherwise oriented (rotated 90 degrees or at other orientations) and the spatially relative descriptors used herein interpreted accordingly. Similarly, the terms “upwardly”, “downwardly”, “vertical”, “horizontal” and the like are used herein for the purpose of explanation only unless specifically indicated otherwise.

(279) Although the terms “first” and “second” may be used herein to describe various features/elements (including steps), these features/elements should not be limited by these terms, unless the context indicates otherwise. These terms may be used to distinguish one feature/element from another feature/element. Thus, a first feature/element discussed below could be termed a second feature/element, and similarly, a second feature/element discussed below could be termed a first feature/element without departing from the teachings of the present invention.

(280) Throughout this specification and the claims which follow, unless the context requires otherwise, the word “comprise”, and variations such as “comprises” and “comprising” means various components can be co-jointly employed in the methods and articles (e.g., compositions and apparatuses including device and methods). For example, the term “comprising” will be understood to imply the inclusion of any stated elements or steps but not the exclusion of any other elements or steps.

(281) As used herein in the specification and claims, including as used in the examples and unless otherwise expressly specified, all numbers may be read as if prefaced by the word “about” or “approximately,” even if the term does not expressly appear. The phrase “about” or “approximately” may be used when describing magnitude and/or position to indicate that the value



and/or position described is within a reasonable expected range of values and/or positions. For example, a numeric value may have a value that is  $\pm 0.1\%$  of the stated value (or range of values),  $\pm 1\%$  of the stated value (or range of values),  $\pm 2\%$  of the stated value (or range of values),  $\pm 5\%$  of the stated value (or range of values),  $\pm 10\%$  of the stated value (or range of values), etc. Any numerical values given herein should also be understood to include about or approximately that value, unless the context indicates otherwise. For example, if the value “10” is disclosed, then “about 10” is also disclosed. Any numerical range recited herein is intended to include all sub-ranges subsumed therein. It is also understood that when a value is disclosed that “less than or equal to” the value, “greater than or equal to the value” and possible ranges between values are also disclosed, as appropriately understood by the skilled artisan. For example, if the value “X” is disclosed the “less than or equal to X” as well as “greater than or equal to X” (e.g., where X is a numerical value) is also disclosed. It is also understood that the throughout the application, data is provided in a number of different formats, and that this data, represents endpoints and starting points, and ranges for any combination of the data points. For example, if a particular data point “10” and a particular data point “15” are disclosed, it is understood that greater than, greater than or equal to, less than, less than or equal to, and equal to 10 and 15 are considered disclosed as well as between 10 and 15. It is also understood that each unit between two particular units are also disclosed. For example, if 10 and 15 are disclosed, then 11, 12, 13, and 14 are also disclosed. (282) While preferred embodiments of the present invention have been shown and described herein, it will be obvious to those skilled in the art that such embodiments are provided by way of example only. Numerous variations, changes, and substitutions will now occur to those skilled in the art without departing from the invention. It should be understood that various alternatives to the embodiments of the invention described herein may be employed in practicing the invention. It is intended that the following claims define the scope of the invention and that methods and structures within the scope of these claims and their equivalents be covered thereby.

## Claims

1. A method of producing an enhanced excitation volume with histotripsy energy, comprising: positioning a focus of an ultrasound therapy transducer at a first focal location in a first discrete treatment location within a target tissue volume, the first focal location being positioned along a central axis of the ultrasound therapy transducer; delivering a first histotripsy pulse to the first focal location at a therapy pulse repetition frequency (PRF) to produce a first cavitation bubble cloud at the first focal location; electronically beam-steering the focus of the ultrasound therapy transducer to a second focal location in the first discrete treatment location that is at least partially outside of the central axis with a steering period equivalent to the therapy PRF; delivering a second histotripsy pulse to the second focal location at the therapy PRF to produce a second bubble cloud at the second focal location; electronically beam-steering the focus of the ultrasound therapy transducer to a third focal location at least partially inside the central axis with the steering period; delivering a third histotripsy pulse to the third focal location at the therapy PRF to produce a third bubble cloud at the third focal location; repeating electronically beam-steering the focus with the steering period and delivering histotripsy pulses at the therapy PRF to the first focal location, the second focal location, and the third focal location in the first discrete treatment location until a desired dose is applied to the first discrete treatment location; and mechanically positioning the focus of the ultrasound therapy transducer at a first focal location in a second discrete treatment location within the target tissue volume after the desired dose has been applied to the first discrete treatment location.
2. The method of claim 1, wherein a natural focus of the ultrasound therapy transducer array is located along the central axis.
3. The method of claim 2, wherein the first focal location comprises the natural focus.

4. The method of claim 1, wherein the second focal location partially overlaps with the central axis.
5. The method of claim 1, wherein the third focal location comprises is overlapped with the first focal location.
6. The method of claim 1, wherein electronically beam-steering to the second focal location comprises steering laterally.
7. The method of claim 1, wherein electronically beam-steering to the second focal location comprises steering laterally and axially.
8. An ultrasound therapy system, comprising; a robotic positioning system; an ultrasound therapy transducer array connected to the robotic positioning system; a generator operatively coupled to the ultrasound therapy transducer array, the generator and ultrasound therapy transducer array being configured to deliver histotripsy pulses into a subject to generate a cavitation bubble cloud in the subject; at least one processor operatively coupled to the robotic positioning system and the generator, the at least one processor being configured to control the robotic positioning system and the generator to provide histotripsy therapy to the subject according to a treatment plan that includes a plurality of discrete treatment locations within a target tissue volume of the subject by: controlling the robotic positioning system to mechanically position a focus of the ultrasound therapy transducer array at a first focal location in a first discrete treatment location within the treatment plan, the first focal location being positioned along a central axis of the ultrasound therapy transducer array; controlling the generator to deliver at least one histotripsy pulse at a therapy pulse repetition frequency (PRF) with the ultrasound therapy transducer array to form a first cavitation bubble cloud at the first focal location; controlling the generator to electronically beam-steer the focus of the ultrasound therapy transducer to a second focal location in the first discrete treatment location that is at least partially outside the central axis with a steering period equivalent to the therapy PRF; controlling the generator to deliver at least one histotripsy pulse to the second focal location at the therapy PRF to produce a second bubble cloud at the second focal location; controlling the generator to electronically beam-steer the focus of the ultrasound therapy transducer to a third focal location in the first discrete treatment location that is at least partially inside the central axis with the steering period; and controlling the generator to deliver at least one histotripsy pulse to the third focal location to produce a third bubble cloud at the third focal location at the therapy PRF; repeating controlling the generator to electronically beam-steer the focus with the steering period and controlling the generator to deliver histotripsy pulses at the therapy PRF to the first focal location, the second focal location, and the third focal location in the first discrete treatment location until a desired dose is applied to the first discrete treatment location; and controlling the robotic positioning system to mechanically position the focus of the ultrasound therapy transducer at a first focal location in a second discrete treatment location within the target tissue volume after the desired dose has been applied to the first discrete treatment location.
9. The system of claim 8, wherein a natural focus of the ultrasound therapy transducer array is located along the central axis.
10. The system of claim 9, wherein the first focal location comprises the natural focus.
11. The system of claim 8, wherein the second focal location partially overlaps with the central axis.
12. The system of claim 8, wherein the third focal location is overlapped with the first focal location.
13. The system of claim 8, wherein controlling the generator to electronically beam-steer to the second focal location comprises steering laterally.
14. The system of claim 8, wherein controlling the generator to electronically beam-steer to the second focal location comprises steering laterally and axially.
15. An ultrasound therapy system, comprising; an ultrasound therapy transducer array; a generator operatively coupled to the ultrasound therapy transducer array, the generator and ultrasound

therapy transducer array being configured to deliver histotripsy pulses into a subject to generate a cavitation bubble cloud in the subject; at least one processor operatively coupled to the generator, the at least one processor being configured to control the generator to provide histotripsy therapy to the subject according to a treatment plan that includes a plurality of discrete treatment locations within a target tissue volume of the subject by: controlling the generator to deliver at least one histotripsy pulse with the ultrasound therapy transducer array at a therapy pulse repetition frequency (PRF) to form a first cavitation bubble cloud at a first focal location in a first discrete treatment location within the treatment plan, the first focal location being positioned along a central axis of the ultrasound therapy transducer array; controlling the generator to electronically beam-steer the focus of the ultrasound therapy transducer to a second focal location in the first discrete treatment location that is at least partially outside the central axis with a steering period equivalent to the therapy PRF; controlling the generator to deliver at least one histotripsy pulse to the second focal location at the therapy PRF to produce a second bubble cloud at the second focal location; controlling the generator to electronically beam-steer the focus of the ultrasound therapy transducer to a third focal location in the first discrete treatment location that is at least partially inside the central axis with the steering period; and controlling the generator to deliver at least one histotripsy pulse to the third focal location to produce a third bubble cloud at the third focal location at the therapy PRF; repeating controlling the generator to electronically beam-steer the focus with the steering period and controlling the generator to deliver histotripsy pulses at the therapy PRF to the first focal location, the second focal location, and the third focal location in the first discrete treatment location until a desired dose is applied to the first discrete treatment location; and controlling the robotic positioning system to mechanically position the focus of the ultrasound therapy transducer at a first focal location in a second discrete treatment location within the target tissue volume after the desired dose has been applied to the first discrete treatment location.

16. The system of claim 15, wherein a natural focus of the ultrasound therapy transducer array is located along the central axis.

17. The system of claim 15, wherein the third focal location is overlapped with the first focal location.

---