

# US Patent & Trademark Office

## Patent Public Search | Text View

---

United States Patent	12390117
Kind Code	B2
Date of Patent	August 19, 2025
Inventor(s)	Bechtel; Kate LeeAnn et al.

---

### Optical sensor module for speckleplethysmography (SPG) and photoplethysmography (PPG)

---

#### Abstract

An optical sensor module for measuring both speckleplethysmography (SPG) and photoplethysmography (PPG) signals at human or animal tissue, the optical sensor module comprising: a first light source, for illuminating the tissue for use with SPG measurements, the first light source comprising a laser; a second light source, for illuminating the tissue for use with PPG measurements; and one or more optical sensor(s) for receiving light from the illuminated tissue.

---

**Inventors:** Bechtel; Kate LeeAnn (Pleasant Hill, CA), McMillan; James (Santa Monica, CA), Afshinmanesh; Farzaneh (Pasadena, CA), Dunn; Cody (Costa Mesa, CA)

**Applicant:** ROCKLEY PHOTONICS LIMITED (Altrincham, GB)

**Family ID:** 1000008764110

**Assignee:** Rockley Photonics Limited (Altrincham, GB)

**Appl. No.:** 17/711974

**Filed:** April 01, 2022

#### Prior Publication Data

Document Identifier	Publication Date
US 20230148885 A1	May. 18, 2023

#### Related U.S. Application Data

us-provisional-application US 63279932 20211116

---

#### Publication Classification

**Int. Cl.:** A61B5/024 (20060101); A61B5/00 (20060101); A61B5/0295 (20060101)

**U.S. Cl.:**

**CPC** A61B5/02433 (20130101); A61B5/02438 (20130101); A61B5/0295 (20130101);  
A61B5/681 (20130101);

## Field of Classification Search

**USPC:** None

---

## References Cited

### U.S. PATENT DOCUMENTS

Patent No.	Issued Date	Patentee Name	U.S. Cl.	CPC
5243983	12/1992	Tarr et al.	N/A	N/A
5830132	12/1997	Robinson	N/A	N/A
6246892	12/2000	Chance	N/A	N/A
6560478	12/2002	Alfano et al.	N/A	N/A
6919549	12/2004	Bamji et al.	N/A	N/A
7035679	12/2005	Addison et al.	N/A	N/A
7113817	12/2005	Winchester, Jr. et al.	N/A	N/A
7250317	12/2006	Heideman	N/A	N/A
7295783	12/2006	Singh et al.	N/A	N/A
7375812	12/2007	Atia et al.	N/A	N/A
7505128	12/2008	Zribi et al.	N/A	N/A
7761126	12/2009	Gardner et al.	N/A	N/A
7865225	12/2010	Kaltschmidt et al.	N/A	N/A
7922664	12/2010	Elliott	N/A	N/A
7925056	12/2010	Presura et al.	N/A	N/A
8237927	12/2011	Reeve et al.	N/A	N/A
8277384	12/2011	Fine	N/A	N/A
8313439	12/2011	McCombie et al.	N/A	N/A
8343062	12/2012	Fortin et al.	N/A	N/A
8343063	12/2012	Borgos	N/A	N/A
8376955	12/2012	Baker, Jr.	N/A	N/A
8398556	12/2012	Sethi et al.	N/A	N/A
8868149	12/2013	Eisen et al.	N/A	N/A
8920332	12/2013	Hong et al.	N/A	N/A
8923942	12/2013	Bernreuter	N/A	N/A
8945017	12/2014	Venkatraman et al.	N/A	N/A
8948832	12/2014	Hong et al.	N/A	N/A
8956303	12/2014	Hong et al.	N/A	N/A
8998815	12/2014	Venkatraman et al.	N/A	N/A
9005129	12/2014	Venkatraman et al.	N/A	N/A
9113794	12/2014	Hong et al.	N/A	N/A
9113795	12/2014	Hong et al.	N/A	N/A
9149216	12/2014	Eisen et al.	N/A	N/A

9155480	12/2014	Thakor et al.	N/A	N/A
9226673	12/2015	Ferguson, Jr. et al.	N/A	N/A
9237855	12/2015	Hong et al.	N/A	N/A
9307917	12/2015	Hong et al.	N/A	N/A
9494567	12/2015	Islam	N/A	N/A
9687162	12/2016	Vetter et al.	N/A	N/A
9704050	12/2016	Lee et al.	N/A	N/A
9730622	12/2016	Eisen et al.	N/A	N/A
9772280	12/2016	Cerussi et al.	N/A	N/A
9804027	12/2016	Fish et al.	N/A	N/A
9846126	12/2016	Gunn, III et al.	N/A	N/A
9848787	12/2016	White et al.	N/A	N/A
9877681	12/2017	Silverman	N/A	N/A
9931040	12/2017	Homyk et al.	N/A	N/A
9970955	12/2017	Homyk et al.	N/A	N/A
10004406	12/2017	Yuen et al.	N/A	N/A
10058256	12/2017	Chen et al.	N/A	N/A
10178959	12/2018	Homyk et al.	N/A	N/A
10178973	12/2018	Venkatraman et al.	N/A	N/A
10194808	12/2018	Thompson et al.	N/A	N/A
10206576	12/2018	Shcherbakov et al.	N/A	N/A
10215698	12/2018	Han et al.	N/A	N/A
10241033	12/2018	Uematsu et al.	N/A	N/A
10271740	12/2018	Ward et al.	N/A	N/A
10314532	12/2018	Ward et al.	N/A	N/A
10326035	12/2018	Lu et al.	N/A	N/A
10326036	12/2018	Sweeney et al.	N/A	N/A
10349847	12/2018	Kwon et al.	N/A	N/A
10352768	12/2018	Simpkin et al.	N/A	N/A
10357165	12/2018	Yoon	N/A	N/A
10422693	12/2018	Fish et al.	N/A	N/A
10451537	12/2018	Nakaji	N/A	N/A
10463286	12/2018	Schenkman et al.	N/A	N/A
10492684	12/2018	Khachaturian et al.	N/A	N/A
10506926	12/2018	Khachaturian et al.	N/A	N/A
10506955	12/2018	Tholl et al.	N/A	N/A
10568527	12/2019	Yoon et al.	N/A	N/A
10588519	12/2019	Yuen et al.	N/A	N/A
10602987	12/2019	Khachaturian et al.	N/A	N/A
10627849	12/2019	Scofield et al.	N/A	N/A
10641962	12/2019	Nykänen et al.	N/A	N/A
10643903	12/2019	Drake et al.	N/A	N/A
10667688	12/2019	Khachaturian et al.	N/A	N/A
10677989	12/2019	Abediasl et al.	N/A	N/A
10681259	12/2019	Ichiki et al.	N/A	N/A
10681283	12/2019	Nakashima et al.	N/A	N/A
10694997	12/2019	Kim et al.	N/A	N/A
10718668	12/2019	Gu et al.	N/A	N/A
10722177	12/2019	Homyk et al.	N/A	N/A
10739256	12/2019	Rickman et al.	N/A	N/A

10750956	12/2019	Zalevsky et al.	N/A	N/A
10775239	12/2019	Lee et al.	N/A	N/A
10813597	12/2019	Rice et al.	N/A	N/A
10820858	12/2019	Yoon et al.	N/A	N/A
10842422	12/2019	Yu et al.	N/A	N/A
10871503	12/2019	Homyk et al.	N/A	N/A
10966616	12/2020	De Morree et al.	N/A	N/A
10973422	12/2020	Pantelopoulos et al.	N/A	N/A
11022751	12/2020	Bauters et al.	N/A	N/A
11045103	12/2020	Shchekin et al.	N/A	N/A
11079364	12/2020	Leger et al.	N/A	N/A
11096601	12/2020	Hong et al.	N/A	N/A
11096608	12/2020	Van Dorpe et al.	N/A	N/A
11129544	12/2020	Zalevsky et al.	N/A	N/A
11202582	12/2020	Verkruijssse et al.	N/A	N/A
11213217	12/2021	Han et al.	N/A	N/A
11278220	12/2021	Tucker et al.	N/A	N/A
11298035	12/2021	Huijbregts et al.	N/A	N/A
11369275	12/2021	Song et al.	N/A	N/A
11445922	12/2021	Naima	N/A	N/A
11553851	12/2022	Kim et al.	N/A	N/A
11583185	12/2022	Homyk et al.	N/A	N/A
11666238	12/2022	Rege et al.	N/A	N/A
11666277	12/2022	Yoon et al.	N/A	N/A
11684281	12/2022	Pantelopoulos et al.	N/A	N/A
11690513	12/2022	Hu et al.	N/A	N/A
11696693	12/2022	Wong	N/A	N/A
11709120	12/2022	Rice et al.	N/A	N/A
11744491	12/2022	Dunn et al.	N/A	N/A
11751811	12/2022	Sun et al.	N/A	N/A
11759116	12/2022	White et al.	N/A	N/A
11759121	12/2022	Mccann et al.	N/A	N/A
11771343	12/2022	Sacha	N/A	N/A
11800990	12/2022	White et al.	N/A	N/A
11857301	12/2023	Homyk et al.	N/A	N/A
11883134	12/2023	Leabman	N/A	N/A
11890081	12/2023	Jang	N/A	N/A
11980451	12/2023	Albert	N/A	N/A
2002/0068859	12/2001	Knopp	600/322	A61B 5/1455
2005/0249509	12/2004	Nagarajan et al.	N/A	N/A
2006/0124829	12/2005	Song et al.	N/A	N/A
2006/0204175	12/2005	Laurent-Lund et al.	N/A	N/A
2008/0204752	12/2007	Dorvee et al.	N/A	N/A
2008/0220512	12/2007	Koh et al.	N/A	N/A
2008/0316567	12/2007	Grasser et al.	N/A	N/A
2011/0054277	12/2010	Pinter et al.	N/A	N/A

2011/0082355	12/2010	Eisen et al.	N/A	N/A
2011/0196244	12/2010	Ribas Ripoll et al.	N/A	N/A
2012/0130215	12/2011	Fine et al.	N/A	N/A
2013/0131475	12/2012	Eisen et al.	N/A	N/A
2013/0190630	12/2012	Borgos	N/A	N/A
2014/0094666	12/2013	Fine	N/A	N/A
2014/0200423	12/2013	Eisen et al.	N/A	N/A
2014/0316286	12/2013	Addison et al.	N/A	N/A
2014/0376001	12/2013	Swanson	N/A	N/A
2015/0157224	12/2014	Carmon et al.	N/A	N/A
2015/0196251	12/2014	Outwater et al.	N/A	N/A
2015/0201854	12/2014	Hong et al.	N/A	N/A
2016/0066790	12/2015	Shcherbakov et al.	N/A	N/A
2016/0106327	12/2015	Yoon et al.	N/A	N/A
2016/0157736	12/2015	Huang et al.	N/A	N/A
2016/0161685	12/2015	Xu et al.	N/A	N/A
2016/0266337	12/2015	Feng	N/A	N/A
2016/0278676	12/2015	Eisen et al.	N/A	N/A
2016/0282265	12/2015	Su et al.	N/A	N/A
2016/0287107	12/2015	Szabados et al.	N/A	N/A
2017/0014037	12/2016	Coppola et al.	N/A	N/A
2017/0065184	12/2016	Barak	N/A	N/A
2017/0108439	12/2016	Stievater et al.	N/A	N/A
2017/0138789	12/2016	Ivanov	N/A	N/A
2017/0231513	12/2016	Presura	600/301	A61B 5/024
2017/0315292	12/2016	Mullen et al.	N/A	N/A
2018/0045566	12/2017	Fish et al.	N/A	N/A
2018/0064399	12/2017	Buettgen	N/A	A61B 5/02416
2018/0110423	12/2017	Presura et al.	N/A	N/A
2018/0160913	12/2017	Fine	N/A	N/A
2018/0238794	12/2017	Kangas et al.	N/A	N/A
2018/0263519	12/2017	Gu	N/A	N/A
2018/0283950	12/2017	Ge et al.	N/A	N/A
2018/0296168	12/2017	Rice et al.	N/A	N/A
2019/0005351	12/2018	Zhou	N/A	G01N 15/1459
2019/0046056	12/2018	Khachaturian et al.	N/A	N/A
2019/0053721	12/2018	Boas et al.	N/A	N/A
2019/0094009	12/2018	Aizawa et al.	N/A	N/A
2019/0167118	12/2018	Vilenskii et al.	N/A	N/A
2019/0175030	12/2018	Verkruijse et al.	N/A	N/A
2019/0336006	12/2018	Horstmeyer et al.	N/A	N/A
2019/0343442	12/2018	Aung et al.	N/A	N/A
2019/0343456	12/2018	Kahlert et al.	N/A	N/A
2019/0387972	12/2018	Hu et al.	N/A	N/A
2019/0391243	12/2018	Nicolaescu	N/A	N/A
2019/0391702	12/2018	Jo et al.	N/A	N/A
2020/0003619	12/2019	Hu et al.	N/A	N/A
2020/0069225	12/2019	Vizbaras et al.	N/A	N/A

2020/0158548	12/2019	Rice et al.	N/A	N/A
2020/0196874	12/2019	Rozental et al.	N/A	N/A
2020/0214602	12/2019	Narumi et al.	N/A	N/A
2020/0359948	12/2019	Dunn et al.	N/A	N/A
2020/0397351	12/2019	Miyata	N/A	N/A
2021/0000385	12/2020	Warren et al.	N/A	N/A
2021/0022623	12/2020	Rice	N/A	A61B 5/02116
2021/0028602	12/2020	Cao et al.	N/A	N/A
2021/0267471	12/2020	Bonomi et al.	N/A	N/A
2021/0386310	12/2020	Hong et al.	N/A	N/A
2022/0015649	12/2021	Ikuta et al.	N/A	N/A
2022/0019861	12/2021	Durr et al.	N/A	N/A
2022/0039679	12/2021	Califa et al.	N/A	N/A
2022/0265158	12/2021	Tokura	N/A	N/A
2022/0370010	12/2021	Zilkie et al.	N/A	N/A
2022/0413143	12/2021	Parsa et al.	N/A	N/A
2023/0003938	12/2022	Zilkie et al.	N/A	N/A
2023/0039055	12/2022	Gardner et al.	N/A	N/A
2023/0087295	12/2022	Dunn et al.	N/A	N/A
2023/0148885	12/2022	Bechtel et al.	N/A	N/A
2023/0148886	12/2022	Bechtel et al.	N/A	N/A
2023/0277075	12/2022	Pantelopoulos et al.	N/A	N/A
2023/0320598	12/2022	Khine et al.	N/A	N/A
2023/0347029	12/2022	Corso et al.	N/A	N/A
2023/0397818	12/2022	Newhouse et al.	N/A	N/A
2023/0401747	12/2022	Dunn et al.	N/A	N/A
2024/0041342	12/2023	Lai et al.	N/A	N/A
2024/0074667	12/2023	Rick et al.	N/A	N/A
2024/0108289	12/2023	Bechtel et al.	N/A	N/A
2024/0115212	12/2023	Jang	N/A	N/A
2024/0156355	12/2023	O'Brien et al.	N/A	N/A

## FOREIGN PATENT DOCUMENTS

Patent No.	Application Date	Country	CPC
108709847	12/2017	CN	N/A
110301896	12/2018	CN	N/A
211094079	12/2019	CN	N/A
211131004	12/2019	CN	N/A
3 002 568	12/2015	EP	N/A
2 395 958	12/2016	EP	N/A
3 384 841	12/2017	EP	N/A
3 558 119	12/2019	EP	N/A
3 886 686	12/2020	EP	N/A
WO 2018/029123	12/2017	WO	N/A
WO 2019/149815	12/2018	WO	N/A
WO 2019/233903	12/2018	WO	N/A
WO 2020/030641	12/2019	WO	N/A

WO 2020/114989	12/2019	WO	N/A
WO 2021/058338	12/2020	WO	N/A
WO 2021/094473	12/2020	WO	N/A
WO 2021/116766	12/2020	WO	N/A
WO 2021/116766	12/2020	WO	N/A
WO 2023/245149	12/2022	WO	N/A
WO 2024/052289	12/2023	WO	N/A

## OTHER PUBLICATIONS

Akram, M. N. et al., “Laser speckle reduction due to spatial and angular diversity introduced by fast scanning micromirror”, *Applied Optics*, Jun. 4, 2010, pp. 3297-3304, vol. 49, No. 17, Optical Society of America. cited by applicant

Baek, H. J. et al., “The Effect of Optical Crosstalk on Accuracy of Reflectance-Type Pulse Oximeter for Mobile Healthcare”, *Journal of Healthcare Engineering*, Oct. 21, 2018, 9 pages, vol. 2018, Article ID 3521738, Hindawi, <https://doi.org/10.1155/2018/3521738>. cited by applicant

Baets, R. et al., “Spectroscopy-on-chip applications of silicon photonics”, *Proc. Of SPIE*, 2013, pp. 86270I-1 through 86270I-10, vol. 8627, SPIE. cited by applicant

Berger, A. J. et al., “Feasibility of measuring blood glucose concentration by near-infrared Raman spectroscopy”, *Spectrochimica Acta Part A*, 1997, pp. 287-292, Elsevier Science B.V. cited by applicant

Bi, R. et al., “A speckle-based method for fast blood flow measurement in deep tissue”, *Proceedings of SPIE, Optical Biopsy XIX: Toward Real-Time Spectroscopic Imaging and Diagnosis*, Mar. 5, 2021, pp. 1163606-1 through 1163606-5, vol. 11636, SPIE. cited by applicant

Biswas, A. et al., “Fast diffuse correlation spectroscopy with a low-cost, fiber-less embedded diode laser”, *Biomedical Optics Express*, Oct. 4, 2021, pp. 6686-6700, vol. 12, No. 11, Optical Society of America. cited by applicant

Brouckaert, J. et al., “Silicon-on-Insulator Microspectrometer”, *Proceedings Symposium IEEE/LEOS Benelux Chapter*, 2008, pp. 7-10, IEEE. cited by applicant

Cole, D. B. et al., “Integrated heterodyne interferometer with on-chip modulators and detectors”, *Optics Letters*, Jun. 25, 2015, pp. 3097-3100, vol. 40, No. 13, Optical Society of America. cited by applicant

Epping, J. P. et al., “High power, tunable, narrow linewidth dual gain hybrid laser”, *Laser Congress*, Oct. 3, 2019, pp. 1-2. cited by applicant

Fu, D. et al., “In Vivo Metabolic Fingerprinting of Neutral Lipids with Hyperspectral Stimulated Raman Scattering Microscopy”, *Journal of the American Chemical Society*, May 28, 2014, pp. 8820-8828, American Chemical Society Publications. cited by applicant

Fukui, T. et al., “Single-Pixel Imaging Using Multimode Fiber and Silicon Photonic Phased Array”, *Journal of Lightwave Technology*, Jul. 14, 2020, pp. 839-844, vol. 39, No. 3, IEEE. cited by applicant

Gottschling, K. et al., “Molecular Insights into Carbon Dioxide Sorption in Hydrazone-Based Covalent Organic Frameworks with Tertiary Amine Moieties”, *Chemistry of Materials*, Feb. 13, 2019, pp. 1946-1955, American Chemical Society. cited by applicant

Hashimoto, Y. et al., “Fabrication of an Anti-Reflective and Super-Hydrophobic Structure by Vacuum Ultraviolet Light-Assisted Bonding and Nanoscale Pattern Transfer”, *Micromachines*, Apr. 15, 2018, pp. 1-11, [www.mdpi.com/journal/micromachines](http://www.mdpi.com/journal/micromachines). cited by applicant

Hollis, V. S. et al., “Non-invasive monitoring of brain tissue temperature by near-infrared spectroscopy”, *Proceedings of SPIE, Optical Tomography and Spectroscopy of Tissue IV*, Jun. 29, 2001, pp. 470-481, vol. 4250, SPIE, <https://www.spiedigitallibrary.org/conference-proceedings-of-spie/4250/1/Noninvasive-monitoring-of-brain-tissue-temperature-by-near-infrared-spectroscopy/10.1117/12.434506.short?SSO=1>. cited by applicant

International Search Report and Written Opinion of the International Searching Authority, Mailed Mar. 11, 2021, Corresponding to PCT/IB2020/001037, 13 pages. cited by applicant

International Search Report and Written Opinion of the International Searching Authority, mailed Apr. 4, 2023, corresponding to PCT/EP2022/082341, 33 pages. cited by applicant

International Search Report and Written Opinion of the International Searching Authority, mailed Feb. 1, 2022, corresponding to PCT/IB2021/000649, 18 pages. cited by applicant

International Search Report and Written Opinion of the International Searching Authority, mailed Nov. 15, 2021, corresponding to PCT/IB2021/000517, 15 pages. cited by applicant

International Search Report and Written Opinion of the International Searching Authority, mailed Oct. 10, 2022, corresponding to PCT/IB2022/000373, 16 pages. cited by applicant

International Search Report and Written Opinion of the International Searching Authority, mailed Oct. 19, 2022, corresponding to PCT/EP2022/071467, 16 pages. cited by applicant

Invitation to Pay Additional Fees and Partial Search Report mailed Feb. 14, 2023 in related International Application No. PCT/EP2022/082341, 18 pages. cited by applicant

Izutsu, M. et al., "Integrated Optical SSB Modulator/Frequency Shifter", IEEE Journal of Quantum Electronics, Nov. 1981, pp. 2225-2227, vol. QE-17, No. 11, IEEE. cited by applicant

Kang, J. W. et al., "Direct observation of glucose fingerprint using in vivo Raman spectroscopy," Science Advances, Jan. 24, 2020, pp. 1-8, American Association for the Advancement of Science. cited by applicant

Karlsson, C. J. et al., "All-fiber multifunction continuous-wave coherent laser radar at 1.55  $\mu\text{m}$  for range, speed, vibration, and wind measurements", Applied Optics, Jul. 20, 2000, pp. 3716-3726, vol. 39, No. 21, Optical Society of America. cited by applicant

Lai, N. et al., "CO<sub>2</sub> Capture With Absorbents of Tertiary Amine Functionalized Nano-SiO<sub>2</sub>", Frontiers in Chemistry, Feb. 28, 2020, pp. 1-9, vol. 8, Article 146, www.frontiersin.org. cited by applicant

Lapchuk, A. et al., "Investigation of speckle suppression beyond human eye sensitivity by using a passive multimode fiber and a multimode fiber bundle", Applied Optics, Feb. 21, 2020, pp. 6820-6834, vol. 28, No. 5, Optical Society of America. cited by applicant

Loi, R. et al., "Transfer Printing of AlGaInAs/InP Etched Facet Lasers to Si Substrates", IEEE Photonics Journal, Nov. 11, 2016, 11 pages, vol. 8, No. 6, IEEE. cited by applicant

Mehta, D. S. et al., "Laser speckle reduction by multimode optical fiber bundle with combined temporal, spatial, and angular diversity", Applied Optics, Apr. 11, 2012, pp. 1894-1904, vol. 51, No. 12, Optical Society of America. cited by applicant

Merritt, S. et al., "Monitoring temperature non-invasively using broadband Diffuse Optical Spectroscopy", OSA/FIO, 2004, 1 page, Optical Society of America, <https://opg.optica.org/abstract.cfm?URI=FiO-2004-FTuK4>. cited by applicant

Noriki, A. et al., "45-degree curved micro-mirror for vertical optical I/O of silicon photonics chip", Optics Express, Jul. 1, 2019, pp. 19749-19757, vol. 27, No. 14, Optical Society of America, <https://doi.org/10.1364/OE.27.019749>. cited by applicant

Poulton, C. V. et al., "Frequency-modulated Continuous-wave LIDAR Module in Silicon Photonics", OFC, 2015, 4 pages, Optical Society of America. cited by applicant

Redding, B. et al., "Compact spectrometer based on a disordered photonic chip", Nature Photonics, Jul. 28, 2013, pp. 746-751, vol. 7, Macmillan Publishers Limited. cited by applicant

Redding, B. et al., "Evanescently coupled multimode spiral spectrometer", Optica, Aug. 25, 2016, pp. 956-962, vol. 3, No. 9, Optical Society of America. cited by applicant

Robinson, M. B., "Interferometric diffuse correlation spectroscopy improves measurements at long source-detector separation and low photon count rate", Journal of Biomedical Optics, Sep. 30, 2020, pp. 097004-1 through 097004-12, vol. 25, No. 9, SPIE. cited by applicant

Roelkens, G. et al., "Transfer printing for silicon photonics transceivers and interposers", 2018 IEEE Optical Interconnects Conference, Jun. 4, 2018, pp. 13-14, IEEE. cited by applicant



Ryckeboer, E., “Spectroscopic Detection of Glucose with a Silicon Photonic Integrated Circuit”, Universiteit Gent, Jan. 1, 2014, 263 pages, ISBN 978-90-8578-688-7, [http://www.photonics.intec.ugent.be/download/phd\\_206.pdf](http://www.photonics.intec.ugent.be/download/phd_206.pdf). cited by applicant

Schneider, S. et al., “Optical coherence tomography system mass-producible on a silicon photonic chip”, Optics Express, Jan. 20, 2016, pp. 1573-1586, vol. 24, No. 2, Optical Society of America. cited by applicant

Shimotsu, S. et al., “Single Side-Band Modulation Performance of a LiNbO<sub>3</sub> Integrated Modulator Consisting of Four-Phase Modulator Waveguides”, IEEE Photonics Technology Letters, Apr. 2001, pp. 364-366, vol. 13, No. 4, IEEE. cited by applicant

Subramanian, A. Z. et al., “Silicon and silicon nitride photonic circuits for spectroscopic sensing on-a-chip [Invited]”, Photon. Res., Aug. 28, 2015, pp. B47-B59, vol. 3, No. 5, Chinese Laser Press. cited by applicant

Timm, U. et al., “Non-Invasive Optical Real-time Measurement of Total Hemoglobin Content”, Procedia Engineering, 2010, pp. 488-491, Elsevier Ltd. cited by applicant

Tran, T-T-K. et al., “Speckle reduction in laser projection displays through angle and wavelength diversity”, Applied Optics, Feb. 16, 2016, pp. 1267-1274, vol. 55, No. 6, Optical Society of America. cited by applicant

Tuchin, V., “Chapter 8: Coherent Effects at the Interaction of Laser Radiation with Tissues and Cell Flows”, Tissue Optics Light Scattering Methods and Instruments for Medical Diagnostics, 3rd Edition, 2015, pp. 359-417, SPIE. cited by applicant

U.S. Appl. No. 18/019,085, filed Jan. 31, 2023. cited by applicant

U.S. Appl. No. 18/027,881, filed Mar. 22, 2023. cited by applicant

U.S. Appl. No. 18/056,666, filed Nov. 17, 2022. cited by applicant

U.S. Notice of Allowance from U.S. Appl. No. 17/757,130, dated Jan. 18, 2023, 10 pages. cited by applicant

Valley, G.C. et al., “Multimode waveguide speckle patterns for compressive sensing”, Optics Letters, May 23, 2016, pp. 2529-2532, vol. 41, No. 11, Optical Society of America. cited by applicant

Van Gastel, M. et al., “Camera-based pulse-oximetry—validated risks and opportunities from theoretical analysis”, Biomedical Optics Express, Dec. 5, 2017, pp. 102-119, vol. 9, No. 1, Optical Society of America. cited by applicant

Website: “Optical Solutions”, Molex, dated 2023, printed May 10, 2023, 13 pages, Molex, LLC, <https://www.molex.com/en-us/products/optical-solutions>. cited by applicant

Website: “Track Your SpO<sub>2</sub> to Uncover Changes in Your Wellbeing”, Fitbit News, dated Sep. 7, 2020, printed Apr. 17, 2023, 7 pages, Fitbit, Inc., <https://blog.fitbit.com/track-your-spo2/>.626. cited by applicant

Wenz, J. J., “Examining water in model membranes by near infrared spectroscopy and multivariate analysis”, BBA—Biomembranes, Dec. 9, 2017, pp. 673-682, Elsevier B.V., <https://www.sciencedirect.com/science/article/pii/S0005273617303905>. cited by applicant

Xu, M. et al., “Laser Speckle Reduction Using a Motionless Despeckle Element Based on Random Mie Scattering”, Journal of Display Technology, Nov. 12, 2013, pp. 151-156, vol. 10, No. 2, IEEE. cited by applicant

Yamakoshi, Y. et al., “Side-scattered finger-photoplethysmography: experimental investigations toward practical noninvasive measurement of blood glucose”, Journal of Biomedical Optics, Jun. 2017, pp. 067001-1 through 067001-11, vol. 22, No. 6, SPIE. cited by applicant

Yao, Z. et al., “Integrated Silicon Photonic Microresonators: Emerging Technologies”, IEEE Journal of Selected Topics in Quantum Electronics, Jun. 11, 2018, 24 pages, vol. 24, No. 6, IEEE. cited by applicant

Zhang, J. et al., “III-V-on-Si photonic integrated circuits realized using micro-transfer-printing”, APL Photonics, Nov. 4, 2019, pp. 110803-1 through 110803-10. cited by applicant

Zijlstra, W. G. et al., "Absorption Spectra of Human Fetal and Adult Oxyhemoglobin, De-Oxyhemoglobin, Carboxyhemoglobin, and Methemoglobin", Clinical Chemistry, Sep. 1991, pp. 1633-1638, vol. 37, No. 9, <https://academic.oup.com/clinchem/article-abstract/37/9/1633/5649610?redirectedFrom=fulltext>. cited by applicant

Zilkie, A. J. et al., "Multi-Micron Silicon Photonics Platform for Highly Manufacturable and Versatile Photonic Integrated Circuits", IEEE Journal of Selected Topics in Quantum Electronics, Apr. 15, 2019, 13 pages, vol. 25, No. 5, IEEE. cited by applicant

Zilkie, A. J. et al., "Power-efficient III-V/Silicon external cavity DBR lasers", Optics Express, Sep. 27, 2012, pp. 23456-23462, vol. 20, No. 21, Optical Society of America. cited by applicant

U.S. Appl. No. 18/667,608, filed May 17, 2024. cited by applicant

U.S. Office Action from U.S. Appl. No. 17/934,502, dated Aug. 31, 2023, 6 pages. cited by applicant

U.S. Office Action from U.S. Appl. No. 17/934,502, dated Feb. 1, 2024, 6 pages. cited by applicant

Ge, Z. et al., "Dynamic laser speckle analysis using the event sensor", Applied Optics, Dec. 23, 2020, pp. 172-178, vol. 60, No. 1, Optical Society of America. cited by applicant

Ghijsen, M. et al., "Wearable speckle plethysmography (SPG) for characterizing microvascular flow and resistance", Biomedical Optics Express, Jul. 30, 2018, pp. 3937-3952, vol. 9, No. 8, Optical Society of America under the terms of the OSA Open Access Publishing Agreement. cited by applicant

International Search Report and Written Opinion of the International Searching Authority, mailed Feb. 14, 2023, corresponding to PCT/EP2022/082162, 33 pages. cited by applicant

Lai, M. et al., "Perfusion Monitoring by Contactless Photoplethysmography Imaging", 2019 IEEE 16th International Symposium on Biomedical Imaging (ISBI 2019), Venice, Italy, Apr. 8-11, 2019, pp. 1778-1782, IEEE. cited by applicant

Liu, X. et al., "Simultaneous measurements of tissue blood flow and oxygenation using a wearable fiber-free optical sensor", Journal of Biomedical Optics, Jan. 29, 2021, pp. 012705-1 through 012705-15, vol. 26, No. 1, SPIE. cited by applicant

Lu, H. et al., "Single-trial estimation of the cerebral metabolic rate of oxygen with imaging photoplethysmography and laser speckle contrast imaging", Optics Letters, Mar. 17, 2015, pp. 1193-1196, vol. 40, No. 7, Optical Society of America. cited by applicant

U.S. Appl. No. 17/934,502, filed Sep. 22, 2022. cited by applicant

U.S. Office Action from U.S. Appl. No. 17/934,502, dated Jun. 5, 2024, 7 pages. cited by applicant

---

*Primary Examiner:* Kuo; Jonathan T

*Attorney, Agent or Firm:* Womble Bond Dickinson (US) LLP

---

## **Background/Summary**

CROSS-REFERENCE TO RELATED APPLICATIONS (1) The present application claims priority to and the benefit of U.S. Provisional Application No. 63/279,932, filed Nov. 16, 2021, entitled "A COMBINED OPTICAL SENSOR MODULE", the entire content of which is incorporated herein by reference.

### **FIELD**

(1) Embodiments of the present invention relate to an optical sensor, particularly to an optical sensor module for measuring both speckleplethysmography (SPG) and photoplethysmography (PPG) signals at human tissue.

### **BACKGROUND**

(2) Photoplethysmography (PPG) signals are currently utilized to extract cardiovascular parameters such as heart rate (HR), pulsatile oxygen saturation (SpO<sub>2</sub>) and blood pressure (BP). PPG signals arise from the change in tissue absorption caused by blood volume variation during pulsatile flow. These AC signals are generally of low magnitude as compared to a DC background. This is especially true on the wrist, where the pulsatile signal may be on the order of 0.4% or less than the measured intensity at red wavelengths. Red and near infra-red (NIR) light sources are required to obtain SpO<sub>2</sub> values. Most wearables that provide SpO<sub>2</sub> at wrist do so only during a person's quiescent periods, such as at night, where many pulses can be averaged in order to overcome low signal quality. To obtain daily beat-to-beat heartbeat values, many PPG modules include a green LED that has higher responsivity to changes in blood absorption. However, the green light does not penetrate very deeply and the signal quality is impacted by pressure that reduces circulation in the microvasculature. It is also undesirable for many people to observe a bright green light in a wearable device during the night.

(3) Furthermore, it is possible that blood pressure may be extracted with more accuracy and reliability from blood flow data (speckleplethysmography, SPG) or a combination of blood flow and PPG than from PPG signals alone. For this reason, it is advantageous to simultaneously collect both SPG and PPG signals for the purposes of obtaining HR, HRV, SpO<sub>2</sub>, and BP measurements. To obtain SpO<sub>2</sub>, either or both of SPG and PPG may be collected from at least two wavelengths. For co-oximetry measurements in which more than one form of modified hemoglobin is measured, such as methemoglobin and carboxyhemoglobin, many more than 2 wavelengths may be utilized, often 6 or more.

(4) It may be desirable to obtain a host of cardiovascular parameters from one compact, body-worn module. This may be achieved through the combination of SPG and PPG signals. Previously, researchers have demonstrated a combined system for measuring blood flow and tissue oxygenation by utilizing two laser diodes (red, NIR/IR) and a CMOS image sensor [Liu et al. J. Biomed. Opt. 26(1) 012705-1, 2021]. The calculation of SPG via spatial contrast measurements involves measuring both the standard deviation and intensity of the speckle image and calculating the ratio. Therefore, the calculation uses a PPG signal, which is directly related to the measured image intensity. However, this method is not optimal for resolving the pulsatile PPG signal at medium to low blood volumes, such as would be the case for a consumer wearable device. Indeed, the authors only demonstrated tissue oxygen saturation, StO<sub>2</sub>, which is not the resolved pulsatile SpO<sub>2</sub> signal.

(5) Systems that are optimized for PPG measurements in consumer wearables include DC and ambient light subtraction prior to amplifying the AC signal in order to improve dynamic range and sensitivity. Additionally, the photodiodes used are generally large format (e.g. 2×3 mm) to increase the number of detected photons. Finally, it is typical to use LEDs which, unlike lasers, do not produce speckle noise in intensity measurements.

## SUMMARY

(6) Accordingly, embodiments of the present invention aim to solve the above problems by providing, according to one or more embodiments of a first aspect, an optical sensor module for measuring both speckleplethysmography (SPG) and photoplethysmography (PPG) signals at human tissue, the optical sensor module comprising: a first light source, for illuminating the human tissue for use with SPG measurements, the first light source comprising a laser; a second light source, for illuminating the human tissue for use with PPG measurements; and one or more optical sensor(s) for receiving light from the illuminated human tissue.

(7) Herein, when it is stated that the optical sensor module is for measuring SPG and PPG signals “at human tissue”, it should be understood that the light from the first and second laser sources is incident on the human tissue, and that the one or more optical sensors are configured to receive, and measure, light which is reflected and/or scattered from the surface of the human tissue, or from components of the tissue beneath the surface, and which is transmitted back through the tissue in a

reflectance measurement geometry, or light that is only transmitted through the tissue in a transmission measurement geometry. In some cases, the one or more optical sensors may be arranged, in use, to be in contact with a user's skin, in which case the light from the first light source and the second light source may enter the user's tissue through the skin, and be reflected or scattered from components of the tissue beneath the skin, and then travel back through the tissue, whereupon, on being transmitted back through the skin, it may be detected by the one or more optical sensors. In alternative cases, the one or more optical sensors may be configured to be spaced from the user's skin in use, in which case the received light may further include light which is reflected from the surface of the user's skin.

(8) In other cases, the one or more optical sensors may be configured to receive light which has been transmitted through the tissue. In those cases, the one or more optical sensors may still be arranged, in use, either to be in contact with the user's skin or spaced from the user's skin.

Transmission may, for example, be transmission from one side of a finger, ear, or toe to the other.

(9) Some embodiments of the invention relate to illumination of organic tissue. This may be "human tissue" or "animal tissue". Herein, "human tissue" or "animal tissue" may refer to blood (e.g. blood cells or components thereof, such as the cell membranes), and elements of the vasculature (e.g. arteries, veins, capillaries, or walls thereof). It will be appreciated that different physiological parameters may be measured or otherwise determined based on illumination of different types of human tissue—this is discussed later in the application.

(10) Optional features of the invention will now be set out. These are applicable singly or in any combination with any aspect of the invention.

(11) Optionally, the first light source is a laser with a wavelength of operation lying within a red wavelength, which may be thought of as the range of 600 nm to 1000 nm, or the range from 620 nm to 1000 nm.

(12) Optionally, the wavelength of operation of the laser is 660 nm or 760 nm.

(13) Optionally, the second light source is an LED.

(14) Optionally, the second light source is an LED operating at infra-red (IR) wavelengths (e.g. >800 nm).

(15) Optionally, the second light source is an LED operating at a red wavelength, e.g. a wavelength within the range of 620 nm to 1000 nm.

(16) Optionally, the first and second light sources are both lasers and are located on the same photonic integrated circuit (PIC).

(17) Optionally, the first light source is a laser having a first wavelength, the second light source is an LED operating at a second wavelength, the optical sensor further comprising a third light source, the third light source comprising an LED operating at the first wavelength or a similar wavelength to the first wavelength.

(18) Optionally, the first light source is a laser having a red wavelength within the range of 620 to 800 nm, the second light source is an LED operating at an IR wavelength within the range of 800 to 1000 nm, the optical sensor further comprising a third light source, the third light source comprising an LED operating at red wavelength within the range of 620 to 800 nm.

(19) Optionally, the first and second light sources and the one or more optical sensor(s) are configured to carry out SPG and PPG measurements simultaneously or near-simultaneously.

(20) Optionally, the one or more optical sensor(s) comprises an image sensor.

(21) Optionally, the same image sensor is used to extract measurements from both the first light source and the second light source.

(22) Optionally, the one or more optical sensor(s) is configured to carry out one or more of the following: in-pixel ambient/DC subtraction; near pixel ambient DC subtraction; pixel block statistics calculation; and/or pixel array statistics calculation.

(23) Optionally, the one or more optical sensor(s) includes a processor configured to process captured data in-device and generate PPG and/or SPG output data.

(24) Optionally, the one or more optical sensor(s) comprises an event-based image sensor.

(25) Optionally, the one or more optical sensor(s) comprises a photodiode and separate sensor (e.g. CMOS).

(26) Optionally, the optical sensor module further comprises one or more processors configured to convert optical measurement(s) at the one or more optical sensor(s) to measurements of one or more of the following: blood pressure, SpO<sub>2</sub>, arterial stiffness, heart rate, heart rate variability, atrial fibrillation, bradycardia, tachycardia, and/or movement such as steps taken or gestures.

(27) Optionally, the optical sensor module may be located or locatable on a consumer wearable, typically understood to have a small form factor.

(28) Optionally, the optical sensor may be located on or as part of a module. The module may be part of a strap or attached to a strap such that measurements are taken over the radial or ulnar arteries of the wrist.

(29) Optionally, the optical sensor module may be located on a wrist strap of a wearable device.

(30) Optionally, when the wearable device is located on the wrist of a user, one or more of the optical sensor(s) are located over the radial artery of the user.

(31) Optionally, the wearable device includes a timepiece, and a strap that connects to the timepiece, and the entire optical sensor module is located on the strap. In this way, the smart strap may advantageously be used in combination with analogue timepieces. That is to say, the operation of the smart strap can be completely separate from the operation of the timepiece. This may be advantageous since there is limited space on the back of the wrist combined with user tolerance for stack height.

(32) In addition, there remains a strong desire for analog timepieces, often in the higher price point market. Consumers must forgo health-related benefits to enjoy such timepieces, or wear two watches; a smartwatch that provides wellness/cardiovascular metrics and a quality analog timepiece.

(33) Some embodiments of the present invention also provide for physiological benefits over prior art devices which may incorporate optical sensor(s) onto the timepiece itself to be located at the back of the wrist. The back of the wrist is actually not the ideal place from which to acquire biophotonic measurements owing to low vascularization. By moving the optical sensors to a radial or ulnar site, it is possible to better utilise PPG signals at red wavelengths and obtain SpO<sub>2</sub> sensors with stronger performance. This results in more accurate measurements, for example of heart rate, where prior art monitors located at the back of the wrist typically use green light to access the surface capillaries and utilize the higher absorption of hemoglobin in the green wavelength range. Although red wavelengths are known for SpO<sub>2</sub> measurements, this typically takes place at the fingertip of a user, because of the significantly higher vascularization in that location.

(34) A smart strap according to one or more embodiments of the present invention would provide many cardiovascular parameters such as Heart Rate (HR), Heart Rate Variability (HRV), SpO<sub>2</sub>, Blood Pressure (BP). The strap may rely on a small form factor PIC, application specific integrated circuit (ASIC) and flexible electronic substrate. It is envisioned that the data for measuring the parameters can be obtained with just 2 or 3 laser wavelengths in the red and NIR regions by combining SPG and PPG information. Space is required in the strap for Bluetooth, battery, and other features normal in such a wearable device.

(35) Optionally, the optical sensor may further comprise one or more additional light sources. In this way, the light source is configured to be capable of operating at more than two wavelengths. By providing multiple wavelengths at the sensor, it would be possible to carry out co-oximetry readings. A co-oximeter may use six or more wavelengths to measure the oxygen carrying state of haemoglobin in the blood of a user.

(36) According to one or more embodiments of a second aspect of the present invention, there is provided, a wearable device comprising an optical sensor module according to any one of the embodiments herein.

- (37) Optionally, the wearable device may further comprise one or more additional optical sensor(s), the one or more optical sensors corresponding to the optical sensor(s) of any one of the embodiments described herein. This may, for example provide a device capable of performing co-oximetry.
- (38) According to one or more embodiments of a third aspect of the present invention, there is provided, a strap (a “smart strap”) comprising an optical sensor module according to any one or more of the embodiments described herein.
- (39) According to one or more embodiments of a fourth aspect of the present invention, there is provided a wearable device comprising two or more bio-monitoring circuits, each biomonitoring circuit comprising a respective light source and sensor. Such a wearable device could incorporate any one or more of the optional features described herein.
- (40) According to one or more embodiments of a fifth aspect of the present invention, there is provided a strap for a wearable device, the strap comprising two or more bio-monitoring circuits, each biomonitoring circuit comprising a respective light source and sensor. Such a strap could incorporate any one or more of the optional features described herein.
- (41) Further optional features of embodiments of the invention are set out below.
- 

## Description

### BRIEF DESCRIPTION OF THE DRAWINGS

- (1) Embodiments of the invention will now be described by way of example with reference to the accompanying drawings in which:
- (2) FIG. 1 shows a schematic diagram of an optical sensor module according to an embodiment of the present invention;
- (3) FIG. 2 shows a schematic diagram of an optical sensor module according to a further embodiment of the present invention;
- (4) FIG. 3 shows a schematic diagram of an optical sensor module according to a further embodiment of the present invention;
- (5) FIGS. 4A and 4B each show a schematic diagram of three further optical sensor modules according to further embodiments of the present invention;
- (6) FIG. 5 shows a representation of module operation according to a first example;
- (7) FIG. 6 shows a representation of module operation according to a second example;
- (8) FIG. 7 shows a representation of module operation according to a third example;
- (9) FIG. 8 depicts (a) an illustration of the operation of in-pixel ambient/DC removal, and (b) an illustration of the operation of an event image sensor;
- (10) FIG. 9 depicts SPG and ECG data from a first subject;
- (11) FIG. 10 depicts SPG and ECG data from a second subject;
- (12) FIG. 11 depicts SPG data from a third subject;
- (13) FIG. 12 depicts SPG data from a fourth subject;
- (14) FIG. 13 depicts an example process followed where the optical sensor module of some embodiments of the present invention is configured to carry out pixel array statistics;
- (15) FIG. 14 depicts an example process followed where the optical sensor module of some embodiments of the present invention is configured to carry out event-based detection;
- (16) FIG. 15 depicts a smart strap comprising an optical sensor according to some embodiments of the present invention, the smart strap being made of a non-conductive material; and
- (17) FIG. 16 depicts a smart strap comprising an optical sensor according to some embodiments of the present invention, portions of the smart strap being made of a conductive material.

### DETAILED DESCRIPTION

- (18) SPG signals are obtained from blood flow speed and are generally of larger magnitude than

PPG signals at the same wavelength. Additionally, because the SPG signal is obtained primarily from deeper lying vasculature with faster flow, it is less impacted by applied pressure. The SPG signal is comprised of very sharp peaks when collected at >20 Hz and preferably >100 Hz frame rate and so provides an excellent means of measuring HR. A cardiovascular module that combines SPG and PPG signals may provide numerous cardiovascular metrics including blood pressure, SpO<sub>2</sub>, arterial stiffness, heart rate, heart rate variability, atrial fibrillation, bradycardia, and tachycardia with just red and “IR” sources. Therefore, a green LED and/or dedicated detector for visible light is not required, which saves both battery consumption and space. Furthermore, the blood flow (SPG) measurement is very sensitive to motion and provides signals that may be interpreted as steps, gestures, or other movement-related phenomena. Thus, the module may or may not include an accelerometer.

(19) Since the hemoglobin absorption band is quite steep in the region of 600-700 nm, SpO<sub>2</sub> accuracy depends on the accuracy of the wavelength of the red light. Devices that utilize LEDs, with resulting broadband emission and sensitivity to temperature, may come out of calibration and give erroneous results. SPG signals require a coherent laser as light source. Lasers have much narrower emissions and less sensitivity to temperature, thereby providing a more accurate SpO<sub>2</sub> over a wide range of conditions. Note that the “IR” light source (>800 nm) may be broader because haemoglobin absorption is relatively flat in the region 830-900 nm.

(20) The image sensor can incorporate methods of in-pixel cancellation of DC light signals (from non-pulsatile light) or light from the ambient environment. In some embodiments, an event camera sensor may be used in place of a conventional image sensor for the detection of the SPG and PPG signals. Such a sensor offers advantages in lower power usage and processing requirements due to its capability of providing pixel change updates instead of full frame updates.

(21) In some embodiments, the invention includes, but is not limited to a cardiovascular module that provides blood pressure, SpO<sub>2</sub>, arterial stiffness, heart rate, heart rate variability, atrial fibrillation, bradycardia, and tachycardia parameters, in a wearable, compact form factor. This module is optimized for both PPG and SPG signal collection to obtain robust measurements.

(22) FIG. 1 shows an optical sensor module, **101** comprising a first light source **1**, for illuminating the human tissue for use with SPG measurements. In this embodiment, the first light source is a laser having sufficient coherence length to acquire SPG signals (e.g. vertical-cavity surface-emitting laser (VCSEL), distributed feedback (DFB) diode, distributed Bragg reflector (DBR) diode, volume holographic (VHG) diode).

(23) The optical sensor module also includes a second light source **2** for illuminating the human tissue for use with PPG measurements. In the embodiment shown, the second light source takes the form of an infra-red LED (e.g. having a center wavelength of 830-980 nm).

(24) Including the second light source in the form of an LED may be advantageous as it is less susceptible to error due to wavelength drift owing to the flat hemoglobin spectrum in that region.

(25) A plurality of optical sensor(s) is also located on the optical module for receiving light from the illuminated human tissue. In this embodiment, these sensors include a photodiode, **3** and a CMOS image sensor, **4**. The photodiode may take the form of a large area photodiode with DC subtraction electronics. Electronics for the photodiode are present and may be optimized for PPG data collection to obtain SpO<sub>2</sub> whereas the CMOS sensor may be utilized for the SPG signal collection to obtain BP and other cardiovascular parameters. This embodiment does not contain a green LED, although it would be possible to adapt it (not shown) to include a green or other wavelength LED or laser either in addition, or as a replacement to the components shown in FIG. **1**.

(26) An optical sensor module, **201** according to a second embodiment is described below in relation to FIG. **2**, the second embodiment, differing from the embodiment of FIG. **2** in that the photodiode **3** is removed, and the PPG signal is instead obtained using an image sensor.

(27) The image sensor may be a charge-coupled device (CCD), a CMOS image sensor (CIS), or an

implementation of a CCD or CIS incorporating an in-pixel DC (non-pulsatile) or ambient light subtraction method such as but not limited to auto-zeroing and chopping, common mode reset, minimum charge transfer, or dual transfer gate architecture.

(28) Alternatively, the image sensor may be an event image sensor (EIS) which produces asynchronous pixel updates according to defined pixel intensity changes as opposed to conventional synchronous frame-based CCD or CIS sensors.

(29) A third embodiment of an optical sensor module **301** is described below with reference to FIG. **3**. In this embodiment, both the first and second light source are lasers. The module comprises only the image sensor described in the second embodiment and a photonic integrated chip (PIC) **5** containing at least two lasers, one laser having a red wavelength and one laser having a NIR/IR wavelength. Red wavelength may be taken to be light with a wavelength from 620 nm to 800 nm. NIR/IR may be taken to be light with a wavelength of 800 nm to 1000 nm. The PIC may, in addition, contain multiple laser wavelengths to also perform CO-oximetry (carboxyhemoglobin, methemoglobin, etc). As with the embodiment of FIG. **2**, a single image sensor is used to detect both PPG and SPG signals.

(30) Any one or more of the embodiments described herein may utilize temporary speckle mitigation techniques for the collection of the PPG signal obtained from a photodiode, such as a deformable mirror to rapidly adjust optical pathlengths, an optical phase array, raster scanning, angle scanning, wavelength scanning or broadening, or any combination thereof. This may improve SNR as it reduces speckle noise from the intensity signal, which is more important as the size of the sensor/detector is reduced. Additionally, or alternatively, multiple photodiode acquisitions of the laser signal may be collected and averaged to improve SNR.

(31) Any one or more of the embodiments may also include a multi-aperture array in front of the image sensor to improve SNR while maintaining the appropriate speckle to pixel ratio by virtue of the aperture diameters and distance from the sensor. Such a sensor does not require a lens to obtain the appropriate speckle to pixel ratio, although a lens or lens array may be used in conjunction with the aperture plate.

(32) It is to be understood that the embodiments described are not limitations, e.g. green, blue, yellow, or other wavelength LEDs or lasers may be included and any LED may be replaced with a laser. The SPG signal requires, at minimum, one laser of any wavelength be present in the system along with at least one image sensor. Alternatively, in place of an image sensor that is required for speckle spatial contrast measurements, the SPG signal may be obtained by diffuse correlation spectroscopy (DCS) or interferometric diffuse correlation spectroscopy (iDCS). DCS or iDCS requires either a photodiode or balanced receiver with >500 kHz sampling rate or a single photon counting avalanche detector (SPAD) detector.

(33) Additional modalities may be added to this module, such as ECG (electrocardiography) sensor(s), which may be utilized in addition to the SPG and/or PPG signals to calculate pulse arrival times to aid in BP estimations. ECG may also be utilized for certain arrhythmia detections as part of the suite of cardiovascular parameters.

(34) Three further optical modules are shown in FIGS. **4A** and **4B**. FIG. **4B** differs from FIG. **4A** in that it shows the link between each light source and its respective detector. Each module ((i), (ii), (iii)) includes an image sensor **403** (e.g. a CMOS image sensor (CIS)), a second sensor **404** (e.g. a photodiode), a laser (e.g. a VCSEL), and one or more LEDs (e.g. a red and an infra-red LED). A module may take the form of discrete components mounted to a printed circuit board assembly (PCBA). Example dimensions are shown and are approximate, based on discreet component sizes. However, it is important to note that these are only illustrative examples, and that other sizes would be possible.

(35) Placement may be over an artery, **15**, such as the radial or arterial artery. The arrangement shown in FIG. **4A** (iii) (and corresponding FIG. **4B** (iii)) is advantageous as the artery passes in the middle of the space, or “banana” (**401**, **402**) formed between the source and detector for both the



SPG and PPG modules. In the particular embodiment shown, the SPG and PPG modules are located separately on the same printed circuit board assembly (PCBA), with the PPG module corresponding to the LED(s) and photodiode, and the SPG module corresponding to the laser (e.g. VCSEL) and the image sensor (e.g. CIS).

(36) It has been found to be advantageous for the modules to be stacked vertically with respect to an artery (i.e. along the artery when in use) in order to maximize light of similar source-detector separation. Again, such an arrangement is shown in FIG. 4A (iii) and FIG. 4B (iii).

(37) A first example of module operation is described below in relation to FIG. 5. In this example, a first laser has a wavelength of 785 nm and separate LEDs having red (660 nm) and IR (900 or 940 nm) wavelengths respectively are also present. This gives a total of three different wavelengths used to interrogate the surface. In the example shown, the LED data acquisition is fitted within the period after camera integration ( $t_{\text{sub.INT}}$ ) and before max frame rate, assuming that the camera integration time is less than the period required for max frame rate operation. A time delay  $t_{\text{sub.SETTLE}}$  is also shown on FIG. 5, this time delay being caused by laser or light intensity stabilisation.

(38) A further example is shown in FIG. 6, the optical module comprising a single laser and a single LED, thereby operating at two wavelengths. The laser generates light of a red (e.g. 660 nm) wavelength and the LED generates light of an IR (e.g. 900 or 940 nm) wavelength.

(39) For convenience, operation parameters can be chosen to keep sampling rate of the IR signal the same for the PD and to use the photodiode to acquire multiple collections of the red laser in order to improve SNR due to speckle noise. Depending on the integration times used, it may be possible to carry out two or more red PD acquisitions.

(40) A third example of an operation of an optical module is shown in FIG. 7. In this embodiment, the optical module includes a laser module or PIC configured to generate light of two different wavelengths (R and IR), and a single sensor such as a CMOS image sensor. Integration of the signal ( $t_{\text{sub.INT}}$ ) by the sensor starts at a time ( $t_{\text{sub.SETTLE}}$ ) after the start of the irradiation of the sample by the laser. The PIC may contain two lasers to generate both red (660 nm) and IR (960 nm) light. It is possible that more than two wavelengths may be generated by the PIC, in which case the operation shown in FIG. 7 can be applied, and each wavelength cycled through in a given order.

(41) A method by which the sensor can mitigate the effect of ambient background light is shown in FIG. 8.(a). The sensor can sample (SAMPLE-B) the ambient background light with the laser drive off (LASERDRV). The sensor can subsequently sample with the laser on (SAMPLE-S), the resulting reading being a combination of the signal and the ambient background light. Subtraction of the two readings results in a readout (READOUT) consisting of only the contribution of the signal. The subtraction can be accomplished by, but not limited to, in pixel charge subtraction or post-photoconversion voltage subtraction.

(42) An example of a speckle pattern measured by an event based imaging device is shown in FIG. 8.(b). The event based image sensor functions by the detection of discrete changes in pixel intensity (events). The temporal fluctuation of the set of speckle pebbles ( $x_i, y_i$ ) translates to the generation of an asynchronous series of pixel events. These events can then be processed by either, but not limited to, integration over a fixed time interval to create a corresponding frame of pixel intensities or in an online fashion by which a histogram of pixel update event rates relates to the temporal fluctuation of the speckle pattern.

(43) Examples of SPG (top) and ECG (second from top) data from two subjects are shown in FIGS. 9 and 10 respectively, the data showing equivalency in performance of measuring HR (third from top) and HRV (bottom).

(44) Examples of SPG signals (top) contain both high and low-frequency components of accelerometer data (X, Y, and Z axes in second, third, and fourth plots from the top) are shown in FIGS. 11 and 12 for two respective subjects. Motion detection and tracking may be achieved via

the SPG signal and without the need for an accelerometer.

(45) The processing of data from one or more sensor pixel arrays can be better understood with reference to the flow diagrams shown in FIGS. **13** and **14**.

(46) In the embodiment shown in FIG. **13**, a sensor pixel array is first used to detect ambient background light with the light source turned off (s**131**). Speckle light is then detected (s**132**) by the sensor pixel array, when the light source is turned on. Ambient background light can then be subtracted and/or a DC set value can be subtracted from the speckle pixel array data (s**133**). For each pixel individually, or in a column fashion, amplification (s**134**) may be carried out to maximize the dynamic range of the data before digitization (s**135**). At digitization, pixel voltage is converted to digital unit via an analogue to digital converter. A processor may then calculate (s**136**) the sensor pixel array statistics, and may do so in blocks of a set size. Once these statistics have been calculated, a single data point can be obtained (s**137**), the single data point corresponding to SPG and PPG signals.

(47) Calculation (s**136**) of the sensor pixel array statistics may include the steps of: subdividing (s**138**) the pixel array into  $N \times N$  blocks; for each block, calculating (s**139**) the pixel value mean and variance and calculating speckle contrast (K) for the block; and calculating (s**140**) averages of speckle contrast and intensity across all blocks.

(48) A process including event-based detection is described below in relation to FIG. **14**. The process differs from that described above in relation to FIG. **13** in that the first two steps (s**131**, s**132**) are replaced by three new steps (s**141**, s**142**, s**143**). Firstly, ambient background light is detected (s**141**) through the sensor pixel array with the light source off, then speckle light is detected (s**142**) through the pixel array with the light source on. This may be achieved by pixel event-based detection. Next, speckle pixel events are accumulated (s. **143**) over a specified time, to create a sparse speckle pixel array.

(49) In one or more embodiments of the present invention, the optical sensor module is located on a wrist strap **1501** of a wearable device. In some of these embodiments, the optical sensor module is entirely located on a smart strap, the smart strap being a strap that includes all of the electronics and processing required by the optical sensor and can thus function completely separately from any timepiece that is to be connected to the strap. In this way, the smart strap may be used in combination with (e.g. by retrofitting onto) any pre-existing timepiece including analogue timepieces.

(50) A first smart strap is shown in FIG. **15**, the smart strap **1501** being connectable to a timepiece **1502**. This embodiment is typical of non-metal (e.g. leather) straps. The optical sensor module may be located within a region **1503**, **1504** on one or both sides of the strap such that the optical sensors lie over the radial and/or ulnar artery (arteries) of the user.

(51) In an alternative smart strap shown in FIG. **16**, the smart strap **1601** is made of a conductive material such as metal. Again, components of the optical sensor module may be located along the strap, and typically along the entire region **1603** of the strap. Any optical sensors of the optical sensor module may be placed so that when the conductive strap (or portionally conductive strap) is in use, being worn by the user, the one or more sensors are located above the radial and/or ulnar artery (arteries) of the user.

(52) On a conductive (or portionally conductive), metal strap, the possibility exists for the sensing and electronic elements such as battery, Bluetooth, etc to be spread out, using the clasp (volar wrist) or other contact points along the strap as an electrical connection

(53) It is possible to optimally design a strap using any external material (leather, silicone, plastic, metal) with sensing and electronics fully enclosed within and using the clasp for electrical attachment. The attachment area (dorsal wrist to timepiece/smartwatch) would be standard lug or compression spring or other common attachment mechanism.

(54) While the invention has been described in conjunction with the exemplary embodiments described above, many equivalent modifications and variations will be apparent to those skilled in

the art when given this disclosure. Accordingly, the exemplary embodiments of the invention set forth above are considered to be illustrative and not limiting. Various changes to the described embodiments may be made without departing from the spirit and scope of the invention.

(55) All references referred to above are hereby incorporated by reference.

## Claims

1. A wearable device comprising an optical sensor module, the optical sensor module comprising: a light source comprising a laser, the light source being configured to selectively apply speckle mitigation techniques to light generated by the laser and to illuminate tissue; an image sensor configured to selectively receive, through the tissue, a first light signal from the light source when the light source does not apply the speckle mitigation techniques; a photodiode configured to selectively receive, through the tissue, a second light signal from the light source when the light source does apply the speckle mitigation techniques; and a processor configured to process first data, corresponding to the first light signal, to generate a speckleplethysmography (SPG) measurement, and to process second data, corresponding to the second light signal, to generate a photoplethysmography (PPG) measurement.
2. The wearable device of claim 1, wherein the image sensor is configured to carry out in-pixel DC subtraction.
3. The wearable device of claim 1, wherein the image sensor is configured to carry out a pixel block statistics calculation or a pixel array statistics calculation.
4. The wearable device of claim 1, wherein the light source is configured to apply the speckle mitigation techniques via a deformable mirror, an optical phase array, raster scanning, angle scanning, wavelength scanning, wavelength broadening, or any combination thereof.
5. The wearable device of claim 1, including a wrist strap, the optical sensor module being coupled to the wrist strap.
6. A wearable device comprising an optical sensor module, the optical sensor module comprising: a light source comprising a laser and being configured to illuminate tissue; one or more optical sensors configured to respectively receive, through the tissue, one or more light signals from the light source, the one or more optical sensors including an image sensor configured to carry out in-pixel DC light subtraction to subtract out a non-pulsatile component of a first light signal of the one or more light signals; and a processor configured to process data to generate a speckleplethysmography (SPG) measurement and a photoplethysmography (PPG) measurement, the data corresponding to the one or more light signals.
7. The wearable device of claim 6, wherein the image sensor is configured to receive, through the tissue, the one or more light signals from the light source.
8. The wearable device of claim 6, wherein the light source is configured to selectively apply speckle mitigation techniques to light generated by the laser.
9. The wearable device of claim 8, wherein the one or more optical sensors are configured to selectively receive the one or more light signals.
10. The wearable device of claim 6, the optical sensor module comprising: a coherent optical output configured to illuminate tissue with coherent light; and an incoherent optical output configured to illuminate the tissue with incoherent light, wherein the laser is configured to generate light at the coherent optical output, and the optical sensor module is configured to apply one or more speckle mitigation techniques to light generated by the laser to provide light at the incoherent optical output, wherein the image sensor is configured to receive, through the tissue, the first light signal from the coherent optical output and a second light signal from the incoherent optical output, and wherein the optical sensor module is configured to process first data and second data to respectively generate (SPG) the SPG measurement and the PPG measurement, the first data corresponding to the first light signal, and the second data corresponding to the second light signal.

11. The wearable device of claim 10, wherein the image sensor includes a CMOS image sensor.
  12. The wearable device of claim 10, wherein the image sensor includes an event-based camera.
  13. The wearable device of claim 10, wherein the image sensor is configured to carry out a pixel block statistics calculation or a pixel array statistics calculation.
-