

US Patent & Trademark Office

Patent Public Search | Text View

United States Patent	12414217
Kind Code	B2
Date of Patent	September 09, 2025
Inventor(s)	Ru; Guoyun et al.

Systems and methods for adaptively controlling filament current in an X-ray tube

Abstract

Systems and methods of adaptively controlling filament current in an x-ray tube of an imaging system include the x-ray tube having a filament being calibrated. Calibration data from the calibration of the x-ray tube is stored at the imaging system, the calibration data including a filament current value that determines a tube current value for a tube voltage value at a plurality of stations. A resistance value of the filament over a period of time is monitored. A change in the resistance value of the filament over the period of time is determined, and the filament current value of at least one of the plurality of stations is adjusted based on the changed resistance value.

Inventors: Ru; Guoyun (Farmington, CT), Aizer; David (Danbury, CT), Fraley; J. Austin (Danbury, CT), Nonnweiler; Edward (Brookfield, CT)

Applicant: Hologic, Inc. (Marlborough, MA)

Family ID: 85175662

Assignee: Hologic, Inc. (Marlborough, MA)

Appl. No.: 18/100683

Filed: January 24, 2023

Prior Publication Data

Document Identifier	Publication Date
US 20230251210 A1	Aug. 10, 2023

Related U.S. Application Data

us-provisional-application US 63307311 20220207

Publication Classification

Int. Cl.: H05G1/26 (20060101); A61B6/58 (20240101); H05G1/34 (20060101); H05G1/54 (20060101); A61B6/50 (20240101)

U.S. Cl.:

CPC H05G1/34 (20130101); A61B6/582 (20130101); H05G1/265 (20130101); H05G1/54 (20130101); A61B6/502 (20130101)

Field of Classification Search

CPC: H05G (1/265); H05G (1/34); H05G (1/54); A61B (6/502); A61B (6/582)

References Cited

U.S. PATENT DOCUMENTS

Patent No.	Issued Date	Patentee Name	U.S. Cl.	CPC
3365575	12/1967	Strax	N/A	N/A

3502878	12/1969	Stewart	N/A	N/A
3863073	12/1974	Wagner	N/A	N/A
3971950	12/1975	Evans et al.	N/A	N/A
4160906	12/1978	Daniels et al.	N/A	N/A
4310766	12/1981	Finkenzeller et al.	N/A	N/A
4334153	12/1981	Stehman	N/A	N/A
4380086	12/1982	Vagi	N/A	N/A
4496557	12/1984	Malen et al.	N/A	N/A
4513433	12/1984	Weiss et al.	N/A	N/A
4542521	12/1984	Hahn et al.	N/A	N/A
4559641	12/1984	Caugant et al.	N/A	N/A
4593371	12/1985	Grajewski	N/A	N/A
4662379	12/1986	Macovski	N/A	N/A
4706269	12/1986	Reina et al.	N/A	N/A
4721856	12/1987	Saotome et al.	N/A	N/A
4744099	12/1987	Huettenrauch et al.	N/A	N/A
4752948	12/1987	MacMahon	N/A	N/A
4760589	12/1987	Siczek	N/A	N/A
4763343	12/1987	Yanaki	N/A	N/A
4773086	12/1987	Fujita et al.	N/A	N/A
4773087	12/1987	Plewes	N/A	N/A
4799248	12/1988	Furbee	N/A	N/A
4819258	12/1988	Kleinman et al.	N/A	N/A
4821727	12/1988	Levene et al.	N/A	N/A
4901335	12/1989	Ferlic	N/A	N/A
4969174	12/1989	Scheid et al.	N/A	N/A
4989227	12/1990	Tirelli et al.	N/A	N/A
4998270	12/1990	Scheid et al.	N/A	N/A
5018176	12/1990	Romeas et al.	N/A	N/A
RE33634	12/1990	Yanaki	N/A	N/A
5029193	12/1990	Saffer	N/A	N/A
5051904	12/1990	Griffith	N/A	N/A
5077773	12/1990	Sammon	N/A	N/A
5078142	12/1991	Siczek et al.	N/A	N/A
5081664	12/1991	Lie et al.	N/A	N/A
5129911	12/1991	Siczek et al.	N/A	N/A
5142557	12/1991	Toker	N/A	N/A
5163075	12/1991	Lubinsky et al.	N/A	N/A
5164976	12/1991	Scheid et al.	N/A	N/A
5199056	12/1992	Darrah	N/A	N/A
5212637	12/1992	Saxena	N/A	N/A
5219351	12/1992	Teubner	N/A	N/A
5240011	12/1992	Assa	N/A	N/A
5256370	12/1992	Slattery	N/A	N/A
5274690	12/1992	Burke	N/A	N/A
5289520	12/1993	Pellegrino et al.	N/A	N/A
5291539	12/1993	Thumann et al.	N/A	N/A
5313510	12/1993	Ebersberger	N/A	N/A
5359637	12/1993	Webber	N/A	N/A
5365562	12/1993	Toker	N/A	N/A
5415169	12/1994	Siczek et al.	N/A	N/A
5426685	12/1994	Pellegrino et al.	N/A	N/A
5451789	12/1994	Wong	N/A	N/A
5452367	12/1994	Bick et al.	N/A	N/A
5479927	12/1995	Shmulewitz	N/A	N/A
5483072	12/1995	Coe	N/A	N/A
5506877	12/1995	Niklason et al.	N/A	N/A
5526394	12/1995	Siczek et al.	N/A	N/A
5528658	12/1995	Hell	N/A	N/A
5539797	12/1995	Heidsieck et al.	N/A	N/A
5553111	12/1995	Moore et al.	N/A	N/A
5592562	12/1996	Rooks	N/A	N/A
5594769	12/1996	Pellegrino et al.	N/A	N/A
5596200	12/1996	Sharma et al.	N/A	N/A
5598454	12/1996	Franetzke et al.	N/A	N/A

5606589	12/1996	Pellegrino et al.	N/A	N/A
5609152	12/1996	Pellegrino et al.	N/A	N/A
5627869	12/1996	Andrew et al.	N/A	N/A
5657362	12/1996	Giger et al.	N/A	N/A
5668844	12/1996	Webber	N/A	N/A
5668889	12/1996	Hara	N/A	N/A
5706327	12/1997	Adamkowski et al.	N/A	N/A
5719952	12/1997	Rooks	N/A	N/A
5735264	12/1997	Siczek et al.	N/A	N/A
5769086	12/1997	Ritchart et al.	N/A	N/A
5773832	12/1997	Sayed et al.	N/A	N/A
5803912	12/1997	Siczek et al.	N/A	N/A
5818898	12/1997	Tsukamoto et al.	N/A	N/A
5828722	12/1997	Ploetz et al.	N/A	N/A
5841829	12/1997	Dolazza	N/A	N/A
5844242	12/1997	Jalink, Jr.	N/A	N/A
5844965	12/1997	Galkin	N/A	N/A
5864146	12/1998	Karellas	N/A	N/A
5872828	12/1998	Niklason	N/A	N/A
5878104	12/1998	Ploetz	N/A	N/A
5896437	12/1998	Ploetz	N/A	N/A
5901197	12/1998	Khutoryansky	N/A	N/A
5930330	12/1998	Wolfe	N/A	N/A
5941832	12/1998	Tumey et al.	N/A	N/A
5970118	12/1998	Sokolov	N/A	N/A
5983123	12/1998	Shmulewitz	N/A	N/A
5986662	12/1998	Argiro et al.	N/A	N/A
5999836	12/1998	Nelson et al.	N/A	N/A
6005907	12/1998	Ploetz	N/A	N/A
6022325	12/1999	Siczek et al.	N/A	N/A
6075879	12/1999	Roehrig et al.	N/A	N/A
6081577	12/1999	Webber	N/A	N/A
6091841	12/1999	Rogers et al.	N/A	N/A
6101236	12/1999	Wang et al.	N/A	N/A
6137527	12/1999	Abdel-Malek et al.	N/A	N/A
6141398	12/1999	He et al.	N/A	N/A
6149301	12/1999	Kautzer et al.	N/A	N/A
6151383	12/1999	Xue	N/A	N/A
6167115	12/1999	Inoue	N/A	N/A
6175117	12/2000	Komardin et al.	N/A	N/A
6196715	12/2000	Nambu et al.	N/A	N/A
6207958	12/2000	Giakos	N/A	N/A
6212256	12/2000	Miesbauer	N/A	N/A
6216540	12/2000	Nelson et al.	N/A	N/A
6219059	12/2000	Argiro	N/A	N/A
6233473	12/2000	Shepherd et al.	N/A	N/A
6243441	12/2000	Zur	N/A	N/A
6244507	12/2000	Garland	N/A	N/A
6256369	12/2000	Lai	N/A	N/A
6256370	12/2000	Yavuz	N/A	N/A
6269176	12/2000	Barski	N/A	N/A
6272207	12/2000	Tang	N/A	N/A
6282264	12/2000	Smith	N/A	N/A
6289235	12/2000	Webber et al.	N/A	N/A
6292530	12/2000	Yavus et al.	N/A	N/A
6292531	12/2000	Hsieh et al.	N/A	N/A
6293282	12/2000	Lemelson	N/A	N/A
6327336	12/2000	Gingold et al.	N/A	N/A
6341156	12/2001	Baetz et al.	N/A	N/A
6345194	12/2001	Nelson et al.	N/A	N/A
6351517	12/2001	Guru	N/A	N/A
6375352	12/2001	Hewes et al.	N/A	N/A
6399951	12/2001	Paulus	N/A	N/A
6411836	12/2001	Patel et al.	N/A	N/A
6415015	12/2001	Nicolas et al.	N/A	N/A

6418189	12/2001	Schafer	N/A	N/A
6442288	12/2001	Haerer et al.	N/A	N/A
6453009	12/2001	Berezowitz	N/A	N/A
6454460	12/2001	Ramanathan	N/A	N/A
6459925	12/2001	Nields et al.	N/A	N/A
6463181	12/2001	Duarte	N/A	N/A
6480565	12/2001	Ning	N/A	N/A
6490476	12/2001	Townsend et al.	N/A	N/A
6496557	12/2001	Wilson	N/A	N/A
6501819	12/2001	Unger et al.	N/A	N/A
6542575	12/2002	Schubert	N/A	N/A
6553096	12/2002	Zhou et al.	N/A	N/A
6556655	12/2002	Chichereau et al.	N/A	N/A
6574304	12/2002	Hsieh et al.	N/A	N/A
6574629	12/2002	Cooke, Jr. et al.	N/A	N/A
6597762	12/2002	Ferrant et al.	N/A	N/A
6611575	12/2002	Alyassin et al.	N/A	N/A
6620111	12/2002	Stephens et al.	N/A	N/A
6626849	12/2002	Huitema et al.	N/A	N/A
6633626	12/2002	Trotter	N/A	N/A
6633674	12/2002	Barnes et al.	N/A	N/A
6638235	12/2002	Miller et al.	N/A	N/A
6647092	12/2002	Eberhard et al.	N/A	N/A
6674835	12/2003	Kaufhold	N/A	N/A
6702459	12/2003	Barnes et al.	N/A	N/A
6744848	12/2003	Stanton et al.	N/A	N/A
6748044	12/2003	Sabol et al.	N/A	N/A
6748046	12/2003	Thayer	N/A	N/A
6748047	12/2003	Gonzalez	N/A	N/A
6751285	12/2003	Eberhard et al.	N/A	N/A
6758824	12/2003	Miller et al.	N/A	N/A
6813334	12/2003	Koppe et al.	N/A	N/A
6846289	12/2004	Besson	N/A	N/A
6882700	12/2004	Wang et al.	N/A	N/A
6885724	12/2004	Li et al.	N/A	N/A
6895076	12/2004	Halsmer	N/A	N/A
6901132	12/2004	Eberhard	N/A	N/A
6909790	12/2004	Tumey et al.	N/A	N/A
6909792	12/2004	Carrott et al.	N/A	N/A
6912319	12/2004	Barnes et al.	N/A	N/A
6931093	12/2004	Op De Beek et al.	N/A	N/A
6940943	12/2004	Claus	N/A	N/A
6950492	12/2004	Besson	N/A	N/A
6950493	12/2004	Besson	N/A	N/A
6957099	12/2004	Arnone et al.	N/A	N/A
6960020	12/2004	Lai	N/A	N/A
6970531	12/2004	Eberhard et al.	N/A	N/A
6970586	12/2004	Baertsch	N/A	N/A
6978040	12/2004	Berestov	N/A	N/A
6987831	12/2005	Ning	N/A	N/A
6999554	12/2005	Mertelmeier	N/A	N/A
7001071	12/2005	Deuringer	N/A	N/A
7016461	12/2005	Rotondo	N/A	N/A
7023960	12/2005	Chretien	378/112	H05G 1/34
7092482	12/2005	Besson	N/A	N/A
7104690	12/2005	Radley	378/207	H05G 1/54
7110490	12/2005	Eberhard	N/A	N/A
7110502	12/2005	Tsuji	N/A	N/A
7116749	12/2005	Besson	N/A	N/A
7123684	12/2005	Jing et al.	N/A	N/A
7127091	12/2005	Op De Beek et al.	N/A	N/A
7142633	12/2005	Eberhard et al.	N/A	N/A
7190758	12/2006	Hagiwara	N/A	N/A
7206462	12/2006	Betke	N/A	N/A
7218766	12/2006	Eberhard	N/A	N/A

7244063	12/2006	Eberhard	N/A	N/A
7245694	12/2006	Jing et al.	N/A	N/A
7263214	12/2006	Uppaluri	N/A	N/A
7286645	12/2006	Freudenberger	N/A	N/A
7302031	12/2006	Hjarn et al.	N/A	N/A
7315607	12/2007	Ramsauer	N/A	N/A
7319734	12/2007	Besson	N/A	N/A
7319735	12/2007	Defreitas et al.	N/A	N/A
7319736	12/2007	Rotondo	N/A	N/A
7323692	12/2007	Rowlands et al.	N/A	N/A
7331264	12/2007	Ozawa	N/A	N/A
7356113	12/2007	Wu	N/A	N/A
7430272	12/2007	Jing et al.	N/A	N/A
7433507	12/2007	Jabri	N/A	N/A
7443949	12/2007	Defreitas et al.	N/A	N/A
7466795	12/2007	Eberhard et al.	N/A	N/A
7577282	12/2008	Gkanatsios et al.	N/A	N/A
7583786	12/2008	Jing et al.	N/A	N/A
7609806	12/2008	Defreitas et al.	N/A	N/A
7609808	12/2008	Tornai	N/A	N/A
7616731	12/2008	Pack	N/A	N/A
7616801	12/2008	Gkanatsios et al.	N/A	N/A
7630531	12/2008	Chui	N/A	N/A
7630533	12/2008	Ruth et al.	N/A	N/A
7688940	12/2009	Defreitas et al.	N/A	N/A
7697660	12/2009	Ning	N/A	N/A
7702142	12/2009	Ren et al.	N/A	N/A
7760853	12/2009	Jing et al.	N/A	N/A
7760924	12/2009	Ruth et al.	N/A	N/A
7792241	12/2009	Wu	N/A	N/A
7792245	12/2009	Hitzke et al.	N/A	N/A
7831296	12/2009	Defreitas et al.	N/A	N/A
7839979	12/2009	Hauttmann	N/A	N/A
7869563	12/2010	Defreitas et al.	N/A	N/A
7869862	12/2010	Seppi	N/A	N/A
7881428	12/2010	Jing et al.	N/A	N/A
7885384	12/2010	Mannar	N/A	N/A
7894646	12/2010	Shirahata et al.	N/A	N/A
7916915	12/2010	Gkanatsios et al.	N/A	N/A
7949091	12/2010	Jing et al.	N/A	N/A
7986765	12/2010	Defreitas et al.	N/A	N/A
7991106	12/2010	Ren et al.	N/A	N/A
8031834	12/2010	Ludwig	N/A	N/A
8131049	12/2011	Ruth et al.	N/A	N/A
8155421	12/2011	Ren et al.	N/A	N/A
8170320	12/2011	Smith et al.	N/A	N/A
8175219	12/2011	Defreitas et al.	N/A	N/A
8275090	12/2011	Ren	378/21	A61B 6/025
8285020	12/2011	Gkanatsios et al.	N/A	N/A
8416915	12/2012	Jing et al.	N/A	N/A
8452379	12/2012	DeFreitas et al.	N/A	N/A
8457282	12/2012	Baorui et al.	N/A	N/A
8515005	12/2012	Ren et al.	N/A	N/A
8532745	12/2012	DeFreitas et al.	N/A	N/A
8559595	12/2012	Defreitas et al.	N/A	N/A
8565372	12/2012	Stein et al.	N/A	N/A
8565374	12/2012	DeFreitas et al.	N/A	N/A
8565860	12/2012	Kimchy	N/A	N/A
8571289	12/2012	Ruth et al.	N/A	N/A
8712127	12/2013	Ren et al.	N/A	N/A
8767911	12/2013	Ren et al.	N/A	N/A
8787522	12/2013	Smith et al.	N/A	N/A
8831171	12/2013	Jing et al.	N/A	N/A
8853635	12/2013	O'Connor	N/A	N/A
8873716	12/2013	Ren et al.	N/A	N/A

8923484	12/2013	Zou	378/126	H01J 35/28
9042612	12/2014	Gkanatsios et al.	N/A	N/A
9066706	12/2014	Defreitas et al.	N/A	N/A
9226721	12/2015	Ren et al.	N/A	N/A
9338868	12/2015	Yabugami	N/A	H05G 1/32
9460508	12/2015	Gkanatsios et al.	N/A	N/A
9498175	12/2015	Stein et al.	N/A	N/A
9502148	12/2015	Ren et al.	N/A	N/A
9549709	12/2016	DeFreitas et al.	N/A	N/A
9851888	12/2016	Gkanatsios et al.	N/A	N/A
9895115	12/2017	Ren	N/A	N/A
10108329	12/2017	Gkanatsios et al.	N/A	N/A
10194875	12/2018	DeFreitas et al.	N/A	N/A
10296199	12/2018	Gkanatsios	N/A	N/A
10413255	12/2018	Stein	N/A	N/A
10452252	12/2018	Gkanatsios et al.	N/A	N/A
10614996	12/2019	Yonezawa	N/A	H01J 37/244
10638994	12/2019	DeFreitas	N/A	N/A
10719223	12/2019	Gkanatsios	N/A	N/A
10753969	12/2019	Xu	N/A	H05G 1/54
10881359	12/2020	Williams	N/A	N/A
11510306	12/2021	Ru	N/A	N/A
11751316	12/2022	Duncan	378/104	H05G 1/54
2001/0038681	12/2000	Stanton et al.	N/A	N/A
2002/0012450	12/2001	Tsujii	N/A	N/A
2002/0048343	12/2001	Launay et al.	N/A	N/A
2002/0050986	12/2001	Inouc et al.	N/A	N/A
2002/0070970	12/2001	Wood et al.	N/A	N/A
2002/0075997	12/2001	Unger et al.	N/A	N/A
2002/0090055	12/2001	Zur et al.	N/A	N/A
2002/0094062	12/2001	Dolazza	N/A	N/A
2002/0113681	12/2001	Byram	N/A	N/A
2002/0122533	12/2001	Marie et al.	N/A	N/A
2002/0126798	12/2001	Harris	N/A	N/A
2003/0007598	12/2002	Wang et al.	N/A	N/A
2003/0010923	12/2002	Zur	N/A	N/A
2003/0018272	12/2002	Treado et al.	N/A	N/A
2003/0026386	12/2002	Tang et al.	N/A	N/A
2003/0058989	12/2002	Rotondo	N/A	N/A
2003/0072409	12/2002	Kaufhold et al.	N/A	N/A
2003/0072417	12/2002	Kaufhold et al.	N/A	N/A
2003/0073895	12/2002	Nields et al.	N/A	N/A
2003/0095624	12/2002	Eberhard et al.	N/A	N/A
2003/0097055	12/2002	Yanof et al.	N/A	N/A
2003/0149364	12/2002	Kapur	N/A	N/A
2003/0169847	12/2002	Karellas et al.	N/A	N/A
2003/0194050	12/2002	Eberhard	N/A	N/A
2003/0194051	12/2002	Wang et al.	N/A	N/A
2003/0194121	12/2002	Eberhard et al.	N/A	N/A
2003/0210254	12/2002	Doan et al.	N/A	N/A
2003/0212327	12/2002	Wang et al.	N/A	N/A
2003/0215120	12/2002	Uppaluri et al.	N/A	N/A
2004/0008809	12/2003	Webber	N/A	N/A
2004/0066882	12/2003	Eberhard et al.	N/A	N/A
2004/0066884	12/2003	Claus et al.	N/A	N/A
2004/0066904	12/2003	Eberhard et al.	N/A	N/A
2004/0070582	12/2003	Smith et al.	N/A	N/A
2004/0094167	12/2003	Brady et al.	N/A	N/A
2004/0101095	12/2003	Jing et al.	N/A	N/A
2004/0109529	12/2003	Eberhard	N/A	N/A
2004/0146221	12/2003	Siegel et al.	N/A	N/A
2004/0171986	12/2003	Tremaglio, Jr. et al.	N/A	N/A
2004/0190682	12/2003	Deuringer	N/A	N/A
2004/0213378	12/2003	Zhou et al.	N/A	N/A
2004/0247081	12/2003	Halsmer	N/A	N/A

2004/0264627	12/2003	Besson	N/A	N/A
2004/0267157	12/2003	Miller et al.	N/A	N/A
2005/0025278	12/2004	Hagiwara	N/A	N/A
2005/0049497	12/2004	Krishnan	N/A	N/A
2005/0049521	12/2004	Miller et al.	N/A	N/A
2005/0063509	12/2004	DeFreitas et al.	N/A	N/A
2005/0078797	12/2004	Danielsson et al.	N/A	N/A
2005/0084070	12/2004	Chretien	378/111	H05G 1/34
2005/0089205	12/2004	Kapur	N/A	N/A
2005/0105679	12/2004	Wu et al.	N/A	N/A
2005/0113681	12/2004	Defreitas	N/A	N/A
2005/0113715	12/2004	Schwindt et al.	N/A	N/A
2005/0117694	12/2004	Francke	N/A	N/A
2005/0129172	12/2004	Mertelmeier	N/A	N/A
2005/0133706	12/2004	Eberhard	N/A	N/A
2005/0135555	12/2004	Claus et al.	N/A	N/A
2005/0135664	12/2004	Kaufhold et al.	N/A	N/A
2005/0157849	12/2004	Radley	N/A	N/A
2005/0226375	12/2004	Eberhard et al.	N/A	N/A
2005/0248347	12/2004	Damadian	N/A	N/A
2006/0009693	12/2005	Hanover et al.	N/A	N/A
2006/0030784	12/2005	Miller et al.	N/A	N/A
2006/0034426	12/2005	Freudenberger	N/A	N/A
2006/0074288	12/2005	Kelly	N/A	N/A
2006/0098855	12/2005	Gkanatsios et al.	N/A	N/A
2006/0109951	12/2005	Popescu	N/A	N/A
2006/0126780	12/2005	Rotondo	N/A	N/A
2006/0129062	12/2005	Nicoson et al.	N/A	N/A
2006/0155209	12/2005	Miller et al.	N/A	N/A
2006/0210016	12/2005	Francke	N/A	N/A
2006/0257009	12/2005	Wang	N/A	N/A
2006/0262898	12/2005	Partain	N/A	N/A
2006/0269041	12/2005	Mertelmeier	N/A	N/A
2006/0291618	12/2005	Eberhard et al.	N/A	N/A
2007/0030949	12/2006	Jing et al.	N/A	N/A
2007/0036265	12/2006	Jing et al.	N/A	N/A
2007/0076844	12/2006	Defreitas et al.	N/A	N/A
2007/0078335	12/2006	Horn	N/A	N/A
2007/0140419	12/2006	Souchay	N/A	N/A
2007/0189463	12/2006	Deuringer	N/A	N/A
2007/0223651	12/2006	Wagenaar et al.	N/A	N/A
2007/0225600	12/2006	Weibrecht et al.	N/A	N/A
2007/0242800	12/2006	Jing et al.	N/A	N/A
2008/0019581	12/2007	Gkanatsios et al.	N/A	N/A
2008/0045833	12/2007	Defreitas et al.	N/A	N/A
2008/0056436	12/2007	Pack	N/A	N/A
2008/0101537	12/2007	Sendai	N/A	N/A
2008/0112534	12/2007	Defreitas	N/A	N/A
2008/0118023	12/2007	Besson	N/A	N/A
2008/0130979	12/2007	Ren	N/A	N/A
2008/0198966	12/2007	Hjam	N/A	N/A
2008/0212861	12/2007	Durgan et al.	N/A	N/A
2008/0285712	12/2007	Kopans	N/A	N/A
2008/0317196	12/2007	Imai	N/A	N/A
2008/0317205	12/2007	Inuga et al.	N/A	N/A
2009/0003519	12/2008	Defreitas et al.	N/A	N/A
2009/0010384	12/2008	Jing et al.	N/A	N/A
2009/0080594	12/2008	Brooks et al.	N/A	N/A
2009/0080602	12/2008	Brooks et al.	N/A	N/A
2009/0080604	12/2008	Shores et al.	N/A	N/A
2009/0135997	12/2008	Defreitas et al.	N/A	N/A
2009/0141859	12/2008	Gkanatsios et al.	N/A	N/A
2009/0143674	12/2008	Nields	N/A	N/A
2009/0177495	12/2008	Abousy	N/A	N/A
2009/0213987	12/2008	Stein et al.	N/A	N/A

2009/0237924	12/2008	Ladewig	N/A	N/A
2009/0238424	12/2008	Arakita et al.	N/A	N/A
2009/0268865	12/2008	Ren et al.	N/A	N/A
2009/0268876	12/2008	Crucs	N/A	N/A
2009/0281867	12/2008	Sievenpiper	N/A	N/A
2009/0296882	12/2008	Gkanatsios	N/A	N/A
2009/0304147	12/2008	Jing et al.	N/A	N/A
2010/0020937	12/2009	Hautmann	N/A	N/A
2010/0020938	12/2009	Koch	N/A	N/A
2010/0034450	12/2009	Mertelmeier	N/A	N/A
2010/0054400	12/2009	Ren	N/A	N/A
2010/0054401	12/2009	Blendl	N/A	N/A
2010/0086188	12/2009	Ruth	N/A	N/A
2010/0091940	12/2009	Ludwig et al.	N/A	N/A
2010/0150306	12/2009	Defreitas et al.	N/A	N/A
2010/0189227	12/2009	Mannar	N/A	N/A
2010/0195882	12/2009	Ren	N/A	N/A
2010/0226475	12/2009	Smith	N/A	N/A
2010/0290585	12/2009	Eliasson	N/A	N/A
2010/0303202	12/2009	Ren	N/A	N/A
2010/0313196	12/2009	De Atley et al.	N/A	N/A
2011/0026667	12/2010	Poorter	N/A	N/A
2011/0069809	12/2010	Defreitas et al.	N/A	N/A
2011/0087132	12/2010	DeFreitas et al.	N/A	N/A
2011/0121969	12/2010	Mercer	N/A	N/A
2011/0164724	12/2010	Ohta et al.	N/A	N/A
2011/0178389	12/2010	Kumar et al.	N/A	N/A
2011/0188624	12/2010	Ren	N/A	N/A
2011/0234630	12/2010	Batman et al.	N/A	N/A
2011/0237927	12/2010	Brooks et al.	N/A	N/A
2011/0268246	12/2010	Dafni	N/A	N/A
2012/0033868	12/2011	Ren	N/A	N/A
2012/0039437	12/2011	Ren	N/A	N/A
2012/0051502	12/2011	Ohta et al.	N/A	N/A
2012/0236987	12/2011	Ruimi	N/A	N/A
2012/0238870	12/2011	Smith et al.	N/A	N/A
2013/0028374	12/2012	Gkanatsios et al.	N/A	N/A
2013/0077748	12/2012	Althoff	N/A	N/A
2013/0211261	12/2012	Wang	N/A	N/A
2013/0223594	12/2012	Sprong	N/A	N/A
2013/0272494	12/2012	DeFreitas et al.	N/A	N/A
2013/0315378	12/2012	Yabugami	378/98	H05G 1/32
2014/0044230	12/2013	Stein et al.	N/A	N/A
2014/0044231	12/2013	Defreitas et al.	N/A	N/A
2014/0064456	12/2013	Zou	N/A	N/A
2014/0086471	12/2013	Ruth et al.	N/A	N/A
2014/0098935	12/2013	Defreitas et al.	N/A	N/A
2014/0232752	12/2013	Ren et al.	N/A	N/A
2014/0314198	12/2013	Ren et al.	N/A	N/A
2014/0321607	12/2013	Smith	N/A	N/A
2014/0376690	12/2013	Jing et al.	N/A	N/A
2015/0049859	12/2014	DeFreitas et al.	N/A	N/A
2015/0117617	12/2014	Ishihara	N/A	N/A
2015/0157880	12/2014	Dalbow	N/A	N/A
2015/0160848	12/2014	Gkanatsios et al.	N/A	N/A
2015/0310611	12/2014	Gkanatsios et al.	N/A	N/A
2015/0347693	12/2014	Lam	N/A	N/A
2015/0352376	12/2014	Wiggers	N/A	N/A
2016/0106383	12/2015	Ren et al.	N/A	N/A
2016/0120497	12/2015	Nasir	N/A	N/A
2016/0188828	12/2015	Klingenberg	N/A	N/A
2016/0220207	12/2015	Jouhikainen	N/A	N/A
2016/0256125	12/2015	Smith	N/A	N/A
2016/0270742	12/2015	Stein et al.	N/A	N/A
2016/0331339	12/2015	Guo	N/A	N/A

2017/0024113	12/2016	Gkanatsios et al.	N/A	N/A
2017/0128028	12/2016	DeFreitas et al.	N/A	N/A
2017/0135650	12/2016	Stein et al.	N/A	N/A
2017/0135653	12/2016	Ren	N/A	N/A
2017/0215831	12/2016	Nagano	N/A	N/A
2017/0319167	12/2016	Goto	N/A	N/A
2017/0372863	12/2016	Price	N/A	N/A
2018/0005796	12/2017	Iida	N/A	N/A
2018/0068066	12/2017	Bronkalla	N/A	N/A
2018/0130201	12/2017	Bernard	N/A	N/A
2018/0177476	12/2017	Jing et al.	N/A	N/A
2018/0188937	12/2017	Gkanatsios et al.	N/A	N/A
2018/0289347	12/2017	DeFreitas et al.	N/A	N/A
2018/0315579	12/2017	Yonezawa	N/A	N/A
2018/0344276	12/2017	DeFreitas et al.	N/A	N/A
2019/0059830	12/2018	Williams	N/A	N/A
2019/0095087	12/2018	Gkanatsios et al.	N/A	N/A
2019/0138693	12/2018	Meller	N/A	N/A
2019/0159747	12/2018	Zanca	N/A	N/A
2019/0188848	12/2018	Madani	N/A	N/A
2019/0200942	12/2018	DeFreitas	N/A	N/A
2019/0221304	12/2018	Ionasec	N/A	N/A
2019/0295248	12/2018	Nakamura	N/A	N/A
2019/0304736	12/2018	Matsuura	N/A	N/A
2019/0317144	12/2018	Xu	N/A	H01J 35/025
2019/0320994	12/2018	Lemaitre	N/A	N/A
2019/0336794	12/2018	Li	N/A	N/A
2020/0012417	12/2019	Gkanatsios	N/A	N/A
2020/0029927	12/2019	Wilson	N/A	N/A
2020/0043600	12/2019	Glottmann	N/A	N/A
2020/0085393	12/2019	Zhang	N/A	N/A
2020/0167920	12/2019	Hall	N/A	N/A
2020/0286613	12/2019	Rego	N/A	N/A
2020/0348835	12/2019	Gkanatsios	N/A	N/A
2020/0352531	12/2019	Smith	N/A	N/A
2021/0136900	12/2020	Duncan	N/A	N/A
2021/0176850	12/2020	Ru	N/A	N/A
2021/0298700	12/2020	Williams	N/A	N/A
2021/0303078	12/2020	Wells	N/A	N/A
2021/0370098	12/2020	Flint	N/A	N/A
2022/0037006	12/2021	Zan	N/A	N/A
2023/0251210	12/2022	Ru	378/21	H05G 1/265
2024/0206044	12/2023	Pronk	N/A	H05G 1/54

FOREIGN PATENT DOCUMENTS

Patent No.	Application Date	Country	CPC
108492874	12/2017	CN	N/A
4104166	12/1991	DE	N/A
102004051401	12/2005	DE	N/A
102004051820	12/2005	DE	N/A
102010027871	12/2010	DE	N/A
102011007215	12/2011	DE	N/A
0775467	12/1996	EP	N/A
0982001	12/1999	EP	N/A
1028451	12/1999	EP	N/A
1428473	12/2003	EP	N/A
1623672	12/2005	EP	N/A
1759637	12/2006	EP	N/A
1569556	12/2011	EP	N/A
2732764	12/2013	EP	N/A
2602743	12/2013	EP	N/A
2819145	12/2013	EP	N/A
3143935	12/2016	EP	N/A
3569148	12/2018	EP	N/A
415709	12/1933	GB	N/A
53151381	12/1977	JP	N/A

H 05-329143	12/1992	JP	N/A
H07-230778	12/1994	JP	N/A
2000-287960	12/1999	JP	N/A
2001-346786	12/2000	JP	N/A
2002219124	12/2001	JP	N/A
2004-511884	12/2003	JP	N/A
2004-188200	12/2003	JP	N/A
2004-528682	12/2003	JP	N/A
2005-142160	12/2004	JP	N/A
2006-519625	12/2005	JP	N/A
2006-231054	12/2005	JP	N/A
2007-50264	12/2006	JP	N/A
2007-054528	12/2006	JP	N/A
2007-521911	12/2006	JP	N/A
2007229269	12/2006	JP	N/A
2008-67933	12/2007	JP	N/A
2008086471	12/2007	JP	N/A
2008-159317	12/2007	JP	N/A
2009500048	12/2008	JP	N/A
4-261864	12/2008	JP	N/A
2011-516116	12/2010	JP	N/A
2019-125460	12/2018	JP	N/A
90/05485	12/1989	WO	N/A
9803115	12/1997	WO	N/A
98/16903	12/1997	WO	N/A
00/51484	12/1999	WO	N/A
2000068863	12/1999	WO	N/A
03/020114	12/2002	WO	N/A
03037046	12/2002	WO	N/A
2003/057564	12/2002	WO	N/A
2004/043535	12/2003	WO	N/A
2005/051197	12/2004	WO	N/A
2005/110230	12/2004	WO	N/A
2005/112767	12/2004	WO	N/A
2006/004185	12/2005	WO	N/A
2006055830	12/2005	WO	N/A
2006/058160	12/2005	WO	N/A
2007129244	12/2006	WO	N/A
2008072144	12/2007	WO	N/A
2009122328	12/2008	WO	N/A
2009136349	12/2008	WO	N/A
2010/070554	12/2009	WO	N/A
2013/184213	12/2012	WO	N/A
2019/030410	12/2018	WO	N/A
2016/057960	12/2018	WO	N/A

OTHER PUBLICATIONS

“Essentials for life: Senographe Essential Full-Field Digital Mammography system”, GE Health-care Brochure, MM-0132-05.06-EN-US, 2006, 12 pgs. cited by applicant

“Filtered Back Projection,” (NYGREN) published May 8, 2007;

URL:<http://web.archive.org/web/19991010131715/http://www.owl.net.rice.edu/~.about.e1ec539/Projects97/cult/node2.html>., 2 pgs. cited by applicant

“Lorad Selenia” Document B-BI-SEO US/Intl (May 2006) copyright Hologic 2006, 12 pgs. cited by applicant

ACRIN website, located at

<https://www.acrin.org/PATIENTS/ABOUTIMAGINGEXAMSANDAGENTS/ABOUTMAMMOGRAPHYANDTOMOSYNTHESIS.asp>

“About Mammography and Tomosynthesis”, obtained online on Dec. 8, 2015, 5 pgs. cited by applicant

American College of Radiology website, located at <http://www.acr.org/FAQs/DBT-FAQ>, “Digital Breast Tomosynthesis FAQ for Insurers” obtained online on Dec. 8, 2015, 2 pages. cited by applicant

Arfelli, F. et al., “Mammography with synchrotron radiation: phase-detection techniques”, Apr. 2000, retrieved at:

<https://www.ncbi.nlm.nih.gov/pubmed/10751500>, 8 pages. cited by applicant

Aslund, Magnus, “Digital Mammography with a Photon Counting Detector in a Scanned Multislit Geometry”, Doctoral Thesis, Dept of Physics, Royal Institute of Technology, Stockholm, Sweden, Apr. 2007, 51 pages. cited by applicant

Boone, J. et al., “Dedicated Breast CT: Radiation Dose and Image Quality Evaluation”, Dec. 31, 2001, retrieved at:

<http://pubs.rsna.org/doi/abs/10.1148/radiol.2213010334>, 11 pages. cited by applicant

Chan, Heang-Ping et al., “ROC study of the effect of stereoscopic imaging on assessment of breast lesions”, Medical Physics, vol. 32, No. 4, Apr. 2005, 7 pgs. cited by applicant

Choi, Bareum et al., "Surgical-tools detection based on Convolutional Neural Network in Laparoscopic Robot-Assisted Surgery", 2017 39th Annual Int'l. Conference of the IEEE Engineering in Medicine and Biology Society (EMBC), IEEE, Jul. 11, 2017, pp. 1756-1759. cited by applicant

Cole, Elodia, et al., "The Effects of Gray Scale Image Processing on Digital Mammography Interpretation Performance", *Academic Radiology*, vol. 12, No. 5, pp. 585-595, May 2005. cited by applicant

Digital Clinical Reports, Tomosynthesis, GE Brochure 98-5493, Nov. 1998, 8 pgs. cited by applicant

Dobbins, James T., "Digital x-ray tomosynthesis: current state of the art and clinical potential," *Physics in Medicine and Biology*, Taylor and Francis LTD, London GB, vol. 48, No. 19, Oct. 7, 2003, 42 pages. cited by applicant

Grant, David G., "Tomosynthesis: a three-dimensional imaging technique", *IEEE Trans. Biomed. Engineering*, vol. BME-19, #1, Jan. 1972, pp. 20-28. cited by applicant

Hamberg, Leena M., "Tomosynthesis breast imaging: early detection and characterization of breast cancer", prepared by Massachusetts General Hospital for the U.S. Army Medical Research and Material Command Fort Detrick, Maryland, Jul. 2000, 20 pages. cited by applicant

Han et al., "MatchNet: Unifying Feature and Metric Learning for Patch-Based Matching", 2015 IEEE Conference on Computer Vision and Pattern Recognition (CVPR), Boston, MA, 2015, pp. 3279-3286. cited by applicant

Kachelriess, Marc et al., "Flying Focal Spot (FFS) in Cone-Beam CT", 2004 IEEE Nuclear Science Symposium Conference Record, Oct. 16-22, 2004, Rome Italy, vol. 6, pp. 3759-3763. cited by applicant

Kapur, Ajay et al., "Combination of Digital Mammography with Semiautomated 3D Breast Ultrasound", Aug. 1, 2004, retrieved at: <http://journals.sagepub.com/doi/abs/10.1177/153303460400300402>, 10 pages. cited by applicant

Kita et al., "Correspondence between different view breast X-rays using simulation of breast deformation", *Proceedings 1998 IEEE Computer Society Conference on Computer Vision and Pattern Recognition*, Santa Barbara, CA, Jun. 23-25, 1998, pp. 700-707. cited by applicant

Kopans, D., "Development and Clinical Evaluation of Tomosynthesis for Digital Mammography", Oct. 31, 2000, retrieved at: <http://oai.dtic.mil/oai/oai?verb=getRecord&metadataPrefix=html&identifier=ADA387722>, 91 pages. cited by applicant

Kopans, Daniel B., "Breast Imaging", Chapter 26: Future Advances in Breast Imaging, 2nd Edition, Lippincott-Raven Publishers, Philadelphia, 1998, 37 pages. cited by applicant

Lehmann, V. et al., "MEMS techniques applied to the fabrication of anti-scatter grids for X-ray imaging", 2002, retrieved at: https://www.researchgate.net/profile/S_Ronnebeck/publication/222546207_MEMS_techniques_applied_to_the_fabrication_of_anti-scatter_grids_for_Xray_imaging/links/5570136f08aecd777417301/MEMS-techniques-applied-to-the-fabrication-of-anti-scatter-grids-for-X-ray-imaging.pdf, 6 pages. cited by applicant

Mammographic Accreditation Phantom, <http://www.cirsinc.com/pdfs/015cp.pdf>. (2006), 2 pgs. cited by applicant

Niklason et al., "Digital breast tomosynthesis: potentially a new method for breast cancer screening", In *Digital Mammography*, 1998, 6 pages. cited by applicant

Niklason et al., "Digital Breast Imaging: Tomosynthesis and Digital Subtraction Mammography", *Breast Disease*, vol. 10, No. 3-4, pp. 151-164, 1998. cited by applicant

Niklason, Loren T. et al., "Digital Tomosynthesis in Breast Imaging", *Radiology*, Nov. 1997, vol. 205, No. 2, pp. 399-406. cited by applicant

Nykanen, Kirsi et al., "X-ray scattering in full-field digital mammography", Jul. 2003, retrieved at: <http://www.siltanen-research.net/publ/NykanenSiltanen2003.pdf>, 10 pages. cited by applicant

Pediconi, Federica et al., "Color-coded automated signal intensity-curve for detection and characterization of breast lesions: Preliminary evaluation of a new software for MR-based breast imaging", *International Congress Series 1281 (2005) 1081-1086*. cited by applicant

Pisano, Etta D., "Digital Mammography", *Radiology*, vol. 234, No. 2, Feb. 2005, pp. 353-362. cited by applicant

Rolf Behling—Ed-Behling et al., Chapter 6: Diagnostic X-Ray Sources from the Inside, *Modern Diagnostic X-Ray Sources*, Taylor & Francis Group, pp. 177-308, Jan. 1, 2016, retrieved from the internet on Jun. 26, 2015 at: <https://ebookcentral.proquest.com/lib/epo-ebooks/detail.action?docID=2075866>. cited by applicant

Senographe 700 & 800T (GE); 2-page download on Jun. 22, 2006 from www.gehealthcare.com/inen/rad/whe/products/mswh800t.html; Figures 1-7 on 4 sheets re lateral shift compression paddle, 2 pgs. cited by applicant

Smith, A., "Fundamentals of Breast Tomosynthesis", White Paper, Hologic Inc., WP-00007, Jun. 2008, 8 pgs. cited by applicant

Smith, Andrew, PhD, "Full Field Breast Tomosynthesis", Hologic White Paper, Oct. 2004, 6 pgs. cited by applicant

Suryanarayanan, S. et al., "Comparison of tomosynthesis methods used with digital mammography", Dec. 31, 2000, retrieved at: <http://www.sciencedirect.com/science/article/pii/S1076633200800616>, 13 pages. cited by applicant

Suryanarayanan, S. et al., "Evaluation of Linear and Nonlinear Tomosynthetic Reconstruction Methods in Digital Mammography", Mar. 2001, retrieved at: <http://www.sciencedirect.com/science/article/pii/S1076633203805305>, 6 pages. cited by applicant

Thurfjell, "Mammography screening: one versus two views and independent double reading", *Acta Radiologica* 35, No. 4, 1994, pp. 345-350. cited by applicant

Webber, Richard, "A controlled evaluation of tuned-aperture computed tomography applied to digital spot mammography", Feb. 2000, retrieved at: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC3453191/>, 8 pages. cited by applicant

Wheeler F. W., et al. "Micro-Calcification Detection in Digital Tomosynthesis Mammography", *Proceedings of SPIE, Conf-Physics of Semiconductor Devices*, Dec. 11, 2001 to Dec. 15, 2001, Delhi, SPIE, US, vol. 6144, Feb. 13, 2006, 12 pgs. cited by applicant

Wu, T. et al., "A comparison of reconstruction algorithms for breast tomosynthesis", Aug. 26, 2004, retrieved at: <http://onlinelibrary.wiley.com/doi/10.1118/1.1786692/full>. cited by applicant

Wu, Tao, et al. "Tomographic Mammography Using a Limited Number of Low-Dose Cone-Beam Projection Images" *Medical Physics*, AIP, Melville, NY, vol. 30, No. 3, Mar. 1, 2003, p. 365-380. cited by applicant

European Extended Search Report in Application 23155076.5, mailed Jul. 7, 2023, 8 pages. cited by applicant

Primary Examiner: Artman; Thomas R

Attorney, Agent or Firm: Merchant & Gould P.C.

Background/Summary

CROSS-REFERENCE TO RELATED APPLICATIONS (1) This application claims priority to and the benefit of U.S. Provisional Application No. 63/307,311, filed Feb. 7, 2022, which is hereby incorporated by reference herein in its entirety.

BACKGROUND

(1) Imaging based on the use of x-rays is commonplace in medical imaging technology, such as, but not limited to, mammography or tomosynthesis systems. The x-rays used in such imaging technology are often generated through the use of an x-ray tube. Inside the x-ray tube there is an anode and a cathode, and within the cathode there is a filament that emits electrons. The electrons are accelerated by an electrical field generated by applying a high voltage potential across the anode and cathode terminals. As the filament wears with use, the filament material evaporates under high temperature. This evaporation results in a thinning of the material and an increase in filament resistance. When the resistance of the filament increases, power and temperature at the filament are increased, thereby resulting in an increased radiation output.

SUMMARY

(2) In one aspect, the technology relates to a method of adaptively controlling filament current in an x-ray tube of an imaging system, the method including: calibrating the x-ray tube having a filament; storing calibration data from the calibration of the x-ray tube at the imaging system, wherein the calibration data includes a filament current value that determines a tube current value for a tube voltage value at a plurality of stations; monitoring a resistance value of the filament over a period of time; determining a change in the resistance value of the filament over the period of time; and adjusting the filament current value of at least one of the plurality of stations based on the changed resistance value.

(3) In an example, monitoring the resistance value of the filament over the period of time includes measuring the resistance value after a predetermined number of sequence exposures and storing each measured resistance value. In another example, measuring the resistance value after the predetermined number of sequence exposures includes applying a constant filament current value to the filament after each of the predetermined number of sequence exposures. In yet another example, the method further includes waiting a predetermined time period between an end of the predetermined number of sequence exposures and prior to applying the constant filament current value to the filament for resistance value measurement. In still another example, the filament current value of each of the plurality of station is updated based on the changed resistance value. In an example, determining the change in the resistance value includes comparing a difference between resistance values over the period of time to a predetermined benchmark.

(4) In another example, the method further includes measuring a resulted tube current value from the x-ray tube at a sequence exposure having a stored tube current value and tube voltage value for a station of the plurality of stations; and comparing a difference between the measured resulted tube current value and the stored tube current value for the respective station. In yet another example, based on the difference between the measured resulted tube current value and the stored tube current value, the period of time is at least partially defined. In still another example, measuring resulted tube current value is performed pre-exposure or post-exposure of the sequence exposure. In an example, the method further includes determining a life-cycle period of the filament based at least partially on the monitored resistance value of the filament.

(5) In another aspect, the technology relates to an imaging system includes: an x-ray tube having a filament; a current control circuit coupled to the x-ray tube and configured to channel current through the filament; at least one processor communicatively coupled to the current control circuit; and memory communicatively coupled to the at least one processor, the memory including computer executable instructions that, when executed by the at least one processor, performs a method including: calibrating the x-ray tube; storing calibration data from the calibration of the x-ray tube at the imaging system, wherein the calibration data includes a filament current value that determines a tube current value for a tube voltage value at a plurality of stations; monitoring a resistance value of the filament over a period of time; determining a change in the resistance value of the filament over the period of time; and adjusting the filament current value of at least one of the plurality of stations based on the changed resistance value.

(6) In an example, monitoring the resistance value of the filament over the period of time includes measuring the resistance value after a predetermined number of sequence exposures and storing each measured resistance value. In another example, measuring the resistance value after the predetermined number of sequence exposures includes applying a constant filament current value to the filament after each of the predetermined number of sequence exposures. In yet another example, the method further includes waiting a predetermined time period between an end of the predetermined number of sequence exposures and prior to applying the constant filament current value to the filament for resistance value measurement. In still another example, the filament current value of each of the plurality of station is updated based on the changed resistance value. In an example, determining the change in the resistance value includes comparing a difference between resistance values over the period of time to a predetermined benchmark.

(7) In another example, the method further includes: measuring a resulted tube current value from the x-ray tube at a sequence exposure having a stored tube current value and tube voltage value for a station of the plurality of stations; and comparing a difference between the measured resulted tube current value and the stored tube current value for the respective station. In yet

another example, based on the difference between the measured resulted tube current value and the stored tube current value, the period of time is at least partially defined. In still another example, measuring resulted tube current value is performed pre-exposure or post-exposure of the sequence exposure. In an example, the method further includes determining a life-cycle period of the filament based at least partially on the monitored resistance value of the filament.

(8) This Summary is provided to introduce a selection of concepts in a simplified form that are further described below in the Detailed Description. This Summary is not intended to identify key features or essential features of the claimed subject matter, nor is it intended to be used to limit the scope of the claimed subject matter. Additional aspects, features, and/or advantages of examples will be set forth in part in the description which follows and, in part, will be apparent from the description, or may be learned by practice of the disclosure.

Description

BRIEF DESCRIPTION OF THE DRAWINGS

- (1) FIG. 1 is a schematic view of an exemplary imaging system.
- (2) FIG. 2 is a perspective view of the imaging system of FIG. 1.
- (3) FIG. 3 is a schematic view of an x-ray tube.
- (4) FIG. 4 is a graph depicting filament resistance change over a life of an x-ray tube.
- (5) FIG. 5 is a graph depicting filament current change over the life of an x-ray tube.
- (6) FIG. 6 is a graph depicting filament power change over the life of an x-ray tube.
- (7) FIG. 7 depicts a flowchart illustrating a method of adaptively controlling filament current in an x-ray tube.
- (8) FIG. 8 depicts an example of a suitable operating environment for use with the present examples.

DETAILED DESCRIPTION

(9) As discussed above, x-ray tubes in medical imaging systems have limited lifetimes. The limited lifetime of x-ray tubes is often due to the high heat and high voltages that are generally required for the operation of the x-ray tube. The high heat and voltages cause the components of the x-ray tube to break down, and in some components, change performance characteristics. When the x-ray tube changes performance characteristics, the x-ray tube needs to be recalibrated, and in some instances replaced. Recalibration and/or replacement costs of the x-ray tube are often significant, and according, improvements to the x-ray tube are desired.

(10) Based on analysis of x-ray tube filaments, as the filament degrades with use, the filament material evaporates and reduces its thickness. This results in the resistive properties of the filament changing and increasing. As the resistance of the filament increases, the power and the temperature generated at the filament increases during operation, causing an increase in radiation output from the x-ray tube.

(11) The present technology relates to a feedback mechanism to adjust the filament current needed to maintain the required or desired x-ray radiation at each tube output current and tube voltage station as the filament resistance changes over time. For example, after calibration of the x-ray tube, the filament power (e.g., based on the filament current generated) for each tube output current and tube voltage station is calculated and saved by the control system. Over the operational life of the x-ray tube, the filament resistance is monitored. When the filament resistance changes by a predetermined amount, the filament current may be adjusted appropriately by the control system to get the filament power the same as the stored value for each tube output current and tube voltage station. As such, the x-ray tube can deliver the desired x-ray radiation even with the filament changing resistive characteristics. Additionally, recalibration procedures will be reduced as the x-ray tube degrades with use.

(12) FIG. 1 is a schematic view of an exemplary imaging system 100. FIG. 2 is a perspective view of the imaging system 100. Referring concurrently to FIGS. 1 and 2, not every element described below is depicted in both figures. The imaging system 100 immobilizes a patient's breast 102 for x-ray imaging (either or both of mammography, tomosynthesis, or other imaging modalities) via a breast compression immobilizer unit 104 that includes a breast support platform 106 and a moveable compression paddle 108. The breast support platform 106 and the compression paddle 108 each have a compression surface 110 and 112, respectively, that move towards each other to compress, immobilize, stabilize, or otherwise hold and secure the breast 102 during imaging procedures. In known systems, the compression surface 110, 112 is exposed so as to directly contact the breast 102. The platform 106 also houses an image receptor 116 and, optionally, a tilting mechanism 118, and optionally an anti-scatter grid (not depicted, but disposed above the image receptor 116). The immobilizer unit 104 is in a path of an imaging beam 120 emanating from x-ray source 122, such that the beam 120 impinges on the image receptor 116.

(13) The immobilizer unit 104 is supported on a first support arm 124 and the compression paddle 108 is supported on the first support arm 124 by a paddle mount 126, which is configured to be raised and lowered along the support arm 124. The x-ray source 122 is supported on a second support arm, also referred to as a tube head 128. For mammography, support arms 124 and 128 can rotate as a unit about an axis 130 between different imaging orientations such as CC and MLO, so that the system 100 can take a mammogram projection image at each orientation. In operation, the image receptor 116 remains in place relative to the platform 106 while an image is taken. The immobilizer unit 104 releases the breast 102 for movement of arms 124, 128 to a different imaging orientation. For tomosynthesis, the support arm 124 stays in place, with the breast 102 immobilized and remaining in place, while at least the second support arm 128 rotates the x-ray source 122 relative to the immobilizer unit 104 and the compressed breast 102 about the axis 130. The system 100 takes plural tomosynthesis projection images of the breast 102 at respective angles of the beam 120 relative to the breast 102.

(14) Concurrently and optionally, the image receptor 116 may be tilted relative to the breast support platform 106 and in sync with the rotation of the second support arm 128. The tilting can be through the same angle as the rotation of the x-ray source 122, but may also be through a different angle selected such that the beam 120 remains substantially in the same position on the image receptor 116 for each of the plural images. The tilting can be about an axis 132, which can but need not be in the image plane of the image receptor 116. The tilting mechanism 118 that is coupled to the image receptor 116 can drive the image receptor 116 in a

tilting motion. For tomosynthesis imaging and/or CT imaging, the breast support platform **106** can be horizontal or can be at an angle to the horizontal, e.g., at an orientation similar to that for conventional MLO imaging in mammography. The system **100** can be solely a mammography system, a CT system, or solely a tomosynthesis system, other modalities such as ultrasound, or a “combo” system that can perform multiple forms of imaging. An example of a system has been offered by the assignee hereof under the trade name Selenia Dimensions.

(15) Whether operating in a mammography or a tomosynthesis mode, the system **100** images the breast **102** by emitting the x-ray beam **120** from the x-ray source **122**. The x-ray beam **120** passes through the breast **102** where it is detected by the image receptor **116**. The image receptor **116** may include a plurality of pixels that detect the intensity of the x-ray beam **120** at a plurality of locations after the x-ray beam **120** has passed through the breast **102**. The attenuation of the x-ray beam **120** as it passes through the breast **102** changes depending on the structures of the breast **102**. Accordingly, images of the breast **102** may be produced from the detected x-ray beam **120**. For instance, the image receptor **116** produces imaging information in the form of electric signals, and supplies that imaging information to an image processor **134** for processing and generating x-ray images of the breast **102**. A system control and work station unit **136** including software controls the operation of the system **100** and interacts with the operator to receive commands and deliver information including processed-ray images. The system control and work station unit **136** may also include software for controlling the operation of the x-ray source **122**.

(16) The imaging system **100** includes a floor mount or base **138** for supporting the imaging system **100** on a floor. A gantry **140** extends upwards from the floor mount **138** and rotatably supports both the tube head **128** and the support arm **124**. The tube head **128** and support arm **124** are configured to rotate discretely from each other and may also be raised and lowered along a face **142** of the gantry **140** so as to accommodate patients of different heights. The x-ray source **122** is disposed within the tube head **128**. Together, the tube head **128** and support arm **124** may be referred to as a C-arm **144**. A number of interfaces and display screens are disposed on the imaging system **100**. These include a foot display screen **146**, a gantry interface **148**, a support arm interface **150**, and a compression arm interface **152**. In general the various interfaces **146**, **148**, **150**, and **152** may include one or more tactile buttons, knobs, switches, as well as one or more display screens, including capacitive touch screens with graphic user interfaces (GUIs) so as to enable user interaction with and control of the imaging system **100**.

(17) FIG. **3** is a schematic view of an x-ray tube **200**. The x-ray tube **200** may be included as at least part of the x-ray source **122** discussed above. The x-ray tube **200** includes a tube body **202** housing a cathode **204** and an anode **206**. The cathode **204** includes a filament **208** and, in some examples, a focusing cup **210**. The filament **208** can be placed adjacent the focusing cup **210** and between the focusing cup **210** and the anode **206**. The filament **208** may be formed from a material with a high melting point, such as tungsten. A voltage or signal $V_{sub.1}$ may be applied across the filament **208** via wires connected to each end of the filament **208**, indicated by the $1+$ for the positive connection to the filament **208** and the $1-$ for the negative connection to the filament **208**. When the signal or voltage is applied across the filament **208**, a current $I_{sub.1}$ flows through the filament **208** which heats the filament **208** and causes electrons to be emitted from the filament **208**. Due to a high voltage potential $V_{sub.2}$ applied between the cathode **204** and the anode **206** as indicated by the $2+$ for the positive connection to the anode **206** and the $2-$ for the negative connection to the cathode **204**, the electrons emitted from the filament **208** are accelerated towards the anode **206**. The accelerated electrons form an electron beam **212** that travels along an electron beam path between the cathode **204** and the anode **206**. The electron beam **212** impacts the anode **206** at a focal spot **214** and causes the emission of x-rays **216** from the anode **206**. The x-rays **216** exit the x-ray tube body **202** through a tube window **218**. The x-rays **216** that exit from the body **202** form the x-ray beam that is used for imaging, such as the x-ray beam **120** discussed above with reference to FIGS. **1** and **2**.

(18) The area that the electron beam **212** impacts the anode **206** is referred to as the focal spot **214**. This size of the focal spot **214** relates to the resolution required or desired for the imaging process. The location of the focal spot **214** on the anode **206**, as well as, the angle of the anode **206**, have an effect on the direction of the x-rays **216** produced from the anode **206**. The size and location of the focal spot **214** may be controlled or modified by the focusing cup **210**. Additionally, the electron beam **212** is produced by thermionic effect from the filament **208** being heated by the electric current $I_{sub.1}$. This current value of the filament **208**, via applied voltage $V_{sub.1}$ across the filament **208**, determines a tube output value $I_{sub.2}$, typically measured in milliamperes (mA), for a given tube high voltage potential $V_{sub.2}$, typically measured in kilovoltage (kV).

(19) In the example, the x-ray tube **200** is included in a high voltage control circuit **220** (also known as a current control circuit) that is configured to control operation of the x-ray tube **200**. The high voltage control circuit **220** may include a micro-computer **222** that stores operational data and processes for the x-ray tube **200** and facilitates operation of the x-ray tube **200**. For example, the high voltage control circuit **220**, via the micro-computer **222** coupled in communication with the x-ray tube **200**, applies voltage $V_{sub.1}$ across the filament **208** (e.g. via channeling current through the filament) based on stored calibration data so as to emit x-rays **216** having a required or desired radiation amount (e.g., radiation dose). In x-ray imaging, control of the amount of x-ray radiation for increased image quality is desired. In an aspect, the x-ray dose is proportional to the tube output value $I_{sub.2}$ per second (mAs) for a given kV tube voltage potential $V_{sub.2}$.

(20) In operation, when a new x-ray tube **200** is installed in an x-ray imaging system, the current applied $I_{sub.1}$ (e.g., via voltage $V_{sub.1}$ generation) at the filament **208** is calibrated for a plurality of kV tube voltage $V_{sub.2}$ and mA tube output $I_{sub.2}$ stations. In an aspect, every x-ray tube **200** can have unique performance characteristics, and as such, each x-ray tube may have a slightly different calibration. Additionally, the x-ray tube **200** can also be re-calibrated as required or desired as performance characteristics change. As used herein, a station corresponds to discrete kV tube voltage $V_{sub.2}$ and mA tube output $I_{sub.2}$ values that produce, and reproduce, the same quantity of radiation during an exposure. In an aspect, power applied at the filament **208** is based at least partially on filament current $I_{sub.1}$ and controls the tube output current values. However, as the filament **208** wears and degrades with use, the filament material evaporates under high temperature, which results in a thinning of the filament material and in turn, an increase of filament resistance R . As the resistance R increases in the filament **208**, during operation of the x-ray tube, the power (e.g., via the voltage $V_{sub.1}$ applied) generated at the filament **208** increases and the temperature of the filament **208** increases, thereby resulting in an increased radiation output from the same voltage $V_{sub.1}$ being applied across the filament **208**.

(21) In some systems, the high voltage control circuit **220** may measure the mA tube output I.sub.2 of the x-ray tube after every exposure sequence (e.g. via an ammeter or the like). As used herein, an exposure sequence is the x-ray emissions during a modality procedure. Accordingly, in some modalities, such as mammography, an exposure sequence may be a single emission and the mA tube output I.sub.2 measurement occurs prior to firing the x-ray. In other modalities, such as tomosynthesis, an exposure sequence may be a plurality of emissions and the mA tube output I.sub.2 measurement occurs after firing the x-rays. This measured actual mA output value I.sub.2 of the exposure sequence may then be compared to the desired mA output value I.sub.2 for the exposure sequence. If the difference between the actual mA output value I.sub.2 and the desired mA output value I.sub.2 exceeds a predetermined threshold, then the control circuit **220** may prompt for a new filament calibration. In an aspect, the predetermined threshold may be based on how much different the actual output is from the desired output while still maintaining accurate and usable x-ray images. In some examples, the predetermined threshold may be between about 35%-20%. In other examples, the predetermined threshold may be about 33%. However, if the difference between the actual mA output value I.sub.2 and the desired mA output value I.sub.2 does not exceed the predetermined threshold, then the control circuit **220** can adjust the voltage V.sub.1 being applied to the filament **208** for the next exposure sequence. This voltage adjustment by the control circuit **220** enables the subsequent measured mA output values I.sub.2 to be closer to the desired mA output values I.sub.2 and increase x-ray imaging performance.

(22) Adjusting the voltage V.sub.1 applied to the filament **208** in the process described above, however, only adjusts the voltage V.sub.1 for the active kV tube voltage V.sub.2 and mA tube output I.sub.2 station relating to the specific the exposure sequence. The other stations (e.g., inactive stations) that are not in use do not gain the benefit of the voltage adjustment algorithm. Furthermore, the mA tube output I.sub.2 measurements only account for the fact that the mA output has changed in the x-ray tube **200** and does not address the underlying tube characteristics that caused the changed mA output, and as such, the accuracy of the adjustments are reduced. Accordingly and as described above, it is the resistive properties R of the filament **208** that change over time and result in the changing mA output values I.sub.2 of the x-ray tube **200**. Thus, and in the examples described herein, the high voltage control circuit **220** is configured to monitor the resistive properties R of the filament **208**. As such, the current I.sub.1 applied to the filament **208** (e.g., via voltage V.sub.1) may be adjusted based on the monitoring of the resistance properties R of the filament **208**. By using the resistive properties R of the filament **208**, each and every kV tube voltage V.sub.2 and mA tube output I.sub.2 station can be adjusted, thereby increasing x-ray imaging performance. Additionally, the accuracy of the adjustments from the x-ray tube calibration baseline values are increased because the adjustments are based on the underlying changing characteristics of the filament **208** that cause the change in mA output values I.sub.2.

(23) In the example, the high voltage control circuit **220** can include a monitoring instrument **224** operatively controlled by the micro-computer **222** to monitor the resistance values R of the filament **208**. In an aspect, the monitoring instrument **224** may be a multimeter or the like that is configured to measure electrical properties of the filament **208**, such as one or more of voltage, resistance, and current. Additionally or alternatively, the monitoring instrument **224** can calculate resistance R of the filament **208** using Ohm's law when the current and voltage applied across the filament **208** are known. The control circuit **220** also may be configured with a clock or other time keeping instrument whereby elapsed time may be monitored and certain operations can be performed at predetermined time sampling periods, intervals, or cycles. Accordingly, when the resistance R of the filament **208** changes by a certain amount over time, an algorithm for the control circuit **220** can adjust the filament current I.sub.1, via the voltage V.sub.1, to get the filament power the same as the stored calibration value for each kV tube voltage V.sub.2 and mA tube output I.sub.2 station. This leads to the x-ray tube **200** being capable of delivering the required or desired mA tube output value I.sub.2 even with changing filament resistance values R.

(24) FIG. **4** is a graph **300** depicting filament resistance change over a life of the x-ray tube **200** (shown in FIG. **3**). As described above, the change in the resistance properties of the filament **208** (shown in FIG. **3**) changes over the operational life of the x-ray tube **200** due to evaporation of filament material under high temperatures. The graph **300** has an x-axis **302** that counts the number of exposures of the x-ray tube **200**, and thus, is a time component for the filament **208**. In the example, the x-ray exposures are at a 33 kV tube voltage and a 200 mA tube current with 60 mAs. The mAs is the product of tube current in mA and time in seconds. A y-axis **304** charts the resistance in Ohms of the filament **208**. As illustrated in the graph **300**, three different x-ray tubes **306**, **308**, **310** are shown to have their filament resistance change during the operational use of the x-ray tubes. Each x-ray tube **306**, **308**, **310** is slightly different because all x-ray tubes are unique due to the design and manufacture of the individual x-ray tubes. However, all of the x-ray tubes **306**, **308**, **310** demonstrate that filament resistance values generally increase over time. This time period of change is over a period ten-thousand plus exposure counts, and thus, filament resistance does not change noticeably after every exposure count.

(25) As used herein, the term “life” or “life-cycle” of the x-ray tube does not necessarily mean the life cycle to a point of physical failure. Rather, “life” or “life cycle” refers to a period of usage after which the x-ray tube no longer performs as required or desired.

(26) FIG. **5** is a graph **330** depicting filament current change over the life of the x-ray tube **200** (shown in FIG. **3**). As discussed in reference to FIG. **4** above, the resistance properties of the filament **208** (shown in FIG. **3**) changes during operational use of the x-ray tube **200**. Accordingly, as the filament resistance changes, the current applied at the filament **208** has to be changed during the operational use of the x-ray tube **200** to get desired tube output. The graph **330** has an x-axis **332** that counts the number of exposures of the x-ray tube **200**, and thus, is a time component for the filament **208**. A y-axis **334** charts the current in amperes of the filament **208**. The graph **330** shows the filament current of the same three x-ray tubes **306**, **308**, **310** as FIG. **4** and at 33 kV tube voltage and 200 mA tube output. The curves illustrate the change in the filament current needed to maintain a constant 200 mA tube output as the filament resistance value increases over time.

(27) Typically, when a new x-ray tube is installed in an imaging system, the filament current is calibrated for the required or desired kV tube voltage and mA tube current stations and these calibration values are stored within the system so that imaging can occur with the correct amount of x-ray radiation. This calibration process also occurs during recalibration of the x-ray tube **200**. However, as shown in the graphs **300**, **330**, the resistive properties of the filament **208** changes and the filament current being

applied at the filament **208** has to be reduced in order to deliver the required or desired mA tube output value. This is because as the filament resistance rises, the power at the filament **208** increases (e.g., power $P=I_{\text{sub.2}} \times R$) and the temperature of the filament increases if filament current is not adjusted, resulting in increased radiation output at the x-ray tube **200**. As described herein, monitoring the change in resistance values of the filament **208** and generating feedback control of the x-ray tube **200** based on the filament resistance allows the filament current to be adjusted to maintain a required or desired x-ray radiation at each kV tube voltage and mA tube output station.

(28) FIG. **6** is a graph **360** depicting filament power change over the life of the x-ray tube **200** (shown in FIG. **3**). The graph **360** has an x-axis **362** that counts the number of exposures of the x-ray tube **200**, and thus, is a time component for the filament **208** (shown in FIG. **3**). A y-axis **364** charts the power in watts of the filament **208**. The graph **360** shows the filament power of the same three x-ray tubes **306**, **308**, **310** as FIGS. **4** and **5**, and at a 33 kV tube voltage and 200 mA tube output. These curves demonstrate that to maintain the same mA tube output over the life of the x-ray tube **200** it is required to maintain the same filament power as right after filament current calibration for each kV tube voltage and mA tube output station. As such, to keep mA tube output within a required or desired value with filament resistance increasing over time, the filament current must be adjusted periodically to operate at a proper filament current for each station.

(29) Referring to FIGS. **4-6**, the graphs illustrate that filament current adjustment over life of the filament is needed to get a required or desired tube current mA. While FIGS. **4-6** are illustrative of a single station, the algorithm described herein can adjust filament current across active and inactive stations within the imaging system.

(30) FIG. **7** depicts a flowchart illustrating a method **400** of adaptively controlling filament current in an x-ray tube. The example methods and operations can be implemented or performed by the systems and devices described herein (e.g., the imaging system **100** and x-ray tube **200** shown in FIGS. **1-3**). The method **400** begins with calibrating (or re-calibrating) the x-ray tube having a filament therein (operation **402**) and storing the calibration data at the imaging system (operation **404**). In an aspect, the calibration data includes a filament current value that determines a tube output value for a tube voltage value at a plurality of stations that is generated from the calibration of the x-ray tube. Each station is defined by the tube output value in mA and the tube voltage value in kV so that a required or desired x-ray radiation amount is generated by the x-ray tube. In some examples, the filament current value may include a filament voltage value that is applied across the filament to generate the current value. The x-ray calibration procedure can be performed by methods that are currently known (e.g., a substitution method) or developed in the future. Because every x-ray tube has slightly different performance characteristics, calibration enables accurate operation of the imaging system as described herein.

(31) The method **400** continues with monitoring a resistance value of the filament over a period of time (operation **406**). As described above, the filament material evaporates under high temperatures and results in a thinning of the filament material, thereby increasing filament resistance over the operational lifetime of the x-ray tube. In an aspect, monitoring the resistance value of the filament may include measuring the resistance value after a predetermined number of sequence exposures and storing each measured resistance value. Based on the stored resistance values a resistance curve can be generated. In an example, the period of time for monitoring the resistance value can be based on the number of exposure counts and for example resistance values can be measured every 100 sequence exposures since it is known that resistance changes in an order many times this predetermined number. In other examples, the resistance value can be measured after every sequence exposure and stored as required or desired. As such, measuring and storing the resistance value is performed periodically.

(32) In the example, measuring the resistance value includes applying a constant and known filament current value to the filament so as to measure the resistance. In an aspect, filament voltage is applied so that the filament receives about 2.5 amps to measure the resistance since voltage, current, and resistance are related via Ohm's law. Because each sequence exposure generates heat at the filament, in some examples, prior to measuring the resistance value, the filament is allowed to cool so as to increase accuracy of the resistance value measurement. In an example, cooling the filament can include waiting a predetermined time period between an end of the sequence exposures and prior to applying the constant filament current value for measuring the resistance value of the filament.

(33) Turning back to the method **400**, once the resistance value of the filament is being monitored (operation **406**), a change in the resistance value of the filament is determined over the period of time (operation **408**). Because the resistive characteristics of the filament change slowly after every exposure, and the tolerance for each exposure does allow for some change to the tube current output, the filament current value does not need to be revised or adjusted for every exposure count. As such, the change in resistance value can be determined by comparing a difference between resistance values over a period of time to a predetermined benchmark. For example, the predetermined benchmark may be a 4% resistance value change over at least 10,000 exposure counts. It should be appreciated that other predetermined benchmark values are also contemplated herein. For example, but not limiting, the resistance value change may be 2%, 5%, 10%, or the like, and the exposure count may be 5,000, 8,000, 12,000, or the like as required or desired.

(34) Once it has been determined that the resistance value of the filament has changed (operation **408**), then the filament current value is adjusted for at least one of the plurality of stations based on the changed resistance value (operation **410**). By adjusting the filament current applied at the filament, the power at the filament during operation can be the same as the calibrated data for the tube current value and the tube voltage value for the station. Accordingly, the x-ray tube can deliver the required or desired tube current value even with an ever changing filament resistance value. In an aspect, the filament current value of each of the plurality of stations is updated based on the changed resistance value. As such, the calibration data can be adjusted based on the current resistive properties of the filament, and less active stations are still configured for a more accurate tube current value output the next time those stations are used for imaging procedures. In an aspect, adjusting the filament current value is based on a power model, whereby the filament power is maintained as right after initial filament calibration.

(35) In the example, once the filament current value is adjusted the process may repeat itself with monitoring the resistance value over a new and reset period of time, determining change in the resistance value, and further adjust the filament current value as required or desired. This feedback algorithm for the resistance value enables the imaging system to operate longer between

maintenance operations that calibrate/re-calibrate the x-ray tube.

(36) In aspects, the method **400** may further include measuring a resulted tube current value from the x-ray tube at a sequence exposure having a stored tube current value and tube voltage value for a station of the plurality of stations (operation **412**), and comparing a difference between the measured resulted tube current value and the stored tube current value for the respective station (operation **414**). By measuring the resulted mA tube output current from the x-ray tube during operation and comparing the measured value to the stored desired output, operational data of the x-ray tube may be generated. For example, the accuracy of the feedback algorithm can be verified by this operational data, and if the adjusted filament current value results in less accurate tube current output values, a new filament calibration may be required. In another example, if the difference between the measured tube output current and the desired output is higher than a predetermined threshold, a new filament calibration may be required.

(37) In other examples, based on the difference between the measured resulted tube current value and the stored tube current value, the period of time is at least partially defined. For example, instead of monitoring the resistance value based on the number of exposure counts, the resistance value can be monitored based on the operational results of the tube head, and if the difference between the measured tube current and the stored desired tube current exceeds a predetermined threshold, the change in resistance value can be used to adjust the filament current value as described above. In aspects, the measurement of the resulted tube current value may be performed pre-exposure or post-exposure of the sequence exposure. For example, in tomosynthesis imaging modalities the tube current measurements may be performed post-exposure, while in mammography imaging modalities the tube current measurement may be performed pre-exposure.

(38) The method **400** may also include using the monitored resistance value of the filament to determine a life-cycle period of the filament (operation **416**). For example, because filament resistance increases during the operational life of the filament due to filament material evaporating, an upper limit of filament resistance can be defined that relates to the filament material being too thin and triggering a tube replacement and a new filament calibration for the new x-ray tube.

(39) FIG. **8** illustrates an exemplary suitable operating environment **500** for controlling an x-ray tube. This operating environment may be incorporated directly into the high voltage control circuit disclosed herein, or may be incorporated into a computer system discrete from, but used to control the breast imaging systems described herein. Such computer system may be, for example, the system control and work station **136** depicted in FIG. **1**. This is only one example of a suitable operating environment and is not intended to suggest any limitation as to the scope of use or functionality. Other well-known computing systems, environments, and/or configurations that can be suitable for use include, but are not limited to, imaging systems, personal computers, server computers, hand-held or laptop devices, multiprocessor systems, microprocessor-based systems, programmable consumer electronics such as smart phones, network PCs, minicomputers, mainframe computers, tablets, distributed computing environments that include any of the above systems or devices, and the like.

(40) In its most basic configuration, operating environment **500** typically includes at least one processing unit **502** and memory **504**. Depending on the exact configuration and type of computing device, memory **504** (storing, among other things, instructions to read from data storage devices or sensors, or perform other methods disclosed herein) can be volatile (such as RAM), non-volatile (such as ROM, flash memory, etc.), or some combination of the two. This most basic configuration is illustrated in FIG. **8** by dashed line **506**. Further, environment **500** can also include storage devices (removable, **508**, and/or non-removable, **510**) including, but not limited to, magnetic or optical disks or tape. Similarly, environment **500** can also have input device(s) **514** such as touch screens, keyboard, mouse, pen, voice input, etc., and/or output device(s) **516** such as a display, speakers, printer, etc. Also included in the environment can be one or more communication connections **512**, such as LAN, WAN, point to point, Bluetooth, RF, etc. In embodiments, the connections may be operable to facility point-to-point communications, connection-oriented communications, connectionless communications, etc.

(41) Operating environment **500** typically includes at least some form of computer readable media. Computer readable media can be any available media that can be accessed by processing unit **502** or other devices having the operating environment. By way of example, and not limitation, computer readable media can include computer storage media and communication media. Computer storage media includes volatile and nonvolatile, removable and non-removable media implemented in any method or technology for storage of information such as computer readable instructions, data structures, program modules or other data. Computer storage media includes, RAM, ROM, EEPROM, flash memory or other memory technology, CD-ROM, digital versatile disks (DVD) or other optical storage, magnetic cassettes, magnetic tape, magnetic disk storage or other magnetic storage devices, solid state storage, or any other tangible medium which can be used to store the desired information. Computer storage media does not include communication media.

(42) Communication media embodies computer readable instructions, data structures, program modules, or other data in a modulated data signal such as a carrier wave or other transport mechanism and includes any information delivery media. The term “modulated data signal” means a signal that has one or more of its characteristics set or changed in such a manner as to encode information in the signal. By way of example, and not limitation, communication media includes wired media such as a wired network or direct-wired connection, and wireless media such as acoustic, RF, infrared and other wireless media. Combinations of the any of the above should also be included within the scope of computer readable media. A computer-readable device is a hardware device incorporating computer storage media.

(43) The operating environment **500** can be a single computer operating in a networked environment using logical connections to one or more remote computers. The remote computer can be a personal computer, a server, a router, a network PC, a peer device or other common network node, and typically includes many or all of the elements described above as well as others not so mentioned. The logical connections can include any method supported by available communications media. Such networking environments are commonplace in offices, enterprise-wide computer networks, intranets and the Internet.

(44) The embodiments described herein may be employed using software, hardware, or a combination of software and hardware to implement and perform the systems and methods disclosed herein. Although specific devices have been recited throughout the disclosure as performing specific functions, one of skill in the art will appreciate that these devices are provided for illustrative purposes, and other devices may be employed to perform the functionality disclosed herein without departing from the scope of

the disclosure. In possible aspects of the present disclosure are described above with reference to block diagrams and/or operational illustrations of systems and methods according to aspects of this disclosure. The functions, operations, and/or acts noted in the blocks may occur out of the order that is shown in any respective flowchart. For example, two blocks shown in succession may in fact be executed or performed substantially concurrently or in reverse order, depending on the functionality and implementation involved.

(45) This disclosure describes some examples of the present technology with reference to the accompanying drawings, in which only some of the possible examples were shown. Other aspects may, however, be embodied in many different forms and should not be construed as limited to the examples set forth herein. Rather, these examples were provided so that this disclosure was thorough and complete and fully conveyed the scope of the possible examples to those skilled in the art. Further, as used herein and in the claims, the phrase “at least one of element A, element B, or element C” is intended to convey any of: element A, element B, element C, elements A and B, elements A and C, elements B and C, and elements A, B, and C. Additionally, one having skill in the art will understand the degree to which terms such as “about” or “substantially” convey in light of the measurement techniques utilized herein. To the extent such terms may not be clearly defined or understood by one having skill in the art, the term “about” shall mean plus or minus ten percent.

(46) Although specific examples were described herein, the scope of the technology is not limited to those specific examples. One skilled in the art will recognize other examples or improvements that are within the scope and spirit of the present technology. Therefore, the specific structure, acts, or media are disclosed only as illustrative examples. Examples according to the technology may also combine elements or components of those that are disclosed in general but not expressly exemplified in combination, unless otherwise stated herein. The scope of the technology is defined by the following claims and any equivalents therein.

Claims

1. A method of adaptively controlling filament current in an x-ray tube of an imaging system, the method comprising: calibrating the x-ray tube having a filament; storing calibration data from the calibration of the x-ray tube at the imaging system, wherein the calibration data includes a filament current value that determines a tube current value for a tube voltage value at a plurality of exposure values; monitoring a resistance value of the filament over a period of time, wherein monitoring the resistance value includes measuring the resistance value of the filament after a predetermined number of sequence exposures and storing each measured resistance value; determining a change in the resistance value of the filament over the period of time, wherein the determined changed resistance value is based on a difference between the stored measured resistance values over a number of exposure counts; and adjusting the filament current value of at least one of the plurality of exposure values based on the determined changed resistance value.
2. The method of claim 1, wherein measuring the resistance value after the predetermined number of sequence exposures includes applying a constant filament current value to the filament after each of the predetermined number of sequence exposures.
3. The method of claim 2, further comprising waiting a predetermined time period between an end of the predetermined number of sequence exposures and prior to applying the constant filament current value to the filament for resistance value measurement.
4. The method of claim 1, wherein the filament current value of each of the plurality of exposure values is updated based on the changed resistance value.
5. The method of claim 1, wherein determining the change in the resistance value includes comparing a difference between resistance values over the period of time to a predetermined benchmark.
6. The method of claim 1, further comprising: measuring a resulted tube current value from the x-ray tube at a sequence exposure having a stored tube current value and tube voltage value for an exposure value of the plurality of exposure values; and comparing a difference between the measured resulted tube current value and the stored tube current value for the respective exposure value.
7. The method of claim 6, wherein based on the difference between the measured resulted tube current value and the stored tube current value, the period of time is at least partially defined.
8. The method of claim 6, wherein measuring resulted tube current value is performed pre-exposure or post-exposure of the sequence exposure.
9. The method of claim 1, further comprising determining a life-cycle period of the filament based at least partially on the monitored resistance value of the filament.
10. An imaging system comprising: an x-ray tube having a filament; a current control circuit coupled to the x-ray tube and configured to channel current through the filament; at least one processor communicatively coupled to the current control circuit; and memory communicatively coupled to the at least one processor, the memory comprising computer executable instructions that, when executed by the at least one processor, performs a method comprising: calibrating the x-ray tube; storing calibration data from the calibration of the x-ray tube at the imaging system, wherein the calibration data includes a filament current value that determines a tube current value for a tube voltage value at a plurality of exposure values; monitoring a resistance value of the filament over a period of time, wherein monitoring the resistance value includes measuring the resistance value of the filament after a predetermined number of sequence exposures and storing each measured resistance value; determining a change in the resistance value of the filament over the period of time, wherein the determined changed resistance value is based on a difference between the stored measured resistance values over a number of exposure counts; and adjusting the filament current value of at least one of the plurality of stations exposure values based on the determined changed resistance value.
11. The imaging system of claim 10, wherein measuring the resistance value after the predetermined number of sequence exposures includes applying a constant filament current value to the filament after each of the predetermined number of sequence exposures.
12. The imaging system of claim 11, wherein the method further comprises waiting a predetermined time period between an end of the predetermined number of sequence exposures and prior to applying the constant filament current value to the filament for resistance value measurement.

13. The imaging system of claim 11, wherein the filament current value of each of the plurality of exposure values is updated based on the changed resistance value.

14. The imaging system of claim 10, wherein determining the change in the resistance value includes comparing a difference between resistance values over the period of time to a predetermined benchmark.

15. The imaging system of claim 10, wherein the method further comprises: measuring a resulted tube current value from the x-ray tube at a sequence exposure having a stored tube current value and tube voltage value for an exposure value of the plurality of exposure values; and comparing a difference between the measured resulted tube current value and the stored tube current value for the respective exposure value.

16. The imaging system of claim 15, wherein based on the difference between the measured resulted tube current value and the stored tube current value, the period of time is at least partially defined.

17. The imaging system of claim 15, wherein measuring resulted tube current value is performed pre-exposure or post-exposure of the sequence exposure.

18. The imaging system of claim 10, wherein the method further comprises determining a life-cycle period of the filament based at least partially on the monitored resistance value of the filament.

19. A method of adaptively controlling filament current in an x-ray tube of an imaging system, the method comprising: calibrating the x-ray tube having a filament; storing calibration data from the calibration of the x-ray tube at the imaging system, wherein the calibration data includes a filament current value that determines a tube current value for a tube voltage value at a plurality of exposure values; monitoring a resistance value of the filament over a period of time; determining a change in the resistance value of the filament over the period of time; adjusting the filament current value of at least one of the plurality of exposure values based on the changed resistance value; measuring a resulted tube current value from the x-ray tube at a sequence exposure having a stored tube current value and tube voltage value for an exposure value of the plurality of exposure values; and comparing a difference between the measured resulted tube current value and the stored tube current value for the respective exposure value, wherein based on the difference between the measured resulted tube current value and the stored tube current value, the period of time is at least partially defined.
