

US Patent & Trademark Office

Patent Public Search | Text View

United States Patent	12383234
Kind Code	B2
Date of Patent	August 12, 2025
Inventor(s)	Park; Ji Young et al.

Ultrasonic imaging apparatus and method for controlling the same

Abstract

Provided is an ultrasonic imaging apparatus capable of obtaining a more accurate shear wave elastic image by removing inefficient areas such as blood vessels using a shear wave observation signal in a section where an object is not affected by a push pulse, and a method for controlling the same. The ultrasonic imaging apparatus according to an embodiment includes: an ultrasonic probe configured to irradiate a push pulse, irradiate a first observation signal to the object after the push pulse is irradiated, and irradiate a second observation signal to the object; a controller configured to determine an inefficiency region of the object, and generate a shear wave elastic image based on the determined inefficiency region for each region of the object; and a display configured to display the generated shear wave elastic image.

Inventors:	Park; Ji Young (Hwaseong-si, KR), Choi; Ki Wan (Anyang-si, KR), Choi; Seong Hyeon (Seoul, KR)
Applicant:	SAMSUNG MEDISON CO., LTD. (Gangwon-do, KR)
Family ID:	72427610
Assignee:	SAMSUNG MEDISON CO., LTD. (Gangwon-Do, KR)
Appl. No.:	17/430563
Filed (or PCT Filed):	July 25, 2019
PCT No.:	PCT/KR2019/009210
PCT Pub. No.:	WO2020/184790
PCT Pub. Date:	September 17, 2020

Prior Publication Data

Document Identifier	Publication Date
US 20220125412 A1	Apr. 28, 2022

Foreign Application Priority Data

KR

10-2019-0029265

Mar. 14, 2019

Publication Classification

Int. Cl.: A61B8/00 (20060101); A61B8/06 (20060101)

U.S. Cl.:

CPC A61B8/485 (20130101); A61B8/06 (20130101); A61B8/4477 (20130101); A61B8/5207 (20130101); A61B8/54 (20130101);

Field of Classification Search

CPC: A61B (34/10); A61B (34/20); A61B (34/25); A61B (90/37); A61B (2034/102); A61B (2034/105); A61B (2034/2065); A61B (2090/365); A61B (2090/376); A61B (2017/00207); A61B (90/36); A61B (2034/2048); A61B (2090/3612); A61B (2090/372); A61B (2090/3762); A61B (2090/502); G06N (3/045); G06N (3/08)

References Cited

U.S. PATENT DOCUMENTS

Patent No.	Issued Date	Patentee Name	U.S. Cl.	CPC
8852102	12/2013	Miyachi	N/A	N/A
9072493	12/2014	Yoshikawa	N/A	N/A
11504087	12/2021	Kong	N/A	A61B 8/42
2013/0317361	12/2012	Tabaru et al.	N/A	N/A
2016/0350503	12/2015	Jun	N/A	A61B 8/465
2016/0367223	12/2015	Honjo et al.	N/A	N/A
2017/0112471	12/2016	Toji	N/A	A61B 8/4254
2020/0060653	12/2019	Kong et al.	N/A	N/A

FOREIGN PATENT DOCUMENTS

Patent No.	Application Date	Country	CPC
2016200737	12/2016	AU	A61M 25/00
106233333	12/2015	CN	G06T 19/00
2964099	12/2015	EP	N/A
2012-100997	12/2011	JP	N/A
2012-183261	12/2011	JP	N/A
2013-244162	12/2012	JP	N/A
5501999	12/2013	JP	N/A
2015-058193	12/2014	JP	N/A
2015-126955	12/2014	JP	N/A

2017-079977	12/2016	JP	N/A
6230891	12/2016	JP	N/A
10-2018-0054360	12/2017	KR	N/A
10-2018-0109500	12/2017	KR	N/A
10-1931748	12/2018	KR	N/A
2012/105152	12/2011	WO	N/A
2014/136018	12/2013	WO	N/A
WO-2017091013	12/2016	WO	A61N 7/00

OTHER PUBLICATIONS

Cinthio et. al., Longitudinal movements and resulting shear strain of the arterial wall Feb. 10, 2006 (Year: 2006). cited by examiner

Extended European Search Report dated Feb. 11, 2022 issued in European Patent Application No. 19918564.6. cited by applicant

International Search Report dated Dec. 11, 2019 issued in International Patent Application No. KR2019/009210 (with English translation). cited by applicant

Office Action issued Feb. 14, 2024 for European Patent Application No. 19918564.6. cited by applicant

Office Action issued Feb. 19, 2024 for Korean Patent Application No. 10-2019-0029265 (See English Translation). cited by applicant

Notice of Allowance dated Oct. 10, 2024, issued in corresponding Korean Patent Application No. 10-2019-0029265 with an English translation. cited by applicant

Primary Examiner: Mattson; Sean D

Attorney, Agent or Firm: Morgan Lewis & Bockius LLP

Background/Summary

CROSS-REFERENCE OF RELATED APPLICATIONS

(1) This application is the U.S. National Phase under 35 U.S.C. § 371 of International Patent Application No. PCT/KR2019/009210, filed on Jul. 25, 2019, which in turn claims the benefit of Korean Application No. 10-2019-0029265, filed on Mar. 14, 2019, the entire disclosures of which applications are incorporated by reference herein.

TECHNICAL FIELD

(2) The present disclosure relates to an ultrasonic imaging apparatus for obtaining an image inside an object using ultrasound and a method for controlling the same.

BACKGROUND ART

(3) The ultrasonic imaging apparatus irradiates an ultrasonic signal generated from a transducer of a probe to an object, and receives information of an echo signal reflected from the object to obtain an image of a region inside the object.

(4) The ultrasonic imaging apparatus is widely used in the medical diagnosis field because the ultrasonic imaging apparatus has no radiation exposure, has higher stability than an X-ray imaging device, may display images in real time, and is inexpensive and portable compared to a magnetic resonance imaging device.

(5) Unlike normal tissue, when factors such as cancer or cirrhosis occur and the tissue strength changes, the ultrasonic imaging apparatus may generate a shear wave in tissue using a push pulse, may diagnose factors such as cancer or liver cirrhosis more accurately by calculating the elastic

value of the shear wave using the change in tissue displacement caused by the shear wave. However, a region such as a blood vessel in which displacement change occurs by itself or a bone in which displacement change does not occur due to high strength exists inside the body, and in such region, an accurate shear wave elastic value may not be calculated. Also, a problem in that the shear wave elastic value of the region of interest in the object is inaccurately calculated due to movement of the body, such as shaking of the probe, movement of blood vessels, respiration, and heartbeat, may be occurred. In order to solve such a problem, an ultrasonic imaging apparatus that calculates an accurate shear wave elastic value by calculating the displacement of an object tissue using a push pulse is required.

DISCLOSURE

Technical Problem

(6) An ultrasonic imaging apparatus capable of obtaining a more accurate shear wave elastic image by removing inefficient areas such as blood vessels using a shear wave observation signal in a section where an object is not affected by a push pulse, and a method for controlling the same are provided.

Technical Solution

(7) An aspect of the present disclosure provides a ultrasonic imaging apparatus comprising an ultrasonic probe configured to irradiate a push pulse to generate a shear wave to an object, irradiate a first observation signal to the object after the push pulse is irradiated, and irradiate a second observation signal to the object before the push pulse is irradiated or after the first observation signal is irradiated; a controller configured to determine an inefficiency region of the object based on displacement data for each region of the object obtained from the irradiated second observation signal, and generate a shear wave elastic image based on the determined inefficiency region and displacement data for each region of the object obtained from the irradiated first observation signal; and a display configured to display the generated shear wave elastic image.

(8) The first observation signal and the second observation signal may include a plurality of ultrasound signals having a predetermined time interval, and a time interval of each of the plurality of ultrasound signals included in the second observation signal may be longer than or equal to a time interval of each of the plurality of ultrasound signals included in the first observation signal.

(9) The second observation signal may be irradiated for a shorter or equal time than the first observation signal.

(10) The time interval of each of the plurality of ultrasound signals included in the second observation signal may be less than or equal to 500 microsecond.

(11) The second observation signal may be irradiated for more than 10 millisecond.

(12) The inefficiency region may include at least one of a blood vessel region in which blood flow exists and a region in which tissue movement occurs.

(13) The controller may determine a region in which a displacement value or a shear wave velocity value is greater than or equal to a predetermined value among regions of the object, and determine the determined region as an inefficiency region, based on the displacement data for each area of the object obtained from the irradiated second observation signal.

(14) The controller may determine a shear wave elastic value for each area of the object based on the displacement data for each area of the object obtained from the irradiated first observation signal, generate a shear wave elastic image of the object based on the determined shear wave elastic value for each region, and generate a shear wave elastic image so that the shear wave elastic value of the determined inefficiency region is not displayed.

(15) The controller may determine a shear wave elastic value for each area of the object based on the displacement data for each area of the object obtained from the irradiated first observation signal, generate a shear wave elastic image of the object based on the determined shear wave elastic value for each region, and generate a shear wave elastic image by changing a reliability measurement index (RMI) of the shear wave elastic value of the determined inefficiency region.

- (16) The controller may control the display to display a shear wave elastic image in which the determined shear wave elastic value of the inefficient region is not displayed.
- (17) The controller may control the display to display a shear wave elastic image to which the changed reliability measurement index (RMI) is applied.
- (18) The controller may control the display to indicate the changed reliability measurement index (RMI).
- (19) The controller may control the display to display the determined inefficiency area.
- (20) Another aspect of the present disclosure provides a method for controlling ultrasonic imaging apparatus, comprising irradiating an object with a second observation signal; irradiating a push pulse for generating a shear wave to the object after the second observation signal is irradiated to the object; irradiating a first observation signal to the object after the push pulse is irradiated; determining an inefficiency region of the object based on displacement data for each region of the object obtained from the irradiated second observation signal; generating a shear wave elastic image based on the displacement data for each region of the object obtained from the irradiated first observation signal and the determined inefficiency region; and displaying the generated shear wave elastic image.
- (21) The first observation signal and the second observation signal may include a plurality of ultrasound signals having a predetermined time interval, and a time interval of each of the plurality of ultrasound signals included in the second observation signal may be longer than or equal to a time interval of each of the plurality of ultrasound signals included in the first observation signal.
- (22) The second observation signal may be irradiated for a shorter or equal time than the first observation signal.
- (23) The time interval of each of the plurality of ultrasound signals included in the second observation signal may be less than or equal to 500 microsecond.
- (24) The second observation signal may be irradiated for more than 10 millisecond.
- (25) The inefficiency region may include at least one of a blood vessel region in which blood flow exists and a region in which tissue movement occurs.
- (26) The determining of an inefficiency region of the object based on displacement data for each region of the object obtained from the irradiated second observation signal may include determining a region in which a displacement value or a shear wave velocity value is greater than or equal to a predetermined value among regions of the object, and determining the determined region as an inefficiency region based on the displacement data for each area of the object obtained from the irradiated second observation signal.
- (27) The generating a shear wave elastic image based on the displacement data for each region of the object obtained from the irradiated first observation signal and the determined inefficiency region, may include determining a shear wave elastic value for each area of the object based on the displacement data for each area of the object obtained from the irradiated first observation signal; and generating a shear wave elastic image based on the determined shear wave elastic value for each region and generating the shear wave elastic image so that the determined shear wave elastic value of the inefficient region is not displayed.
- (28) The generating a shear wave elastic image based on the displacement data for each region of the object obtained from the irradiated first observation signal and the determined inefficiency region, may include determining a shear wave elastic value for each area of the object based on the displacement data for each area of the object obtained from the irradiated first observation signal; generating the shear wave elastic image based on the determined shear wave elastic value for each region; and generating a shear wave elastic image by changing a reliability measurement index (RMI) of the shear wave elastic value of the determined inefficiency region.
- (29) The method for controlling ultrasonic imaging apparatus according to an embodiment, may further include displaying a shear wave elastic image in which the shear wave elastic value of the determined inefficiency region is not displayed.

(30) The method for controlling ultrasonic imaging apparatus according to an embodiment, may further include displaying the changed reliability measurement index (RMI).

(31) The method for controlling ultrasonic imaging apparatus according to an embodiment, may further include displaying the determined inefficiency area.

(32) The method for controlling ultrasonic imaging apparatus according to an embodiment, the method may include irradiating a push pulse to the object to generate a shear wave to an object; irradiating a first observation signal to the object after the push pulse is irradiated; irradiating a second observation signal to the object after the first observation signal is irradiated; determining an inefficiency region of the object based on displacement data for each region of the object obtained from the irradiated second observation signal; generating a shear wave elastic image based on the displacement data for each region of the object obtained from the irradiated first observation signal and the determined inefficiency region; and displaying the generated shear wave elastic image.

Advantageous Effects

(33) According to the ultrasonic imaging apparatus and its control method according to one aspect, since it is possible to calculate a more accurate shear wave elastic value by removing the inaccurate measurement area such as blood vessels using the shear wave observation signal in the section where the object is not affected by the push pulse, a highly reliable shear wave elastic image can be provided to the user.

Description

DESCRIPTION OF DRAWINGS

(1) These and/or other aspects of the invention will become apparent and more readily appreciated from the following description of the embodiments, taken in conjunction with the accompanying drawings of which:

(2) FIG. 1 is an external view of an ultrasonic imaging apparatus according to an embodiment.

(3) FIG. 2 is a block diagram of an ultrasonic imaging apparatus according to an embodiment.

(4) FIG. 3 is a view for explaining a process in which a conventional ultrasonic imaging apparatus obtains a shear wave elastic image.

(5) FIG. 4A is a view illustrating a push pulse and an observation signal irradiated by a conventional ultrasonic imaging apparatus.

(6) FIG. 4B is a view illustrating the shear wave velocity for each region obtained according to an observation signal irradiated by a conventional ultrasonic imaging apparatus over time.

(7) FIG. 5A is a view for explaining a case in which the shear wave velocity for each region exists in a second half.

(8) FIG. 5B is a view illustrating an ultrasound image obtained when the shear wave velocity for each region exists in a second half.

(9) FIG. 6A is a view for explaining an ultrasonic irradiation process of an ultrasonic imaging apparatus according to an embodiment.

(10) FIG. 6B is a view for explaining an ultrasonic irradiation process of an ultrasonic imaging apparatus according to another embodiment.

(11) FIG. 7A is a view illustrating a push pulse, a first observation signal, and a second observation signal radiated by the ultrasonic imaging apparatus according to an embodiment.

(12) FIG. 7B is a view illustrating a push pulse, a first observation signal, and a second observation signal that are irradiated by the ultrasonic imaging apparatus according to another embodiment.

(13) FIG. 8 is a view illustrating the shear wave velocity for each region obtained by irradiating the second observation signal over time.

(14) FIG. 9 is a view illustrating a display on which a shear wave elastic image is displayed,

according to an exemplary embodiment.

(15) FIG. **10** is a view illustrating a display unit on which a shear wave elastic image is displayed according to another exemplary embodiment.

(16) FIG. **11** is a control flowchart of an ultrasonic imaging apparatus according to an embodiment.

MODE OF THE DISCLOSURE

(17) Like numerals refer to like elements throughout the specification. Not all elements of embodiments of the present disclosure will be described, and description of what are commonly known in the art or what overlap each other in the embodiments will be omitted.

(18) The terms as used throughout the specification, such as “~ part”, “~ module”, “~ member”, “~ block”, etc., may be implemented in software and/or hardware, and a plurality of “~ parts”, “~ modules”, “~ members”, or “~ blocks” may be implemented in a single element, or a single “~ part”, “~ module”, “~ member”, or “~ block” may include a plurality of elements.

(19) It will be further understood that the term “connect” or its derivatives refer both to direct and indirect connection, and the indirect connection includes a connection over a wireless communication network.

(20) It will be further understood that the terms “comprises” and/or “comprising,” when used in this specification, specify the presence of stated features, integers, steps, operations, elements, and/or components, but do not preclude the presence or addition of one or more other features, integers, steps, operations, elements, components, and/or groups thereof, unless the context clearly indicates otherwise.

(21) As used herein, the singular forms “a,” “an” and “the” are intended to include the plural forms as well, unless the context clearly indicates otherwise.

(22) Reference numerals used for method steps are just used for convenience of explanation, but not to limit an order of the steps. Thus, unless the context clearly dictates otherwise, the written order may be practiced otherwise.

(23) A term ‘object’ may include a person or animal, or a part of a person or animal. For example, the object may include organs, such as liver, heart, uterus, brain, breast, abdomen, or blood vessels as well as a mass. Also, in the specification of the present disclosure, a term ‘user’ may be a doctor, a nurse, a clinical pathologist, a medical imaging expert, etc., and may be a technician who develops and repairs a medical device, but is not limited thereto.

(24) Terms ‘Ultrasound image’ and ‘object image’ mean an image of an object obtained using ultrasound.

(25) A term ‘shear wave elastic image’ refers to an image in which the shear wave elastic value for each area of an object obtained using ultrasound is displayed.

(26) A term ‘inefficiency region’ refers to a region in which a self-signal exists without a push pulse, such as a blood vessel region in which blood flow exists and a region in which tissue movement occurs.

(27) Hereinafter, an embodiment of an ultrasonic imaging apparatus and a method for controlling the same according to an aspect will be described in detail with reference to accompanying drawings.

(28) FIG. **1** is an external view of an ultrasonic imaging apparatus according to an embodiment.

(29) Referring to FIG. **1**, the ultrasonic imaging apparatus **1** includes an ultrasonic probe **100** and a main body **200**. The ultrasonic probe **100** may transmit an ultrasonic signal to an object to be diagnosed and receive an ultrasonic echo signal reflected from the object. The ultrasonic probe **100** receives the ultrasonic echo signal reflected from the object and converts it into an electrical signal (hereinafter, referred to as an ultrasonic signal).

(30) The ultrasonic probe **100** is connected to the main body **200** of the ultrasonic imaging apparatus **1** through a cable **120**, and may receive various signals necessary for controlling the ultrasonic probe **100** from the main body **200**. Also, the ultrasonic probe **100** may transmit an analog signal or a digital signal corresponding to the ultrasonic echo signal to the main body **200**.

(31) Meanwhile, the ultrasonic probe **100** may be implemented as a wireless probe, and may transmit and receive signals through a network formed between the probe **100** and the main body **200**.

(32) The main body **200** may include a PSA board (Probe Select Assembly, **250**), a control panel **260**, and displays **270**; **270-1** and **270-2**. The PSA board **250** includes a port connected to the ultrasonic probe **100**. The PSA board **250** may activate the ultrasonic probe **100** according to a command input of a user through the control panel **260** and the control of the controller **300**. One end of the cable **120** includes a connector **130** connectable to a port of the PSA board **250**.

(33) The control panel **260** is a device that receives a command for operating the ultrasonic imaging apparatus **1** from the user. The control panel **260** may receive setting information regarding the probe **100** and may receive various control commands related to an operation of the main body **200**.

(34) The control panel **260** may include a keyboard. The keyboard may include a button, a switch, a knob, a touch pad, a trackball, and the like. Also, the control panel **260** may include a first display **270-1**. The first display **270-1** may display a graphic user interface (GUI) for controlling an operation of the ultrasonic imaging apparatus **1**. The first display **270-1** may display related information such as a menu for optimizing an ultrasound image or an auxiliary image.

(35) The first display **270-1** may include a touch panel and may receive a touch input of the user for the graphic user interface. The first display **270-1** may display a graphic user interface having the same shape as a button included in the keyboard. The user may input a command for controlling the ultrasonic imaging apparatus **1** through a touch input to the first display **270-1**.

(36) The second display **270-2** may display an ultrasound image. The ultrasound image may be a 2D ultrasound image or a 3D stereoscopic ultrasound image, and various ultrasound images may be displayed according to an operation mode of the ultrasonic imaging apparatus **1**. In addition, the second display **270-2** may display a menu necessary for ultrasound diagnosis, guide information, information about an operating state of the probe **100**, and the like.

(37) The second display **270-2** may display a shear wave elastic image by overlapping or matching the reference ultrasound image.

(38) The second display **270-2** may also include a touch panel and receive a user's touch input for the graphic user interface. The user may input a command for controlling the ultrasonic imaging apparatus **1** through a touch input to the second display **270-2**.

(39) The display **270** may be implemented with various display devices, such as a liquid crystal display (LCD), a light emitting diode (LED), a plasma display panel (PDP), and an organic light emitting diode (OLED).

(40) FIG. **2** is a block diagram of an ultrasonic imaging apparatus according to an embodiment.

(41) Referring to FIG. **2**, the ultrasonic probe **100** may be a linear array probe, a curved array probe, a phased array probe, or a volume probe. The ultrasonic probe **100** is not limited thereto, and may include various probes such as an endocavity probe, a convex probe, a matrix probe, and/or a 3D probe.

(42) The main body **200** of the ultrasonic imaging apparatus **1** may further include beamformers **281** and **282**, an image processor **290**, and a controller **300**.

(43) The beamformer may be divided into a transmission beamformer **281** and a reception beamformer **282**. In obtaining an image using an ultrasound signal, a beamforming technology is applied to increase the resolution of the image. The transmission beamformer **281** may apply a transmit pulse to the ultrasonic probe **100**. The transmission beamformer **281** may apply an appropriate time delay so that ultrasound signals to be transmitted by a plurality of transducer elements are simultaneously focused on one focal point, and may generate a transmission beam. The transducer array **110** may irradiate a transmission beam to a target portion within the object.

(44) For convenience of description, such a transmission beam may be referred to as an 'observation signal' so as to be contrasted with a push pulse to be described later.

(45) Also, the transmission beamformer **281** may generate a push pulse transmitted along a push line. The push pulse is irradiated to a region of interest of the object to induce tissue displacement and may induce a shear wave. The tissue displacement is used to measure the shear wave elastic value, which will be described later. The push pulse may be a focused beam having a relatively high focusing velocity.

(46) The observation signal irradiated to the object may be reflected from the object and incident on the transducer array **110** of the ultrasonic probe **100** again. The reflected ultrasound signal may be defined as an ultrasound echo signal.

(47) The reception beamformer **282** analog/digitizes the ultrasound echo signal received from the ultrasonic probe **100** and performs reception beamforming. The reception beamformer **282** may apply a time delay to the ultrasound echo signals that are reflected from the focal point and return to the transducer element, and may be summed at the same time.

(48) Meanwhile, the beamformers **281** and **282** may be provided in the ultrasonic probe **100**. For example, when the ultrasonic probe **100** is a wireless probe, the ultrasonic probe **100** may include beamformers **281** and **282**.

(49) The image processor **290** filters a noise component from the received beam to improve the quality of the ultrasound image, performs envelope detection processing for detecting the strength of the received signal, and generates digital ultrasound image data.

(50) The image processor **290** may perform scan conversion for converting a scan line of the digital ultrasound image data so that the digital ultrasound image data may be displayed on the display **270**. Also, the image processor **290** may perform image processing on the ultrasound echo signal to generate an A-mode image, a B-mode image, a D-mode image, an E-mode image, an M-mode image, a Doppler image, and/or a 3D ultrasound image. The image processor **290** performs RGB processing on ultrasound image data and transmits the ultrasound image data to the display **270** so that the ultrasound image may be displayed through the display **270**.

(51) Also, the image processor **290** may perform image processing to display various types of additional information on the ultrasound image.

(52) Meanwhile, although the image processor **290** is illustrated separately from the controller **300** in FIG. 2, the controller **300** may include the image processor **290**.

(53) The display **270** may display an ultrasound image and various information processed by the ultrasonic imaging apparatus **1**. The display **270** may display various graphic user interfaces for adjusting the generated ultrasound image.

(54) The controller **300** may control the operation of the ultrasonic imaging apparatus **1** and the signal flow between internal components of the ultrasonic imaging apparatus **1**. The controller **300** may include a processor **310** and a memory **320**. The controller **300** may be implemented as a processing board in which the processor **310** and the memory **320** are installed on a circuit board. The processor **310** and the memory **320** may be connected through a bus. One or a plurality of processors **310** may be provided.

(55) The controller **300** may be implemented with a plurality of logic gates, or may be implemented as a combination of a general-purpose micro-processor and a memory **320** in which a program executable in the microprocessor is stored.

(56) The memory **320** refers to a storage medium for storing algorithms and data necessary for the operation of each component of the ultrasonic imaging apparatus **1**. The memory **320** may include a high-speed random access memory, a magnetic disk, an SRAM, a DRAM, a ROM, and the like. Also, the memory **320** may be detachable from the ultrasonic imaging apparatus **1**. The memory **320** may include, but is not limited to, a compact flash card (CF card), a secure digital card (SD card), a smart media card (SM card), a multimedia card (MMC), or a memory stick.

(57) The controller **300** may be electrically connected to each of the PSA board **250**, the control panel **260**, the display **270**, and the beamformers **281** and **282**, and generate a control signal for controlling each component of the probe **100** and the main body **200**.

(58) Hereinafter, a process in which a conventional ultrasonic imaging apparatus obtains a shear wave elastic image will be described with reference to FIGS. 3 to 5.

(59) FIG. 3 is a view for explaining a process in which a conventional ultrasonic imaging apparatus obtains a shear wave elastic image. FIG. 4A is a view illustrating a push pulse and an observation signal irradiated by a conventional ultrasonic imaging apparatus. FIG. 4B is a view illustrating the shear wave velocity for each region obtained according to an observation signal irradiated by a conventional ultrasonic imaging apparatus over time. FIG. 5A is a view for explaining a case in which the shear wave velocity for each region exists in a second half. FIG. 5B is a view illustrating an ultrasound image obtained when the shear wave velocity for each region exists in a second half.

(60) The content to be described later with reference to FIGS. 3 to 5 relates to a conventional ultrasonic imaging apparatus, but since the ultrasonic imaging apparatus according to an embodiment includes a configuration of the conventional ultrasonic imaging apparatus, for convenience of description, it may be described as the ultrasonic imaging apparatus 1 according to an embodiment.

(61) Referring to FIG. 3, the ultrasonic probe 100 according to an embodiment may generate a shear wave 532 to the object by irradiating the push pulse 521 to the object along the push line under the control of the ultrasonic imaging apparatus 1. The push pulse 521 may be irradiated to the focal point 520 in the region of interest of the object to induce displacement of the object and generate the shear wave 532. The push pulse 521 is a focused beam having a relatively high focusing velocity and may have a narrow beam profile.

(62) When the push pulse 521 is irradiated to the focal point 520 in the region of interest of the object, the shear wave 532 may be generated. That is, when a force is applied to the object of the focal point 520 in the depth direction by the push pulse, the object of the focal point 520 moves in the depth direction. The distance the object moves in the depth direction may be defined as displacement. Since the tissue of the object has a certain degree of elasticity and the adjacent tissues are organically connected, the movement of the object located at the focal point 520 also affects the adjacent tissues.

(63) The displacement of the adjacent tissue is induced by the movement of the object located at the focal point 520. The shear wave 532 may propagate in a direction perpendicular to the depth direction. The shear wave 532 propagates in both directions from the focal point 520. The shear wave 532 changes its speed according to the vibratory properties of a medium. Therefore, the shear wave elastic value of the object may be obtained by estimating the velocity of the shear wave 532.

(64) In addition, since the change in displacement or displacement, that is, the elastic value that occurs when a constant force is applied according to the characteristics of the medium, is different, so the shear wave elastic value of the object may be estimated based on the change or displacement of the displacement for each area of the object. In other words, the shear wave elastic value for each area of the object may be estimated based on the displacement data for each area of the object.

(65) In order to estimate the velocity of the shear wave, that is, in order to obtain the shear wave elastic value of the object, the ultrasonic probe 100 according to an embodiment may irradiate the observation signal 541 to the object under the control of the ultrasonic imaging apparatus 1 after irradiating the push pulse 521. That is, the ultrasonic probe 100 may irradiate an observation signal 541 having a wide beam profile to the region of interest in which the shear wave 532 proceeds due to tissue displacement, and receive an ultrasonic echo signal in which the observation signal 541 is reflected from the region of interest. The ultrasonic imaging apparatus 1 may detect tissue displacement based on the echo signal of the observation signal 541.

(66) For example, the controller 300 may obtain a shear wave elastic image at a high frame rate, and may detect tissue displacement by comparing successive shear wave elastic image frames.

(67) However, when measuring the elastic value of the tissue using the shear wave 532, the accurate shear wave elastic value of the object may not be obtained due to blood vessels located in the region of interest of the object.

(68) Specifically, when displacement data exists by itself due to blood flow without irradiation of the push pulse **521**, such as blood vessels, or when there is a displacement change due to movement of the tissue without irradiation of the push pulse **521**, it may be difficult to obtain an accurate shear wave elastic value of the tissue.

(69) As described above, the elastic value for each area of the object may be determined based on the displacement data for each area of the object based on the displacement data for each area of the object from which the displacement, change of displacement, speed, etc. of the object by the shear wave **532** generated by the push pulse **521** may be estimated.

(70) That is, determining the elastic value for each area of the object based on the displacement data may mean determining the elastic value for each area of the object based on the displacement, and determining the elastic value for each area of the object based on the displacement change per time, that is, the speed of the shear wave for each area of the object.

(71) Hereinafter, with reference to FIGS. **4A** to **4B**, a process in which the ultrasonic imaging apparatus **1** according to the conventional art obtains an accurate shear wave elastic value of an object excluding a region in which its own signal such as a blood vessel exists will be described.

(72) Referring to FIG. **4A**, the order in which the conventional ultrasonic imaging apparatus irradiates the push pulse **521** and the observation signal **541** according to time may be confirmed. The ultrasonic imaging apparatus irradiates the push pulse **521** to the region of interest of the object, and then may irradiate the observation signal **541** to the object in order to observe the occurrence of displacement of the adjacent tissue.

(73) That is, the push pulse **521** may be irradiated to the focal point **520** in the region of interest of the object, and the observation signal **541** may be irradiated to the region of interest of the object in order to observe the displacement data of the tissue resulting therefrom. In this case, for convenience of explanation, the observation signal **541** for observing the displacement data of the tissue due to the push pulse **521** may be defined as a 'first observation signal **541**'.

(74) In this case, the first observation signal **541** may include a plurality of ultrasound signals having a predetermined time interval. In this case, the predetermined time interval may be determined enough to observe the tissue displacement data by the push pulse **521**.

(75) That is, the first observation signal **541** is always irradiated for a predetermined time after the push pulse **521** is irradiated, so that the tissue displacement data by the push pulse **521** may be observed. In this case, the predetermined time may be determined as the time until the tissue displacement by the push pulse **521** disappears. As shown in FIG. **4A**, in order to obtain a more accurate shear wave elastic value of an object, the ultrasonic imaging apparatus **1** may irradiate the push pulse and the first observation signal several times.

(76) Although not shown in the figure, the controller **300** may determine the shear wave elastic value for each area of the object based on the displacement data for each area of the object obtained from the irradiated first observation signal **541**, and obtain a shear wave elastic image based on the shear wave elastic value for each region of the object.

(77) Referring to FIG. **4B**, the velocity value of the particle in the object region corresponding to the tissue displacement data obtained from the irradiated first observation signal **541** may be confirmed.

(78) Referring to part (a), which shows the particle velocity in the object region over time, it may be seen that the speed of particles in the object region is generated by the push pulse **521** in the first half **410** not long after the push pulse **521** is irradiated, but it may be confirmed that the particle velocity in the object region disappears in the second half **420** after a certain time has elapsed after the push pulse **521** is irradiated. That is, it may be confirmed that the shear wave **532** generated in the tissue disappears.

(79) On the other hand, referring to part (b) showing the particle velocity in the blood vessel region over time, it may be seen that the particle velocity in the second half **420** does not disappear after a certain time has passed after the push pulse **521** is irradiated. That is, it may be confirmed that the

particle velocity in the blood vessel region does not disappear.

(80) Since the displacement of the blood vessel region occurs due to factors such as the movement of blood flow regardless of whether the push pulse **521** is irradiated or not, displacement data by the push pulse **521** may not be obtained, and thus accurate shear wave elastic value may not be obtained.

(81) According to the fact that only the signal due to the change in displacement of the blood vessel region exists in the second half **420**, the ultrasonic imaging apparatus according to the conventional art determines the region where displacement exists in the second half **420** as the blood vessel region, and displays the image so that the shear wave elastic value of the blood vessel region is not displayed.

(82) Hereinafter, with reference to FIGS. **5A** to **5B**, a problem according to a method of determining a region in which displacement exists in the second half **420** as a blood vessel region will be described.

(83) Referring to FIG. **5A**, it may be confirmed that there is a change in displacement of the liver region in the second half **420** after a predetermined time has passed after the push pulse is irradiated.

(84) When the first observation signal **541** is irradiated for a longer period of time, the displacement change in the liver region will disappear in the second half **420**, but as described above, in order to obtain a shear wave elastic image at a high frame rate, it may be advantageous to shorten the irradiation time of the first observation signal **541** as much as possible.

(85) The conventional ultrasonic imaging apparatus determines that the particle velocity is present in the second half **420** despite the liver region as a blood vessel region, and displays an image so that the shear wave elastic value of the region in which the particle velocity exists despite the liver region was not displayed.

(86) That is, since particle velocities exist in the second half **420** of the blood vessel region and also in the second half **420** of the liver region, there is a problem that such a difference in particle velocities may not be distinguished.

(87) Referring to FIG. **5B**, a shear wave elastic image obtained by a conventional ultrasonic imaging apparatus may be confirmed. Looking at the shear wave elastic image obtained according to the conventional ultrasonic imaging apparatus, there is a portion **421** in which the elastic value is not displayed due to the presence of particle velocity in the latter part **420** as a blood vessel region, whereas although it is the liver region, there may be a portion **422** in which an elastic value is not displayed because particle velocity is present in the second half **420**.

(88) Due to these problems, the conventional ultrasonic imaging apparatus may not provide an accurate shear wave elastic image to the user and may reduce reliability of the user of the ultrasonic imaging apparatus.

(89) Hereinafter, an ultrasonic imaging apparatus **1** according to an embodiment that may solve the above-described problems will be described with reference to FIGS. **6** to **8**.

(90) FIG. **6A** is a view for explaining an ultrasonic irradiation process of an ultrasonic imaging apparatus according to an embodiment. FIG. **6B** is a view for explaining an ultrasonic irradiation process of an ultrasonic imaging apparatus according to another embodiment.

(91) FIG. **7A** is a view illustrating a push pulse, a first observation signal, and a second observation signal radiated by the ultrasonic imaging apparatus according to an embodiment. FIG. **7B** is a view illustrating a push pulse, a first observation signal, and a second observation signal that are irradiated by the ultrasonic imaging apparatus according to another embodiment.

(92) FIG. **8** is a view illustrating the shear wave velocity for each region obtained by irradiating the second observation signal over time.

(93) Referring to FIG. **6A**, the ultrasonic probe **100** according to an exemplary embodiment may irradiate the second observation signal **542** before the push pulse **521** is irradiated. That is, by using the second observation signal **542**, displacement data for each region of the object in a state in

which the shear wave **532** does not occur in the tissue may be obtained.

(94) In this case, the controller **300** may determine the inefficiency region of the object based on the displacement data for each region of the object obtained from the irradiated second observation signal **542**.

(95) Referring to FIG. **8**, it may be confirmed that the particle velocity in the liver region obtained from the irradiated second observation signal **542** is very small, and it may be confirmed that the particle velocity in the blood vessel region obtained from the irradiated second observation signal **542** exists.

(96) The blood vessel region is a region included in the inefficiency region, and the inefficiency region may include at least one of a blood vessel region in which blood flow exists and a region in which tissue movement occurs.

(97) That is, the inefficiency region may mean a region in which displacement changes even without the push pulse **52**.

(98) Accordingly, the controller **300**, based on the displacement data for each area of the object obtained from the irradiated second observation signal **542**, may determine a region in which a displacement value or a shear wave velocity value is greater than or equal to a predetermined value among regions of the object, and determine the determined region as an inefficiency region.

(99) Referring back to FIG. **6A**, as described with reference to FIG. **3**, the ultrasonic probe **100** according to an embodiment may irradiate the push pulse **521** after the second observation signal **542** is irradiated, and obtain displacement data for each area of the object by irradiating the first observation signal **541** after the push pulse **521** is irradiated.

(100) The controller **300** may generate a shear wave elastic image based on the determined inefficiency region and the displacement data for each region of the object obtained from the irradiated first observation signal **541**, which will be described later with reference to FIGS. **9** to **10**.

(101) Referring to FIG. **6B**, the ultrasonic probe **100** according to an embodiment may irradiate a first observation signal **541** after irradiating a push pulse **521** to obtain tissue displacement by the shear wave **532**, and irradiate the second observation signal **542** after the first observation signal **541** is irradiated. After the push pulse **521** is irradiated and the first observation signal **541** is irradiated, the shear wave **532** caused by the push pulse **521** may disappear. Accordingly, by using the second observation signal **542**, displacement data for each area of the object in a state in which the shear wave **532** is not generated by the push pulse **521** may be obtained.

(102) In this case, the controller **300** may determine the inefficiency region based on the displacement data for each region of the object obtained from the irradiated second observation signal **542**.

(103) Specifically, the controller **300**, based on the displacement data for each area of the object obtained from the irradiated second observation signal **542**, may determine a region in which the displacement value or shear wave velocity value is greater than or equal to a predetermined value among regions of the object, and determine the determined region as an inefficiency region.

(104) Referring to FIG. **8**, it may be confirmed that the particle velocity in the liver region obtained from the irradiated second observation signal **542** is very small, and it may be confirmed that the particle velocity in the blood vessel region obtained from the irradiated second observation signal exists.

(105) Referring to FIG. **7A**, as shown in FIG. **6A**, the ultrasonic probe **100** according to an embodiment may irradiate the second observation signal **542** before the push pulse **521** is irradiated. The second observation signal **542** may be irradiated to obtain displacement data for each area of the object without the influence of the push pulse **521**, and then the push pulse **521** may be irradiated to generate the shear wave **532** in the object, and the first observation signal **541** may be irradiated to obtain displacement data for each area of the object generated by the generated shear wave **532**.

(106) Referring to FIG. 7B, as shown in FIG. 6B, the ultrasonic probe **100** according to an embodiment may generate the shear wave **532** on the object by irradiating a push pulse **521**, and irradiate the first observation signal **541** to obtain displacement data for each region of the object generated by the generated shear wave **532**. After irradiating the first observation signal **541**, the second observation signal **542** may be irradiated to obtain displacement data for each area of the object that is not affected by the push pulse **521**.

(107) FIGS. 7A to 7B, the first observation signal **541** and the second observation signal **542** may include a plurality of ultrasound signals having predetermined time intervals.

(108) In general, since the frequency range of blood vessels is about 500 Hz or less, in order to observe a signal of 500 Hz or less, the sampling frequency must be 2 kHz or more. Therefore, the signal for observing the displacement data of the blood vessel can be irradiated with a time interval of up to 500 microsecond. That is, it may be less than or equal to 500 microsecond.

(109) In addition, in order to obtain the displacement data of the blood vessel, $\frac{1}{2}$ cycle of the blood vessel needs to be observed, so the signal for observing the displacement data of the blood vessel may be irradiated for a time of at least 10 millisecond. That is, it may be irradiated for a time of 10 millisecond or more.

(110) On the other hand, in general, in order to observe the displacement data for each area of the object by the push pulse **521**, the sampling frequency must be 2 kHz or less, and the signal for observing the displacement data for each area of the object may be irradiated at a time interval shorter than a maximum of 500 microsecond. For example, it may be irradiated at an interval of at least 200 microsecond or less.

(111) In addition, in order to obtain displacement data for each area of the object by the push pulse **521**, since it is necessary to observe the displacement data for each area of the object until the displacement or displacement data by the push pulse **521** disappears, or until the shear wave generated by the push pulse **521** is transmitted to the tip of the tissue and the displacement data of the tip of the tissue may be obtained, a signal for obtaining displacement data for each area of the object by the push pulse **521** may be irradiated for a time longer than 10 millisecond. For example, it may be irradiated for at least 50 millisecond or longer.

(112) Accordingly, the time interval **545** of each of the plurality of ultrasound signals included in the second observation signal **542** may be longer than or equal to the time interval **547** of each of the plurality of ultrasound signals included in the first observation signal **541**, and the time **544** for which the second observation signal **542** is irradiated may be shorter than or equal to the time **546** for which the first observation signal **541** is irradiated.

(113) Also, the second observation signal **542** may be irradiated for a time **544** of 10 millisecond or more, and a time interval **545** of each of the plurality of ultrasound signals included in the second observation signal **542** may be 500 microsecond or less.

(114) Hereinafter, a process of generating a shear wave elastic image by the controller of the ultrasonic imaging apparatus **1** according to an embodiment and displaying it on the display **270** will be described with reference to FIGS. **9** to **10**.

(115) FIG. **9** is a view illustrating a display on which a shear wave elastic image is displayed, according to an exemplary embodiment. FIG. **10** is a view illustrating a display unit on which a shear wave elastic image is displayed according to another exemplary embodiment.

(116) As described above, the controller **300** according to an embodiment may determine the inefficiency region based on the displacement data for each region of the object obtained from the second observation signal **542** irradiated by the ultrasonic probe **10**.

(117) In addition, the controller **300** according to an embodiment may generate a shear wave elastic image based on the determined inefficiency region and the displacement data for each region of the object obtained from the first observation signal **541** irradiated by the ultrasonic probe **100**, and display the generated shear wave elastic image on the display. Specifically, the controller **300** may determine the shear wave elastic value for each area of the object based on the displacement data

for each area of the object obtained from the irradiated first observation signal **541**, and generate the shear wave elastic image based on the determined shear wave elastic value for each region.

(118) Referring to FIG. **9**, the controller **300** according to an embodiment may determine the shear wave elastic value for each region of the object based on the displacement data for each region of the object obtained from the irradiated first observation signal **541**, generate the shear wave elastic image based on the determined shear wave elastic value for each region, and generate a shear wave elastic image so that the determined shear wave elastic value of the inefficiency region **271** is not displayed.

(119) Also, the controller **300** may control the display **270** to display a shear wave elastic image in which the shear wave elastic value of the determined inefficiency region **271** is not displayed.

(120) When the ultrasound image is generated in B-mode, the inefficiency region **271** such as blood vessels is displayed in black on the display **270**, so when the shear wave elastic value of the inefficiency region **271** marked in black is not displayed, the user can check only the area marked in black, so that it may be determined as the inefficiency region **271**.

(121) Also, the controller **300** according to an embodiment may control the display **270** to display the determined inefficiency region **271**. For example, the controller **300** may control the display **270** to display a border **273** indicating the inefficiency region **271** with respect to the inefficiency region **271**, but is not limited thereto.

(122) Referring to FIG. **10**, the controller **300** according to an embodiment determines the shear wave elastic value for each region of the object based on the displacement data for each region of the object obtained from the irradiated first observation signal **541**, generate the shear wave elastic image based on the determined shear wave elastic value for each region, and generate a shear wave elastic image by changing a reliability measurement index (RMI) of the determined shear wave elastic value of the inefficiency region **272**.

(123) In addition, the controller **300** may control the display **270** to display a shear wave elastic image to which a changed reliability measurement index (RMI) is applied as shown in FIG. **10**, although not shown in the drawing, the display **270** may be controlled to display the changed reliability measurement index (RMI).

(124) For example, the reliability measurement index of the shear wave elastic value of the inefficiency region **272** may be changed to 0, which means that the shear wave elastic value of the inefficiency region **272** is not displayed as described above in FIG. **9**.

(125) Also, for example, as shown in FIG. **10**, the controller **300** may display the shear wave elastic value by changing the reliability measurement index of the shear wave elastic value of the inefficiency region **272** to a low value to obtain a smooth shear wave elastic image.

(126) Hereinafter, a method of controlling the ultrasonic imaging apparatus **1** according to an embodiment will be described with reference to FIG. **11**. FIG. **11** is a control flowchart of an ultrasonic imaging apparatus according to an embodiment.

(127) The ultrasonic probe **100** according to an embodiment may irradiate a push pulse for generating a shear wave to the object to the object (**1000**). As described above, the push pulse may be irradiated to the focal point **520** in the region of interest of the object to induce displacement of the object and to induce a shear wave.

(128) When the push pulse is irradiated to the focal point within the region of interest of the object, a shear wave may be generated. That is, when a force is applied to the object at the focal point in the depth direction by the push pulse, the object at the focal point moves in the depth direction. The distance the object moves in the depth direction may be defined as displacement. Since the tissue of the object has a certain elasticity and the adjacent tissues are organically connected, the movement of the object located at the focal point also affects the adjacent tissues.

(129) In addition, the ultrasonic probe **100** according to an embodiment may irradiate the first observation signal after the push pulse is irradiated in order to obtain displacement data for each region of the object by the generated shear wave (**1100**).

(130) Also, the ultrasonic probe **100** according to an embodiment may radiate the second observation signal after the first observation signal is irradiated (**1200**), and irradiate the second observation signal before (between 1000 and the start) the push pulse is irradiated to the object regardless of the order of the flowchart. That is, the step **1200** of irradiating the second observation signal may be performed after the first observation signal is irradiated (after **1100**) or before the push pulse is irradiated to the object (between 1000 and the start).

(131) Thereafter, the controller **300** according to an embodiment may determine the inefficiency region of the object based on the displacement data for each region of the object obtained from the irradiated second observation signal (**1300**).

(132) At this time, determining of the inefficiency region based on the displacement data for each region of the object obtained from the irradiated second observation signal, may include determining a region of the object region in which a displacement value or a shear wave velocity value is greater than or equal to a predetermined value, based on the displacement data for each region of the object obtained from the second observation signal irradiated, and determining the determined region as an inefficiency region.

(133) As described above, the inefficiency region may include at least one of a blood vessel region in which blood flow exists and a region in which tissue movement occurs.

(134) Thereafter, the controller **300** according to an embodiment may generate a shear wave elastic image based on the displacement data for each region of the object obtained from the irradiated first observation signal and the determined inefficiency region (**1400**).

(135) At this time, generating of the shear wave elastic image may include determining the shear wave elastic value for each area of the object based on the displacement data for each area of the object obtained from the irradiated first observation signal, generating the shear wave elastic image based on the determined shear wave elastic value for each region, and generating a shear wave elastic image so that the shear wave elastic value of the determined inefficiency region is not displayed.

(136) In addition, generating of the shear wave elastic image may include determining the shear wave elastic value for each area of the object based on the displacement data for each area of the object obtained from the irradiated first observation signal, generating a shear wave elastic image based on the determined shear wave elastic value for each region, and generating a shear wave elastic image by changing a reliability measurement index (RMI) of the shear wave elastic value of the determined inefficiency region.

(137) Thereafter, the display **270** according to an embodiment may display the generated shear wave elastic image (**1500**). At this time, a shear wave elastic image in which the shear wave elastic value of the determined inefficiency region is not displayed may be displayed.

(138) In addition, a shear wave elastic image to which a changed reliability measurement index (RMI) is applied may be displayed, and a changed reliability measurement index (RMI) may be displayed.

(139) In addition, the determined inefficiency region may be displayed by displaying a border indicating that the inefficiency region is the determined inefficiency region.

(140) Meanwhile, the disclosed embodiments may be embodied in the form of a recording medium storing instructions executable by a computer. The instructions may be stored in the form of program code and, when executed by a processor, may generate a program module to perform the operations of the disclosed embodiments. The recording medium may be embodied as a computer-readable recording medium.

(141) The computer-readable recording medium includes all kinds of recording media in which instructions which may be decoded by a computer are stored, for example, a Read Only Memory (ROM), a Random Access Memory (RAM), a magnetic tape, a magnetic disk, a flash memory, an optical data storage device, and the like.

(142) Although exemplary embodiments of the present invention have been described for

illustrative purposes, those skilled in the art will appreciate that various modifications, additions and substitutions are possible, without departing from the scope and spirit of the present invention. Therefore, exemplary embodiments of the present invention have not been described for limiting purposes.

Claims

1. An ultrasonic imaging apparatus, comprising: an ultrasonic probe configured to irradiate a push pulse to generate a shear wave to an object, irradiate a first observation signal for observing a displacement of the object caused by the push pulse to the object after the push pulse is irradiated, and irradiate a second observation signal for observing a displacement of a blood vessel that occurs regardless of the push pulse to the object before the push pulse is irradiated or after the first observation signal is irradiated; a controller configured to determine an inefficiency region of the object containing blood flow based on displacement data obtained from the irradiated second observation signal that indicates the displacement of the blood vessel that occurs regardless of the push pulse, and generate a shear wave elastic image based on the determined inefficiency region and displacement data for each region of the object obtained from the irradiated first observation signal; and a display configured to display the generated shear wave elastic image, wherein the second observation signal is irradiated for a time more than 10 milliseconds for observing more than half cycle of the blood vessel and shorter than a time for which the first observation signal is irradiated, wherein the first observation signal and the second observation signal include a plurality of ultrasound signals having a predetermined time interval, wherein a time interval of each of the plurality of ultrasound signals included in the second observation signal is less than or equal to 500 microseconds for observing the displacement of the blood vessel, a time interval of each of the plurality of ultrasound signals included in the first observation signal is less than or equal to 200 microseconds for observing the displacement of the object caused by the push pulse, and the time interval of each of the plurality of ultrasound signals included in the second observation signal is longer than the time interval of each of the plurality of ultrasound signals included in the first observation signal, wherein the controller is configured to determine a region in which a displacement value or a shear wave velocity value is greater than or equal to a predetermined value among regions of the object based on the displacement data for each region of the object obtained from the irradiated second observation signal, and determine the determined region as the inefficiency region.

2. The ultrasonic imaging apparatus according to claim 1, wherein the inefficiency region further includes a region in which tissue movement occurs regardless of the push pulse.

3. The ultrasonic imaging apparatus according to claim 1, wherein the controller is configured to: determine a shear wave elastic value for each region of the object based on the displacement data for each region of the object obtained from the irradiated first observation signal, and generate the shear wave elastic image of the object based on the determined shear wave elastic value for each region, wherein the shear wave elastic image is generated to prevent the shear wave elastic value of the determined inefficiency region from being displayed.

4. The ultrasonic imaging apparatus according to claim 3, wherein the controller is configured to control the display to display the shear wave elastic image in which the shear wave elastic value of the determined inefficiency region is prevented from being displayed.

5. The ultrasonic imaging apparatus according to claim 1, wherein the controller is configured to: determine a shear wave elastic value for each region of the object based on the displacement data for each region of the object obtained from the irradiated first observation signal, and generate the shear wave elastic image of the object based on the determined shear wave elastic value for each region, wherein the shear wave elastic image is generated by changing a reliability measurement index (RMI) of the shear wave elastic value of the determined inefficiency region.

6. The ultrasonic imaging apparatus according to claim 5, wherein the controller is configured to control the display to display the shear wave elastic image to which the changed RMI is applied.

7. The ultrasonic imaging apparatus according to claim 5, wherein the controller is configured to control the display to display the RMI.

8. The ultrasonic imaging apparatus according to claim 1, wherein the controller is configured to control the display to display the determined inefficiency region.

9. A method for controlling an ultrasonic imaging apparatus, the method comprising: irradiating a second observation signal to an object for observing a displacement of a blood vessel that occurs regardless of a push pulse; irradiating the push pulse for generating a shear wave to the object after the second observation signal is irradiated to the object; irradiating a first observation signal for observing a displacement of the object caused by the push pulse to the object after the push pulse is irradiated; determining an inefficiency region of the object containing blood flow based on displacement data obtained from the irradiated second observation signal that indicates the displacement of the blood vessel that occurs regardless of the push pulse; generating a shear wave elastic image based on displacement data for each region of the object obtained from the irradiated first observation signal and the determined inefficiency region; and displaying the generated shear wave elastic image, wherein the second observation signal is irradiated for a time more than 10 milliseconds for observing more than half cycle of the blood vessel and shorter than a time for which the first observation signal is irradiated, wherein the first observation signal and the second observation signal include a plurality of ultrasound signals having a predetermined time interval, wherein a time interval of each of the plurality of ultrasound signals included in the second observation signal is less than or equal to 500 microseconds for observing the displacement of the blood vessel, a time interval of each of the plurality of ultrasound signals included in the first observation signal is less than or equal to 200 microseconds for observing the displacement of the object caused by the push pulse, and the time interval of each of the plurality of ultrasound signals included in the second observation signal is longer than the time interval of each of the plurality of ultrasound signals included in the first observation signal, wherein determining the inefficiency region of the object comprises: determining a region in which a displacement value or a shear wave velocity value is greater than or equal to a predetermined value among regions of the object based on the displacement data for each region of the object obtained from the irradiated second observation signal, and determining the determined region as the inefficiency region.

10. A method for controlling an ultrasonic imaging apparatus, the method comprising: irradiating a push pulse to an object to generate a shear wave in the object; irradiating a first observation signal for observing a displacement of the object caused by the push pulse to the object after the push pulse is irradiated; irradiating a second observation signal for observing a displacement of a blood vessel that occurs regardless of the push pulse to the object after the first observation signal is irradiated; determining an inefficiency region of the object containing blood flow based on displacement data obtained from the irradiated second observation signal that indicates movement the displacement of the blood vessel that occurs regardless of the push pulse; generating a shear wave elastic image based on displacement data for each region of the object obtained from the irradiated first observation signal and the determined inefficiency region; and displaying the generated shear wave elastic image, wherein the second observation signal is irradiated for a time more than 10 milliseconds for observing more than half cycle of the blood vessel and shorter than a time for which the first observation signal is irradiated, wherein the first observation signal and the second observation signal include a plurality of ultrasound signals having a predetermined time interval, wherein a time interval of each of the plurality of ultrasound signals included in the second observation signal is less than or equal to 500 microseconds for observing the displacement of the blood vessel, a time interval of each of the plurality of ultrasound signals included in the first observation signal is less than or equal to 200 microseconds for observing the displacement of the object caused by the push pulse, and the time interval of each of the plurality of ultrasound signals

included in the second observation signal is longer than the time interval of each of the plurality of ultrasound signals included in the first observation signal, wherein determining the inefficiency region of the object comprises: determining a region in which a displacement value or a shear wave velocity value is greater than or equal to a predetermined value among regions of the object based on the displacement data for each region of the object obtained from the irradiated second observation signal, and determining the determined region as the inefficiency region.
