

# US Patent & Trademark Office

## Patent Public Search | Text View

---

United States Patent Application Publication

20250281342

Kind Code

A1

Publication Date

September 11, 2025

Inventor(s)

Donval; Jacques Antoine et al.

---

### DETECTION OF ACTUATOR QUICK RELEASE ACTIVATION

---

#### Abstract

A quick release system for a patient support apparatus includes an arm that moves a cable to release an actuator so that a frame of the patient support apparatus moves to a CPR position. A detector detects a position of the arm.

---

**Inventors:** Donval; Jacques Antoine (Pluvigner, FR), Cou; Sebastien Jean Henri (Erdeven, FR), Guyot; David Jean Francois (Landaul, FR), Morel; Thomas (Auray, FR)

**Applicant:** Hill-Rom Services, Inc. (Batesville, IN)

**Family ID:** 1000008481052

**Assignee:** Hill-Rom Services, Inc. (Batesville, IN)

**Appl. No.:** 19/062239

**Filed:** February 25, 2025

#### Related U.S. Application Data

us-provisional-application US 63563561 20240311

---

#### Publication Classification

**Int. Cl.:** A61G13/10 (20060101); A61G13/04 (20060101); A61G13/12 (20060101); A61H31/00 (20060101)

**U.S. Cl.:**

**CPC** A61G13/105 (20130101); A61G13/04 (20130101); A61G13/1265 (20130101); A61G13/1285 (20130101); A61H31/008 (20130101); A61G2203/30 (20130101)

---

## Background/Summary

CROSS-REFERENCE TO RELATED APPLICATIONS [0001] This application claims priority under 35 U.S.C. § 119 (e) to U.S. Provisional Patent Application Ser. No. 63/563,561, filed Mar. 11, 2024, which is expressly incorporated by reference herein.

### BACKGROUND

[0002] The present disclosure relates to an actuator quick release and, in particular, to detection of an actuator quick release.

[0003] Hospital beds typically include a CRP handle that is activated to move the support sections of the bed into flat positions that are level to the ground. When the CPR handle is activated, a quick release system having a cable connected to a clutch of the CPR handle releases at least one actuator responsible for positioning the various support sections of the bed. Upon release of the actuator, the support sections return to the flat configuration.

[0004] Generally, to detect the activation of the quick release system of an actuator, access to the clutch is achieved by drilling into a cover of the system and a mechanical detection device with a pin and a mechanical switch is inserted at the clutch. This solution potentially leads to a loss of the systems IP rating, which is a measure of the resistance of the system to dust or liquids. This can lead to mechanical blocking in the system due to dust insertion. Such complex mechanical adjustments to the parts can also degrade the actuator functions.

### SUMMARY

[0005] The present disclosure includes one or more of the features recited in the appended claims and/or the following features which, alone or in any combination, may comprise patentable subject matter.

[0006] According to a first aspect of the disclosed embodiments, a patient support apparatus includes a frame having an actuator that moves the frame to various positions. A handle is moveable between a deactivated position and an activated position. In the activated position, the handle releases the actuator of the frame to move the frame to a CPR position. An arm is coupled to the handle and moveable with the handle. In the activated position, the arm releases the actuator. A magnet is coupled to the arm. A magnetic field detector detects the magnet when the handle is in the deactivated position. The magnetic field detector does not detect the magnet when the handle is in the activated position.

[0007] In some embodiments of the first aspect a housing may be provided. The arm and the magnet may be positioned inside the housing. The magnetic field detector may be positioned outside the housing. The magnet may generate a magnetic field that is detectable by the magnetic field detector. The magnetic field detector may transmit an electrical current in the presence of the magnetic field. A controller may be provided. When the handle is in the deactivated position, the controller may receive the electrical current. When the handle is in the activated position, the controller may not receive the electrical current. A user interface may indicate to a user whether the handle is in the activated position or the deactivated position based on whether the controller receives the electrical current. The user interface may be at least one of a mobile device and a remote device.

[0008] Optionally, in the first aspect, the arm may rotate about an axel. The magnet may be positioned on an end of the arm so that the magnet rotates about the axel. A cable may attach the arm to the handle.

[0009] It may be desired, in the first aspect, that, in response to the handle being moved to the activated position, at least one motor may move the frame to lie substantially parallel to the ground. In response to the handle being moved to the activated position, the frame may be moved into a Trendelenburg position. In response to the handle being moved to the activated position, a mattress

positioned on the frame may be deflated. In response to the handle being moved to the activated position, a mattress positioned on the frame may be inflated.

[0010] According to a second aspect of the disclosed embodiments, a quick release system for a patient support apparatus includes an arm that is moveable between a deactivated position and an activated position. In the activated position, the arm releases an actuator so that a frame of the patient support apparatus moves to a CPR position. A magnet is coupled to the arm. A magnetic field detector detects the magnet when the arm is in the deactivated position. The magnetic field detector does not detect the magnet when the arm is in the activated position.

[0011] In some embodiments of the second aspect a housing may be provided. The arm and the magnet may be positioned inside the housing. The magnetic field detector may be positioned outside the housing. The magnet may generate a magnetic field that is detectable by the magnetic field detector. The magnetic field detector may transmit an electrical current in the presence of the magnetic field. A controller may be provided. When the arm is in the deactivated position, the controller may receive the electrical current. When the arm is in the activated position, the controller may not receive the electrical current. A user interface may indicate to a user whether the arm is in the activated position or the deactivated position based on whether the controller receives the electrical current. The user interface may be at least one of a mobile device and a remote device.

[0012] Optionally, in the second aspect, the arm may rotate about an axel. The magnet may be positioned on an end of the arm so that the magnet rotates about the axel. The axel may couple to the actuator.

[0013] According to a third aspect of the disclosed embodiments, a quick release system for a patient support apparatus includes a housing coupled to the patient support apparatus. An arm is positioned in the housing. The arm is moveable about an axel between a deactivated position and an activated position. In the activated position, the arm releases a frame of the patient support apparatus into a CPR position. A magnet is coupled to an end of the arm and configured to move about the axel between the deactivated position and the activated position. A magnetic field detector detects a magnetic field generated by the magnet when the arm is in the deactivated position. The magnetic field detector does not detect the magnet field when the arm is in the activated position. The magnetic field detector transmits an electrical current in the presence of the magnetic field.

[0014] In some embodiments of the third aspect, a controller may be provided. When the arm is in the deactivated position, the controller may receive the electrical current. When the arm is in the activated position, the controller may not receive the electrical current.

[0015] Additional features, which alone or in combination with any other feature(s), such as those listed above and those listed in the claims, may comprise patentable subject matter and will become apparent to those skilled in the art upon consideration of the following detailed description of various embodiments exemplifying the best mode of carrying out the embodiments as presently perceived.

---

## Description

### BRIEF DESCRIPTION OF THE DRAWINGS

[0016] The detailed description particularly refers to the accompanying figures in which:

[0017] FIG. 1 is a perspective view of an exemplary hospital bed in accordance with an embodiment.

[0018] FIG. 2 is a view of a quick release system for a CPR handle, in accordance with an embodiment, wherein a cover of the system is removed to show an inside of the system, and

[0019] wherein a reed switch is attached to an outside of the system.

[0020] FIG. 3 is a schematic view of the quick release system shown in FIG. 2 in a deactivated position.

[0021] FIG. 4 is a schematic view of the quick release system shown in FIG. 2 in an activated position.

#### DETAILED DESCRIPTION

[0022] While the concepts of the present disclosure are susceptible to various modifications and alternative forms, specific exemplary embodiments thereof have been shown by way of example in the drawings and will herein be described in detail. It should be understood, however, that there is no intent to limit the concepts of the present disclosure to the particular forms disclosed, but on the contrary, the intention is to cover all modifications, equivalents, and alternatives falling within the spirit and scope of the invention as defined by the appended claims.

[0023] Referring to FIG. 1, a patient support apparatus is embodied as a hospital bed **10**. The bed **10** includes a frame **12** that is movable between an operating position and a CPR position. The frame **12** included in the bed **10** is supported by wheels or casters **28**. The bed **10** includes a mattress **30** supported by the frame **12**, a number of siderails **32**, a footboard **34**, and a headboard **33**. The bed **10** also includes multiple support sections such as a head support section **36**, a midsection support section **38**, and a foot support section **40**. An actuator **60** of the frame **12** is adjustably operable to place the bed support sections in multiple positions relative to one another.

[0024] In an exemplary embodiment, the frame **12** is adjustable to place the bed **10** in a CPR position in which the head support section **36**, the midsection support section **38**, and the foot support section **40** lie substantially parallel to one another and ground **42**. The patient resting on the bed **10** while the bed **10** is in the CPR position receives head support from the head support section **36**, midsection support from the midsection support section **38**, and leg and foot support from the foot support section **40** at substantially the same distance above ground **42**.

[0025] The CPR handle **44** is coupled to the frame **12** and is used to adjust the frame **12** to release the actuator **60** and move the frame **12** toward the CPR position. In some embodiments, a CPR handle **44** is provided on both sides of the frame **12**. That is, when the CPR handle **44** is activated, the actuator **60** is released so that the frame **12** of the bed **10** is moved into the CPR position. When the CPR handle **44** is activated, the head support section **36** rapidly drops to the substantially parallel position. Additionally, when the CPR handle **44** is activated, at least one motor moves the midsection support section **38** and the foot support section **40** to lie substantially parallel to one another and the ground **42**. In some embodiments, in response to the CPR handle **44** being activated, the bed **10** is moved into the Trendelenburg position. In some embodiments, in response to the CPR handle **44** being activated, bladders in the mattress **30** are completely deflated to provide a hard surface for performing CPR. In some embodiments, in response to the CPR handle **44** being activated, bladders in the mattress **30** are inflated to a maximum inflation to provide a hard surface for performing CPR. In some embodiments, the maximum inflation is within a range of 20 to 55 inches of water. When the CPR handle is deactivated, the frame **12** of the bed **10** remains in a selected position. The CPR handle **44** is movable relative to the frame **12** and may be activated to move the bed **10** toward the CPR position until the bed **10** is placed in the CPR position. The CPR handle **44** may also be deactivated to allow the frame **12** to remain in the operating position. As described in more detail below, a signal is transmitted to a controller **62** of the bed **10** to indicate whether the CPR handle **44** is activated. The controller **62** is in communication with a user interface **52** that displays indicia indicating whether the CPR handle **44** is activated or deactivated. As described in more detail below, the user interface **52** indicates to a user whether the CPR handle **44** is activated or deactivated based on whether the controller **62** receives the electrical current. In some embodiments, the user interface **52** is located on the frame **12** and/or within the vicinity of a patient room. In some embodiments, the user interface **52** is a remote interface, for example, at a nurse's station or elsewhere in a healthcare facility. In some embodiments, the user interface **52** is provided on a mobile device.

[0026] Referring now to FIG. 2, quick release system **100** for the bed **10** includes a housing **102** coupled to the patient support apparatus **10**. A magnetic field detector **110** is positioned outside the housing **102**. It will be appreciated that in some embodiments the magnetic field detector **110** is positioned inside the housing **102**. In an exemplary embodiment, the magnetic field detector **110** is embodied as a reed switch that transmits an electrical current in the presence of a magnetic field. An arm **120** is positioned inside the housing **102**. The arm **120** is coupled to the CPR handle **44** and operable to move with the CPR handle **44** between an activated position and a deactivated position. That is, when the CPR handle **44** is activated to release the actuator **60** of the frame **12**, the arm **120** is moved to an activated position, and, when the CPR handle **44** is deactivated to maintain an operating position of the frame **12**, the arm **120** is moved to a deactivated position.

[0027] FIG. 3 illustrates the arm **120** in the deactivated position **104**, and FIG. 4 illustrates the arm **120** in the activated position **106**. Referring to FIGS. 3 and 4, the arm **120** is coupled to the CPR handle **44** by a CPR cable **124** that extends between the arm **120** and the CPR handle **44**. The arm **120** is moveable with the CPR handle **44**. In an embodiment having a CPR handle **44** on each side of the frame **12**, two CPR cables **124** are provided, wherein each CPR cable **124** extends from the arm **120** to one of the CPR handles **44**. . . . The CPR cable **124** attaches to the arm **120** between an end **126** of the arm **120** and an axel **122**. It will be appreciated that the CPR cable **124** is attachable to any position of the arm **120** between the end **126** and the axel **122**. In some embodiments, the CPR cable **124** is attached to the end **126** of the arm **120**. The arm **120** rotates about the axel **122** between the deactivated position **104** and the activated position **106**. The axel **122** is coupled to the actuator **60**. When in the activated position **106**, the arm **120** rotates the axel **122** to release the actuator **60** so that the frame **12** of the bed **10** moves into the CPR position.

[0028] A magnet **140** is coupled to the arm **120**. In the exemplary embodiment, the magnet **140** is coupled to the end **126** of the arm **120**. In the exemplary embodiment, the end **126** of the arm **120** includes a concave surface that forms a cup, and the magnet **140** is held in the cup. The magnet **140** is configured to move when the arm **120** is moved between the deactivated position **104** and the activated position **106**. In the exemplary embodiment, the magnet **140** rotates about the axel **122** when the arm **120** is rotated between the deactivated position **104** and the activated position **106**. The magnet **140** generates a magnetic field **142** that is detectable by the magnetic field detector **110**.

[0029] The magnetic field detector **110** detects the magnetic field **142**, when the magnetic field **142** is within a predetermined distance to the magnet field detector **110**. When the arm **120** is in the deactivated position **104**, the magnet **140** is positioned proximate to the magnetic field detector **110** so that the magnetic field **142** is within the predetermined distance and detected by the magnetic field detector **110**. The magnetic field detector **110** transmits an electrical current in the presence of the magnetic field **142**. That is, when the arm **120** is in the deactivated position **104**, the magnetic field detector **110** transmits the electrical current. Accordingly, when the CPR handle **44** is in the deactivated position **104**, the arm **120** positions the magnet **140** so that the magnetic field detector **110** transmits the electrical current, which is received at the controller **62**. The controller **62** causes the user interface **52** to indicate to the user that the CPR handle **44** is in the deactivated position **104** when the controller **62** receives the electrical current.

[0030] When the arm **120** is in the activated position **106**, the magnet **140** is moved away from the magnetic field detector **110** so that the magnetic field **142** is not within the predetermined distance and is not detected by the magnetic field detector **110**. The magnetic field detector **110** stops transmitting the electrical current when the magnetic field **142** is not detected. That is, when the arm **120** is in the activated position **106**, the magnetic field detector **110** does not transmit the electrical current. Accordingly, when the CPR handle **44** is in the activated position **106**, the arm **120** positions the magnet **140** so that the magnetic field detector **110** does not transmit the electrical current. The controller **62** causes the user interface **52** to indicate to the user that the CPR handle **44** is in the activated position **106** when the controller **62** does not receive the electrical current.

[0031] In some embodiments, when the CPR handle 44 is in the activated position 106, the controller 62 causes at least one motor to move the midsection support section 38 and the foot support section 40 to lie substantially parallel to one another and the ground 42, in response to the controller 62 not receiving the electrical current. In some embodiments, when the CPR handle 44 is in the activated position 106, the controller 62 causes the bed 10 to be moved into the Trendelenburg position, in response to the controller 62 not receiving the electrical current. In some embodiments, when the CPR handle 44 is in the activated position 106, the controller 62 causes bladders in the mattress 30 to be completely deflated to provide a hard surface for performing CPR, in response to the controller 62 not receiving the electrical current. In some embodiments, when the CPR handle 44 is in the activated position 106, the controller 62 causes bladders in the mattress 30 to be inflated to a maximum inflation to provide a hard surface for performing CPR, in response to the controller 62 not receiving the electrical current.

[0032] The disclosed embodiments ensure an optimum reliability of the actuator without any change to the design of the actuator parts. The quick release system is linked to critical functions and requires a maximum level of reliability. The disclosed embodiments, provide a cost effective solution to detection of the quick release mechanism. The disclosed embodiments also allow a live detection of the clutching system without any other factors in the quick release mechanism, for example the cables or the handles.

[0033] Any theory, mechanism of operation, proof, or finding stated herein is meant to further enhance understanding of principles of the present disclosure and is not intended to make the present disclosure in any way dependent upon such theory, mechanism of operation, illustrative embodiment, proof, or finding. It should be understood that while the use of the word preferable, preferably or preferred in the description above indicates that the feature so described can be more desirable, it nonetheless cannot be necessary and embodiments lacking the same can be contemplated as within the scope of the disclosure, that scope being defined by the claims that follow.

[0034] In reading the claims it is intended that when words such as “a,” “an,” “at least one,” “at least a portion” are used there is no intention to limit the claim to only one item unless specifically stated to the contrary in the claim. When the language “at least a portion” and/or “a portion” is used, the item can include a portion and/or the entire item unless specifically stated to the contrary.

[0035] It should be understood that only selected embodiments have been shown and described and that all possible alternatives, modifications, aspects, combinations, principles, variations, and equivalents that come within the spirit of the disclosure as defined herein or by any of the following claims are desired to be protected. While embodiments of the disclosure have been illustrated and described in detail in the drawings and foregoing description, the same are to be considered as illustrative and not intended to be exhaustive or to limit the disclosure to the precise forms disclosed. Additional alternatives, modifications and variations can be apparent to those skilled in the art. Also, while multiple inventive aspects and principles have been presented, they need not be utilized in combination, and many combinations of aspects and principles are possible in light of the various embodiments provided above.

## Claims

1. A patient support apparatus comprising: a frame having an actuator that moves the frame to various positions, a handle that is moveable between a deactivated position and an activated position, wherein, in the activated position, the handle releases the actuator of the frame to move the frame to a CPR position, an arm coupled to the handle and moveable with the handle, wherein, in the activated position, the arm releases the actuator, a magnet coupled to the arm, and a magnetic field detector to detect the magnet when the handle is in the deactivated position, wherein the reed switch does not detect the magnet when the handle is in the activated position.

2. The patient support apparatus of claim 1, further comprising a housing, wherein the arm and the magnet are positioned inside the housing, wherein the magnetic field detector is positioned outside the housing.
3. The patient support apparatus of claim 1, wherein the magnet generates a magnetic field that is detectable by the magnetic field detector, and wherein the magnetic field detector transmits an electrical current in the presence of the magnetic field.
4. The patient support apparatus of claim 3, further comprising a controller, wherein, when the handle is in the deactivated position, the controller receives the electrical current, and, when the handle is in the activated position, the controller does not receive the electrical current.
5. The patient support apparatus of claim 4, further comprising a user interface that indicates to a user whether the handle is in the activated position or the deactivated position based on whether the controller receives the electrical current.
6. The patient support apparatus of claim 5, wherein the user interface is at least one of a mobile device and a remote device.
7. The patient support apparatus of claim 1, wherein: the arm rotates about an axel, and the magnet is positioned on an end of the arm so that the magnet rotates about the axel.
8. The patient support apparatus of claim 1, wherein a cable attaches the arm to the handle.
9. The patient support apparatus of claim 1, wherein, in response to the handle being moved to the activated position, at least one motor moves the frame to lie substantially parallel to the ground.
10. The patient support apparatus of claim 1, wherein, in response to the handle being moved to the activated position, the frame is moved into a Trendelenburg position.
11. The patient support apparatus of claim 1, wherein, in response to the handle being moved to the activated position, a mattress positioned on the frame is at least one of deflated or inflated.
12. A quick release system for a patient support apparatus comprising: an arm that is moveable between a deactivated position and an activated position, wherein, in the activated position, the arm releases an actuator so that a frame of the patient support apparatus moves to a CPR position, a magnet coupled to the arm, and a magnetic field detector to detect the magnet when the arm is in the deactivated position, wherein the magnetic field detector does not detect the magnet when the arm is in the activated position.
13. The quick release system of claim 12, further comprising a housing, wherein the arm and the magnet are positioned inside the housing, wherein the magnetic field detector is positioned outside the housing.
14. The quick release system of claim 12, wherein the magnet generates a magnetic field that is detectable by the magnetic field detector, and wherein the magnetic field detector transmits an electrical current in the presence of the magnetic field.
15. The quick release system of claim 14, further comprising a controller, wherein, when the arm is in the deactivated position, the controller receives the electrical current, and, when the arm is in the activated position, the controller does not receive the electrical current.
16. The quick release system of claim 15, further comprising a user interface that indicates to a user whether the arm is in the activated position or the deactivated position based on whether the controller receives the electrical current.
17. The quick release system of claim 16, wherein the user interface is at least one of a mobile device and a remote device.
18. The quick release system of claim 12, wherein: the arm rotates about an axel, and the magnet is positioned on an end of the arm so that the magnet rotates about the axel.
19. The quick release system of claim 18, wherein the axel couples to the actuator.
20. A quick release system for a patient support apparatus comprising: a housing coupled to the patient support apparatus, an arm positioned in the housing, wherein the arm is moveable about an axel between a deactivated position and an activated position, wherein, in the activated position, the arm releases a frame of the patient support apparatus into a CPR position, a magnet coupled to

an end of the arm and configured to move about the axel between the deactivated position and the activated position, a magnetic field detector to detect a magnetic field generated by the magnet when the arm is in the deactivated position, wherein the magnetic field detector does not detect the magnet field when the arm is in the activated position, wherein the magnetic field detector transmits an electrical current in the presence of the magnetic field, and a controller, wherein, when the arm is in the deactivated position, the controller receives the electrical current, and, when the arm is in the activated position, the controller does not receive the electrical current.

---