

US Patent & Trademark Office

Patent Public Search | Text View

United States Patent Application Publication

20250285221

Kind Code

A1

Publication Date

September 11, 2025

Inventor(s)

Zeller; Mario et al.

Super-Resolution with Non-Rectangular Acquisitions

Abstract

A method and associated interpolation device is described for increasing the resolution of magnetic resonance (MR) image data of an examination object based on MR raw data acquired using a non-rectangular acquisition scheme. The method may include extracting a rectangular portion of the acquired MR raw data, transforming the extracted rectangular portion into image data space, generating high resolution image data based on the image data, transforming the high resolution image data into k-space, partly replacing the high resolution raw data by original raw data assigned to the non-rectangular portion, and transforming the consistent high resolution raw data into image data space.

Inventors: Zeller; Mario (Erlangen, DE), Zhou; Kun (Shenzhen, CN), Nickel; Marcel Dominik (Herzogenaurach, DE)

Applicant: Siemens Healthineers AG (Forchheim, DE)

Family ID: 1000008478963

Assignee: Siemens Healthineers AG (Forchheim, DE)

Appl. No.: 19/072021

Filed: March 06, 2025

Foreign Application Priority Data

EP	24162003.8	Mar. 07, 2024
----	------------	---------------

Publication Classification

Int. Cl.: G06T3/4053 (20240101); G01R33/56 (20060101); G06T3/4007 (20240101)

U.S. Cl.:

Background/Summary

CROSS REFERENCE TO RELATED APPLICATIONS

[0001] This patent application claims priority to European (EP) Patent Application No. 24162003.8, filed Mar. 7, 2024, which is incorporated herein by reference in its entirety.

BACKGROUND

Field

[0002] The disclosure relates to a method for increasing the resolution of magnetic resonance image data of an examination object based on magnetic resonance raw data being acquired using a non-rectangular acquisition scheme. The disclosure also relates to an interpolator. Further, the disclosure relates to a magnetic resonance imaging system.

Related Art

[0003] Imaging systems that are based on a method of magnetic resonance measurement, in particular of nuclear spins, so-called magnetic resonance tomographs, have been successfully established and proven through a wide range of applications. In this type of image acquisition, a static basic magnetic field B_0 , which is used for the initial alignment and homogenization of magnetic dipoles to be examined, is usually superimposed with a quickly switched magnetic field, the so-called gradient field, for spatial resolution of the imaging signal. To determine the material properties of an examination object to be imaged, the dephasing or relaxation time is determined after the magnetization is deflected from the initial orientation, so that various material-typical relaxation mechanisms or relaxation times can be identified. The deflection is usually carried out by a number of HF pulses (the abbreviation HF stands for high frequency), also referred to as excitation pulses, and the spatial resolution is based on a time-determined manipulation of the deflected magnetization with the help of the gradient field in a so-called measurement sequence or control sequence, which determines an exact temporal sequence of HF pulses, the change in the gradient field (by sending out a switching sequence of gradient pulses) and the acquisition of measured values.

[0004] Typically, an association between measured magnetization—from which the material properties mentioned can be derived—and a spatial coordinate of the measured magnetization in the spatial space in which the object under examination is arranged is carried out using an intermediate step. In this intermediate step, acquired magnetic resonance raw data, also referred to as k-space data, are arranged at readout points in the so-called “k-space”, whereby the coordinates of the k-space are encoded as a function of the gradient field. The amount of magnetization (in particular the transverse magnetization in a plane transverse to the basic magnetic field described above) at a specific location on the examination object can be determined from the data of the readout point using a Fourier transformation, which consists of a signal strength (amount of magnetization), which is assigned to a specific frequency (the spatial frequency) or phase position, calculates a signal strength of the signal in the spatial space.

[0005] However, often only a partial scan of the k-space is carried out in order to save time. However, the reduced sampling of the k-space leads to a reduction in the image information reconstructed on the basis of the sampled k-space data, in particular to a reduction in the image resolution. To compensate for this loss of information, there are methods that compensate for undersampling. Such methods can be associated with an augmentation of the k-space data or can be aimed at directly increasing the resolution in the image data space.

[0006] Super-resolution is a means of improving the image quality in magnetic resonance imaging in the sense of improving the image resolution in the image data space. Super-resolution can be

achieved through machine learning, with the goal being to increase the resolution of an image, often by four times or more.

[0007] Deep Resolve Sharp (DRS) is a method for interpolating MR images using a neural super-resolution network (see e.g. Zhang et al., “Residual Dense Network for Image Super-Resolution”, C VPR paper in Open Access version provided by the Computer Vision Foundation). Compared to conventional interpolation methods such as bicubic interpolation or k-space-based zero filling, images interpolated with DRS generally have higher image sharpness because the network has been trained on a variety of edge types.

[0008] In a product setting, the method works as such that a rectangular complex-valued image is fed into the interpolation network which interpolates each image dimension by a factor of 2. For data consistency reasons, the upsampled image is then transformed into k-space and the center is replaced by the original image's k-space. The final image is then obtained by transforming back into image space as it is illustrated in FIG. 1.

[0009] However, the method can currently not be applied to non-rectangular input data or data with Partial Fourier or asymmetric sampling acquisitions as the data consistency step results in zero-valued areas or sharp edges, which result in ringing artifacts after Fourier transform in image space and lead therefore to a deteriorated image quality. Such zero-valued areas are shown in FIG. 2.

[0010] The data to be used for consistency are not able to be specified as a subset of the input data because we have to treat the network used for super-resolution as a “black box”.

Description

BRIEF DESCRIPTION OF THE DRAWINGS/FIGURES

[0011] The accompanying drawings, which are incorporated herein and form a part of the specification, illustrate the embodiments of the present disclosure and, together with the description, further serve to explain the principles of the embodiments and to enable a person skilled in the pertinent art to make and use the embodiments.

[0012] FIG. 1 is a schematic diagram illustrating an example conventional method for increasing the resolution of magnetic resonance image data.

[0013] FIG. 2 shows examples of k-spaces after a data consistency step for an acquisition with asymmetric sampling in readout direction and a BLADE acquisition.

[0014] FIG. 3 is a schematic diagram illustrating a method for increasing the resolution of magnetic resonance image data according to one or more embodiments of the disclosure.

[0015] FIG. 4 is a schematic diagram illustrating a method for increasing the resolution of magnetic resonance image data according to one or more embodiments of the disclosure.

[0016] FIG. 5 is a flowchart illustrating a method for increasing the resolution of magnetic resonance image data according to one or more embodiments of the disclosure.

[0017] FIG. 6 is a schematic view of an interpolator according to one or more embodiments of the disclosure.

[0018] FIG. 7 is a schematic view of a magnetic resonance imaging system according to one or more embodiments of the disclosure.

[0019] The exemplary embodiments of the present disclosure will be described with reference to the accompanying drawings. Elements, features and components that are identical, functionally identical and have the same effect are—insofar as is not stated otherwise—respectively provided with the same reference character.

DETAILED DESCRIPTION

[0020] In the following description, numerous specific details are set forth in order to provide a thorough understanding of the embodiments of the present disclosure. However, it will be apparent to those skilled in the art that the embodiments, including structures, systems, and methods, may be

practiced without these specific details. The description and representation herein are the common means used by those experienced or skilled in the art to most effectively convey the substance of their work to others skilled in the art. In other instances, well-known methods, procedures, components, and circuitry have not been described in detail to avoid unnecessarily obscuring embodiments of the disclosure. The connections shown in the figures between functional units or other elements can also be implemented as indirect connections, wherein a connection can be wireless or wired. Functional units can be implemented as hardware, software or a combination of hardware and software.

[0021] An object of the disclosure is to provide a magnetic resonance imaging technique to improve the image quality when scanning k-space with a non-rectangular acquisition scheme.

[0022] This object may be accomplished by a method for increasing the resolution of magnetic resonance image data of an examination object based on magnetic resonance raw data being acquired using a non-rectangular acquisition scheme according to the disclosure, by an interpolator according to the disclosure, and by a magnetic resonance imaging system according to the disclosure.

[0023] According to the method for increasing the resolution of magnetic resonance image data of an examination object based on magnetic resonance raw data being acquired using a non-rectangular acquisition scheme a rectangular portion of the acquired magnetic resonance raw data is extracted, wherein the rectangular portion may comprise a completely scanned k-space portion. "A completely scanned k-space portion" means that there are no sub-portions in the scanned k-space portion not being sampled due to the geometry of the non-rectangular acquisition scheme. In particular, there are no edge sub-portions being not sampled due to the determined dimensions for the rectangular portion. "Raw data" are measurement data arranged in k-space. In contrast thereto, image data are represented in image data space and can be determined by transforming raw data from k-space into image data space using a Fourier transform, in particular a Fast Fourier transform.

[0024] Then, the raw data assigned to the extracted rectangular portion are transformed into image data space, wherein image data based on reduced rectangular k-space are generated. The transformation into image data space is realized, because the super-resolution method is basically performed in image data space.

[0025] Further, high resolution image data are generated based on the image data by applying a super-resolution method to the image data which are based on k-space data assigned to the extracted rectangular portion. Super-resolution, in particular using a Residual Dense Network, is described in Zhang et al. being already cited above.

[0026] After increase of resolution, the high resolution image data are transformed back into k-space using inverse Fourier transform, wherein high resolution raw data are generated.

[0027] After transformation back into k-space, high resolution raw data are partly replaced by the original raw data comprising raw data assigned to the non-rectangular portion of the acquired magnetic resonance raw data. In this manner, consistent high resolution raw data are generated.

[0028] After generation of consistency, the consistent high resolution raw data are transformed into image data space using a Fourier transformation, wherein consistent high resolution image data are generated.

[0029] Advantageously, by extracting a rectangular portion of the k-space data, which is a real subset of the non-rectangular portion, the input of "empty" k-space into a super-resolution network is avoided and by inserting k-space data assigned to the non-rectangular portion which does not include empty k-space portions, ringing artifacts etc. are reduced dramatically compared to conventional techniques.

[0030] The interpolator according to the disclosure may comprise an extractor for extracting a rectangular portion of acquired magnetic resonance raw data, wherein the rectangular portion may comprise a completely scanned k-space portion. That means that the rectangular portion does not

include any “empty portion” of k-space. As mentioned above, the exclusion of any empty portion enables to avoid artifacts in the step of increasing the resolution, because the super-resolution algorithm is adapted to rectangular k-space sampling methods and rectangular k-space portions without any empty portion.

[0031] The interpolator also may comprise a transformer for transforming the extracted rectangular portion into image data space, wherein image data based on reduced rectangular k-space are generated.

[0032] The interpolator also may comprise a super-resolution unit for generating high resolution image data based on the image data by applying a super-resolution method to the generated image data.

[0033] Part of the interpolator **60** is also a high resolution transformer for transforming the high resolution image data into k-space, wherein high resolution raw data are generated.

[0034] Further, the interpolator **60** includes a replacing unit for partly replacing the high resolution raw data by the original raw data assigned to the non-rectangular k-space portion of the acquired magnetic resonance raw data, wherein consistent high resolution raw data are generated.

[0035] Furthermore, the interpolator **60** may comprise a consistent data transformer for transforming the consistent high resolution raw data into image data space, wherein consistent high resolution image data are generated.

[0036] The interpolator **60** according to the disclosure shares the advantages of the method for increasing the resolution of magnetic resonance image data of an examination object based on magnetic resonance raw data being acquired using a non-rectangular acquisition scheme according to the disclosure.

[0037] The magnetic resonance imaging system according to the disclosure may include a scanner (scan unit) configured to acquire magnetic resonance raw data from an examination object. The magnetic resonance imaging system may also include a controller configured to control the acquisition and for reconstructing image data based on the acquired magnetic resonance raw data. Further, the magnetic resonance imaging system may comprise an interpolator according to the disclosure, which is arranged for generating high resolution image data.

[0038] The magnetic resonance imaging system according to the disclosure shares the advantages of the interpolator according to the disclosure, which is arranged for generating high resolution image data.

[0039] Some units or modules of the interpolator mentioned above can be completely or partially realized as software modules running on a processor of a respective computing system, e.g. of a controller of a magnetic resonance imaging system. A realization largely in the form of software modules can have the advantage that applications already installed on an existing computing system can be updated, with relatively little effort, to install and run these units of the present application. The object of the disclosure is also achieved by a computer program product with a computer program that is directly loadable into the memory of a computing system, and which may comprise program units to perform the steps of the inventive method for increasing the resolution of magnetic resonance image data of an examination object based on magnetic resonance raw data being acquired using a non-rectangular acquisition scheme, when the program is executed by the computing system. In addition to the computer program, such a computer program product can also comprise further parts such as documentation and/or additional components, also hardware components such as a hardware key (dongle etc.) to facilitate access to the software.

[0040] A computer readable medium such as a memory stick, a hard-disk or other transportable or permanently-installed carrier can serve to transport and/or to store the executable parts of the computer program product so that these can be read from a processor unit (processor) of a computing system. A processor can comprise one or more microprocessors or their equivalents.

[0041] One or more aspects of the disclosure can also be developed analogously to one or more other aspects of the disclosure. In addition, within the scope of the disclosure, the various features

of different exemplary embodiments can also be combined to form new exemplary embodiments.

[0042] In an exemplary embodiment of the method according to the disclosure, the extracted rectangular portion of the acquired magnetic resonance raw data may comprise a real subset of the non-rectangular k-space portion. Since the real subset is an area in k-space, which is completely sampled, the extracted rectangular portion can be input to the super-resolution network without any empty edge portions and therefore, ringing artifacts resulting from the super-resolution step can be avoided.

[0043] In an exemplary embodiment, the non-rectangular acquisition scheme of the method according to the disclosure may comprise a two-dimensional (2D) non-rectangular acquisition scheme. A 2D acquisition scheme is much more efficient than a three-dimensional (3D) acquisition scheme due to a huge reduction of amount of raw data and need for reconstruction time.

[0044] In an exemplary embodiment, the non-rectangular acquisition scheme may comprise the following types of acquisition schemes: [0045] a radial acquisition scheme, [0046] a PROPELLER acquisition scheme, [0047] a 2D BLADE acquisition scheme, [0048] a spiral imaging acquisition scheme, and/or [0049] an elliptical acquisitions scheme.

[0050] In radial scanning, k-space is scanned by a radial k-space trajectory, which consists of multiple radial profiles, each passing through the center of k-space. Radial trajectories are less susceptible to motion artifacts and are therefore often used to record dynamic physiological processes. In addition, compared to Cartesian trajectories, radial trajectories are more robust against a rotationally incomplete scan, which is used in time-resolved imaging to improve temporal resolution.

[0051] The PROPELLER technique (PROPELLER is an acronym for “Periodically Rotated Overlapping Parallel Lines with Enhanced Reconstruction”) was developed in the late 1990s as a method for reducing motion. The basic idea was to rotate k-space using a series of radially directed strips or “rotor blades”.

[0052] Each “rotor blade” consists of several parallel phase-encoded lines that can be detected using fast spin-echo or gradient-echo methods. In common practice, 8 to 32 rotor blade lines are recorded in a single recording. The rotor blades are then rotated through a small angle (10° to 20°), during which a second data set is recorded. The process continues until image data from the entire k-space circle has been collected.

[0053] The PROPELLER trajectory through k-space offers some unique advantages. The center of k-space (which contains the highest signal amplitude and contributes the most to image contrast) is oversampled, meaning that the signal-to-noise ratio and contrast-to-noise ratio are high.

Oversampling in this area also provides redundancy of information, meaning that the data for each new blade can be compared to data from previous blades for consistency. As the patient moves between acquiring k-space data from different blades, the data for the next blade can be corrected (or even discarded entirely) depending on how anomalous its central information appears.

[0054] In an exemplary embodiment of the method for increasing the resolution of magnetic resonance image data of an examination object based on magnetic resonance raw data assigned to a non-rectangular k-space portion, being acquired using a non-rectangular acquisition scheme according to the disclosure, the consistent high resolution raw data are used for another increasing of resolution based on super-resolution. That means that the method for increasing the resolution is repeated once again, wherein the result of the first loop pass is used as input data for the second loop pass. Advantageously, in the second loop pass, there is no empty space in the input data and therefore, the extracted rectangular portion of the acquired magnetic resonance raw data can be expanded above the non-rectangular acquisition scheme. That means that in the second loop pass, the area of the rectangular portion can be increased and also the resolution of image data based on reduced rectangular k-space, which is the input for the super-resolution algorithm can be increased. Hence, the resolution of the result of the super-resolution algorithm is increased with every loop pass without generating the above-mentioned artifacts caused by empty k-space portions.

[0055] In an exemplary embodiment of the method according to the disclosure, the non-rectangular acquisition scheme may comprise a 3D non-rectangular acquisition scheme. 3D acquisition schemes are suitable for acquiring data not in layers but in volumes. This makes it easier, for example, to create arbitrary layer orientations when viewing.

[0056] In an exemplary embodiment, such a 3D-non-rectangular acquisition scheme may comprise a stack-of-stars acquisition scheme. Such a stack-of-stars scheme is appropriate also for 3D imaging, with scanning taking place radially. It is also just an example of a non-Cartesian 3D recording technique.

[0057] In an exemplary embodiment, the method according to the disclosure is iteratively repeated for further increasing the resolution of the image data, until a predetermined resolution value is achieved. Advantageously, any desired resolution can be achieved for the reconstructed image data.

[0058] In an exemplary embodiment, for each iteration lap the rectangular portion is increased compliant with the consistent high resolution raw data of the last iteration lap. Advantageously, the resolution is increased in each lap without generating ringing artifacts. It is conceivable to apply a larger cropped volume such that the edges of the sequences that are inserted into the super-resolution are not acquired. You may accept a certain amount of data not being included at the edge, so for example the rectangle expands a little up and down a little, even if 5% of the data is missing at the corners, for example.

[0059] In a very preferred variant of the method according to the disclosure, the acquired magnetic resonance raw data cover a circle in the 2D k-space and the extracted rectangular portion may comprise a square which is a real subset of the circle, since the format of the input data of the super-resolution algorithm is predetermined as a rectangle. Advantageously, if the acquisition schema is dimensioned as a circle, a square is optimal format compared to other rectangle formats.

[0060] In an exemplary embodiment, the square may comprise the maximum possible area in the circle such that as much information as possible is input into the super-resolution network.

[0061] FIG. 1 is a schematic diagram 10 illustrating a conventional method for increasing the resolution of magnetic resonance image data. The method works as such that a rectangular complex-valued image BD is fed into an interpolation network SR which interpolates each image dimension by a factor 2 such that high resolution image data HD-BD are generated. For data consistence reasons, the upsampled image data HD-BD are then transformed into k-space data HD-KD using an inverse Fourier transform IFT and the k-space center is replaced in a replacement step RP by the original k-space data KD. The final consistent high resolution image data K-HD-BD are then obtained by transforming the consistent high resolution raw data K-HD-KD into image space using a Fourier transform FT.

[0062] However, the conventional method can currently not be applied to non-rectangular sampled raw data or data with Partial Fourier acquisition schemes or asymmetric sampling acquisition schemes since the data consistency step results in zero-valued areas or sharp edges, which result in ringing artifacts after Fourier transformation in image space.

[0063] In FIG. 2, a diagram 20 is illustrated including examples of k-spaces after a data consistency step for an acquisition with asymmetric sampling in readout direction and a BLADE acquisition. On the left side, a k-space diagram PA representing an acquisition with asymmetric sampling in readout direction is illustrated. On the right side, a k-space diagram BLA of a BLADE acquisition after consistency step is illustrated.

[0064] As can be seen in FIG. 2, a vertical strip VS in the replacing portion in the k-space including no raw data exists which cause ringing artefacts in high resolution image data. In the right image, the replacing portion in k-space may comprise on the edge area empty sub-portions ESP, which also do not include any raw data.

[0065] In FIG. 3, a schematic diagram 30 is shown, illustrating a method for increasing the resolution of magnetic resonance image data according to an embodiment of the disclosure. The example shown in FIG. 3 refers to a BLADE acquisition scheme. Since the original k-space data

have been acquired as rotated blades, k-space is only available as a quasi-circular shape. To avoid results as illustrated in FIG. 2, the k-space of the input image BD is restricted to the quadratic white area which amounts $2/\pi=64\%$ of the original k-space data KD such that extracted k-space data EXT-KD are generated. Then, image data RD-BD, which are based on the extracted k-space data EXT-KD, are generated. These image data RD-BD are further fed into a super-resolution network SR. The super-resolution network SR generates high resolution image data HD-BD. After that, the high resolution image data HD-BD are transformed into k-space such that high resolution k-space data HD-KD are generated. For consistency, the original circular area of the initial raw data KD is inserted into the k-space of the high resolution k-space data HD-KD such that consistent high resolution k-space data K-HD-KD are generated. Based on the consistent high resolution k-space data K-HD-KD, consistent high resolution image data K-HD-BD are finally generated by transforming the consistent high resolution k-space data K-HD-KD into image space.

[0066] Due to the reduced matrix size of the restricted and extracted k-space data EXT-KD used for super resolution, the upsampling factor compared to the original matrix size is smaller than for an original rectangular dataset. Because the black areas of original k-space data KD are filled as well, the ratio of k-space points of interpolated data versus original data is still $8/\pi=2.55$ compared to 4 in a conventional rectangular use case.

[0067] In FIG. 4, a schematic diagram 40 is shown, illustrating a method for increasing the resolution of magnetic resonance image data according to an alternative embodiment of the disclosure. In the method illustrated in FIG. 4, the consistent high resolution k-space data K-HD-KD generated by the method illustrated in FIG. 3 is used as input data. Then, another super resolution step is carried out. Extracted k-space data EXT-KD are generated such that the size is just the size of the original k-space data KD used as input data in the method shown in FIG. 3. In contrast to the original k-space data KD used as input data in the method of FIG. 3, the areas outside of the circular portion are now (in FIG. 4) filled with data and can be used as rectangular input for the super-resolution network SR. After generating high resolution k-space data HD-KD, a consistency step is carried out again, wherein the original k-space data KD are inserted in the center of the high resolution k-space data HD-KD such that consistent high resolution k-space data K-HD-KD are generated. The last step of transforming consistent high resolution k-space data K-HD-KD into image data is left out in FIG. 4 for the sake of simplicity.

[0068] In FIG. 5, a flow chart 500 is shown, illustrating a method for increasing the resolution of magnetic resonance image data according to an embodiment of the disclosure.

[0069] In step 5.I, a rectangular portion EXT-KD of the acquired magnetic resonance raw data KD is extracted. The rectangular portion EXT-KD may comprise a completely scanned k-space portion.

[0070] In step 5.II, the extracted rectangular portion EXT-KD is transformed into image data space using a Fourier transform FT and image data RD-BD, based on reduced rectangular k-space, are generated.

[0071] In step 5.III, high resolution image data HD-BD are generated based on the image data RD-BD by applying a super resolution method to the image data RD-BD.

[0072] In step 5.IV, the high resolution image data HD-BD are transformed into k-space using an inverse Fourier transform IFT and high resolution raw data HD-KD are generated.

[0073] In step 5.V, the high resolution raw data HD-KD are partly replaced by the rectangular portion of the acquired magnetic resonance raw data KD and consistent high resolution raw data K-HD-KD are generated.

[0074] In step 5.VI, the consistent high resolution raw data K-HD-KD are transformed into image data space, wherein consistent high resolution image data K-HD-BD are generated.

[0075] In FIG. 6, a schematic view on an interpolator 60 according to an embodiment of the disclosure is shown.

[0076] The interpolator 60 may comprise an extractor 61 for extracting a rectangular portion EXT-KD of acquired magnetic resonance raw data KD including a non-rectangular k-space portion,

wherein the rectangular portion EXT-KD may comprise a real subset of the non-rectangular k-space portion.

[0077] Part of the interpolator **60** is also a transformer **62** for transforming the extracted rectangular portion EXT-KD into image data space, wherein image data RD-BD based on a reduced rectangular k-space are generated.

[0078] The interpolator **60** may comprise a super resolution generator **63** configured to generate high resolution image data HD-BD based on the image data RD-BD, which are generated based on a reduced rectangular k-space, by applying a super resolution method to the image data RD-BD.

[0079] Further, the interpolator **60** may comprise a high resolution transformer **64** for transforming the high resolution image data HD-BD into k-space, wherein high resolution raw data HD-KD are generated.

[0080] Furthermore, the interpolator **60** includes a replacing unit (replacer) **65** for partly replacing the high resolution raw data HD-KD by the non-rectangular k-space portion of the acquired magnetic resonance raw data KD, wherein consistent high resolution raw data K-HD-KD are generated.

[0081] The interpolator **60** may also may comprise a consistent data transformer **66** for transforming the consistent high resolution raw data K-HD-KD into image data space, wherein consistent high resolution image data K-HD-BD are generated.

[0082] In FIG. 7, a schematic view on a magnetic resonance imaging system according to an embodiment of the disclosure is shown.

[0083] FIG. 7 shows a roughly schematic representation of a magnetic resonance system **70** according to the disclosure (hereinafter referred to as “MR system” for short). On the one hand, it includes the actual magnetic resonance scanner **102** with an examination room **103** or patient tunnel, in which, on a couch **108**, a patient O, in whose body there is, for example, a specific object to be imaged like an organ is located, can be retracted.

[0084] The magnetic resonance scanner **102** is equipped in the usual way with a basic field magnet system **104**, a gradient system **106** as well as an HF transmitting antenna system **105** and an HF receiving antenna system **107**. In the exemplary embodiment shown, the HF transmitting antenna system **105** is a whole-body coil permanently installed in the magnetic resonance scanner **102**, whereas the HF receiving antenna system **107** consists of local coils to be arranged on the patient or test subject (in FIG. 7 only by a single local coil symbolized). In principle, however, the whole body coil **105** can also be used as an HF receiving antenna system and the local coils **107** as an HF transmitting antenna system, provided that these coils can each be switched to different operating modes.

[0085] The MR system **70** may also include a controller (e.g. a central control device) **113**, which is configured to control the MR system **70**. In one or more exemplary embodiments, the controller **113** may include processing circuitry that is configured to perform one or more operations and/or functions of the controller **113**. One or more components of the controller **113** may include processing circuitry that is configured to perform one or more respective operations and/or functions of the component(s).

[0086] This controller **113** may include a sequence controller **114** for pulse sequence control. This is used to control the temporal sequence of high-frequency pulses (HF pulses) and gradient pulses depending on a selected imaging sequence (PS) according to a pulse sequence scheme (PSS). Such an imaging sequence PS or the pulse sequence scheme PSS on which the imaging sequence PS is based can be specified, for example, within a measurement or control protocol P. Usually, different control protocols P for different measurements are stored in a memory **119** and can be selected by an operator (and changed if necessary) and then used to carry out the measurement.

[0087] To output the individual HF pulses, the controller **113** has a high-frequency transmitting device **115**, which generates the HF pulses, amplifies them and feeds them into the HF transmitting antenna system **105** via a suitable interface (not shown in detail). To control the gradient coils of

the gradient system **106**, the controller **113** has a gradient system interface **116**. The sequence controller **114** communicates in a suitable manner, e.g. by sending out sequence control data SD, with the high-frequency transmitting device **115** and the gradient system interface **116** for sending out the pulse sequences PS. The controller **113** also has a high-frequency receiving device **117** (which also communicates in a suitable manner with the sequence controller **114**) in order to be coordinated by the HF transmitting antenna system **107** to acquire received magnetic resonance signals.

[0088] The controller **113** may also include an interpolator **60** according to the disclosure, which may have the structure illustrated in detail in FIG. **6** in one or more exemplary embodiments.

[0089] The interpolator **60** also takes over the acquired data RD after demodulation and digitization as raw data or k-space data RD and reconstructs consistent high resolution magnetic resonance image data K-HD-BD from it. These magnetic resonance image data, optimized in terms of its resolution are then stored in a memory **119**, for example.

[0090] The controller **113** can be operated via a terminal with an input unit **111** and a display unit **109**, via which the entire MR system **70** can also be operated by an operator. MR images can also be displayed on the display unit **109**, and measurements can be planned and started using the input unit **111**, if necessary in combination with the display unit **109**, and in particular suitable control protocols with suitable ones measuring sequences can be selected as explained above and modified if necessary.

[0091] The MR system **70** according to the disclosure and in particular the controller **113** can also have a large number of other components that are not shown in detail here but are usually present on such devices, such as a network interface to connect the entire system to a network and to be able to exchange raw data RD and/or image data K-HD-BD or parameter cards, but also other data, such as patient-relevant data or control protocols.

[0092] Finally, it should be pointed out once again that the detailed methods and structures described above are exemplary embodiments and that the basic principle can also be varied in wide areas by the person skilled in the art without leaving the scope of the disclosure, insofar as it is specified by the claims. For the sake of completeness, it should also be noted that the use of the indefinite articles “a” or “an” does not exclude the fact that the characteristics in question can be present multiple times. Likewise, the term “unit” does not exclude the fact that it consists of several components, which may also be spatially distributed. Further, independent of the grammatical term usage, individuals with male, female or other gender identities are included within the term.

[0093] To enable those skilled in the art to better understand the solution of the present disclosure, the technical solution in the embodiments of the present disclosure is described clearly and completely below in conjunction with the drawings in the embodiments of the present disclosure. Obviously, the embodiments described are only some, not all, of the embodiments of the present disclosure. All other embodiments obtained by those skilled in the art on the basis of the embodiments in the present disclosure without any creative effort should fall within the scope of protection of the present disclosure.

[0094] It should be noted that the terms “first”, “second”, etc. in the description, claims and abovementioned drawings of the present disclosure are used to distinguish between similar objects, but not necessarily used to describe a specific order or sequence. It should be understood that data used in this way can be interchanged as appropriate so that the embodiments of the present disclosure described here can be implemented in an order other than those shown or described here. In addition, the terms “comprise” and “have” and any variants thereof are intended to cover non-exclusive inclusion. For example, a process, method, system, product or equipment comprising a series of steps or modules or units is not necessarily limited to those steps or modules or units which are clearly listed, but may comprise other steps or modules or units which are not clearly listed or are intrinsic to such processes, methods, products or equipment.

[0095] References in the specification to “one embodiment,” “an embodiment,” “an exemplary

embodiment,” etc., indicate that the embodiment described may include a particular feature, structure, or characteristic, but every embodiment may not necessarily include the particular feature, structure, or characteristic. Moreover, such phrases are not necessarily referring to the same embodiment. Further, when a particular feature, structure, or characteristic is described in connection with an embodiment, it is submitted that it is within the knowledge of one skilled in the art to affect such feature, structure, or characteristic in connection with other embodiments whether or not explicitly described.

[0096] The exemplary embodiments described herein are provided for illustrative purposes, and are not limiting. Other exemplary embodiments are possible, and modifications may be made to the exemplary embodiments. Therefore, the specification is not meant to limit the disclosure. Rather, the scope of the disclosure is defined only in accordance with the following claims and their equivalents.

[0097] Embodiments may be implemented in hardware (e.g., circuits), firmware, software, or any combination thereof. Embodiments may also be implemented as instructions stored on a machine-readable medium, which may be read and executed by one or more processors. A machine-readable medium may include any mechanism for storing or transmitting information in a form readable by a machine (e.g., a computer). For example, a machine-readable medium may include read only memory (ROM); random access memory (RAM); magnetic disk storage media; optical storage media; flash memory devices; electrical, optical, acoustical or other forms of propagated signals (e.g., carrier waves, infrared signals, digital signals, etc.), and others. Further, firmware, software, routines, instructions may be described herein as performing certain actions. However, it should be appreciated that such descriptions are merely for convenience and that such actions in fact results from computing devices, processors, controllers, or other devices executing the firmware, software, routines, instructions, etc. Further, any of the implementation variations may be carried out by a general-purpose computer.

[0098] The various components described herein may be referred to as “modules,” “units,” or “devices.” Such components may be implemented via any suitable combination of hardware and/or software components as applicable and/or known to achieve their intended respective functionality. This may include mechanical and/or electrical components, processors, processing circuitry, or other suitable hardware components, in addition to or instead of those discussed herein. Such components may be configured to operate independently, or configured to execute instructions or computer programs that are stored on a suitable computer-readable medium. Regardless of the particular implementation, such modules, units, or devices, as applicable and relevant, may alternatively be referred to herein as “circuitry,” “controllers,” “processors,” or “processing circuitry,” or alternatively as noted herein.

[0099] For the purposes of this discussion, the term “processing circuitry” shall be understood to be circuit(s) or processor(s), or a combination thereof. A circuit includes an analog circuit, a digital circuit, data processing circuit, other structural electronic hardware, or a combination thereof. A processor includes a microprocessor, a digital signal processor (DSP), central processor (CPU), application-specific instruction set processor (ASIP), graphics and/or image processor, multi-core processor, or other hardware processor. The processor may be “hard-coded” with instructions to perform corresponding function(s) according to aspects described herein. Alternatively, the processor may access an internal and/or external memory to retrieve instructions stored in the memory, which when executed by the processor, perform the corresponding function(s) associated with the processor, and/or one or more functions and/or operations related to the operation of a component having the processor included therein.

[0100] In one or more of the exemplary embodiments described herein, the memory is any well-known volatile and/or non-volatile memory, including, for example, read-only memory (ROM), random access memory (RAM), flash memory, a magnetic storage media, an optical disc, erasable

programmable read only memory (EPROM), and programmable read only memory (PROM). The memory can be non-removable, removable, or a combination of both.

Claims

1. A method for increasing the resolution of magnetic resonance image data of an examination object based on magnetic resonance raw data assigned to a non-rectangular k-space portion, being acquired using a non-rectangular acquisition scheme, the method comprising: extracting a rectangular portion of acquired magnetic resonance raw data, wherein the rectangular portion of the acquired magnetic resonance raw data comprises a completely sampled k-space portion of the non-rectangular k-space portion; transforming the extracted rectangular portion into image data space to generate image data based on a reduced rectangular k-space; generating high resolution image data based on the image data by applying a super resolution method to the image data; transforming the high resolution image data into k-space to generate high resolution raw data; partly replacing the high resolution raw data by original raw data assigned to the non-rectangular k-space portion of the acquired magnetic resonance raw data to generate consistent high resolution raw data; and transforming the consistent high resolution raw data into image data space to generate consistent high resolution image data.
2. The method according to claim 1, wherein the rectangular portion of the acquired magnetic resonance raw data comprises a real subset of the non-rectangular k-space portion.
3. The method according to claim 1, wherein the non-rectangular acquisition scheme comprises a two-dimensional (2D) non-rectangular acquisition scheme.
4. The method according to claim 3, wherein the 2D non-rectangular acquisition scheme comprises one of the following types of acquisition schemes: a 2D BLADE acquisition scheme; a radial acquisition scheme; or a spiral imaging acquisition scheme.
5. The method according to claim 1, wherein the non-rectangular acquisition scheme comprises a three-dimensional (3D) non-rectangular acquisition scheme.
6. The method according to claim 5, wherein the non-rectangular acquisition scheme comprises a stack-of-stars acquisition scheme.
7. The method according to claim 1, wherein the extracting the rectangular portion, transforming the extracted rectangular portion, generating high resolution image data, transforming the high resolution image data, and partly replacing the high resolution raw data are repeated, wherein the consistent high resolution raw data are used in the extracting the rectangular portion instead of acquired magnetic resonance raw data.
8. The method according to claim 7, wherein the extracting the rectangular portion, transforming the extracted rectangular portion, generating high resolution image data, transforming the high resolution image data, and partly replacing the high resolution raw data are iteratively repeated until a predetermined resolution value is achieved.
9. The method according to claim 8, wherein for each iteration, the rectangular portion is increased compliant with the consistent high resolution raw data of a last iteration.
10. The method according to claim 1, wherein the acquired magnetic resonance raw data cover a circle in two-dimensional (2D) k-space and the extracted rectangular portion comprises a square which is a real subset of the circle.
11. The method according to claim 10, wherein the square comprises a maximum possible area in the circle.
12. A non-transitory computer-readable storage medium with an executable program stored thereon, that when executed, instructs a processor to perform the method of claim 1.
13. An apparatus comprising: one or more processors; and memory storing instructions that, when executed by the one or more processors, cause the apparatus to perform the method of claim 1.
14. An interpolation device, comprising: an extractor configured to extract a rectangular portion of

acquired magnetic resonance (MR) raw data assigned to a non-rectangular k-space portion, wherein the rectangular portion comprises a completely sampled k-space portion of the non-rectangular k-space portion; a transformer configured to transform the extracted rectangular portion into image data space to generate image data based on a reduced rectangular k-space; a super-resolution generator configured to generate high resolution image data based on the image data by applying a super-resolution method to the image data; a high resolution transformer configured to transform the high resolution image data into k-space to generate high resolution raw data; a replacer configured to partly replace the high resolution raw data by original raw data assigned to the non-rectangular portion of the acquired magnetic resonance raw data to generate consistent high resolution raw data; and a consistent data transformer configured to transform the consistent high resolution raw data into image data space to generate consistent high resolution image data.

15. A magnetic resonance (MR) imaging system, comprising: a scanner configured to acquire MR raw data from an examination object; a controller configured to control the scanner to acquire the MR raw data and to reconstruct image data based on the acquired MR raw data; and the interpolation device (60) according to claim 14 configured to generate a high resolution image data.

16. An interpolator comprising: one or more processors; and memory storing instructions that, when executed by the one or more processors, cause the interpolator to: extract a rectangular portion of acquired magnetic resonance (MR) raw data assigned to a non-rectangular k-space portion, wherein the rectangular portion comprises a completely sampled k-space portion of the non-rectangular k-space portion; transform the extracted rectangular portion into image data space to generate image data based on a reduced rectangular k-space; generate high resolution image data based on the image data by applying a super-resolution method to the image data; transform the high resolution image data into k-space to generate high resolution raw data; partly replace the high resolution raw data by original raw data assigned to the non-rectangular portion of the acquired magnetic resonance raw data to generate consistent high resolution raw data; and transform the consistent high resolution raw data into image data space to generate consistent high resolution image data.

17. A magnetic resonance (MR) imaging system, comprising: a scanner configured to acquire MR raw data from an examination object; a controller configured to control the scanner to acquire the MR raw data and to reconstruct image data based on the acquired MR raw data; and the interpolator according to claim 16 configured to generate a high resolution image data.
